

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems

Appareils électromédicaux –
Partie 2-66: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de correction auditive et des systèmes de correction auditive



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22,000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67,000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –
**Part 2-66: Particular requirements for the basic safety and essential
performance of hearing aids and hearing aid systems**

Appareils électromédicaux –
**Partie 2-66: Exigences particulières pour la sécurité de base et les
performances essentielles des appareils de correction auditive et des systèmes
de correction auditive**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.180.15; 17.140.50

ISBN 978-2-8322-7350-0

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	11
201.4 General requirements	12
201.5 General requirements for testing ME EQUIPMENT	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	15
201.7 ME EQUIPMENT identification, marking and documents	15
201.8 * Protection against electrical HAZARDS from ME EQUIPMENT	21
201.9 * Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10 * Protection against unwanted and excessive radiation HAZARDS	25
201.11 * Protection against excessive temperatures and other HAZARDS	25
201.12 * Accuracy of controls and instruments and protection against hazardous outputs.....	27
201.13 * HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	28
201.14 * PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	30
201.15 * Construction of ME EQUIPMENT.....	31
201.16 * ME SYSTEMS	33
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	33
Annexes	35
Annex E (informative) Examples of the connection of the measuring device (MD) for measurement of the PATIENT LEAKAGE CURRENT and PATIENT AUXILIARY CURRENT	36
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	37
Annex H (informative) PEMS structure, PEMS DEVELOPMENT LIFE-CYCLE and documentation	38
Annex I (informative) ME SYSTEMS aspects	39
Annex J (informative) Survey of insulation paths.....	40
Annex K (informative) Simplified PATIENT LEAKAGE CURRENT diagrams	41
Annex L (normative) Insulated winding wires for use without interleaved insulation.....	42
Annex AA (informative) Particular guidance and rationale	43
Annex BB (informative) Abbreviations	48
Annex CC (informative) Essential performance	49
Annex DD (informative) Electromagnetic compatibility of ME EQUIPMENT	50
Bibliography.....	52
Index of defined terms used in this particular standard.....	54
Figure 201.101 – Measuring circuit for LEAKAGE CURRENT (see 201.8.7.4.7)	22
Table 201.101 – MECHANICAL HAZARDS to be considered	24
Table AA.101 – Summary of the approach of this document	44

Table DD.101 – Example of applicable standards for HEARING AID specific setting for RF radiated emissions	50
Table DD.102 – Example of applicable tests	51

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 60601-2-66:2019](https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019)

<https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-66 has been prepared by IEC technical committee 29: Electroacoustics.

This third edition cancels and replaces the second edition published in 2015. It constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) revision of the definition about ESSENTIAL PERFORMANCE;
- b) revision of the application of IEC 60601-1-2:2014 for electromagnetic disturbances;
- c) correction of the used voltage for HEARING AIDS from 1,6 V to 4,5 V;
- d) correction of the drop test level from 1,5 m to 1,0 m;
- e) correction of the wording of IEC 60601-2-66:2015.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
29/1023/FDIS	29/1030/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications*: italic type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 60601-2-66:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019>

INTRODUCTION

In 1998, the HEARING AID industry represented by the European hearing instrument manufacturers association (EHIMA) attempted to establish a standard with the main purpose of providing MANUFACTURERS with a guide to demonstrate conformity with the European Medical Devices Directive 93/42/EEC.

The draft document prEN 50220 failed CENELEC vote and was published as "EHIMA standard" in June 1998 with almost identical content. EHIMA concluded in 2009 that the requirements of that standard were no longer up to date and an internationally accepted standard for HEARING AID safety published by IEC or ISO to demonstrate compliance with regulatory requirements should be produced.

This particular standard amends and supplements IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, hereinafter referred to as the "general standard".

A general guidance and rationale for the requirements of this particular standard are given in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of this particular standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this document.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC 60601-2-66:2019](https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019)

<https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019>

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY of HEARING AIDS and HEARING AID SYSTEMS, hereafter also referred to as ME EQUIPMENT or ME SYSTEM.

If a clause or subclause is specifically intended to be applicable to HEARING AIDS only, or to HEARING AID SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to HEARING AIDS and to HEARING AID SYSTEMS, as relevant.

iTeh STANDARD PREVIEW

HAZARDS inherent in the intended physiological function of HEARING AIDS or HEARING AID SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 201.7.9.2 and 201.9.6.

<https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019>

NOTE See also 4.2 of the general standard.

ACCESSORIES to HEARING AIDS in the HOME HEALTHCARE ENVIRONMENT (e.g. remote control units, audio streamers, battery chargers, power supplies) can be tested according to the applicable standard, IEC 60065, IEC 60950-1, IEC 62368-1 or other applicable IEC safety standards. Alternatively, the general standard may be applied. HEARING AIDS do not have a MAINS PART intended for connection to AC SUPPLY MAINS. The connection to the SUPPLY MAINS of a HEARING AID SYSTEM is covered by power supply, charger or other types of ACCESSORIES.

ACCESSORIES with FUNCTIONAL CONNECTION to a HEARING AID may form a HEARING AID SYSTEM. HEARING AID related ACCESSORIES that are not physically connected to the HEARING AID during NORMAL USE are not considered to be APPLIED PART, because they do not directly contribute to the INTENDED USE of the HEARING AID.

Wireless programming interfaces are covered by the applicable standard IEC 60065, IEC 60950-1, IEC 62368-1 or other applicable IEC safety standards. Alternatively, the general standard may be applied.

Programming interfaces with wired connection to the HEARING AID are covered by the general standard.

NOTE Detachable parts of HEARING AIDS, even if supplied separately (e.g. ear hooks, domes, wax guards etc.), are not considered as ACCESSORIES, but as component parts.

¹ The general standard is IEC 60601-1 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

This document does not apply to:

- cochlear implants or other implanted HEARING AIDS;
- bone conduction HEARING AIDS;
- educational HEARING AIDS (i.e. group HEARING AIDS, auditory trainers etc.);
- the application of a HEARING AID for the measurement of hearing levels; IEC 60645-1 applies;
- fix installed audio-frequency induction-loop systems or their component parts, as described in IEC 60118-4 and IEC 62489-1;
- the sound generating function of a tinnitus masker.

This document does not address applicable testing for intentional RF radiation of wireless equipment (e.g. maximum radiated output power, modulation bandwidth, etc.).

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY requirements for HEARING AIDS and HEARING AID SYSTEMS as defined in 201.3.202 and 201.3.203.

201.1.3 * Collateral standards

Addition:

iTeh STANDARD PREVIEW

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-3, IEC 60601-1-9 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 and IEC 60601-1:2005/AMD1:2012 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

ITEH STANDARD PREVIEW

201.2 Normative references (standards.iteh.ai)

NOTE Informative references are listed in the [Bibliography](#).

Clause 2 of the general standard applies except as follows:

Replacement:

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

IEC 60950-1:2005, *Information technology equipment – Safety – Part 1: General requirements*

IEC 60950-1:2005/AMD1:2009

IEC 60950-1:2005/AMD2:2013

Addition:

IEC 60118-0:2015, *Electroacoustics – Hearing aids – Part 0: Measurement of the performance characteristics of hearing aids*

IEC 60118-13, *Electroacoustics – Hearing aids – Part 13: Electromagnetic compatibility (EMC)*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 62368-1:2018, *Audio/video, information and communication technology equipment – Part 1: Safety requirements*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1 and IEC 60601-1:2005/AMD1:2012 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE An index of defined terms is found beginning on page 54.

201.3.73

OPERATOR

Addition:

Note 101 to entry Usually equals to PATIENT for HEARING AIDS in a home healthcare environment.

201.3.76

PATIENT

Replacement:

iTeh STANDARD PREVIEW
(standards.iteh.ai)

Note 1 to entry: In this particular standard and in applying the requirements of the general standard, the term PATIENT has the meaning explained in the second paragraph of 4.1 of the general standard. The PATIENT is also usually the OPERATOR.

[IEC 60601-2-66:2019](https://standards.iteh.ai/catalog/standards/sist/7079-d2-2812-1559/iec-60601-2-66-2019)

The term PATIENT is being used in this document in line with the general terminology in the medical product field. It is, however, understood that the user of a HEARING AIDS is typically not an ill person but someone healthy with a hearing impairment in a HOME HEALTHCARE ENVIRONMENT.

201.3.113

SERVICE PERSONNEL

Replacement:

individuals or entity that assemble, maintain or repair HEARING AIDS or HEARING AID SYSTEMS

201.3.132

TYPE B APPLIED PART

Replacement:

APPLIED PART complying with the specified requirements of this particular standard to provide protection against electric shock, particularly regarding allowable PATIENT LEAKAGE CURRENT and PATIENT AUXILIARY CURRENT

Addition:

201.3.201

HEARING AID PROFESSIONAL

audiologically competent person who professionally assesses hearing, selects, fits and delivers HEARING AID systems and rehabilitation services to persons with hearing impairment

201.3.202

HEARING AID

ME EQUIPMENT which picks up sound and delivers processed sound to the ear canal through air-conduction

Note 1 to entry: A HEARING AID includes all detachable parts that are essential for the performance of its INTENDED USE.

201.3.203

HEARING SYSTEM

HEARING AID SYSTEM

combination, as specified by its MANUFACTURER, of items of equipment, at least one of which is a HEARING AID to be inter-connected by specified FUNCTIONAL CONNECTIONS

201.3.204

SOUND PRESSURE LEVEL

SPL

L_p

ten times the logarithm to the base 10 of the ratio of the root-mean-square of the sound pressure in the time domain, p , to the square of the reference value, p_0 , expressed in decibels

$$L_p = 10 \lg \frac{p^2}{p_0^2} \text{ dB}$$

where the reference value, p_0 , is 20 μPa

Note 1 to entry: p^2 is understood to denote the square of a frequency-weighted, frequency-band-limited or time-weighted sound pressure.

If specific frequency and time weightings as specified in IEC 61672-1 and/or specific frequency bands are applied, this should be indicated by appropriate subscripts; e.g. $L_{p,AF}$ denotes the A-weighted SOUND PRESSURE LEVEL with time weighting F.

Note 2 to entry: This definition is technically in accordance with ISO/FDIS 80000-8:2019, 8-22.

Note 3 to entry: This note applies to the French language only.

Note 4 to entry: Any SOUND PRESSURE LEVEL specified in this document is measured in decibels (dB) as described in IEC 60118-0 by using a coupler according to IEC 60318-5.

[SOURCE: ISO/TR 25417:2007, 2.2. modified – The abbreviated term "SPL" has been added, as well as Notes 3 and 4 to entry.]

201.3.205

MAXIMUM OUTPUT SOUND PRESSURE LEVEL

MAXIMUM OSPL90

maximum output SPL of the HEARING AID when measured with full-on setting and 90 dB input SPL, as defined in IEC 60118-0

201.3.206

FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL

FITTED OSPL90

maximum output SPL of a HEARING AID when measured after fitting and 90 dB input SPL

Note 1 to entry: Since the HEARING AID parameters are not full-on, the FITTED OSPL90 is equal or lower than the MAXIMUM OSPL90. Usually, the FITTED OSPL90 is achieved by selection of a HEARING AID with an appropriate MAXIMUM OSPL90 and fine tuning of the maximum output.

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.1 Conditions for application to ME EQUIPMENT or ME SYSTEMS

Replacement:

Unless otherwise specified, the requirements of this document shall apply in NORMAL USE and reasonably foreseeable misuse.

When applying this document to HEARING AID or HEARING AID SYSTEMS, the definitions and requirements that use the term PATIENT shall be considered as applying to the person for whom the HEARING AID or HEARING AID SYSTEMS is intended.

201.4.3 ESSENTIAL PERFORMANCE

Addition:

See Annex CC.

201.4.6 ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT

Subclause 4.6 of the general standard does not apply.

201.4.10 Power supply

Subclause 4.10 of the general standard does not apply.

201.4.11 Power input

Subclause 4.11 of the general standard does not apply.

201.5 General requirements for testing ME EQUIPMENT

Clause 5 of the general standard applies, except as follows:

<https://standards.iteh.ai/catalog/standards/sist/7079a42c-2f42-4550-9314-69d0b9fd1ea6/iec-60601-2-66-2019>

201.5.2 Number of samples

Replacement:

TYPE TESTS are performed on a representative sample of the item being tested. If multiple products are under consideration, which have a similar mechanical and electrical architecture, then an engineering analysis by the MANUFACTURER may justify a single representative sample for a family of products.

201.5.3 Ambient temperature, humidity, atmospheric pressure

Replacement:

After the HEARING AID or HEARING AID SYSTEM to be tested has been set up for NORMAL USE, tests are performed within the range of environmental conditions indicated in the technical description, as specified by the MANUFACTURER.

201.5.4 Other conditions

Addition:

- aa) Inventory stocking conditions are specified by the MANUFACTURER.
- bb) HEARING AID transport conditions are specified by the MANUFACTURER.

201.5.5 Supply voltages, type of current, nature of supply, frequency

Replacement: