



# SLOVENSKI STANDARD

## SIST EN 13624:2004

01-februar-2004

Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of fungicidal activity of chemical disinfectants for instruments used in the medical area - Test method and requirements (phase 2, step 1)

Chemische Desinfektionsmittel und Antiseptika - Quantitativer Suspensionsversuch zur Prüfung der fungiziden Wirkung chemischer Desinfektionsmittel für Instrumente im humanmedizinischen Bereich - Prüfverfahren und Anforderungen (Phase 2, Stufe 1)

**STANDARD PREVIEW**

Antiseptiques et désinfectants chimiques - Essai quantitatif de suspension pour l'évaluation de l'activité fongicide des désinfectants chimiques utilisés pour les instruments en médecine - Méthode d'essai et exigences (phase 2, étape 1)

<https://standards.iteh.ai/catalog/standards/sist/7c89990b-ceee-44de-871d-11eb-0374-2c442c442c44>

**Ta slovenski standard je istoveten z: EN 13624:2003**

### ICS:

11.080.20      Dezinfektanti in antiseptiki      Disinfectants and antiseptics

**SIST EN 13624:2004**

**en,fr,de**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN 13624:2004](#)

<https://standards.iteh.ai/catalog/standards/sist/7c89990b-ceee-44de-871d-344a613a6ff4/sist-en-13624-2004>

EUROPEAN STANDARD

EN 13624

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2003

ICS 11.080.20

English version

Chemical disinfectants and antiseptics - Quantitative suspension test for the evaluation of fungicidal activity of chemical disinfectants for instruments used in the medical area - Test method and requirements (phase 2, step 1)

Antiseptiques et désinfectants chimiques - Essai quantitatif de suspension pour l'évaluation de l'activité fongicide des désinfectants chimiques utilisés pour les instruments en médecine - Méthode d'essai et prescriptions (phase 2, étape 1)

Chemische Desinfektionsmittel und Antiseptika - Quantitativer Suspensionsversuch zur Prüfung der fungiziden Wirkung chemischer Desinfektionsmittel für Instrumente im humanmedizinischen Bereich - Prüfverfahren und Anforderungen (Phase 2, Stufe 1)

This European Standard was approved by CEN on 3 November 2003.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Management Centre has the same status as the official versions.

<https://standards.iteh.ai/catalog/standards/sist/7c89990b-cccc-44de-871d->

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Slovakia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

<b>Contents</b>	<b>Page</b>
Foreword.....	3
Introduction .....	4
<b>1 Scope .....</b>	<b>5</b>
<b>2 Normative references .....</b>	<b>5</b>
<b>3 Terms and definitions.....</b>	<b>5</b>
<b>4 Requirements .....</b>	<b>6</b>
<b>5 Test methods.....</b>	<b>7</b>
5.1 Principle.....	7
5.2 Materials and reagents.....	7
5.3 Apparatus and glassware .....	10
5.4 Preparation of test organism suspensions and product test solutions .....	11
5.5 Procedure for assessing the fungicidal activity of the product.....	14
5.6 Experimental data and calculation.....	19
5.7 Verification of methodology .....	23
5.8 Expression of results and precision.....	24
5.9 Interpretation of results - conclusion.....	25
5.10 Test report .....	25
<b>Annex A (informative) Referenced strains in national collections.....</b>	<b>27</b>
<b>Annex B (informative) Neutralizer and Rinsing liquids.....</b>	<b>28</b>
B.1 Neutralizers.....	28
B.2 Rinsing liquids .....	29
B.3 Neutralizer added to the agar for counting :.....	29
<b>Annex C (informative) Flow diagrams .....</b>	<b>30</b>
C.1 Dilution neutralization method .....	30
C.2 Membrane filtration method.....	31
<b>Annex D (informative) Example of a typical test report.....</b>	<b>34</b>
<b>Annex E (informative) Information on the application and interpretation of European standards on chemical disinfectants and antiseptics .....</b>	<b>36</b>
E.1 Introduction.....	36
E.2 General guidelines for the application and interpretation of test methods in accordance with European Standards for chemical disinfectants and antiseptics.....	36
E.3 Guide to the interpretation of tests for chemical disinfectants and antiseptics.....	37
<b>Annex ZA (informative) Clauses of this European Standard addressing essential requirements or other provisions of EU Directives. ....</b>	<b>38</b>
<b>Bibliography .....</b>	<b>39</b>

## Foreword

This document (EN 13624:2003) has been prepared by Technical Committee CEN/TC 216 "Chemical disinfectants and antiseptics", the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2004, and conflicting national standards shall be withdrawn at the latest by June 2004.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Medical Devices Directive 93/42.

For relationship with EU Directive, see informative annex ZA, which is an integral part of this document.

A collaborative trial has been undertaken to provide a precision annex to this standard.

Other methods to evaluate the efficacy of chemical disinfectants and antiseptics for different applications in the medical field are in preparation.

Annexes A, B, C, D, and E are informative.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Luxembourg, Malta, Netherlands, Norway, Portugal, Slovakia, Spain, Sweden, Switzerland and the United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/7c89990b-ccee-44de-871d-344a613a6ff4/sist-en-13624-2004>

## Introduction

This European Standard describes a suspension test for establishing whether a chemical disinfectant for use on instruments (surgical instruments, anaesthesia material, endoscopes etc.) has or does not have a fungicidal activity in the area described in the scope.

In this laboratory test chosen conditions include contact time, temperature, test organisms and interfering substances, i.e. conditions which may influence the action of chemical disinfectants in practical situations.

The conditions are intended to cover general purposes and to allow reference between laboratories and product types. Each concentration of the chemical disinfectant found by this test corresponds to the chosen experimental conditions. However, for some applications the instructions of use of a product may differ and therefore additional test conditions need to be used.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN 13624:2004](#)

<https://standards.iteh.ai/catalog/standards/sist/7c89990b-ceee-44de-871d-344a613a6ff4/sist-en-13624-2004>

## 1 Scope

This European Standard specifies a test method and the minimum requirements for fungicidal or yeasticidal activity of chemical disinfectant products that form a homogeneous, physically stable preparation when diluted with hard water or - in the case of ready-to-use products - with water. Products can only be tested at a concentration of 80 % or less as some dilution is always produced by adding the test organisms and interfering substance.

This European Standard applies to products that are used in the medical area for disinfecting instruments by immersion - even if they are not covered by the EEC/93/42 Directive on Medical Devices.

This European Standard applies to areas and situations where disinfection is medically indicated. Such indications occur in patient care, for example :

- in hospitals, in community medical facilities and in dental institutions ;
- in clinics of schools, of kindergartens and of nursing homes ;

and may occur in the workplace and in the home. It may also include services such as laundries and kitchens supplying products directly for the patients.

NOTE 1 The method described is intended to determine the activity of commercial formulations or active substances under the conditions in which they are used.

NOTE 2 This method corresponds to a phase 2 step 1 test (see annex E).

iTeH STANDARD PREVIEW  
(standards.iteh.ai)

## 2 Normative references

[SIST EN 13624:2004](#)

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

EN 12353, *Chemical disinfectants and antiseptics – Preservation of microbial strains used for the determination of bactericidal and fungicidal activity.*

ISO 4793, *Laboratory sintered (fritted) filters – Porosity grading, classification and designation.*

ISO 6710, *Single-use containers for venous blood specimen collection.*

## 3 Terms and definitions

For the purposes of this European Standard, the following terms and definitions apply.

### 3.1

#### **product**

chemical agent or formulation used as chemical disinfectant or antiseptic

**EN 13624:2003 (E)****3.2****fungicide**

product which kills fungi and their spores under defined conditions

Adjective : fungicidal

**3.3****fungicidal activity**

capability of a product to produce a reduction in the number of viable vegetative yeast cells and mould spores of relevant test organisms under defined conditions

**3.4****fungistatic activity**

capability of a product to inhibit the growth of fungi under defined conditions

**3.5****yeasticide**

product which kills yeasts under defined conditions

Adjective : yeasticidal

**3.6****yeasticidal activity**

capability of a product to produce a reduction in the number of viable yeast cells of relevant test organisms under defined conditions

**3.7****clean conditions**

conditions representative of surfaces which have received a satisfactory cleaning programme and/or are known to contain minimal levels of organic and/or inorganic substances

**3.8****dirty conditions**

conditions representative of surfaces which are known to or may contain, organic and/or inorganic substances

iTeh STANDARD PREVIEW  
(standards.itteh.ai)

SIST EN 13624:2004

<http://standards.itteh.ai/catalog/standards/sist/7c89990b-cccc-44de-871d-344a613a6ff4/sist-en-13624-2004>

**4 Requirements**

The product, when diluted with hard water or - in the case of ready-to-use products - with water and tested in accordance with clause 5 under simulated clean conditions (0,3 g/l bovine albumin solution) or dirty conditions (3 g/l bovine albumin solution plus 3 ml/l washed sheep erythrocytes) according to its practical applications and under the obligatory test conditions (one or two selected test organisms, 20 °C, 60 min), shall demonstrate at least a decimal log (lg) reduction in counts of 4. It is possible to test also the product as delivered (highest test concentration = 80 %).

The fungicidal activity shall be evaluated using the following two test organisms : *Aspergillus niger* and *Candida albicans*.

The yeasticidal activity shall be evaluated using the following test organism : *Candida albicans*.

Where indicated, additional specific fungicidal or yeasticidal activity shall be determined applying other contact times, temperatures and interfering substances (see 5.5.1.1) in order to take into account intended specific use conditions.

NOTE For these additional conditions, the concentration defined as a result can be lower than the one obtained under the obligatory test conditions.



## 5 Test methods

### 5.1 Principle

**5.1.1** A test suspension of fungi (yeast cells and mould spores) in a solution of an interfering substance is added to a sample of the product as delivered and/or diluted with hard water (for ready to use products: water). The mixture is maintained at  $20\text{ °C} \pm 1\text{ °C}$  for  $60\text{ min} \pm 10\text{ s}$  (obligatory test conditions). At the end of this contact time, an aliquot is taken; the fungicidal and/or the fungistatic action in this portion is immediately neutralized or suppressed by a validated method. The method of choice is dilution-neutralization. If a suitable neutralizer cannot be found, membrane filtration is used. The numbers of surviving fungi in each sample are determined and the reduction is calculated.

**5.1.2** The test is performed using the spores of *Aspergillus niger* and the vegetative cells of *Candida albicans* or only the vegetative cells of *Candida albicans* as test organisms (obligatory test conditions).

**5.1.3** Additional and optional contact times and temperatures are specified. Additional interfering substances can be used.

### 5.2 Materials and reagents

#### 5.2.1 Test organisms

The fungicidal activity shall be evaluated using the following two test organisms<sup>1)</sup> :

— *Candida albicans* ATCC 10231;

— *Aspergillus niger* ATCC 16404.

The yeasticidal activity shall be evaluated using only *Candida albicans*.

NOTE See annex A for corresponding strain reference in some other culture collections.

#### 5.2.2 Culture media and reagents

##### 5.2.2.1 General

The reagents shall be of analytical grade and/or appropriate for microbiological purposes. They shall be free from substances that are toxic or inhibitory to the test organisms.

NOTE To improve reproducibility, it is recommended that commercially available dehydrated material is used for the preparation of culture media. The manufacturer's instructions relating to the preparation of these products should be rigorously followed. For each culture medium and reagent a limitation for use should be fixed.

##### 5.2.2.2 Water

The water shall be freshly glass distilled water and not demineralized water.

Sterilize in the autoclave (see 5.3.1).

NOTE 1 Sterilization is not necessary if the water is used - e.g. for preparation of culture media - and subsequently sterilized.

---

1) The ATCC numbers are the collection numbers of strains supplied by the American Type Culture Collection (ATCC). This information is given for the convenience of users of this standard and does not constitute an endorsement by CEN of the product named.

**EN 13624:2003 (E)**

NOTE 2 If distilled water of adequate quality is not available, water for injections (see EP) can be used.

**5.2.2.3 Malt Extract Agar (MEA)**

Malt extract	30,0 g
Soja peptone, papaic digest or soybean meal	3,0 g
Agar	15,0 g
Water (see 5.2.2.2)	to 1 000 ml

Sterilize in the autoclave (see 5.3.1). After sterilization the pH of the medium shall be equivalent to  $5,6 \pm 0,2$  when measured at 20 °C (see 5.3.2.4).

NOTE In special circumstances (problems with neutralization - see 5.5.1.2 and 5.5.1.3) it may be necessary to add neutralizer to MEA (see B.3).

**5.2.2.4 Diluent**

Tryptone Sodium Chloride Solution :

- tryptone, pancreatic digest of casein 1,0 g ;
- sodium chloride (NaCl) 8,5 g ;
- water (see 5.2.2.2) to 1 000 ml.

Sterilize in the autoclave (see 5.3.1). After sterilization the pH of the diluent shall be equivalent to  $7,0 \pm 0,2$  when measured at 20 °C.

**5.2.2.5 Neutralizer**

[SIST EN 13624:2004  
https://standards.iteh.ai/catalog/standards/sist/7c89990b-ccee-44de-871d-344a613a6ff4/sist-en-13624-2004](https://standards.iteh.ai/catalog/standards/sist/7c89990b-ccee-44de-871d-344a613a6ff4/sist-en-13624-2004)

The neutralizer shall be validated for the product being tested in accordance with 5.5.1 and 5.5.2. The neutralizer shall be sterile.

NOTE Information on neutralizers that have been found to be suitable for some categories of products is given in B.1.

**5.2.2.6 Rinsing liquid (for membrane filtration)**

The liquid shall be sterile, compatible with the filter membrane and capable of filtration through the filter membrane under the test conditions described in 5.5.3.

NOTE Information on rinsing liquids that have been found to be suitable for some categories of products is given in B.2.

**5.2.2.7 Hard water for dilution of products**

Prepare :

- solution A : dissolve 19,84 g anhydrous magnesium chloride (MgCl<sub>2</sub>) or an equivalent of hydrated magnesium chloride and 46,24 g anhydrous calcium chloride (CaCl<sub>2</sub>) or an equivalent of hydrated calcium chloride in water (see 5.2.2.2) and dilute to 1 000 ml . Sterilize in the autoclave (see 5.3.1). Store the solution at 2 °C to 8 °C for no longer than one month ;
- solution B : dissolve 35,02 g sodium hydrogencarbonate (NaHCO<sub>3</sub>) in water (see 5.2.2.2) and dilute to 1 000 ml. Sterilize by membrane filtration (see 5.3.2.7). Store the solution at 2 °C to 8 °C for no longer than one week ;

- hard water : for the preparation of 1 litre, place 600 ml to 700 ml water (see 5.2.2.2) in a 1000 ml volumetric flask (see 5.3.2.12) and add 6,0 ml of solution A, then 8,0 ml of solution B. Mix and dilute to 1 000 ml with water (see 5.2.2.2). The pH of the hard water shall be  $7,0 \pm 0,2$ . If necessary adjust the pH by using a solution of approximately 40 g/L (about 1 mol/L) of sodium hydroxide (NaOH) or approximately 36,5 g/L (about 1 mol/L) of hydrochloric acid (HCl). The hard water shall be freshly prepared under aseptic conditions and used within 12 hours.

NOTE When preparing the product test solutions (see 5.4.2) the addition of the product to this hard water produces a different final water hardness in each test tube. In any case the final hardness is lower than 300 mg/L of calcium carbonate (CaCO<sub>3</sub>) in the test tube.

### 5.2.2.8 Interfering substances

#### 5.2.2.8.1 General

The interfering substance shall be chosen according to the conditions of use laid down for the product.

The interfering substance shall be sterile and prepared at 10 times its final concentration in the test.

The ionic composition e.g. pH, calcium and/or magnesium hardness and chemical composition e.g. mineral substances, protein, carbohydrates, lipids, detergents shall be defined.

NOTE In the following the term "interfering substance" is used even if it contains more than one substance.

#### 5.2.2.8.2 Clean conditions (bovine albumin solution – low concentration)

Dissolve 0,30 g of bovine albumin fraction V (suitable for microbiological purposes) in 100 ml of diluent (see 5.2.2.4).

Sterilize by membrane filtration (see 5.3.2.7), keep in a refrigerator (at 2 °C to 8 °C) and use within 1 month.

The final concentration of the bovine albumin in the test procedure (see 5.5) is 0,3 g/l.

#### 5.2.2.8.3 Dirty conditions (Mixture of bovine albumin solutions – high concentration with sheep erythrocytes)

Dissolve 3,00 g of bovine albumin fraction V (suitable for microbiological purposes) in 97 ml of diluent (see 5.2.2.4).

Sterilize by membrane filtration (see 5.3.2.7).

Prepare at least 8,0 ml fresh sterile defibrinated sheep blood according to ISO 6710 or purchase such blood from a commercial supplier (see 5.2.2.8.4). Centrifuge the erythrocytes at 800 g for 10 min. After discarding the supernatant, resuspend erythrocytes in diluent (see 5.2.2.4). Repeat this procedure at least 3 times, until the supernatant is colourless. Resuspend 3 ml of the packed sheep erythrocytes in the 97 ml of sterilized bovine albumin solution (see above). To avoid later contamination this mixture should be split in portions probably needed per day and kept in separate containers for a maximum of 7 days in a refrigerator at 2 °C to 8 °C.

The final concentration of bovine albumin and sheep erythrocytes in the test procedure (see 5.5) shall be 3 g/l and 3 ml/l respectively.

#### 5.2.2.8.4 Sterile defibrinated sheep blood

The sterile defibrinated sheep blood can be acquired from a commercial supplier or prepared according to ISO 6710.

## EN 13624:2003 (E)

## 5.3 Apparatus and glassware

## 5.3.1 General

Sterilize all glassware and parts of the apparatus that will come into contact with the culture media and reagents or the sample, except those which are supplied sterile, by one of the following methods :

- a) by moist heat, in the autoclave (see 5.3.2.1 a) by maintaining it at  $(121_0^{+3})$  °C for a minimum holding time of 15 min ;
- b) by dry heat, in the hot air oven (see 5.3.2.1 b) by maintaining it at 180 °C for a minimum holding time of 30 min, at 170 °C for a minimum holding time of 1 h or at 160 °C for a minimum holding time of 2 h.

**5.3.2 Usual microbiological laboratory equipment<sup>2)</sup>** and in particular, the following :

## 5.3.2.1 Apparatus for sterilization :

- a) for moist heat sterilization, an autoclave capable of being maintained at  $(121_0^{+3})$  °C for a minimum holding time of 15 min ;
- b) for dry heat sterilization, a hot air oven capable of being maintained at 180 °C for a minimum holding time of 30 min, at 170 °C for a minimum holding time of 1 h or at 160 °C for a minimum holding time of 2 h.

**5.3.2.2 Water baths** capable of being controlled at  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ , at  $45^{\circ}\text{C} \pm 1^{\circ}\text{C}$  (to maintain melted TSA in case of pour plate technique) and at additional test temperatures  $\pm 1^{\circ}\text{C}$  (see 5.5.1).

**5.3.2.3 Incubator**, capable of being controlled at  $30^{\circ}\text{C} \pm 1^{\circ}\text{C}$ .

**5.3.2.4 pH-meter**, having an inaccuracy of calibration of not more than  $\pm 0,1$  pH units at 20 °C.

The additional inaccuracy of the electrodes shall not be more than  $\pm 0,1$  pH units.

NOTE For measuring the pH of the agar-media (see 5.2.2.3) a puncture electrode or a flat membrane electrode should be used.

## 5.3.2.5 Stopwatch

**5.3.2.6 Electromechanical agitator** e.g. Vortex<sup>®</sup> mixer <sup>3)</sup>.

**5.3.2.7 Membrane filtration apparatus** constructed of a material compatible with the product, with a filter holder of at least 50 ml volume, and suitable for use of filters with diameter 47 mm to 50 mm and 0,45 µm pore size (or 0,22 µm pore size for sterilization of hard water - see 5.2.2.7 and albumin – see 5.2.2.8).

The vacuum source used shall give an even filtration flow rate. In order to obtain a uniform distribution of the micro-organisms over the membrane and in order to prevent overlong filtration.

The device shall be set so as to obtain the filtration of 100 ml of rinsing liquid in 20 s to 40 s.

---

2) Disposable equipment is an acceptable alternative to reusable glassware.

3) Vortex<sup>®</sup> is an example of a suitable product available commercially. This information is given for the convenience of users of this standard and does not constitute an endorsement by CEN of this product.

**5.3.2.8 Containers** : test tubes, culture bottles or flasks of suitable capacity.

**5.3.2.9 Graduated pipettes** of nominal capacities 10 ml and 1 ml and 0,1 ml.

Calibrated automatic pipettes may be used.

**5.3.2.10 Petri dishes** of size 90 mm to 100 mm.

**5.3.2.11 Glass beads** (Diameter : 3 mm to 4 mm).

**5.3.2.12 Volumetric flasks**

**5.3.2.13 Orbital mechanical shaker**

**5.3.2.14 Centrifuge** (800 g<sub>N</sub> and 2000 g<sub>N</sub>).

**5.3.2.15 Roux bottles or similar flasks**

**5.3.2.16 Fritted filter**, porosity of 40 µm to 100 µm (see ISO 4793).

## 5.4 Preparation of test organism suspensions and product test solutions

**5.4.1 Test organism suspensions (Test and validation suspension)**

**5.4.1.1 General**

Two different test organism suspensions have to be prepared: the “test suspension” to perform the test and the “validation suspension” to perform the controls and method validation.

**5.4.1.2 Stock culture of test organisms**

The test organisms and their stock cultures shall be prepared and kept in accordance with the requirements of EN 12353.

**5.4.1.3 Working culture of test organisms**

**5.4.1.3.1 *Candida albicans* (yeast)**

In order to prepare the working culture of *Candida albicans* (see 5.2.1), subculture from the stock culture (see 5.4.1.1) by streaking onto MEA slopes or plates (see 5.2.2.3) and incubate (see 5.3.2.3). After 42 h to 48 h prepare a second subculture from the first subculture in the same way and incubate for 42 h to 48 h. From this second subculture, a third subculture may be produced in the same way. The second and/or third subculture is/are the working culture(s).

If it is not possible to prepare the second subculture on a particular day, a 72 h subculture may be used for subsequent subculturing, provided that the subculture has been kept in the incubator (see 5.3.2.3) during the 72 h period.

Never produce and use a fourth subculture.

**5.4.1.3.2 *Aspergillus niger* (mould)**

For *Aspergillus niger* use only the first subculture grown on MEA (see 5.2.2.3) in Roux bottles (see 5.3.2.15) and incubate for 7-9 days. No further subculturing is needed.