

Edition 3.0 2019-02

# **INTERNATIONAL STANDARD**

# NORME **INTERNATIONALE**

#### **GROUP SAFETY PUBLICATION** PUBLICATION GROUPÉE DE SÉCURITÉ

Safety requirements for electrical equipment for FVFW measurement, control, and laboratory use – Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

https://standards.iteh.ai/catalog/standards/sist/7e11f4f0-c308-4aaf-b79c-Exigences de sécurité pouraappareils électriques de mesurage, de régulation et de laboratoire -

Partie 2-081: Exigences particulières pour les appareils de laboratoire, automatiques et semi-automatiques, destinés à l'analyse et à d'autres usages





## THIS PUBLICATION IS COPYRIGHT PROTECTED Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office 3, rue de Varembé CH-1211 Geneva 20 Switzerland Tel.: +41 22 919 02 11 info@iec.ch www.iec.ch

#### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

# IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available <u>online and</u>)-2 once a month by email. https://standards.iteh.ai/catalog/standa

IEC Customer Service Centre - webstore.1ac.ch/csc146/iec-6 If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French-extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

#### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

#### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

#### webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

#### Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.



Edition 3.0 2019-02

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

#### GROUP SAFETY PUBLICATION PUBLICATION GROUPÉE DE SÉCURITÉ

Safety requirements for electrical equipment for EVIEW measurement, control, and laboratory use <u>Teh.ai</u>) Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

https://standards.iteh.ai/catalog/standards/sist/7e11f4f0-c308-4aaf-b79c-

Exigences de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire –

Partie 2-081: Exigences particulières pour les appareils de laboratoire, automatiques et semi-automatiques, destinés à l'analyse et à d'autres usages

INTERNATIONAL ELECTROTECHNICAL COMMISSION

COMMISSION ELECTROTECHNIQUE INTERNATIONALE

ICS 19.080; 71.040.10

ISBN 978-2-8322-6501-7

Warning! Make sure that you obtained this publication from an authorized distributor. Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

 Registered trademark of the International Electrotechnical Commission Marque déposée de la Commission Electrotechnique Internationale

# CONTENTS

FOR	EWORD	3	
1	Scope and object	5	
2	Normative references	6	
3	Terms and definitions	6	
4	Tests	6	
5	Marking and documentation	6	
6	Protection against electric shock	8	
7	Protection against mechanical HAZARDS	8	
8	Resistance to mechanical stresses	8	
9	Protection against the spread of fire	9	
10	Equipment temperature limits and resistance to heat	9	
11	Protection against HAZARDS from fluids and solid foreign objects	9	
12	Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	9	
13	Protection against liberated gases and substances, explosion and implosion	9	
14	Components and subassemblies	9	
15	Protection by interlocks STANDARD PREVIEW	9	
16	HAZARDS resulting from application dama suite h.a.	10	
17	RISK assessment	10	
Ann	Annexes IEC 61010-2-081:2019		
Bibli	ography	11	
Tabl	e 1 – Symbols	6	

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

# Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees. TANDARD PREVIEW
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their sational and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-081 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

It has the status of a group safety publication in accordance with IEC Guide 104.

This third edition cancels and replaces the second edition published in 2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- adaptation of changes introduced by Amendment 1 of IEC 61010-1:2010;
- added tolerance for stability of AC voltage test equipment to Clause 6.

The text of this International Standard is based on the following documents:

CDV	Report on voting
66/652/CDV	66/671A/RVC

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 61010 series, published under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This Part 2-081 is to be used in conjunction with IEC 61010-1. It was established on the basis of the third edition (2010) and its Amendment 1 (2016), hereinafter referred to as Part 1.

This Part 2-081 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes.* 

Where a particular subclause of Part 1 is not mentioned in this Part 2-081, that subclause applies as far as is reasonable. Where this Part 2-081 states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in Part 1 should be adapted accordingly. **Standards.iten.al** 

In this standard:

IEC 61010-2-081:2019 https://standards.iteh.ai/catalog/standards/sist/7e11f4f0-c308-4aaf-b79c-

1) the following print types are used: 7b6146/iec-61010-2-081-2019

- requirements: in roman type;
- NOTES: in smaller roman type;
- conformity and test: in italic type;
- terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS.
- 2) subclauses, figures, tables and notes which are additional to those in Part 1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

# Part 2-081: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes

#### 1 Scope and object

This clause of Part 1 is applicable except as follows:

#### 1.1.1 Equipment included in scope

Replacement:

Replace the text, except the first paragraph, by the following new text:

This part of IEC 61010 applies to automatic and semi-automatic laboratory equipment for analysis and other purposes.

Automatic and semi-automatic laboratory equipment consists of instruments or systems for measuring or modifying one or more characteristics or parameters of samples, performing the complete process or parts of the process without manual intervention. Equipment forming part of such a system is within the scope of this document.

Examples of equipment within the scope of this document include: 44af-b79c-

- analytical equipment; 0a654e7b6146/iec-61010-2-081-2019
- automatic sampler (pipettor, aliquoter);
- equipment for sample replication and amplification.

NOTE 1 In the case of analytical equipment, the complete process usually includes the following steps:

- taking a specific quantity of the sample;
- preparing the sample by chemical, thermal, mechanical or other means;
- measurement;
- display, transmission or printing of the results of measurement.

NOTE 2 If all or part of the equipment falls within the scope of one or more other Part 2 documents of IEC 61010 as well as within the scope of this document, consideration is given to those other Part 2 documents.

#### 1.1.2 Equipment excluded from scope

Addition:

Add the following new item:

aa) IEC 61010-2-101 (in vitro diagnostic (IVD) equipment).

#### 1.2 Object

#### 1.2.1 Aspects included in scope

Addition:

Add the following new items:

- aa) biohazards;
- bb) hazardous chemical substances.

#### 1.2.2 Aspects excluded from scope

Addition:

Add the following new item and note:

aa) handling or manipulation of material outside the equipment.

NOTE Requirements covering these subjects are the responsibility of committees preparing the relevant standards.

- 6 -

## 2 Normative references

This clause of Part 1 is applicable **CANDARD PREVIEW** 

# 3 Terms and definitions (standards.iteh.ai)

This clause of Part 1 is applicable. IEC 61010-2-081:2019

https://standards.iteh.ai/catalog/standards/sist/7e11f4f0-c308-4aaf-b79c-0a654e7b6146/iec-61010-2-081-2019

### 4 Tests

This clause of Part 1 is applicable.

### 5 Marking and documentation

This clause of Part 1 is applicable except as follows:

### Table 1 – Symbols

Addition:

Add the following new symbol to Table 1:

Number		Symbol	Reference	Description
		Background colour – optional	ISO 7000-0659 (2004-01)	Biological risks
101		Symbol colour – optional		
		Outline / outline colour - optional		

Add the following new subclause:

#### 5.1.5.101 Gas and liquid connections

If necessary for safety, the equipment shall be clearly marked near the connector on the equipment with:

- a) a means of identifying the gas or liquid to be used. Where no internationally recognized symbol (including chemical formulae) exists, the equipment shall be marked with symbol 14 of Table 1;
- b) the maximum permitted pressure, or alternatively symbol 14 of Table 1 (see 5.4.3);
- c) flow direction of the gas and liquid, if applicable.

Conformity is checked by inspection.

#### 5.3 Durability of markings

Replacement:

Replace the first paragraph with the following new text:

Markings required by 5.1.2 to 5.2 shall be removable only with a TOOL or by appreciable force and shall remain clear and legible under conditions of NORMAL USE, and resist the effects of temperature and rubbing, and of solvent and reagents likely to be encountered in NORMAL USE, including cleaning and decontaminating agents specified by the manufacturer.

Addition:

# (standards.iteh.ai)

Add the following new paragraph after the second paragraph:

0a654e7b6146/iec-61010-2-081-2019 If a solvent or reagent specified for use with the equipment could affect the durability of a particular marking, that marking is also rubbed for 30 s with the most frequently used and/or aggressive solvent or reagent to which the equipment is likely to be exposed in NORMAL USE. A representative sample of groups of solvents or reagents likely to have a similar effect can optionally be used.

#### 5.4.1 General

Deletion:

Delete Note 2.

#### 5.4.4 Equipment operation

Addition:

Add the following note to item h):

NOTE Manufacturers can find valuable details in the internationally recognized Laboratory Biosafety Manual, published by the World Health Organization. This gives information on decontaminants, their use, dilutions and potential applications. There are also national guidelines that cover these areas.

Addition:

Add the following new subclause:

#### 5.4.101 Removal of equipment from use for repair or disposal

Instructions shall be provided to the RESPONSIBLE BODY for eliminating or reducing HAZARDS involved in removal from use, transportation or disposal, or appropriate contact information shall be provided in the instructions.

NOTE Regional or international requirements can apply.

Conformity is checked by inspection of the documentation.

#### 6 **Protection against electric shock**

This clause of Part 1 is applicable except as follows:

#### 6.8.3.1 The AC voltage test

Replacement:

Replace the first sentence by the following new sentence:

The voltage tester shall be capable of maintaining the test voltage throughout the test within ±5 % of the specified value.

#### Protection against mechanical HAZARDSD PREVIEW 7

(standards.iteh.ai) This clause of Part 1 is applicable except as follows:

IEC 61010-2-081:2019 7.3.2 Exceptions https://standards.iteh.ai/catalog/standards/sist/7e11f4f0-c308-4aaf-b79c-0a654e7b6146/iec-61010-2-081-2019

Replacement:

Replace the text in item b) 3) by the following new text:

there are warning markings prohibiting access by untrained OPERATORS. Markings shall be placed within the area requiring maintenance where they can alert the OPERATOR to the HAZARD. As an alternative, symbol 14 of Table 1 can be used, with the warnings included in the documentation.

#### Addition:

Add the following new item:

b) 4) there are OPERATOR maintenance instructions that specify safe maintenance procedures.

#### Resistance to mechanical stresses 8

This clause of Part 1 is applicable except as follows:

#### 8.1 General

Replacement:

Replace the text of item 3) by the following new text:

3) except for FIXED EQUIPMENT, for equipment with a mass over 100 kg, or for equipment whose size and weight make unintentional movement unlikely and which is not moved in NORMAL USE, the appropriate test of 8.3. The equipment is not operated during the tests.

#### 9 Protection against the spread of fire

This clause of Part 1 is applicable.

#### **10** Equipment temperature limits and resistance to heat

This clause of Part 1 is applicable.

#### **11** Protection against HAZARDS from fluids and solid foreign objects

This clause of Part 1 is applicable.

# 12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure

This clause of Part 1 is applicable.

# 13 Protection against liberated gases and substances, explosion and implosion (standards.iteh.ai)

This clause of Part 1 is applicable except as follows:

IEC 61010-2-081:2019

Addition:

**n:** https://standards.iteh.ai/catalog/standards/sist/7e11f4f0-c308-4aaf-b79c-0a654e7b6146/iec-61010-2-081-2019

Add the following new subclause:

#### 13.101 Biohazardous substances

Equipment that can be potentially biohazardous due to the use of biohazardous substances shall be prominently marked with symbol 101 of Table 1, or the appropriate international symbol, or (if none is available) symbol 14 of Table 1.

At minimum, a biohazard symbol shall be located near the sampling area and be visible in NORMAL USE.

Biohazard symbols shall be located near biohazardous areas accessed during OPERATOR maintenance and visible only during this maintenance.

Any part of the equipment that contains biohazardous waste material which can be removed from the equipment during NORMAL USE or a biohazardous drain connection shall be marked with an appropriate biohazard symbol.

#### **14** Components and subassemblies

This clause of Part 1 is applicable.

#### **15 Protection by interlocks**

This clause of Part 1 is applicable except as follows.