

INTERNATIONAL STANDARD



Photovoltaic devices –
Part 3: Measurement principles for terrestrial photovoltaic (PV) solar devices
with reference spectral irradiance data

Document Preview

IEC 60904-3:2019

<https://standards.iteh.ai/catalog/standards/iec/1f635743-6549-4576-8448-ea6ce64a868b/iec-60904-3-2019>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC 60904-3:2019

<https://standards.iteh.ai/catalog/standards/iec/1f635743-6549-4576-8448-ea6ce64a868b/iec-60904-3-2019>



IEC 60904-3

Edition 4.0 2019-02
REDLINE VERSION

INTERNATIONAL STANDARD



Photovoltaic devices –
Part 3: Measurement principles for terrestrial photovoltaic (PV) solar devices
with reference spectral irradiance data

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

IEC 60904-3:2019

<https://standards.itih.ai/catalog/standards/iec/1f635743-6549-4576-8448-ea6ce64a868b/iec-60904-3-2019>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 27.160

ISBN 978-2-8322-6585-7

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD	3
1 Scope and object	5
2 Normative references	5
3 Terms and definitions	6
4 Measurement principles	6
5 Reference solar spectral and angular irradiance distribution	6
Annex A (informative) Use of SMARTS	64
Bibliography	64

Figure 1 – Global and direct reference solar spectral irradiance distribution listed in Table 1	107
---	-----

Table 1 – Reference solar spectral irradiance distribution	8
--	---

Table A.1 – Input data for generation of reference solar spectral irradiance distribution	108
---	-----

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

IEC 60904-3:2019

<https://standards.itih.ai/catalog/standards/iec/1f635743-6549-4576-8448-ea6ce64a868b/iec-60904-3-2019>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

PHOTOVOLTAIC DEVICES –

**Part 3: Measurement principles for terrestrial photovoltaic (PV)
solar devices with reference spectral irradiance data**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

International Standard IEC 60904-3 has been prepared by IEC technical committee 82: Solar photovoltaic energy systems.

This fourth edition cancels and replaces the third edition published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) all spectral data were recalculated due to some minor calculation and rounding errors in the third edition; the global spectral irradiance returned to exactly the data of the second edition;
- b) the angular distribution of the irradiance was clarified.

This publication contains an attached file in the form of an Excel spreadsheet. This file is intended to be used as a complement and does not form an integral part of the publication.

The text of this International Standard is based on the following documents:

CDV	Report on voting
82/1342/CDV	82/1425/RVC

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 60904 series, published under the general title *Photovoltaic devices*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

PHOTOVOLTAIC DEVICES – Part 3: Measurement principles for terrestrial photovoltaic (PV) solar devices with reference spectral irradiance data

1 ~~Scope and object~~

This part of IEC 60904 applies to the following photovoltaic devices for terrestrial applications:

- solar cells with or without a protective cover;
- sub-assemblies of solar cells;
- modules; and
- systems.

NOTE The term “test specimen” is used to denote any of these devices.

The principles contained in this document cover testing in both natural and simulated sunlight.

Photovoltaic conversion is spectrally selective due to the nature of the semiconductor materials used in PV solar cells and modules. To compare the relative performance of different PV devices and materials a reference standard solar spectral distribution is necessary. This document includes such a reference solar spectral irradiance distribution.

This document also describes basic measurement principles for determining the electrical output of PV devices. The principles given in this document are designed to relate the performance rating of PV devices to a common reference terrestrial solar spectral irradiance distribution.

The reference terrestrial solar spectral irradiance distribution is given in this document in order to classify solar simulators according to the spectral performance requirements contained in IEC 60904-9.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60891, *Photovoltaic devices – Procedures for temperature and irradiance corrections to measured I-V characteristics*

IEC 60904-1, *Photovoltaic devices – Part 1: Measurements of photovoltaic current-voltage characteristics*

IEC 60904-2, *Photovoltaic devices – Part 2: Requirements for photovoltaic reference devices*

IEC 60904-5, *Photovoltaic devices – Part 5: Determination of the equivalent cell temperature (ECT) of photovoltaic (PV) devices by the open-circuit voltage method*

IEC 60904-7, *Photovoltaic devices – Part 7: Computation of the spectral mismatch correction for measurements of photovoltaic devices*

IEC 60904-8, *Photovoltaic devices – Part 8: Measurement of spectral responsivity of a photovoltaic (PV) device*

IEC 60904-9, *Photovoltaic devices – Part 9: Solar simulator performance requirements*

IEC 61853-4, *Photovoltaic (PV) module performance testing and energy rating - Part 4: Standard reference climatic profiles*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Measurement principles

In current practice the photovoltaic performance of a solar cell or module is determined by exposing it at a known temperature to stable sunlight, natural or simulated, and measuring its current-voltage (I-V) characteristic curve while measuring the magnitude of both the incident irradiance and the PV device temperature. Detailed I-V curve measurement procedures are included in IEC 60904-1. The measured performances can then be corrected to standard test conditions (STC) or other desired conditions of irradiance and temperature according to IEC 60891. The corrected power output at the maximum power voltage and STC is commonly referred to as the rated power.

Incident irradiance can be measured by means of a PV reference device (whose spectral responsivity shall be known) or, if measuring under natural sunlight, by means of a thermopile-type irradiance detector (pyranometer). If a PV reference device is used, it shall satisfy the requirements specified in IEC 60904-2. Temperature determination of the PV device under test shall be made according to IEC 60904-1 or IEC 60904-5.

Since a solar cell has a wavelength-dependent ~~response~~ responsivity, its performance is significantly affected by the spectral distribution of the incident radiation, which in natural sunlight varies with factors such as location, weather, time of year, time of day, orientation of the receiving surface, etc., and with a simulator varies with its type and conditions of use according to IEC 60904-9. Regardless of whether the irradiance is measured with either a thermopile-type radiometer (that is not spectrally selective) or with a reference solar device, the spectral irradiance distribution of the incoming light shall be known in order to use IEC 60904-7 to calculate the spectral mismatch between the measured performance and the predicted performance under the global or direct reference solar spectral distribution defined in this document. When performing measurements according to the standard test conditions, the spectral mismatch between the PV device under test and the thermopile is usually much larger than the spectral mismatch between the PV reference device and the PV device under test.

When the spectral responsivity of the PV device is known as determined according to IEC 60904-8, it is also possible to use IEC 60904-7 to compute the performance of that PV device when exposed to light of any known spectral irradiance distribution.

5 Reference solar spectral and angular irradiance distribution

The reference solar spectral distributions ~~AM1.5~~ are given in Table 1 and Figure 1.

The reference angular distribution is *defined* for both, the direct and the global spectral distribution, so that the complete radiation hits the solar device perpendicularly under normal incidence. Other, more realistic geometrical and ambient conditions for different climate zones are defined in IEC 61853-4.

The *calculation* of the spectral distributions was done using the following geometrical and environmental input parameters:

- global distribution (direct + diffuse + albedo) of sunlight (AM1.5 spectrum), corresponding to an integrated irradiance of $1\,000\text{ W}\cdot\text{m}^{-2}$ incident on a sun-facing plane surface tilted at 37° to the horizontal, and
- the direct distribution of sunlight (AM1.5d spectrum), corresponding to an integrated irradiance of $900\text{ W}\cdot\text{m}^{-2}$ incident on a sun-facing plane surface perpendicular to the incident sunlight,

considering the wavelength-dependent albedo of a light bare soil, under the following atmospheric conditions:

- U.S. standard atmosphere with CO_2 concentration increased to current level (370 ppm), a rural aerosol model, and no pollution;
- precipitable water: $1,416\,4\text{ cm}$;
- ozone content: $0,343\,8\text{ atm}\cdot\text{cm}$ (or $343,8\text{ DU}$);
- turbidity (aerosol optical depth): $0,084$ at 500 nm ;
- pressure: $101,325\text{ kPa}$ (i.e. sea level).

Data contained in Table 1 have been generated using the solar spectral model SMARTS, Version 2.9.2. A general description of this model and its suitability to reproduce actual solar spectral irradiance distributions can be found in “Proposed Reference Irradiance Spectra for Solar Energy Systems Testing” by C. A. Gueymard, C. Myers and K. Emery¹, and in the references therein. Table 1 can be obtained using the data contained in Annex A as an input to the model SMARTS Version 2.9.2. The resulting output spectral irradiance values have to be multiplied by a normalization factor (~~$0,997\,08$~~ $0,997\,1$) in order to get an integrated irradiance of $1\,000\text{ W}\cdot\text{m}^{-2}$ in the wavelength range 0 to infinity for the global irradiance. This same scaling factor is applied to the direct spectrum giving an integrated irradiance of $900\text{ W}\cdot\text{m}^{-2}$ in the wavelength range 0 to infinity.

At the time of publication of this document the SMARTS Version 2.9.2 spectral model code is available, free of charge, subject to the author's license agreement, at <http://www.nrel.gov/rredc/smarts>. A copy of the model, not for distribution purposes, is kept under IEC TC 82 control.

The contents of Table 1 are included in an attached file in the form of an Excel spreadsheet.

¹ C. A. Gueymard, C. Myers and K. Emery, “Proposed Reference Irradiance Spectra for Solar Energy Systems Testing”, *Solar Energy*, Vol 73, No. 6, pp. 443-467, 2002.

Table 1 – Reference solar spectral irradiance distribution

Wave-length (nm)	Global spectral irradiance (W·m ⁻² ·nm ⁻¹)	Global photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative global integrated irradiance (W·m ⁻²)	Direct spectral irradiance (W·m ⁻² ·nm ⁻¹)	Direct photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative direct integrated irradiance (W·m ⁻²)
280,0	4,717E-23	6,649E-5	0,00E+0	2,529E-26	3,564E-8	0,00E+0
280,5	1,227E-21	1,733E-3	3,19E-22	1,089E-24	1,537E-6	2,78E-25
281,0	5,673E-21	8,025E-3	2,04E-21	6,107E-24	8,640E-6	2,08E-24
281,5	1,562E-19	2,213E-1	4,25E-20	2,740E-22	3,883E-4	7,21E-23
282,0	1,191E-18	1,691E+0	3,79E-19	2,826E-21	4,012E-3	8,47E-22
282,5	4,530E-18	6,443E+0	2,53E-18	1,323E-20	1,882E-2	6,87E-21
283,0	1,840E-17	2,621E+1	1,04E-17	6,745E-20	9,609E-2	3,51E-20
283,5	3,526E-17	5,032E+1	2,77E-17	1,457E-19	2,080E-1	1,05E-19
284,0	7,246E-16	1,036E+3	3,06E-16	4,969E-18	7,105E+0	2,00E-18
284,5	2,478E-15	3,550E+3	1,41E-15	2,156E-17	3,088E+1	1,13E-17
285,0	7,991E-15	1,147E+4	4,94E-15	8,974E-17	1,288E+2	4,97E-17
285,5	4,249E-14	6,107E+4	2,26E-14	6,424E-16	9,232E+2	3,10E-16
286,0	1,364E-13	1,964E+5	8,33E-14	2,343E-15	3,374E+3	1,34E-15
286,5	8,358E-13	1,205E+6	4,26E-13	1,840E-14	2,654E+4	8,75E-15
287,0	2,729E-12	3,942E+6	1,64E-12	7,234E-14	1,045E+5	4,02E-14
287,5	1,087E-11	1,573E+7	6,30E-12	3,651E-13	5,284E+5	1,93E-13
288,0	6,216E-11	9,011E+7	3,20E-11	2,798E-12	4,057E+6	1,32E-12
288,5	1,711E-10	2,485E+8	1,10E-10	9,039E-12	1,313E+7	5,37E-12
289,0	5,610E-10	8,162E+8	3,56E-10	3,488E-11	5,074E+7	2,04E-11
289,5	2,069E-9	3,015E+9	1,25E-9	1,532E-10	2,233E+8	8,54E-11
290,0	5,999E-9	8,758E+9	3,95E-9	5,130E-10	7,490E+8	3,12E-10
290,5	1,374E-8	2,010E+10	1,03E-8	1,326E-9	1,940E+9	9,18E-10
291,0	3,495E-8	5,120E+10	2,61E-8	3,885E-9	5,691E+9	2,64E-9
291,5	1,088E-7	1,597E+11	7,40E-8	1,438E-8	2,111E+10	8,84E-9
292,0	2,675E-7	3,932E+11	1,97E-7	4,067E-8	5,978E+10	2,72E-8
292,5	4,256E-7	6,267E+11	4,10E-7	7,021E-8	1,034E+11	6,19E-8
293,0	8,621E-7	1,272E+12	8,06E-7	1,571E-7	2,318E+11	1,33E-7
293,5	2,264E-6	3,345E+12	1,82E-6	4,696E-7	6,938E+11	3,40E-7
294,0	4,162E-6	6,160E+12	3,84E-6	9,428E-7	1,395E+12	7,91E-7
294,5	6,572E-6	9,743E+12	7,06E-6	1,592E-6	2,360E+12	1,57E-6
295,0	1,225E-5	1,820E+13	1,28E-5	3,215E-6	4,775E+12	3,05E-6
295,5	2,775E-5	4,127E+13	2,54E-5	7,997E-6	1,190E+13	6,65E-6
296,0	4,776E-5	7,117E+13	4,87E-5	1,469E-5	2,190E+13	1,38E-5
296,5	7,114E-5	1,062E+14	8,39E-5	2,324E-5	3,469E+13	2,52E-5
297,0	9,652E-5	1,443E+14	1,32E-4	3,309E-5	4,947E+13	4,15E-5
297,5	1,855E-4	2,779E+14	2,17E-4	6,771E-5	1,014E+14	7,23E-5
298,0	2,890E-4	4,336E+14	3,59E-4	1,110E-4	1,664E+14	1,27E-4
298,5	3,569E-4	5,362E+14	5,42E-4	1,423E-4	2,138E+14	1,99E-4
299,0	4,907E-4	7,386E+14	7,79E-4	2,026E-4	3,050E+14	2,97E-4

Wave-length (nm)	Global spectral irradiance (W·m ⁻² ·nm ⁻¹)	Global-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative global integrated irradiance (W·m ⁻²)	Direct-spectral irradiance (W·m ⁻² ·nm ⁻¹)	Direct-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative direct integrated irradiance (W·m ⁻²)
299,5	8,582E-4	1,294E+15	1,18E-3	3,728E-4	5,620E+14	4,70E-4
300,0	1,018E-3	1,537E+15	0,00	4,550E-4	6,871E+14	0,00
300,5	1,241E-3	1,878E+15	0,00	5,704E-4	8,629E+14	0,00
301,0	1,924E-3	2,916E+15	0,00	9,166E-4	1,389E+15	0,00
301,5	2,684E-3	4,073E+15	0,00	1,316E-3	1,998E+15	0,00
302,0	2,912E-3	4,428E+15	0,01	1,453E-3	2,209E+15	0,00
302,5	4,272E-3	6,505E+15	0,01	2,185E-3	3,327E+15	0,00
303,0	7,074E-3	1,079E+16	0,01	3,722E-3	5,678E+15	0,01
303,5	8,953E-3	1,368E+16	0,02	4,790E-3	7,319E+15	0,01
304,0	9,443E-3	1,445E+16	0,02	5,082E-3	7,778E+15	0,01
304,5	1,192E-2	1,827E+16	0,03	6,449E-3	9,885E+15	0,01
305,0	1,642E-2	2,520E+16	0,03	8,908E-3	1,368E+16	0,02
305,5	1,866E-2	2,870E+16	0,04	1,016E-2	1,562E+16	0,02
306,0	1,852E-2	2,853E+16	0,05	1,012E-2	1,559E+16	0,03
306,5	2,105E-2	3,247E+16	0,06	1,153E-2	1,780E+16	0,03
307,0	2,777E-2	4,291E+16	0,08	1,520E-2	2,349E+16	0,04
307,5	3,553E-2	5,500E+16	0,09	1,941E-2	3,005E+16	0,05
308,0	3,773E-2	5,850E+16	0,11	2,069E-2	3,208E+16	0,06
308,5	4,131E-2	6,415E+16	0,13	2,268E-2	3,523E+16	0,07
309,0	4,042E-2	6,287E+16	0,16	2,223E-2	3,458E+16	0,08
309,5	4,318E-2	6,728E+16	0,18	2,360E-2	3,678E+16	0,10
310,0	5,079E-2	7,926E+16	0,20	2,775E-2	4,330E+16	0,11
310,5	6,535E-2	1,022E+17	0,23	3,577E-2	5,592E+16	0,13
311,0	8,268E-2	1,294E+17	0,27	4,526E-2	7,086E+16	0,15
311,5	8,384E-2	1,315E+17	0,32	4,602E-2	7,217E+16	0,17
312,0	9,310E-2	1,462E+17	0,36	5,075E-2	7,971E+16	0,20
312,5	9,870E-2	1,553E+17	0,41	5,361E-2	8,434E+16	0,23
313,0	1,070E-1	1,686E+17	0,47	5,815E-2	9,163E+16	0,25
313,5	1,073E-1	1,693E+17	0,52	5,883E-2	9,284E+16	0,28
314,0	1,193E-1	1,886E+17	0,58	6,508E-2	1,029E+17	0,32
314,5	1,302E-1	2,062E+17	0,65	7,027E-2	1,113E+17	0,35
315,0	1,359E-1	2,154E+17	0,71	7,347E-2	1,165E+17	0,39
315,5	1,180E-1	1,875E+17	0,78	6,464E-2	1,027E+17	0,42
316,0	1,231E-1	1,959E+17	0,83	6,689E-2	1,064E+17	0,45
316,5	1,499E-1	2,389E+17	0,91	8,088E-2	1,289E+17	0,49
317,0	1,711E-1	2,730E+17	0,99	9,275E-2	1,480E+17	0,54
317,5	1,819E-1	2,908E+17	1,09	9,942E-2	1,589E+17	0,59
318,0	1,754E-1	2,808E+17	1,17	9,554E-2	1,529E+17	0,64
318,5	1,854E-1	2,972E+17	1,27	9,976E-2	1,600E+17	0,69
319,0	2,041E-1	3,278E+17	1,37	1,094E-1	1,757E+17	0,74
319,5	1,953E-1	3,142E+17	1,47	1,066E-1	1,715E+17	0,80

Wave-length (nm)	Global spectral irradiance (W·m ⁻² ·nm ⁻¹)	Global-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative global integrated irradiance (W·m ⁻²)	Direct-spectral irradiance (W·m ⁻² ·nm ⁻¹)	Direct-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative direct integrated irradiance (W·m ⁻²)
320,0	2,047E-1	3,297E+17	1,57	1,124E-1	1,811E+17	0,85
320,5	2,445E-1	3,945E+17	1,69	1,327E-1	2,140E+17	0,92
321,0	2,495E-1	4,032E+17	1,82	1,338E-1	2,161E+17	0,99
321,5	2,377E-1	3,848E+17	1,94	1,278E-1	2,068E+17	1,05
322,0	2,214E-1	3,589E+17	2,05	1,216E-1	1,972E+17	1,11
322,5	2,165E-1	3,514E+17	2,15	1,194E-1	1,938E+17	1,17
323,0	2,116E-1	3,441E+17	2,26	1,159E-1	1,884E+17	1,23
323,5	2,479E-1	4,037E+17	2,38	1,335E-1	2,175E+17	1,29
324,0	2,746E-1	4,478E+17	2,52	1,481E-1	2,415E+17	1,37
324,5	2,824E-1	4,613E+17	2,66	1,542E-1	2,519E+17	1,45
325,0	2,781E-1	4,550E+17	2,80	1,546E-1	2,529E+17	1,52
325,5	3,234E-1	5,300E+17	2,96	1,788E-1	2,930E+17	1,61
326,0	3,801E-1	6,238E+17	3,15	2,081E-1	3,415E+17	1,71
326,5	4,060E-1	6,674E+17	3,35	2,210E-1	3,632E+17	1,83
327,0	3,969E-1	6,534E+17	3,56	2,177E-1	3,584E+17	1,94
327,5	3,835E-1	6,323E+17	3,75	2,122E-1	3,499E+17	2,04
328,0	3,501E-1	5,781E+17	3,93	1,972E-1	3,255E+17	2,14
328,5	3,706E-1	6,128E+17	4,10	2,062E-1	3,409E+17	2,24
329,0	4,211E-1	6,975E+17	4,31	2,323E-1	3,847E+17	2,36
329,5	4,674E-1	7,753E+17	4,54	2,579E-1	4,278E+17	2,49
330,0	4,700E-1	7,808E+17	4,79	2,612E-1	4,339E+17	2,62
330,5	4,268E-1	7,100E+17	5,00	2,403E-1	3,999E+17	2,74
331,0	4,015E-1	6,689E+17	5,20	2,277E-1	3,794E+17	2,86
331,5	4,168E-1	6,956E+17	5,41	2,357E-1	3,933E+17	2,97
332,0	4,350E-1	7,270E+17	5,62	2,444E-1	4,084E+17	3,09
332,5	4,379E-1	7,330E+17	5,84	2,458E-1	4,115E+17	3,22
333,0	4,282E-1	7,178E+17	6,06	2,419E-1	4,056E+17	3,34
333,5	4,061E-1	6,817E+17	6,26	2,320E-1	3,895E+17	3,46
334,0	4,138E-1	6,957E+17	6,47	2,375E-1	3,994E+17	3,57
334,5	4,438E-1	7,473E+17	6,69	2,536E-1	4,270E+17	3,70
335,0	4,625E-1	7,800E+17	6,92	2,640E-1	4,452E+17	3,83
335,5	4,518E-1	7,631E+17	7,15	2,582E-1	4,361E+17	3,96
336,0	4,140E-1	7,002E+17	7,36	2,374E-1	4,016E+17	4,08
336,5	3,810E-1	6,455E+17	7,55	2,204E-1	3,733E+17	4,19
337,0	3,727E-1	6,323E+17	7,73	2,170E-1	3,682E+17	4,30
337,5	3,993E-1	6,785E+17	7,93	2,337E-1	3,970E+17	4,41
338,0	4,328E-1	7,365E+17	8,14	2,525E-1	4,296E+17	4,54
338,5	4,539E-1	7,735E+17	8,37	2,647E-1	4,511E+17	4,67
339,0	4,622E-1	7,888E+17	8,60	2,702E-1	4,611E+17	4,81
339,5	4,731E-1	8,085E+17	8,84	2,777E-1	4,745E+17	4,95
340,0	5,003E-1	8,564E+17	9,09	2,957E-1	5,062E+17	5,09

Wave-length (nm)	Global spectral irradiance (W·m ⁻² ·nm ⁻¹)	Global-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative global integrated irradiance (W·m ⁻²)	Direct-spectral irradiance (W·m ⁻² ·nm ⁻¹)	Direct-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative direct integrated irradiance (W·m ⁻²)
340,5	4,993E-1	8,558E+17	9,34	2,959E-1	5,072E+17	5,24
341,0	4,700E-1	8,068E+17	9,58	2,785E-1	4,781E+17	5,38
341,5	4,680E-1	8,045E+17	9,81	2,777E-1	4,774E+17	5,52
342,0	4,879E-1	8,400E+17	10,05	2,904E-1	4,999E+17	5,66
342,5	5,062E-1	8,728E+17	10,31	3,021E-1	5,208E+17	5,82
343,0	5,134E-1	8,865E+17	10,56	3,077E-1	5,313E+17	5,97
343,5	4,847E-1	8,381E+17	10,81	2,916E-1	5,043E+17	6,12
344,0	4,172E-1	7,225E+17	11,02	2,528E-1	4,378E+17	6,25
344,5	4,019E-1	6,970E+17	11,22	2,437E-1	4,226E+17	6,37
345,0	4,576E-1	7,948E+17	11,44	2,777E-1	4,824E+17	6,50
345,5	4,879E-1	8,486E+17	11,69	2,967E-1	5,161E+17	6,65
346,0	4,764E-1	8,298E+17	11,93	2,905E-1	5,059E+17	6,80
346,5	4,852E-1	8,463E+17	12,17	2,966E-1	5,174E+17	6,95
347,0	4,926E-1	8,605E+17	12,42	3,023E-1	5,281E+17	7,10
347,5	4,754E-1	8,316E+17	12,66	2,927E-1	5,120E+17	7,24
348,0	4,737E-1	8,299E+17	12,89	2,922E-1	5,119E+17	7,39
348,5	4,820E-1	8,455E+17	13,13	2,980E-1	5,228E+17	7,54
349,0	4,643E-1	8,157E+17	13,37	2,878E-1	5,056E+17	7,68
349,5	4,767E-1	8,386E+17	13,60	2,963E-1	5,214E+17	7,83
350,0	5,264E-1	9,276E+17	13,86	3,282E-1	5,782E+17	7,99
350,5	5,658E-1	9,983E+17	14,14	3,537E-1	6,241E+17	8,17
351,0	5,501E-1	9,720E+17	14,43	3,450E-1	6,096E+17	8,35
351,5	5,287E-1	9,355E+17	14,69	3,329E-1	5,891E+17	8,51
352,0	5,164E-1	9,151E+17	14,95	3,258E-1	5,773E+17	8,67
352,5	4,882E-1	8,663E+17	15,19	3,086E-1	5,477E+17	8,83
353,0	5,189E-1	9,221E+17	15,45	3,288E-1	5,843E+17	8,99
353,5	5,706E-1	1,015E+18	15,73	3,625E-1	6,450E+17	9,17
354,0	6,032E-1	1,075E+18	16,03	3,841E-1	6,845E+17	9,36
354,5	6,098E-1	1,088E+18	16,34	3,893E-1	6,947E+17	9,56
355,0	6,096E-1	1,090E+18	16,65	3,903E-1	6,974E+17	9,76
355,5	5,886E-1	1,053E+18	16,94	3,777E-1	6,759E+17	9,95
356,0	5,523E-1	9,897E+17	17,22	3,552E-1	6,366E+17	10,12
356,5	5,179E-1	9,295E+17	17,48	3,340E-1	5,994E+17	10,29
357,0	4,554E-1	8,184E+17	17,71	2,944E-1	5,291E+17	10,44
357,5	4,608E-1	8,293E+17	17,93	2,986E-1	5,374E+17	10,59
358,0	4,288E-1	7,728E+17	18,15	2,786E-1	5,020E+17	10,73
358,5	3,981E-1	7,185E+17	18,35	2,592E-1	4,678E+17	10,86
359,0	4,682E-1	8,461E+17	18,57	3,056E-1	5,523E+17	11,00
359,5	5,638E-1	1,020E+18	18,85	3,691E-1	6,679E+17	11,18
360,0	5,964E-1	1,081E+18	19,16	3,913E-1	7,091E+17	11,38
360,5	5,637E-1	1,023E+18	19,45	3,706E-1	6,725E+17	11,58

Wave-length (nm)	Global spectral irradiance (W·m ⁻² ·nm ⁻¹)	Global-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative global integrated irradiance (W·m ⁻²)	Direct-spectral irradiance (W·m ⁻² ·nm ⁻¹)	Direct-photon flux (m ⁻² ·s ⁻¹ ·nm ⁻¹)	Cumulative direct integrated irradiance (W·m ⁻²)
361,0	5,187E-1	9,427E+17	19,71	3,418E-1	6,211E+17	11,75
361,5	5,081E-1	9,246E+17	19,96	3,355E-1	6,105E+17	11,91
362,0	5,326E-1	9,707E+17	20,22	3,525E-1	6,423E+17	12,09
362,5	5,834E-1	1,065E+18	20,51	3,869E-1	7,061E+17	12,28
363,0	6,002E-1	1,097E+18	20,81	3,989E-1	7,289E+17	12,48
363,5	5,837E-1	1,068E+18	21,11	3,888E-1	7,115E+17	12,68
364,0	6,045E-1	1,108E+18	21,41	4,035E-1	7,395E+17	12,87
364,5	5,988E-1	1,099E+18	21,71	4,006E-1	7,351E+17	13,08
365,0	6,218E-1	1,143E+18	22,02	4,169E-1	7,660E+17	13,28
365,5	6,843E-1	1,259E+18	22,35	4,598E-1	8,459E+17	13,51
366,0	7,332E-1	1,351E+18	22,72	4,936E-1	9,095E+17	13,76
366,5	7,344E-1	1,355E+18	23,10	4,955E-1	9,142E+17	14,01
367,0	7,207E-1	1,332E+18	23,46	4,873E-1	9,002E+17	14,25
367,5	7,071E-1	1,308E+18	23,81	4,790E-1	8,862E+17	14,49
368,0	6,656E-1	1,233E+18	24,15	4,519E-1	8,371E+17	14,72
368,5	6,612E-1	1,227E+18	24,47	4,497E-1	8,343E+17	14,94
369,0	6,911E-1	1,284E+18	24,81	4,711E-1	8,750E+17	15,18
369,5	7,425E-1	1,381E+18	25,18	5,071E-1	9,432E+17	15,43
370,0	7,529E-1	1,402E+18	25,56	5,152E-1	9,595E+17	15,69
370,5	6,806E-1	1,269E+18	25,92	4,666E-1	8,703E+17	15,93
371,0	6,914E-1	1,291E+18	26,25	4,749E-1	8,869E+17	16,16
371,5	7,184E-1	1,344E+18	26,61	4,944E-1	9,247E+17	16,41
372,0	6,725E-1	1,259E+18	26,95	4,637E-1	8,684E+17	16,64
372,5	6,407E-1	1,201E+18	27,27	4,426E-1	8,300E+17	16,86
373,0	6,171E-1	1,159E+18	27,58	4,271E-1	8,019E+17	17,08
373,5	5,562E-1	1,046E+18	27,86	3,857E-1	7,252E+17	17,27
374,0	5,548E-1	1,045E+18	28,13	3,854E-1	7,256E+17	17,46
374,5	5,507E-1	1,038E+18	28,41	3,832E-1	7,225E+17	17,65
375,0	5,876E-1	1,109E+18	28,70	4,097E-1	7,734E+17	17,85
375,5	6,497E-1	1,228E+18	29,02	4,538E-1	8,579E+17	18,08
376,0	6,728E-1	1,274E+18	29,36	4,708E-1	8,912E+17	18,32
376,5	6,620E-1	1,255E+18	29,69	4,640E-1	8,795E+17	18,55
377,0	7,102E-1	1,348E+18	30,04	4,987E-1	9,464E+17	18,80
377,5	7,922E-1	1,506E+18	30,43	5,573E-1	1,059E+18	19,07
378,0	8,535E-1	1,624E+18	30,86	6,014E-1	1,144E+18	19,37
378,5	8,318E-1	1,585E+18	31,29	5,872E-1	1,119E+18	19,68
379,0	7,417E-1	1,415E+18	31,67	5,246E-1	1,001E+18	19,94
379,5	6,649E-1	1,270E+18	32,00	4,712E-1	9,002E+17	20,18
380,0	6,987E-1	1,337E+18	32,34	4,961E-1	9,489E+17	20,42
380,5	7,486E-1	1,434E+18	32,71	5,324E-1	1,020E+18	20,68
381,0	7,616E-1	1,461E+18	33,09	5,427E-1	1,041E+18	20,96