



Edition 2.0 2019-03

TECHNICAL REPORT

Nuclear medicine instrumentation DRoutine tests EVIEW Part 4: Radionuclide calibrators (standards.iteh.ai)

> IEC TR 61948-4:2019 https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904ba3c6d79f9ba7/iec-tr-61948-4-2019





THIS PUBLICATION IS COPYRIGHT PROTECTED Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office 3, rue de Varembé CH-1211 Geneva 20 Switzerland Tel.: +41 22 919 02 11 info@iec.ch www.jec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished Stay up to date on all new IEC publications. Just Published

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore iecch/csc and collected If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch. IEC TR 61948-4:2019

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904b-

a3c6d79f9ba7/iec-tr-61948-4-2019





Edition 2.0 2019-03

TECHNICAL REPORT

Nuclear medicine instrumentation DRoutine testsEVIEW Part 4: Radionuclide calibrators standards.iteh.ai)

> IEC TR 61948-4:2019 https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904ba3c6d79f9ba7/iec-tr-61948-4-2019

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ICS 11.040.50

ISBN 978-2-8322-6602-1

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD				
1	Scop	e5		
2	Norm	Normative references5		
3	Terms and definitions5			
4	4 Test methods			
	4.1	BACKGROUND RESPONSE		
	4.2	Constancy of instrument response7		
	4.3	SYSTEM LINEARITY		
	4.3.1	General7		
	4.3.2	Decaying source method7		
	4.3.3	Data analysis7		
	4.4	Additional checks7		
	4.5	Frequency of ROUTINE TESTS		
Bi	Bibliography9			
In	Index of defined terms			

Table 1 – Frequency of ROUTINE TESTS NDARD PREVIEW (standards.iteh.ai)

IEC TR 61948-4:2019 https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904ba3c6d79f9ba7/iec-tr-61948-4-2019

INTERNATIONAL ELECTROTECHNICAL COMMISSION

NUCLEAR MEDICINE INSTRUMENTATION – ROUTINE TESTS –

Part 4: Radionuclide calibrators

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committee; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, EC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter. IEC TR 61948-4:2019
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC TR 61948-4, which is a Technical Report, has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2006. This edition constitutes a technical revision.

This edition includes the following significant technical change with respect to the previous edition: the test method to determine SYSTEM LINEARITY has been updated to reflect the technical developments of RADIONUCLIDE CALIBRATORS.

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
62C/715/DTR	62C/727/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

Terms used throughout this document that have been defined in Clause 3 appear in SMALL CAPITALS.

A list of all parts in the IEC 61948 series, published under the general title *Nuclear medicine Instrumentation – Routine tests*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or ANDARD PREVIEW
- amended.

(standards.iteh.ai)

A bilingual version of this publication may be issued at a later date.

IEC TR 61948-4:2019 https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904ba3c6d79f9ba7/iec-tr-61948-4-2019

NUCLEAR MEDICINE INSTRUMENTATION – ROUTINE TESTS –

Part 4: Radionuclide calibrators

1 Scope

This part of IEC 61948 covers the ROUTINE TESTING of RADIONUCLIDE CALIBRATORS used in nuclear medicine. Such devices utilise ionisation chambers of the well type and a direct readout in units of ACTIVITY. Requirements and specific methods to determine performance parameters are described in IEC 61303. These methods are primarily designed for ACCEPTANCE TESTING.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC TR 60788:2004, Medical electrical equipment – Glossary of defined terms

IEC 61303:1994, Medical electrical equipment Radionuclide calibrators – Particular methods for describing performance

IEC TR 61948-4:2019

3 Terms and definitions dards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904ba3c6d79f9ba7/iec-tr-61948-4-2019

For the purpose of this document, the terms and definitions given in IEC TR 60788, IEC 61303, and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at http://www.iso.org/obp

NOTE Defined terms are printed in small capital letters.

3.1

ACCEPTANCE TEST

test carried out after new equipment has been installed, or major modifications have been made to existing equipment, in order to verify compliance with contractual specifications

Note 1 to entry: During or immediately after ACCEPTANCE TEST, REFERENCE DATA are collected to be used as a standard for comparison with future ROUTINE TESTS.

[SOURCE: IEC TR 60788:2004, rm-70-01, modified –A note to entry has been added.]

3.2

BACKGROUND RESPONSE

reading of the instrument without intended RADIOACTIVE SOURCE

Note 1 to entry: The BACKGROUND RESPONSE is caused by external radiation fields, but in addition also by electronic noise and contamination.

3.3

CERTIFIED RADIOACTIVE STANDARD SOURCE

RADIOACTIVE SOURCE that has been calibrated by a laboratory recognized as a country's national standardizing laboratory for radioactivity measurements and has been so certified by the aforementioned laboratory

[SOURCE: IEC 61303:1994, 2.1.1]

3.4

IONISATION CHAMBER TEST SOURCE

RADIOACTIVE SOURCE used for the determination of the long-term stability of an ionisation chamber

Note 1 to entry: The half-life of the source shall be greater than five years and the effects of any radioactive contaminants shall be such that the indication of the device over a period of five years would not deviate by more than 0,5 % after decay correction for the known half-life of the principal radionuclide.

Note 2 to entry: The IONISATION CHAMBER TEST SOURCE is used to test the functionality of a RADIONUCLIDE CALIBRATOR under defined conditions.

[SOURCE: IEC 61303:1994, 2.7, modified – Part of the definition has been moved to a note to entry, and a second note to entry has been added.]

3.5

RADIONUCLIDE CALIBRATOR

device for measuring the ACTIVITY of a radioactive sample FVFW

[SOURCE: IEC 61303:1994, 2.1(standards.iteh.ai)

3.6

IEC TR 61948-4:2019

REFERENCE DATA https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904bset of data measured immediately_after_ACCEPTANCE_TESTING, using test methods designed for ROUTINE TEST

[SOURCE: IEC TR 61948-1:2016, 3.7]

3.7

ROUTINE TEST

test of a piece of equipment or its components which is repeated at specified intervals, to establish and document changes from the initial status described by REFERENCE DATA

Note 1 to entry: A ROUTINE TEST could be carried out by the user with simple methods and equipment.

[SOURCE: IEC TR 61948-1:2016, 3.8]

3.8

SYSTEM LINEARITY

function relating the observed and predicted ACTIVITY values when the ACTIVITY of a specified RADIOACTIVE SOURCE is varied

[SOURCE: IEC 61303:1994, 2.4]

3.9

TRACEABLE RADIOACTIVE STANDARD SOURCE

RADIOACTIVE SOURCE that has been calibrated by comparing it to a CERTIFIED RADIOACTIVE STANDARD SOURCE or to another TRACEABLE RADIOACTIVE STANDARD SOURCE of the same radionuclide

[SOURCE: IEC 61303:1994, 2.1.2]

IEC TR 61948-4:2019 © IEC 2019 - 7 -

4 Test methods

4.1 BACKGROUND RESPONSE

The BACKGROUND RESPONSE is obtained with the setting that corresponds to the most often used nuclide. The measurement is made with the sample holder in place. The reading of the background is reported.

4.2 Constancy of instrument response

The constancy of instrument response of the RADIONUCLIDE CALIBRATOR and, hence, its calibration is obtained for a specified radionuclide setting by inserting the IONISATION CHAMBER TEST SOURCE into the measurement position. The instrument reading is compared to the REFERENCE DATA and reported.

4.3 SYSTEM LINEARITY

4.3.1 General

The test of SYSTEM LINEARITY covers the range of ACTIVITY from 10 MBq up to 120 % of the maximum ACTIVITY injected into a patient as measured with the device under test and at least up to 1 000 MBq.

4.3.2 Decaying source method

SYSTEM LINEARITY is tested using the radioactive decay of a sample of a short-lived

SYSTEM LINEARITY is tested using the radioactive decay of a sample of a short-lived radionuclide, for example ^{99m}Tc. The sample of the radionuclide used is introduced into the measuring position of the device under test and measured at defined points in time. The time intervals between the individual measurements are chosen such that at least two data points are acquired per decade of the investigated <u>ACTIVITY.019</u>

https://standards.iteh.ai/catalog/standards/sist/2c08c32c-d199-4da4-904b-

4.3.3 Data analysis a3c6d79f9ba7/iec-tr-61948-4-2019

The logarithms of the measured data are plotted against the times of measurement. A straight line is fit to the logarithm of the measured data in the range from 10 MBq up to 1 000 MBq. From the fit, the expected ACTIVITY values (not the logarithm of their values) at the time points of measurement are determined. The ratio of the measured ACTIVITY to the expected ACTIVITY is calculated for each measurement point. The maximum percent deviation is reported.

4.4 Additional checks

When additional checks are recommended in the operation manual, they are also performed at the frequency suggested by the MANUFACTURER.

4.5 Frequency of ROUTINE TESTS

ROUTINE TESTS are carried out at the time intervals given in Table 1.