



Designation: E617 – 18

Standard Specification for Laboratory Weights and Precision Mass Standards¹

This standard is issued under the fixed designation E617; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers weights and mass standards used in laboratories, specifically classes 000, 00, 0, 1, 2, 3, 4, 5, 6, and 7. This specification replaces National Bureau of Standards Circular 547, Section 1, which is out of print.

1.2 This specification and calibration method is intended for use by weight manufacturers, national metrology institutes, weight calibration laboratories, accreditation bodies, users of weights, and regulatory bodies.

1.3 This specification contains the principal physical characteristics and metrological requirements for weights that are used.

1.3.1 For the verification of weighing instruments;

1.3.2 For the calibration of weights of a lower class of accuracy; and

1.3.3 With weighing instruments.

1.4 Maximum Permissible Errors (formerly tolerances) and design restrictions for each class are described in order that both individual weights or sets of weights can be chosen for appropriate applications.

1.5 Weight manufacturers must be able to provide evidence that all new weights comply with specifications in this standard (for example, material, density, magnetism, surface finish, mass values, uncertainties). Statements of compliance by calibration laboratories during subsequent calibrations must meet the requirements of ISO/IEC 17025, 5.10.4.2 and indicate on the calibration report which sections have or have not been assessed. Subsequent calibrations must meet all the requirements (including environmental parameters as shown in Table 11, of Sections 7, 8, and 9; and the requirements of ISO/IEC 17025:2005, 5.10.4.2 to make any claim of compliance to Specification E617, Maximum Permissible Errors, weight classes, or metrological traceability.

1.6 The values stated in SI units are to be regarded as standard.

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ISO Standards:*²

ISO/IEC 17025 General Requirements for the Competence of Testing and Calibration Laboratories (2005)

2.2 *NIST Standards:*³

NIST Handbook 143 State Weights and Measures Laboratories Program Handbook (2007)

NIST SP 811 Guide for the Use of the International System of Unit (SI) 2008 Edition

NIST SP 1038 The International System of Units (SI) – Conversion Factors for General Use (May 2006)

NISTIR 5672 Advanced Mass Calibration and Measurement Assurance Program for State Calibration Laboratories (2014)

NISTIR 6969 Selected Laboratory and Measurement Practices to Support Basic Mass Calibrations (2017)

NIST Technical Note 1297 (1994) Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results

2.3 *OIML Standards:*⁴

OIML D 28 Conventional Value of the Result of Weighing in Air (2004)

OIML R111–1e04 Weights of classes E1, E2, F1, F2, M1, M1–2, M2, M2–3 and M3 Part 1: Metrological and Technical Requirements (2004)

2.4 *BIPM Standards:*⁵

VIM: JCGM 200:2012 International Vocabulary of Metrology—Basic and General Concepts and Associated Terms

² Available from International Organization for Standardization (ISO), 1, ch. de la Voie-Creuse, CP 56, CH-1211 Geneva 20, Switzerland, <http://www.iso.org>.

³ Available from National Institute of Standards and Technology (NIST), 100 Bureau Dr., Stop 1070, Gaithersburg, MD 20899-1070, <http://www.nist.gov>.

⁴ Available from Organisation Internationale de Metrologie Legale, 11 Rue Turgot, 75009 Paris, France, <http://www.oiml.org>.

⁵ Available from Bureau International des Poids et Mesures (BIPM), Pavillon de Breteuil, F-92312, Sèvres Cedex, France, <http://www.bipm.org>.

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.06 on Laboratory Instruments and Equipment.

Current edition approved Oct. 1, 2018. Published November 2018. Originally approved in 1978. Last previous edition approved in 2013 as E617 – 13. DOI: 10.1520/E0617-18.

GUM: JCGM 100:2008 Evaluation of Measurement Data—Guide to the Expression of Uncertainty in Measurement

2.5 *EURAMET Standards*.⁶

EURAMET/cg-18/V. 4.0 Guidelines on the Calibration of Non-Automatic Weighing Instruments (2015)

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *accuracy class of weights*—a class of weights that meets certain metrological requirements intended to keep the errors within specified limits.

3.1.2 *balance*—instrument indicating apparent mass that is sensitive to the following forces:

$$F_g = m \cdot g$$

Force due to gravity

$$F_b = v \cdot \rho_a \cdot g = \frac{m}{\rho} \rho_a \cdot g$$

Air buoyancy equal to the weight of the displaced air.

$$F_z = \mu_0 \int \int \int_V (M + \chi H) \frac{\partial H}{\partial z} dV$$

Vertical component of the magnetic interaction between the weight and the balance or the environment, or both.

H and M are vectors; z is the vertical cartesian coordinate. If magnetic effects are negligible, that is, the permanent magnetization (M) of the weight and the magnetic susceptibility (χ) are sufficiently small, and the balance is calibrated with reference weights of well-known mass, the balance can be used to indicate the conventional mass, m_c , of a body under conventionally chosen conditions.

3.1.3 *calibration (of weights)*—the acts of determining the mass difference between a standard of known mass value and an “unknown” test weight or set of weights, establishing the mass value and conventional mass value of the “unknown,” and of determining a quantitative estimate of the uncertainty to be assigned to the stated mass or conventional mass value of the “unknown,” or both, and providing metrological traceability to the “unknown.”

3.1.3.1 *calibration (generally)*—set of operations that establish, under specified conditions, the relationship between values of quantities indicated by a measuring instrument or measuring system, or values represented by a material measure or a reference material, and the corresponding values realized by standards.

3.1.4 *calibration certificate*—certificate issued by calibration laboratories to document the results of a calibration.

3.1.5 *conventional mass*—conventional value of the result of weighing in air, in accordance to International Recommendation OIML D 28. For a weight taken at 20°C, the conventional mass is the mass of a reference weight of a density of 8000 kg/m³ which it balances in air of density of 1.2 kg/m³.

3.1.6 *correction*—mass values are traditionally expressed by two numbers, one being the nominal mass of the weight, and the second being a correction. The mass of the weight is the

assigned nominal value plus the assigned correction. Positive corrections indicate that the weight embodies more mass than is indicated by the assigned nominal value. Negative corrections indicate that the weight embodies less mass than is indicated by the assigned nominal value. The correction is equivalent to the “error.”

3.1.7 *international prototype kilogram*—the platinum-iridium cylinder maintained at the International Bureau of Weights and Measures (BIPM), at Sevres, France with an internationally accepted defined mass of 1 kg.

3.1.8 *magnetism*—effect that generates an attractive or repulsive force.

3.1.8.1 (*volume*) *magnetic susceptibility* (χ)—measure of the ability of a medium to modify a magnetic field. It is related to the magnetic permeability (μ) by the relation: $\mu/\mu_0 = 1 + \chi$. The quantity μ/μ_0 is sometimes referred to as the relative permeability, μ_r .

3.1.8.2 (*permanent*) *magnetization* (M)—parameter that specifies a magnetic state of material bodies such as weights, in the absence of an external magnetic field (most generally, magnetization is a vector whose magnitude and direction are not necessarily constant within the material). The magnetization of a body generates an inhomogeneous magnetic field in space and thus may produce magnetic forces on other materials.

3.1.9 *mass*—physical quantity, which can be ascribed to any material object and which gives a measure of its quantity of matter. The unit of mass is the kilogram.

3.1.10 *maximum permissible errors*—the maximum amount by which the sum of the conventional mass of the weight, its deviation from nominal value, and its associated uncertainty is allowed to deviate from the assigned nominal value.

3.1.11 *metrological traceability*—property of a measurement result whereby the result can be related to a reference through a documented unbroken chain of calibrations, each contributing to the measurement uncertainty. Metrological traceability requires an established calibration hierarchy. Elements for confirming metrological traceability to be an unbroken chain to an international measurement standard or a national measurement standard (IPK or NPS), shall include a documented measurement uncertainty, a documented measurement procedure, accredited technical competence, metrological traceability to the SI, and established calibration intervals (see current VIM: JCGM 200).

3.1.12 *reference standard*—a standard, generally of the highest metrological quality available at a given location, from which measurements made at that location are derived.

3.1.13 *roughness parameter or R-parameter* (R_a or R_z)—parameter that describes the assessed roughness profile of a sample. The letter R is indicative of the type of assessed profile, in this case R for roughness profile. The assessed profile of a sample can be in terms of different profile types: a roughness profile or R-parameter, primary profile or P-parameter, a waviness profile or W-parameter.

3.1.14 *set of weights*—a series of weights, usually presented in a case so arranged to make possible any weighing of all

⁶ Available from Euramet, Bundesallee 100, 38116 Braunschweig, Germany, <http://www.euramet.org>.

loads between the mass of the weight with the smallest nominal value and the sum of the masses of all weights of the series with a progression in which the mass of the smallest nominal value weight constitutes the smallest step of the series.

3.1.15 *temperature (t)*—in degrees Celsius, is related to the absolute thermodynamic temperature scale, called the Kelvin scale, by $t = T - 273.15$ K.

3.1.16 *test weight (m_t)*—weight that is to be tested according to this standard.

3.1.17 *tolerance test*—verification that the conventional mass of the weights and their corresponding uncertainties as tested are correct within the maximum permissible errors of the respective weight class.

3.1.18 *uncertainty*—non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used.

3.1.19 *units*—the units used are: (1) for mass, the milligram (mg), the gram (g) and the kilogram (kg); (2) for density, the kilogram per cubic meter (kg m^{-3}).

3.1.20 *U.S. National prototype standard*—platinum-iridium kilogram identified as K20, maintained at the National Institute of Standards and Technology, with value assigned relative to the International Prototype Kilogram provides the United States access to the mass unit.

3.1.21 *weight*—material measure of mass, regulated in regard to its physical and metrological characteristics: shape, dimensions, material, surface quality, nominal value, density, magnetic properties and maximum permissible error.

NOTE 1—The term “weight” is also used as the physical quantity of the gravitational force of a body. From the context it is usually clear in which sense the term is used. If the sense is not clear, one may use the words “weight force” or “weight piece,” depending on its meaning.

3.2 Symbols:

Symbol	Unit	Definition
A	—	represents weighing the reference weight in a weighing cycle
B	—	represents weighing the test weight in a weighing cycle
C	—	correction factor for air buoyancy
D	kg	difference of balance readings between minimum and maximum values from eccentricity test
d	kg	scale interval
d_1	m	estimated distance between centers of weights during loading
d_2	m	estimated distance from the center of the load receptor to one of the corners
F_b	N	air buoyancy equal to the weight of the displaced air
F_g	N	gravitational force
F_z	N	magnetic force between a mass comparator and a weight in the vertical or z-direction
g	m s^{-2}	gravitational acceleration
H	A m^{-1}	magnetizing field strength
hr	%	relative humidity
I	kg	indication of the weighing instruments (scale division)
ΔI	kg	indication difference of the balance, where $\Delta I = I_t - I_r$
ΔI_1	kg	indication difference using an automatic exchange mechanism with weights in first position

Symbol	Unit	Definition
ΔI_2	kg	indication difference using an automatic exchange mechanism with weights in reversed position
ΔI_s	kg	change in indication of balance due to sensitivity weight
i	—	subscript used as an index in summations
j	—	subscript for number of test weights or number of series of measurements
k	—	coverage factor, typically 2 or 3
M	A m^{-1}	permanent magnetization (see also $\mu_0 M$)
m	kg	mass of a rigid body (weight)
Δm	kg	mass difference, usually between test and reference weight
δm	kg	maximum permissible error on the weights
m_0	kg	mass, nominal value of the weight (e.g. 1 kg)
m_c	kg	conventional mass of the weight
Δm_c	kg	conventional mass difference between test weight and reference weight
$\overline{\Delta m_c}$	kg	average conventional mass difference between test weight and reference weight
m_{cr}	kg	conventional mass of the reference weight
m_{ct}	kg	conventional mass of the test weight
m_s	kg	mass of the sensitivity weight
m_t	kg	mass of the test weight
n	—	subscript for number of measurement sequences
ρ	Pa	barometric pressure
R_a	μm	mean height of roughness profile (R-parameter)
R_z	μm	maximum height of roughness profile (R-parameter)
r	—	subscript for reference weight
s	—	subscript for sensitivity weight
s	kg	standard deviation
s^2	kg^2	variance
T	K	thermodynamic temperature using the International Temperature Scale of 1990 (ITS-90)
ΔT^*	$^{\circ}\text{C}$	initial difference between weight temperature and laboratory temperature
t	—	subscript for test weight
t	$^{\circ}\text{C}$	temperature in degrees Celsius, where $t = T - 273.15$ K
U	kg	uncertainty, expanded uncertainty
u	kg	uncertainty, standard uncertainty
u_b	kg	uncertainty of air buoyancy correction
u_{ba}	kg	uncertainty of the balance
u_c	kg	combined standard uncertainty
u_d	kg	uncertainty due to the display resolution of a digital balance
u_E	kg	uncertainty due to eccentricity
u_F	kg m^{-3}	uncertainty of the formula used to calculate air density
u_{hr}	%	uncertainty in relative humidity
u_{inst}	kg	uncertainty due to instability of the reference weight
u_{ma}	kg	uncertainty due to magnetism
u_p	Pa	uncertainty in barometric pressure
u_s	kg	uncertainty due to the sensitivity of the balance
u_t	$^{\circ}\text{C}$	uncertainty in temperature
u_w	kg	uncertainty due to the weighing process
V	m^3	volume of a solid body (weight)
z	m	vertical cartesian coordinate
μ	N A^{-2}	magnetic permeability
μ_0	N A^{-2}	magnetic constant (magnetic permeability of vacuum), $\mu_0 = 4\pi \times 10^{-7} \text{ N A}^{-2}$
$\mu_0 M$	T	magnetic polarization
μ_r	—	relative magnetic permeability (μ/μ_0)

Symbol	Unit	Definition
V_{eff}	—	effective degrees of freedom
ρ	kg m ⁻³	mass of a rigid body (weight)
ρ_0	kg m ⁻³	density of air as a reference value equal to 1.2 kg m ⁻³
ρ_a	kg m ⁻³	density of moist air
ρ_{al}	kg m ⁻³	density of moist air during the last (previous) calibration of the reference weight
ρ_r	kg m ⁻³	density of a reference weight with mass m_r
ρ_t	kg m ⁻³	density of the weight being tested
χ	—	magnetic susceptibility

4. Maximum Permissible Errors

4.1 For each weight, the expanded uncertainty U at approximately 95 % confidence (See Section 9) of the conventional mass shall be less than or equal to one-third of the maximum permissible error given in Table 1 as defined in Section 9. Subsequent calibrations must meet all the requirements (including environmental parameters as shown in Table 11, of Sections 7, 8, and 9; and the requirements of ISO/IEC 17025:2005, 5.10.4.2 to make any claim of compliance to Specification E617, Maximum Permissible Errors, weight classes, or metrological traceability.

4.1.1 For each weight, the conventional mass, m_c (determined with an expanded uncertainty), shall not differ by more than the difference: maximum permissible error δm minus expanded uncertainty, from the nominal value of the weight, m_o :

$$m_o - (\delta m - U) \leq (m_c) \leq m_o + (\delta m - U) \quad (1)$$

4.2 Maximum permissible errors for classes 000, 00, 0, 1, 2, 3, 4, 5, 6, and 7 are given in Table 1. These maximum permissible errors apply to conventional mass values.

4.3 Maximum Permissible Errors for weights of denomination intermediate between those listed, the maximum permissible error shall be proportional to the values shown.

4.4 For class 000, 00, and 0 weights, which are always accompanied by certificates giving the mass values and uncertainties, the deviation from the nominal value, $m_c - m_o$, shall be taken into account by the user.

5. Physical Characteristics

5.1 Construction:

5.1.1 *Type*—Weights are divided into two types based upon the design:

5.1.1.1 *Type I*—These weights are of one-piece construction and contain no added adjusting material. They must be specified when weights are to be used as standards for the calibration of weights of Classes 000, 00, 0, 1, 2, and 3, and where maximum stability is required. A precise measurement of density can only be made for one-piece weights.

5.1.1.2 *Type II*—Weights of this type can be of any appropriate design such as screw knob, ring, or sealed plug. Adjusting material can be used provided it is of a material at least as stable as the base material and is contained in such a way that it will not become separated from the weight.

5.1.2 Class 000, 00, and 0 shall be Type I, one piece construction. Weights with nominal values less than 1 g shall have unique shapes to differentiate the weights from one

another. See Table 2. The shape of weights smaller than 1 mg shall be discussed and verified with the customer.

5.1.3 Class 1, 2, 3, 4, 5, 6, and 7 may be either Type I or Type II depending on the application.

5.2 *Design*—A weight may have any shape that does not introduce features that reduce the reliability. All weights shall be free of ragged or sharp edges or ends. Both sheet metal and wire weights shall be free of cracks such as may be formed from bending.

5.3 *Surface Area*—For classes 000, 00, 0, 1, 2, 3, and 4 the surface area is not to exceed twice the area of a cylinder of equal height and diameter for weights 1 g and above. Sheet metal weights or wire weights may be used below 1 g. For Classes 5, 6, and 7 the total surface areas should be minimized to the extent possible.

5.4 Material:

5.4.1 *Class 000, 00, 0, 1, 2, 3, 4, and 5 Weights*—The hardness of this material and its resistance to wear and corrosion shall be similar to or better than that of austenitic stainless steel.

5.4.2 *Classes 6 and 7*—Cylindrical class 6 and 7 weights below 5 kg and class 6 and 7 weights below 100 g shall be made of steel or a material whose hardness and resistance to corrosion is similar or better than that of steel. Other class 6 and 7 weights of 5 kg or greater shall be made of grey cast iron or of another material whose brittleness and resistance to corrosion is similar or better than that of grey cast iron. The surface of the weights may be treated with a suitable coating in order to improve their corrosion resistance. This coating shall withstand shocks and outdoor weather conditions.

5.5 *Magnetism*—Weights shall not exceed maximum permissible magnetic properties as listed in Tables 3 and 4 for any portion of the weight. If the values of all local measurements of magnetization and susceptibility are less than these limits, then it may be assumed that the uncertainty components due to the magnetism of the weight are negligible. The maximum permanent magnetization and magnetic susceptibilities given in Tables 3 and 4 are such that, at magnetic fields and magnetic field gradients possibly present on balance pans, they produce a change of the conventional mass of less than 1/10 of the maximum permissible error of the test weight.

NOTE 2—Magnetic susceptibility may be tested in accordance with OIML R 111-1, Annex B. Cast iron cannot have a susceptibility specification of any real value.

5.6 *Density*—Because of the effect of the buoyant force of air on a weight, precision measurements of mass require that the volume of the weight be known, as well as the density of the air in which it is being measured, so that appropriate corrections can be made. For weights of higher precision, the range of density is limited to values at or near the density of well-established standards, such as are used by primary calibration laboratories. For Class 000 and 00, the manufacturer shall provide a measured value for the density of the weights at time of manufacture. Use of a sample taken adjacent to the material from which the weight is manufactured to measure the density is permitted, however an additional uncertainty component equal to 5×10^{-5} must be combined with the standard

TABLE 1 Maximum Permissible Errors

NOTE 1—Maximum Permissible Errors are reported in SI units, typically milligrams.

NOTE 2—The “grain” is the same in avoirdupois, troy, and apothecaries units of mass.

NOTE 3—See NIST SP 811 and NIST SP 1038 for conversion and units of measure.

Denomination	±mg except as noted									
	Class 000	Class 00	Class 0	Class 1	Class 2	Class 3	Class 4	Class 5	Class 6	Class 7
Metric										
5000 kg					25 g	50 g	100 g	250 g	500 g	750 g
3000 kg					15 g	30 g	60 g	150 g	300 g	450 g
2000 kg					10 g	20 g	40 g	100 g	200 g	300 g
1000 kg					5.0 g	10 g	20 g	50 g	100 g	150 g
500 kg					2.5 g	5.0 g	10 g	25 g	50 g	75 g
300 kg					1.5 g	3.0 g	6.0 g	15 g	30 g	45 g
200 kg					1.0 g	2.0 g	4.0 g	10 g	20 g	30 g
100 kg					500 mg	1.0 g	2.0 g	5.0 g	10 g	15 g
50 kg	13 mg	25 mg	63 mg	120 mg	250	500 mg	1.0 g	2.5 g	5.0 g	7.5 g
30 kg	7.5	15	38	75	150	300	600 mg	1.5 g	3.0 g	4.5 g
25 kg	6.25	12.5	31	62	125	250	500	1.2 g	2.5 g	3.8 g
20 kg	5.0	10	25	50	100	200	400	1.0 g	2.0 g	3.8 g
10 kg	2.5	5.0	13	25	50	100	200	500 mg	1.0 g	2.2 g
5 kg	1.3	2.5	6.0	12	25	50	100	250	500 mg	1.4 g
3 kg	0.75	1.5	3.8	7.5	15	30	60	150	300	1.0 g
2 kg	0.5	1.0	2.5	5.0	10	20	40	100	200	750 mg
1 kg	0.25	0.5	1.3	2.5	5.0	10	20	50	100	470
500 g	0.13	0.25	0.60	1.2	2.5	5.0	10	30	50	300
300 g	0.075	0.15	0.38	0.75	1.5	3.0	6.0	20	30	210
200 g	0.05	0.10	0.25	0.50	1.0	2.0	4.0	15	20	160
100 g	0.025	0.05	0.13	0.25	0.50	1.0	2.0	9.0	10	100
50 g	0.015	0.030	0.060	0.12	0.25	0.60	1.2	5.6	7.0	62
30 g	0.014	0.026	0.037	0.074	0.15	0.45	0.90	4.0	5.0	44
20 g	0.013	0.025	0.037	0.074	0.10	0.35	0.70	3.0	3.0	33
10 g	0.010	0.020	0.025	0.050	0.074	0.25	0.50	2.0	2.0	21
5 g	0.005	0.010	0.017	0.034	0.054	0.18	0.36	1.3	2.0	13
3 g	0.005	0.010	0.017	0.034	0.054	0.15	0.30	0.95	2.0	9.4
2 g	0.005	0.010	0.017	0.034	0.054	0.13	0.26	0.75	2.0	7.0
1 g	0.005	0.010	0.017	0.034	0.054	0.10	0.20	0.50	2.0	4.5
500 mg	0.002	0.003	0.005	0.010	0.025	0.080	0.16	0.38	1.0	3.0
300 mg	0.002	0.003	0.005	0.010	0.025	0.070	0.14	0.30	1.0	2.2
200 mg	0.002	0.003	0.005	0.010	0.025	0.060	0.12	0.26	1.0	1.8
100 mg	0.002	0.003	0.005	0.010	0.025	0.050	0.10	0.20	1.0	1.2
50 mg	0.002	0.003	0.005	0.010	0.014	0.042	0.085	0.16		0.88
30 mg	0.002	0.003	0.005	0.010	0.014	0.038	0.075	0.14		0.68
20 mg	0.002	0.003	0.005	0.010	0.014	0.035	0.070	0.12		0.56
10 mg	0.002	0.003	0.005	0.010	0.014	0.030	0.060	0.10		0.40
5 mg	0.002	0.003	0.005	0.010	0.014	0.028	0.055	0.080		
3 mg	0.002	0.003	0.005	0.010	0.014	0.026	0.052	0.070		
2 mg	0.002	0.003	0.005	0.010	0.014	0.025	0.050	0.060		
1 mg	0.002	0.003	0.005	0.010	0.014	0.025	0.050	0.050		
0.5 mg	0.002	0.003	0.005	0.010	0.014	0.025	0.050	0.050		
0.3 mg	0.002	0.003	0.005	0.010	0.014	0.025				
0.2 mg	0.002	0.003	0.005	0.010	0.014					
0.1 mg	0.002	0.003	0.005	0.010						
0.05 mg	0.002	0.003	0.005	0.010						
Avoirdupois			Class 0	Class 1	Class 2	Class 3	Class 4	Class 5	Class 6	Class 7
Pound			mg	mg	mg	g & mg	g & mg	g & mg	g & mg	g & mg
10000 lb						45 g	91 g	230 g	450 g	680 g
5000 lb						23 g	45 g	110 g	230 g	340 g
3000 lb						14 g	27 g	68 g	140 g	200 g
2500 lb						11 g	23 g	57 g	110 g	170 g
2000 lb						9.1 g	18 g	45 g	91 g	140 g
1000 lb						4.5 g	9.1 g	23 g	45 g	68 g
500 lb						2.3 g	4.5 g	11 g	23 g	34 g
300 lb						1.4 g	2.7 g	6.8 g	14 g	20 g
200 lb						910 mg	1.8 g	4.5 g	9.1 g	14 g
100 lb			57	110	230	450	910 mg	2.3 g	4.5 g	6.8 g
50 lb			29	57	110	230	450	1.1 g	2.3 g	4.1 g
30 lb			17	32	68	140	270	680 mg	1.4 g	2.7 g
25 lb			14	28	57	110	230	570	1.1 g	2.4 g
20 lb			12	23	45	91	180	450	910 mg	2.0 g
10 lb			5.5	11	23	45	91	230	450	1.3 g
5 lb			2.7	5.4	11	23	45	110	230	780 mg
3 lb			1.7	3.4	6.8	14	27	68	140	580
2 lb			1.2	2.3	4.5	9.1	18	45	91	440
1 lb			0.55	1.1	2.3	4.5	9.1	23	45	270

TABLE 1 *Continued*

Denomination	±mg except as noted									
	Class 000	Class 00	Class 0	Class 1	Class 2	Class 3	Class 4	Class 5	Class 6	Class 7
0.5 lb			0.27	0.54	1.1	2.3	4.5	16	23	170
0.3 lb			0.17	0.34	0.68	1.4	2.7	11	14	120
0.2 lb			0.12	0.23	0.45	0.91	1.8	8.2	9.1	97
0.1 lb			0.055	0.11	0.23	0.45	1.1	5.4	6.8	59
0.05 lb			0.027	0.054	0.11	0.36	0.77	3.3	4.5	37
0.03 lb			0.017	0.034	0.068	0.32	0.59	2.4	3.2	26
0.02 lb			0.017	0.034	0.045	0.23	0.45	1.9	2.3	20
0.01 lb			0.012	0.023	0.034	0.16	0.34	1.2	1.4	12
0.005 lb			0.0075	0.015	0.024	0.14	0.27	0.86	0.91	7.8
0.003 lb			0.0075	0.015	0.024	0.11	0.22	0.64	0.91	5.4
0.002 lb			0.0075	0.015	0.024	0.091	0.19	0.50	0.91	4.2
0.001 lb			0.0075	0.015	0.024	0.068	0.15	0.36	0.91	2.9
0.0005 lb			0.0045	0.0091	0.023	0.064	0.13	0.27		2.1
0.0003 lb			0.0045	0.0091	0.023	0.054	0.11	0.23		1.5
0.0002 lb			0.0045	0.0091	0.023	0.045	0.095	0.20		1.2
0.0001 lb			0.0045	0.0091	0.023	0.041	0.086	0.16		0.84
0.00005 lb			0.0045	0.0091	0.014	0.036	0.073	0.14		
0.00003 lb			0.0045	0.0091	0.014	0.032	0.064	0.11		
0.00002 lb			0.0045	0.0091	0.014	0.029	0.059	0.091		
0.00001 lb			0.0045	0.0091	0.014	0.027	0.054	0.091		
0.000005 lb			0.0045	0.0091	0.014	0.023	0.054	0.091		
0.000003 lb			0.0045	0.0091	0.014	0.023	0.054	0.091		
0.000002 lb			0.0045	0.0091	0.014	0.023	0.054	0.091		
0.000001 lb			0.0045	0.0091	0.014	0.023	0.054	0.091		
Avoirdupois			Class 0	Class 1	Class 2	Class 3	Class 4	Class 5	Class 6	Class 7
Ounce			mg							
10 oz			0.35	0.70	1.4	2.8	5.4	19	23	200
8 oz			0.27	0.54	1.1	2.3	4.5	16	23	170
5 oz			0.18	0.35	0.71	1.4	2.8	12	11	130
4 oz			0.14	0.28	0.57	1.1	2.3	9.5	11	110
3 oz			0.10	0.21	0.43	0.91	1.8	8.2	5.4	
2 oz			0.070	0.14	0.28	0.64	1.3	5.9	5.4	70
1 oz			0.027	0.054	0.14	0.43	0.86	3.9	3.2	42
1/2 oz			0.017	0.034	0.071	0.30	0.59	2.5	2.3	27
1/4 oz			0.011	0.023	0.034	0.20	0.43	1.6	1.4	17
1/8 oz			0.0075	0.015	0.024	0.16	0.31	1.1	0.91	10
1/16 oz			0.0075	0.015	0.024	0.12	0.24	0.73	0.91	6.5
1/32 oz			0.0075	0.015	0.024	0.095	0.19	0.50		4.2
1/64 oz			0.0075	0.015	0.023	0.077	0.15	0.36		2.8
0.5 oz					0.071	0.30	0.59	2.5	2.3	27
0.3 oz					0.034	0.23	0.45	1.8	1.4	19
0.2 oz					0.034	0.19	0.38	1.4	0.91	14
0.1 oz					0.024	0.14	0.29	0.91	0.91	9.0
0.05 oz					0.024	0.11	0.23	0.64	0.91	5.7
0.03 oz					0.024	0.095	0.19	0.45	0.91	4.1
0.02 oz					0.024	0.077	0.18	0.40	0.91	3.2
0.01 oz					0.023	0.064	0.14	0.30	0.91	2.2
0.005 oz					0.023	0.054	0.11	0.23		1.5
0.003 oz					0.023	0.050	0.095	0.19		1.1
0.002 oz					0.023	0.044	0.086	0.16		0.92
0.001 oz					0.023	0.038	0.077	0.13		0.66
0.0005 oz					0.014	0.032	0.064	0.11		0.47
0.0003 oz					0.014	0.029	0.059	0.095		
0.0002 oz					0.014	0.027	0.054	0.086		
0.0001 oz					0.014	0.026	0.050	0.073		
Troy Ounce			Class 0	Class 1	Class 2	Class 3	Class 4	Class 5	Class 6	Class 7
						mg	mg	g & mg		
1000 oz t						310 mg	620 mg	1.6 g		
500 oz t						160	310	780 mg		
300 oz t						91	190	470		
200 oz t						62	120	310		
100 oz t						31	62	160		
50 oz t						16	31	78		
30 oz t						9.1	19	47		
20 oz t						6.2	12	35		
10 oz t						3.1	6.2	21		
5 oz t						1.6	3.1	12		
3 oz t						0.91	1.9	8.4		
2 oz t						0.71	1.4	6.5		
1 oz t						0.45	0.91	4.2		
0.5 oz t						0.31	0.62	2.6		
0.3 oz t						0.24	0.49	1.9		
0.2 oz t						0.20	0.40	1.5		
0.1 oz t						0.15	0.30	0.97		