

TECHNICAL REPORT



High frequency surgical equipment and high frequency surgical accessories –
Operation and maintenance

(standards.iteh.ai)

[IEC TR 61289:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/06446f02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019>



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

[IEC TR 61289:2019](#)

<https://standards.iec.ch/catalog/standards/sist/06-446/02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019>

TECHNICAL REPORT



High frequency surgical equipment and high frequency surgical accessories –
Operation and maintenance
STANDARD PREVIEW
(standards.iteh.ai)

[IEC TR 61289:2019](https://standards.iteh.ai/catalog/standards/sist/06446f02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019)

<https://standards.iteh.ai/catalog/standards/sist/06446f02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.01

ISBN 978-2-8322-6885-8

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD 4

INTRODUCTION 6

1 Scope 7

2 Normative references 7

3 Terms and definitions 7

4 General information regarding HF SURGICAL EQUIPMENT 12

5 Recommended practices before use 13

5.1 Inspection of HF SURGICAL EQUIPMENT and HF SURGICAL ACCESSORIES before use 13

5.2 Preparation 13

5.3 Handling of NEUTRAL ELECTRODES, HF SURGICAL ACCESSORIES, cables and connections 14

5.4 PATIENTS with active implants (active implantable medical devices) 14

5.5 Simultaneous use of two items of HF SURGICAL EQUIPMENT 15

5.6 Instructions for use 15

6 Recommended practices during use 15

7 Recommended practices after use 16

8 Nature of HAZARDS 16

8.1 General 16

8.2 HF SURGICAL EQUIPMENT related HAZARDS 16

8.2.1 Incompatible combinations 16

8.2.2 Electromagnetic compatibility 16

8.2.3 Misconnection of ACTIVE ACCESSORIES 16

8.2.4 Specialty HF SURGICAL EQUIPMENT 17

8.3 ACTIVE ACCESSORY related HAZARDS 17

8.3.1 Incompatible combinations 17

8.3.2 Environment of use 17

8.3.3 Misuse 18

8.4 OPERATOR-related HAZARDS 18

8.4.1 OPERATOR not reading or following the instructions for use 18

8.4.2 OPERATOR selecting inappropriate power or mode settings 18

8.4.3 OPERATOR using an HF SURGICAL ACCESSORY in an inappropriate manner 18

8.5 NEUTRAL ELECTRODE related HAZARDS 19

8.5.1 General 19

8.5.2 Inadequate contact area of a NEUTRAL ELECTRODE 19

8.5.3 Inappropriate application 19

8.5.4 Surgical procedures utilizing high currents or long duty cycles 19

9 Safety provisions of, and symbols on, HF SURGICAL EQUIPMENT 22

9.1 General 22

9.2 Colours of indicator lights 22

9.3 Markings on HF SURGICAL ELECTRICAL EQUIPMENT 22

9.4 Protection against electric shock and burns 24

9.4.1 Method of protection 24

9.4.2 Degree of protection 24

9.5 HF SURGICAL EQUIPMENT not properly marked 25

9.6 Monitoring the effectiveness of the NEUTRAL ELECTRODE 25

STANDARD PREVIEW

(standards.itech.ai)

IEC TR 61289:2019

standards/sist/06446/f02-99d8-44f1-b737-

7789-706-d11/s01-61289-2019

| | | |
|-----|--|----|
| 9.7 | Output indicators | 25 |
| 10 | Accompanying documents | 26 |
| 11 | Preventive maintenance | 26 |
| | Annex A (informative) Concerns regarding high current surgical procedures or high duty cycles..... | 27 |
| | Bibliography..... | 28 |
| | Figure 1 – Symbol used on Class II equipment | 24 |
| | Figure 2 – Symbol used with an earth referenced PATIENT circuit | 25 |
| | Figure 3 – Symbol used with an HF ISOLATED PATIENT CIRCUIT | 25 |
| | Table 1 – Colours and significance of indicator lights according to IEC 60601-2-2 | 22 |
| | Table 2 – Symbols used on HF SURGICAL EQUIPMENT | 23 |

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

[IEC TR 61289:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/06446f02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**HIGH FREQUENCY SURGICAL EQUIPMENT AND HIGH FREQUENCY
SURGICAL ACCESSORIES – OPERATION AND MAINTENANCE**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 61289, which is a technical report, has been prepared by sub-committee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2011. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) general adaption to IEC 60601-2-2:2017;
- b) refinement and additions to the defined terms;

- c) separation of HF SURGICAL EQUIPMENT and HF SURGICAL ACCESSORIES;
- d) consideration of the HIGH CURRENT MODE;
- e) update of symbols.

The text of this technical report is based on the following documents:

| | |
|---------------|------------------|
| Enquiry draft | Report on voting |
| 62D/1652/DTR | 62D/1662A/RVDTR |

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- TERMS DEFINED IN CLAUSE 3: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

[IEC TR 61289:2019](#)

A bilingual version of this publication may be issued at a later date.

<https://standards.iteh.ai/catalog/standards/sist/06446f02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019>

IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.

INTRODUCTION

This document gives guidelines to personnel in charge of operation of equipment covered by IEC 60601-2-2:2017 to enable them to attain the best conditions of safety for their PATIENTS and themselves.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC TR 61289:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/06446f02-99d8-44f1-b737-778aa7c4adb1/iec-tr-61289-2019>

HIGH FREQUENCY SURGICAL EQUIPMENT AND HIGH FREQUENCY SURGICAL ACCESSORIES – OPERATION AND MAINTENANCE

1 Scope

This document contains guidelines for medical and nursing personnel regarding the safe and effective operation of HIGH FREQUENCY SURGICAL EQUIPMENT and HIGH FREQUENCY SURGICAL ACCESSORIES (also referred to as HF SURGICAL EQUIPMENT in this document). It is also of use to scientific/technical staff who have responsibility for the maintenance of this equipment.

The application guidelines in this document deal with the safe operation of HIGH FREQUENCY SURGICAL EQUIPMENT constructed according to the safety requirements of IEC 60601-1 [1]¹ and IEC 60601-2-2 [4].

Not all existing HIGH FREQUENCY SURGICAL EQUIPMENT meets the minimum requirements of current international standards, however, the guidelines in this document is still helpful in utilizing these devices.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

ACCESSORY

additional part for use with equipment in order to:

- achieve the INTENDED USE,
- adapt it to some special use,
- facilitate its use,
- enhance its performance, or
- enable its functions to be integrated with those of other equipment

[SOURCE: IEC 60601-1:2005, 3.3]

3.2

ACTIVE ACCESSORY

HF SURGICAL ACCESSORY intended for manipulation by the OPERATOR to produce an effect by electrical conduction adjacent to the ACTIVE ELECTRODE at the intended site on the PATIENT,

¹ Numbers in square brackets refer to the Bibliography.

generally comprising an ACTIVE HANDLE, the cord of an ACTIVE ACCESSORY, ACTIVE CONNECTOR and ACTIVE ELECTRODE

[SOURCE: IEC 60601-2-2:2017, 201.3.201]

3.3

ACTIVE CONNECTOR

part of an ACTIVE ACCESSORY intended for connection to an ACTIVE OUTPUT TERMINAL, which may include additional terminals for connection of a FINGERSWITCH to a SWITCH SENSOR

[SOURCE: IEC 60601-2-2:2017, 201.3.202]

3.4

ACTIVE ELECTRODE

part of an ACTIVE ACCESSORY extending from the ACTIVE HANDLE to the surgical site and intended to pass HF current into body tissue

[SOURCE: IEC 60601-2-2:2017, 201.3.203]

3.5

ACTIVE HANDLE

part of an ACTIVE ACCESSORY intended to be held by the OPERATOR

[SOURCE: IEC 60601-2-2:2017, 201.3.205]

3.6

ACTIVE OUTPUT TERMINAL

part of HF SURGICAL EQUIPMENT or ASSOCIATED EQUIPMENT intended for connection to an ACTIVE ACCESSORY and for delivery of HF current thereto

[SOURCE: IEC 60601-2-2:2017, 201.3.206, modified – The notes have been deleted.]

3.7

APPLIED PART

part of ME EQUIPMENT that in normal use necessarily comes into physical contact with the PATIENT for ME EQUIPMENT or an ME SYSTEM to perform its function

[SOURCE: IEC 60601-1:2005, 3.8, modified – The notes have been deleted.]

3.8

ASSOCIATED EQUIPMENT

MEDICAL ELECTRICAL EQUIPMENT other than HF SURGICAL EQUIPMENT that may be electrically connected to the PATIENT circuit

[SOURCE: IEC 60601-2-2:2017, 201.3.207]

3.9

BIPOLAR

method of applying HF current to a PATIENT between two or more ACTIVE ELECTRODES without the need for a separately connected NEUTRAL ELECTRODE (or the need to use the PATIENT'S body capacitance to earth) in which an effect is intended in tissue near one or more ACTIVE ELECTRODES

Note 1 to entry: The BIPOLAR method includes devices energizing pairs of ACTIVE ELECTRODES as well as devices energizing groups of ACTIVE ELECTRODES where the HF current source and return may have different numbers of electrodes.

[SOURCE: IEC 60601-2-2:2017, 201.3.208]

3.10**COAGULATION**

use of HF current to induce a thermal effect, e.g. to control or prevent bleeding, induce tissue destruction, or induce tissue shrinkage

Note 1 to entry: COAGULATION may take the form of contact or non-contact COAGULATION.

Note 2 to entry: FULGURATION, desiccation, spray, forced, swift, soft and argon beam (plasma) COAGULATION are all names of COAGULATION types.

[SOURCE: IEC 60601-2-2:2017, 201.3.210]

3.11**CONTACT QUALITY MONITOR**

circuit in HF SURGICAL EQUIPMENT or ASSOCIATED EQUIPMENT intended for connection to a MONITORING NE providing an alarm in the event that NEUTRAL ELECTRODE (NE) contact with the PATIENT becomes insufficient

Note 1 to entry: A CONTACT QUALITY MONITOR is functional only when used with a MONITORING NE.

[SOURCE: IEC 60601-2-2:2017, 201.3.211]

3.12**CONTINUITY MONITOR**

circuit in HF SURGICAL EQUIPMENT or ASSOCIATED EQUIPMENT intended for connection to an NE, providing an alarm in the event of electrical discontinuity in the NE cable or its connections

[SOURCE: IEC 60601-2-2:2017, 201.3.212]

3.13**CUTTING**

division of body tissue caused by the passage of HIGH FREQUENCY current of high current density at the ACTIVE ELECTRODE(S)

[SOURCE: IEC 60601-2-2:2017, 201.3.214]

3.14**FINGERSWITCH**

device generally included with an ACTIVE ACCESSORY which, when manipulated by the OPERATOR, enables HF output to be produced and, when released disables HF output

[SOURCE: IEC 60601-2-2:2017, 201.3.216]

3.15**HAZARD**

potential source of harm

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.39]

3.16**HEATING FACTOR**

a value equal to $I^2 \times t$ where I is the MONOPOLAR current in amperes and t is the duration of the current flow in s

Note 1 to entry: The HEATING FACTOR is expressed as A²s (amperes squared seconds).

[SOURCE: IEC 60601-2-2:2017, 201.3.218]

3.17

HIGH CURRENT MODE

MONOPOLAR output mode whose INTENDED USE (MAXIMUM OUTPUT CURRENT and maximum DUTY CYCLE) results in a HEATING FACTOR of greater than 30 A²s in any 60 s period

[SOURCE: IEC 60601-2-2:2017, 201.3.219]

3.18

HIGH FREQUENCY

HF

frequencies less than 5 MHz and generally greater than 200 kHz

[SOURCE: IEC 60601-2-2:2017, 201.3.220]

3.19

HIGH FREQUENCY SURGICAL ACCESSORY

HF SURGICAL ACCESSORY

ACCESSORY intended to conduct, supplement or monitor HF energy applied to the PATIENT from HF SURGICAL EQUIPMENT

Note 1 to entry: HF SURGICAL ACCESSORIES include HF surgical application electrodes, including cords and connectors for attachment to HF SURGICAL EQUIPMENT, as well as other ASSOCIATED EQUIPMENT intended for connection to the HF surgical PATIENT circuit.

Note 2 to entry: Not all accessories used with HF surgical equipment are HF SURGICAL ACCESSORIES.

[SOURCE: IEC 60601-2-2:2017, 201.3.223, modified – In Note 1, replacement of "active electrodes" by " HF surgical application electrodes".]

3.20

HIGH FREQUENCY SURGICAL EQUIPMENT

HF SURGICAL EQUIPMENT

MEDICAL ELECTRICAL EQUIPMENT which generates HIGH FREQUENCY currents intended for the performance of surgical tasks such as CUTTING or COAGULATION of biological tissue by means of these HIGH FREQUENCY currents

Note 1 to entry: HF SURGICAL EQUIPMENT is also variously known as surgical diathermy, electrosurgical equipment, electrosurgical generator, RF generator or HF generator.

Note 2 to entry: A footswitch is an example of an associated ACCESSORY that is part of HF SURGICAL EQUIPMENT.

[SOURCE: IEC 60601-2-2:2017, 201.3.224]

3.21

HF SURGICAL MODE

any of a number of OPERATOR selectable HF output characteristics intended to provide a specific effect at a connected ACTIVE ACCESSORY, such as CUTTING, COAGULATION and the like

Note 1 to entry: Each available HF SURGICAL MODE may be provided with an OPERATOR-adjustable output control to set the desired intensity or speed of the effect.

[SOURCE: IEC 60601-2-2:2017, 201.3.225]

3.22

INTENDED USE

INTENDED PURPOSE

use for which a product, process or service is intended according to the specifications, instructions and information provided by the manufacturer

[SOURCE: IEC 60601-1:2005/AMD1:2012, 3.44]