

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

Electrical equipment for measurement, control and laboratory use –  
EMC requirements –

Part 2-6: Particular requirements – In vitro diagnostic (IVD) medical equipment

Matériel électrique de mesure, de commande et de laboratoire –  
Exigences relatives à la CEM –

Partie 2-6: Exigences particulières – Matériel médical de diagnostic in vitro (IVD)



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22,000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67,000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Electrical equipment for measurement, control and laboratory use –  
EMC requirements –  
Part 2-6: Particular requirements – In vitro diagnostic (IVD) medical equipment**

**Matériel électrique de mesure, de commande et de laboratoire –  
Exigences relatives à la CEM –  
Partie 2-6: Exigences particulières – Matériel médical de diagnostic in vitro (IVD)**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 17.220.20; 25.040.40; 33.100.20

ISBN 978-2-8322-8983-9

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD .....	3
1 Scope .....	5
2 Normative references .....	5
3 Terms and definitions .....	5
4 General .....	7
5 EMC test plan .....	7
5.1 General .....	7
5.2 Configuration of EUT during testing .....	7
5.3 Operation conditions of EUT during testing .....	7
5.4 Specification of FUNCTIONAL PERFORMANCE .....	7
5.5 Test description .....	7
6 Immunity requirements .....	7
6.1 Conditions during the tests .....	7
6.2 Immunity test requirements .....	8
6.3 Random aspects .....	11
6.4 Performance criteria .....	11
7 Emission requirements .....	12
8 Test results and test report .....	12
9 Instructions for use .....	12
Annex A (normative) Immunity test requirements for PORTABLE TEST AND MEASUREMENT EQUIPMENT powered by battery or from the circuit being measured .....	14
Annex B (informative) Guide for analysis and assessment for electromagnetic compatibility .....	15
Bibliography .....	16
Table 101 – Immunity test requirements for equipment intended to be used in PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT .....	9
Table 102 – Immunity test requirements for equipment intended to be used in a HOME HEALTHCARE ENVIRONMENT .....	10
Table 103 – Immunity test requirements for equipment intended to be used in a HOME HEALTHCARE ENVIRONMENT .....	11

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ELECTRICAL EQUIPMENT FOR MEASUREMENT,  
CONTROL AND LABORATORY USE –  
EMC REQUIREMENTS –****Part 2-6: Particular requirements –  
In vitro diagnostic (IVD) medical equipment**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61326-2-6 has been prepared by subcommittee 65A: System aspects, of IEC technical committee 65: Industrial-process measurement, control and automation.

This third edition cancels and replaces the second published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical change with respect to the previous edition:

- update of the document with respect to IEC 61326-1:2020.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
65A/979/FDIS	65A/990/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document the following print types are used:

- Terms used throughout this document which have been defined in Clause 3 of this document and of IEC 61326-1:2020: SMALL CAPITALS.

This part of IEC 61326 is to be used in conjunction with IEC 61326-1:2020 and follows the same numbering of clauses, subclauses, tables and figures.

When a particular subclause of IEC 61326-1 is not mentioned in this part, that subclause applies as far as is reasonable. When this standard states “addition”, “modification” or “replacement”, the relevant text in IEC 61326-1 is to be adapted accordingly.

NOTE The following numbering system is used:

- subclauses, tables and figures that are numbered starting from 101 are additional to those in IEC 61326-1;
- unless notes are in a new subclause or involve notes in IEC 61326-1, they are numbered starting from 101 including those in a replaced clause or subclause;
- additional annexes are lettered AA, BB, etc.

A list of all parts of the IEC 61326 series, under the general title *Electrical equipment for measurement, control and laboratory use – EMC requirements*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

# ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE – EMC REQUIREMENTS –

## Part 2-6: Particular requirements – In vitro diagnostic (IVD) medical equipment

### 1 Scope

In addition to the scope of IEC 61326-1, this part of IEC 61326 specifies minimum requirements for immunity and emissions regarding electromagnetic compatibility for IN VITRO DIAGNOSTIC (IVD) MEDICAL EQUIPMENT, taking into account the particularities and specific aspects of this electrical equipment and their electromagnetic environment.

### 2 Normative references

Clause 2 of IEC 61326-1:2020 applies, except as follows:

*Addition:*

IEC 61326-1:2020, *Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements*

ISO 14971:2019, *Medical devices – Application of risk management to medical devices*  
<https://standards.iteh.ai/catalog/standards/sist/52699cc8-6202-4f1d-a34e-33ec75609167/iec-61326-2-6-2020>

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 61326-1 apply, except as follows.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

*Addition:*

#### 3.101

##### **in vitro diagnostic medical equipment**

instruments and apparatus intended for use in the diagnosis of disease or other conditions, including a determination of the state of health, in order to cure, mitigate, treat, or prevent disease

Note 1 to entry: Such instruments or apparatus are intended for use in the collection, preparation, and examination of specimens taken from the human body without direct or wired patient connection with the device.

Note 2 to entry: IVD: In vitro diagnostic.

#### 3.102

##### **professional healthcare facility environment**

environment where professional healthcare is administered

Note 1 to entry: Locations include hospitals, diagnostic laboratories, blood banks, blood donation centres, physician offices, intensive care units, surgical centres, emergency rooms, surgery rooms, clinics, patient rooms, dental offices, limited care facilities, nursing homes, drugstore with trained operator, and first aid rooms.

Note 2 to entry: Most environments and locations in the PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT are considered to have a CONTROLLED ELECTROMAGNETIC ENVIRONMENT with regard to fixed electromagnetic sources. However, mobile communication devices are widely used by healthcare professionals in providing efficient patient care. For this reason, it is more difficult to control the environment for proximity electromagnetic disturbances. Examples of electromagnetic sources that might be used adjacent to IVD MEDICAL EQUIPMENT are:

- high frequency surgical equipment;
- radio frequency identification (RFID) systems;
- wireless local area networks (WLAN);
- handheld mobile radios (e.g. TETRA, two-way radio);
- paging systems;
- other wireless devices (including consumer devices).

Note 3 to entry: It is assumed that IVD MEDICAL EQUIPMENT is not directly connected to the public mains network.

Note 4 to entry: IVD MEDICAL EQUIPMENT should have a suitable level of immunity to ensure the safe and effective performance of the device in its intended use environment. IVD MEDICAL EQUIPMENT that might be used in ambulances, or any ground vehicle or aircraft can require a higher level of immunity than the PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT.

### 3.103

#### home healthcare environment

environment other than a PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT with a much more diverse electromagnetic environment with electromagnetic disturbances that may be more uncontrolled and less well-characterized in terms of amplitude and probability of occurrence

Note 1 to entry: Except in transportation or while operating under battery power, IVD MEDICAL EQUIPMENT is usually connected to the public mains network.

Note 2 to entry: The characteristics of this environment justify higher immunity test levels for basic safety and ESSENTIAL PERFORMANCE. Locations include the home, and any public place such as shops and libraries, offices, transport stations and airports, etc. Examples of electromagnetic sources that might be used near IVD MEDICAL EQUIPMENT in these environments or otherwise expose the IVD MEDICAL EQUIPMENT to intense electromagnetic disturbances are:

- small mains frequency transformers (50 Hz and 60 Hz), e.g. in a clock radio on a bedside table;
- mains disturbances;
- mobile phones (often several);
- fixed radio broadcast stations;
- TV transmitting equipment;
- amateur Radio Equipment;
- mobile radio transmitters (e.g. taxi, police).

### 3.104

#### analyte

constituent of a sample with a measurable property

EXAMPLES In “mass of protein in 24-hour urine”, “protein” is the ANALYTE and “mass” is the property. In “concentration of glucose in plasma”, “glucose” is the ANALYTE and “concentration” is the property. In both cases, the full phrase designates the measured property.

[SOURCE: ISO 18113-1:2009, 3.3, modified – property has been added after measurand]

### 3.105

#### basic safety

freedom from unacceptable risk to the operator directly caused by physical hazards when IVD MEDICAL EQUIPMENT is used under normal condition and single fault condition



**3.106****essential performance**

performance of a clinical function, other than that related to BASIC SAFETY, where loss or degradation beyond the limits specified in the user documentation results in an unacceptable risk

Note 1 to entry: ESSENTIAL PERFORMANCE is most easily understood by considering whether its absence or degradation would result in an unacceptable risk.

**3.2 Abbreviations**

Subclause 3.2 of IEC 61326-1:2020 applies.

**4 General**

Clause 4 of IEC 61326-1:2020 applies.

**5 EMC test plan****5.1 General**

Subclause 5.1 of IEC 61326-1:2020 applies.

**5.2 Configuration of EUT during testing**

Subclause 5.2 of IEC 61326-1:2020 applies.

**5.3 Operation conditions of EUT during testing**

Subclause 5.3 of IEC 61326-1:2020 applies, except as follows:

*Addition:*

**5.3.101 Operational conditions**

The device shall be set to conditions specified in the user documentation in accordance with the intended use.

For each mode of operation, the power option (e.g. battery, AC or DC supply) which represents the most severe condition shall be specified in accordance with the product risk analysis.

**5.4 Specification of FUNCTIONAL PERFORMANCE**

Subclause 5.4 of IEC 61326-1:2020 applies.

**5.5 Test description**

Subclause 5.5 of IEC 61326-1:2020 applies.

**6 Immunity requirements****6.1 Conditions during the tests**

Subclause 6.1 of IEC 61326-1:2020 is replaced as follows.

### 6.101 Conditions during the tests

The configuration and modes of operation during the tests shall be precisely noted in the test report.

Tests shall be applied to the relevant PORTS in accordance with Table 101 or Table 102 of this document, as applicable.

The tests shall be carried out one at a time in accordance with the basic standards. If additional methods are required, the method and rationale shall be documented in the test report.

## 6.2 Immunity test requirements

Subclause 6.2 of IEC 61326-1:2020 and its title are replaced as follows.

### 6.2 Risk assessment and consideration of EMC immunity requirements

Powerful electromagnetic emission sources can lead to malfunctions in nearby medical equipment under certain circumstances. Different types of medical electrical equipment have different levels of risk with a malfunction. IVD MEDICAL EQUIPMENT however is not intended to keep alive or resuscitate patients, so a malfunction would not directly cause the death or serious injury of a patient. Such a malfunction in IVD MEDICAL EQUIPMENT can result in an incorrect reading, which can in turn lead to a wrong therapeutic decision (misdiagnosis). For some ANALYTES and in some circumstances, an incorrect result could result in serious harm to the patient. In the case of larger IVD MEDICAL EQUIPMENT, electromagnetic disturbances can also cause malfunctions that pose a direct threat to the operator, for example through unexpected mechanical movements.

The manufacturer shall perform risk management according to ISO 14971 for guidance in assessing risk associated with direct hazards.

NOTE As a rule, results from IVD MEDICAL EQUIPMENT are checked for plausibility by medical personnel or followed-up by decisions of a healthcare professional. IVD MEDICAL EQUIPMENT for self-testing by lay users is always provided with advice on action to be taken in case of indeterminate results. The users are urged to contact their medical practitioner before making any decision of medical relevance.

For equipment intended to be used in a PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT, the immunity requirements of Table 101 shall be applied.

In addition, depending on the outcome of the risk assessment of the electromagnetic environment, consideration shall be given to include the immunity test requirements of Table 102 and Table 103.

The risk assessment shall include point-of-care testing equipment or other equipment used close to patients, who might have home use electronics such as mobile phones even in the professional environment.

**Table 101 – Immunity test requirements for equipment intended to be used in PROFESSIONAL HEALTHCARE FACILITY ENVIRONMENT**

PORT	Phenomenon	Basic standard	Test value	Performance criterion
ENCLOSURE	Electrostatic discharge	IEC 61000-4-2	± 4 kV contact ± 2 kV, ± 4 kV, ± 8 kV air	B B
	Electromagnetic field	IEC 61000-4-3	3 V/m (80 MHz to 6 GHz)	A
	Power frequency magnetic field	IEC 61000-4-8	3 A/m (50 Hz, 60 Hz)	A
AC power (including protective earth)	Burst	IEC 61000-4-4	± 1 kV (5 kHz or 100 kHz)	B
	Surge	IEC 61000-4-5	± 0,5 kV line-to-line ± 1 kV line-to-ground	B B
	Conducted RF	IEC 61000-4-6	3 V (150 kHz to 80 MHz)	A
	Voltage dip	IEC 61000-4-11	0 % during 0,5 cycles 0 % during 1 cycle	B B
			70 % during 25/30 cycles <sup>a</sup>	C
	Short interruptions	IEC 61000-4-11	0 % during 250/300 cycles <sup>a</sup>	C
DC power <sup>b, c</sup> (including protective earth)	Burst	IEC 61000-4-4	± 1 kV (5 kHz or 100 kHz)	B
	Surge	IEC 61000-4-5	± 0,5 kV line-to-line ± 1 kV line-to-ground	B B
	Conducted RF	IEC 61000-4-6	3 V (150 kHz to 80 MHz)	A
I/O signal/control <sup>c</sup> (including functional earth)	Burst <sup>b</sup>	IEC 61000-4-4	± 0,5 kV (5 kHz or 100 kHz)	B
	Surge <sup>e</sup>	IEC 61000-4-5	± 1 kV line-to-ground	B
	Conducted RF <sup>b</sup>	IEC 61000-4-6	3 V (150 kHz to 80 MHz)	A
I/O signal/control <sup>c</sup> connected directly to mains supply	Burst <sup>b</sup>	IEC 61000-4-4	± 1 kV (5 kHz or 100 kHz)	B
	Surge <sup>d</sup>	IEC 61000-4-5	± 0,5 kV line-to-line ± 1 kV line-to-ground	B B
	Conducted RF <sup>b</sup>	IEC 61000-4-6	3 V (150 kHz to 80 MHz)	A
<sup>a</sup> For example, "25/30 cycles" means "25 cycles for 50 Hz test" or "30 cycles for 60 Hz test". <sup>b</sup> Only in the case of lines > 3 m. <sup>c</sup> DC POWER PORTS intended to be connected to a low voltage DC supply (≤ 60 V), where secondary circuits (isolated from the AC mains supply) are not subject to transient overvoltages (i.e. reliably-grounded, capacitive-filtered DC secondary circuits) shall be regarded as I/O signal/control PORTS. <sup>d</sup> Only in the case of LONG-DISTANCE LINES.				

For equipment intended to be used in a HOME HEALTHCARE ENVIRONMENT, the immunity requirements of Table 102 and Table 103 shall be applied.

**Table 102 – Immunity test requirements for equipment intended to be used in a HOME HEALTHCARE ENVIRONMENT**

PORT	Phenomenon	Basic standard	Test value	Performance criterion
ENCLOSURE	Electrostatic discharge	IEC 61000-4-2	±6 kV contact ±2 kV, ±4 kV, ±8 kV air	B B
	Electrostatic discharge	IEC 61000-4-2	±8 kV contact ±15 kV air	C C
	Electromagnetic field	IEC 61000-4-3	10 V/m (80 MHz to 1 GHz) 3 V/m (1 GHz to 6 GHz)	A A
	Magnetic field	IEC 61000-4-8	30 A/m (50 Hz, 60 Hz)	A
AC power	Voltage dip	IEC 61000-4-11	0 % during 0,5 cycles 0 % during 1 cycle	B B
			70 % during 25/30 cycles <sup>a</sup>	C
	Short interruptions	IEC 61000-4-11	0 % during 250/300 cycles <sup>a</sup>	C
	Burst	IEC 61000-4-4	± 2 kV (5 kHz or 100 kHz)	B
	Surge <sup>e</sup>	IEC 61000-4-5	± 0,5 kV, ± 1 kV line-to-line ± 0,5 kV, ± 1 kV, ± 2 kV line-to-ground	B B
	Conducted RF	IEC 61000-4-6	3 V (150 kHz to 80 MHz) 6 V in ISM and amateur radio bands in the frequency range 150 kHz to 80 MHz <sup>b</sup> 80 % AM at 1 kHz	A A
DC power and/or I/O signal/control ports, including protective earth	Burst	IEC 61000-4-4	± 2 kV (5 kHz or 100 kHz)	B
	Surge <sup>e</sup>	IEC 61000-4-5	± 0,5, ± 1 kV Line to line ± 0,5 kV, ± 1 kV, ± 2 kV Line to ground	B B
	Conducted RF	IEC 61000-4-6	3 V (150 kHz to 80 MHz) 6 V in ISM and amateur radio bands in the frequency range 150 kHz to 80 MHz <sup>e</sup> 80 % AM at 1 kHz	A A

<sup>a</sup> For example "25/30 cycles" means "25 cycles for 50 Hz test" or "30 cycles for 60 Hz test".

<sup>b</sup> The ISM (industrial, scientific and medical) bands between 0,15 MHz and 80 MHz are 6,765 MHz to 6,795 MHz; 13,553 MHz to 13,567 MHz; 26,957 MHz to 27,283 MHz; and 40,66 MHz to 40,70 MHz. The amateur radio bands between 0,15 MHz and 80 MHz are 1,8 MHz to 2,0 MHz, 3,5 MHz to 4,0 MHz, 5,3 MHz to 5,4 MHz, 7 MHz to 7,3 MHz, 10,1 MHz to 10,15 MHz, 14 MHz to 14,2 MHz, 18,07 MHz to 18,17 MHz, 21,0 MHz to 21,4 MHz, 24,89 MHz to 24,99 MHz, 28,0 MHz to 29,7 MHz and 50,0 MHz to 54,0 MHz.

<sup>c</sup> Only in the case of lines > 3 m.

<sup>d</sup> DC connections between parts of equipment/system which are not connected to a DC DISTRIBUTION NETWORK are treated as I/O signal/control PORTS (USB charging and communication combined). Using for example USB connection only as a power source, it falls under DC power.

<sup>e</sup> Only in the case of LONG-DISTANCE LINES.

The frequencies and services listed in Table 103 are representative examples that are based on RF communications equipment in use at the time of publication of this document. The test specification does not attempt to cover every frequency and service used in every country. The risk management process should take the country specific communications services into account. Testing should be performed at the additional frequencies identified that are not represented in Table 103.