



Designation: D5191 – 19

# Standard Test Method for Vapor Pressure of Petroleum Products and Liquid Fuels (Mini Method)<sup>1</sup>

This standard is issued under the fixed designation D5191; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

*This standard has been approved for use by agencies of the U.S. Department of Defense.*

## 1. Scope\*

1.1 This test method covers the use of automated vapor pressure instruments to determine the total vapor pressure exerted in vacuum by air-containing, volatile, liquid petroleum products and liquid fuels, including automotive spark-ignition fuels with or without oxygenates and with ethanol blends up to 85 % (volume fraction) (see **Note 1**). This test method is suitable for testing samples with boiling points above 0 °C (32 °F) that exert a vapor pressure between 7 kPa and 130 kPa (1.0 psi and 18.6 psi) at 37.8 °C (100 °F) at a vapor-to-liquid ratio of 4:1. Measurements are made on liquid sample sizes in the range from 1 mL to 10 mL. No account is made for dissolved water in the sample.

NOTE 1—The precision (see Section 16) using 1 L containers was determined in a 2003 interlaboratory study (ILS);<sup>2</sup> the precision using 250 mL containers was determined in a 2016 ILS.<sup>3</sup>

NOTE 2—Samples can also be tested at other vapor-to-liquid ratios, temperatures, and pressures, but the precision and bias statements need not apply.

NOTE 3—The ILS conducted in 1988, 1991, 2003, and 2016 to determine the precision statements in Test Method D5191 did not include any crude oil in the sample sets. Test Method D6377, as well as IP 481, have been shown to be suitable for vapor pressure measurements of crude oils.

1.1.1 Some gasoline-oxygenate blends may show a haze when cooled to 0 °C to 1 °C. If a haze is observed in 8.5, it shall be indicated in the reporting of results. The precision and bias statements for hazy samples have not been determined (see **Note 15**).

1.2 This test method is suitable for calculation of the dry vapor pressure equivalent (DVPE) of gasoline and gasoline-

oxygenate blends by means of a correlation equation (see **Eq 1** in 14.2). The calculated DVPE very closely approximates the dry vapor pressure that would be obtained on the same material when tested by Test Method D4953.

1.3 The values stated in SI units are to be regarded as standard. The values given in parentheses after SI units are provided for information only and are not considered standard.

1.4 **WARNING**—Mercury has been designated by many regulatory agencies as a hazardous substance that can cause serious medical issues. Mercury, or its vapor, has been demonstrated to be hazardous to health and corrosive to materials. Use Caution when handling mercury and mercury-containing products. See the applicable product Safety Data Sheet (SDS) for additional information. The potential exists that selling mercury or mercury-containing products, or both, is prohibited by local or national law. Users must determine legality of sales in their location.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* For specific safety warning statements, see 7.2 through 7.8.

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

- 2.1 *ASTM Standards*:<sup>4</sup>
- D2892 Test Method for Distillation of Crude Petroleum (15-Theoretical Plate Column)
  - D4057 Practice for Manual Sampling of Petroleum and Petroleum Products

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.08 on Volatility.

Current edition approved Jan. 1, 2019. Published February 2019. Originally approved in 1991. Last previous edition approved in 2018 as D5191 – 18a. DOI: 10.1520/D5191-19.

<sup>2</sup> Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1619. Contact ASTM Customer Service at service@astm.org.

<sup>3</sup> Research Report IP 394 (EN 13016-1) and IP 619 (EN 13016-3) 2016, available from the Energy Institute, 61 New Cavendish Street, London W1G 7AR, UK, email: ILS@energyinst.org.

<sup>4</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

\*A Summary of Changes section appears at the end of this standard

- [D4953 Test Method for Vapor Pressure of Gasoline and Gasoline-Oxygenate Blends \(Dry Method\)](#)
  - [D5798 Specification for Ethanol Fuel Blends for Flexible-Fuel Automotive Spark-Ignition Engines](#)
  - [D6299 Practice for Applying Statistical Quality Assurance and Control Charting Techniques to Evaluate Analytical Measurement System Performance](#)
  - [D6377 Test Method for Determination of Vapor Pressure of Crude Oil: VPCR<sub>x</sub> \(Expansion Method\)](#)
  - [D6378 Test Method for Determination of Vapor Pressure \(VP<sub>x</sub>\) of Petroleum Products, Hydrocarbons, and Hydrocarbon-Oxygenate Mixtures \(Triple Expansion Method\)](#)
  - [D7717 Practice for Preparing Volumetric Blends of Denatured Fuel Ethanol and Gasoline Blendstocks for Laboratory Analysis](#)
- 2.2 *IP Standard:*
- [IP 481 Test Method for Determination of the Air Saturated Vapour Pressure \(ASVP\) of Crude Oil<sup>5</sup>](#)

### 3. Terminology

#### 3.1 Definitions:

3.1.1 *fuel ethanol (Ed75–Ed85), n*—blend of ethanol and hydrocarbon, of which the ethanol portion is nominally 75 to 85 volume % denatured fuel ethanol.

#### 3.2 Definitions of Terms Specific to This Standard:

3.2.1 *absolute vapor pressure, n*—the pressure of the air-free sample. It is calculated from the total vapor pressure of the sample by subtracting out the partial pressure of the dissolved air.

3.2.2 *dry vapor pressure equivalent (DVPE), n*—a value calculated by a correlation equation (see 14.2) from the total vapor pressure.

3.2.2.1 *Discussion*—The DVPE is expected to be equivalent to the value obtained on the sample by Test Method [D4953](#), Procedure A.

3.2.3 *total vapor pressure (P<sub>tot</sub>), n*—the observed pressure measured in the experiment that is the sum of the partial pressure of the sample and the partial pressure of the dissolved air.

#### 3.3 Abbreviations:

3.3.1 *DVPE*—dry vapor pressure equivalent

3.3.2 *P<sub>tot</sub>*—total vapor pressure

### 4. Summary of Test Method

4.1 A known volume of chilled, air-saturated sample is introduced into a thermostatically controlled, evacuated test chamber, or a test chamber with a moveable piston that expands the volume after sample introduction, the internal volume of which is five times that of the total test specimen introduced into the chamber. After introduction into the test chamber, the test specimen is allowed to reach thermal equilibrium at the test temperature, 37.8 °C (100 °F). The resulting rise in pressure in the chamber is measured using a pressure

transducer sensor and indicator. Only total pressure measurements (sum of the partial pressure of the sample and the partial pressure of the dissolved air) are used in this test method, although some instruments can measure the absolute pressure of the sample as well.

4.2 The measured total vapor pressure is converted to a dry vapor pressure equivalent (DVPE) by use of a correlation equation (see [Eq 1](#) in [14.2](#)).

### 5. Significance and Use

5.1 Vapor pressure is a very important physical property of volatile liquids.

5.2 The vapor pressure of gasoline and gasoline-oxygenate blends is regulated by various government agencies.

5.3 Specifications for volatile petroleum products generally include vapor pressure limits to ensure products of suitable volatility performance.

5.4 This test method is more precise than Test Method [D4953](#), uses a small sample size (1 mL to 10 mL), and requires about 7 min to complete the test.

### 6. Apparatus

6.1 *Vapor Pressure Apparatus*—The type of apparatus suitable for use in this test method employs a small volume test chamber incorporating a transducer for pressure measurements and associated equipment for thermostatically controlling the chamber temperature and for evacuating the test chamber prior to sample introduction or expanding the volume after sample introduction by a moveable piston.

6.1.1 The test chamber shall be designed to contain between 5 mL and 50 mL of liquid and vapor and be capable of maintaining a vapor-to-liquid ratio between 3.95 to 1.00 and 4.05 to 1.00.

NOTE 4—The test chamber employed by the instruments used in generating the precision and bias statements were constructed of stainless steel, aluminum, or brass.

NOTE 5—Test chambers exceeding a 15 mL capacity can be used, but the precision and bias statements (see Section 16) are not known to apply.

6.1.2 The pressure transducer shall have a minimum operational range from 0 kPa to 177 kPa (0 psi to 25.7 psi) with a minimum resolution of 0.1 kPa (0.01 psi) and a minimum accuracy of ±0.8 kPa (±0.12 psi). The pressure measurement system shall include associated electronics and readout devices to display the resulting pressure reading.

6.1.3 A thermostatically controlled heater shall be used to maintain the test chamber at 37.8 °C ± 0.1 °C (100 °F ± 0.2 °F) for the duration of the vapor pressure measurement.

6.1.4 A platinum resistance thermometer shall be used for measuring the temperature of the test chamber with a resolution of 0.1 °C (0.2 °F) and an accuracy of 0.1 °C (0.2 °F).

6.1.5 The vapor pressure apparatus shall have provisions for introduction of the test specimen into an evacuated test chamber, or into a test chamber by a moveable piston, and for the cleaning or purging of the chamber following or preceding the test.

6.2 *Vacuum Pump*, capable of reducing the pressure in the test chamber to less than 0.01 kPa (0.001 psi) absolute.

<sup>5</sup> Available from Energy Institute, 61 New Cavendish St., London, W1G 7AR, U.K., <http://www.energyinst.org>.

**TABLE 1 Accepted Reference Value (ARV) and Acceptable Testing Range for Reference Fluids Analyzed in the 2003 ILS<sup>A</sup>**

Reference Fluid	ARV [ $P_{\text{tot}}$ ] $\pm$ Uncertainty, (kPa)	Recommended Instrument Manufacturer Tolerance, (kPa)	Acceptable Testing Range for Reference Fuel [ $P_{\text{tot}}$ ], (kPa)
Pentane	112.8 $\pm$ 0.2	$\pm$ 1.0	112.8 $\pm$ 1.2 (111.6 to 114.0)
2,2 Dimethylbutane	74.1 $\pm$ 0.2	$\pm$ 1.0	74.1 $\pm$ 1.2 (72.9 to 75.3)
2,3 Dimethylbutane	57.1 $\pm$ 0.2	$\pm$ 1.0	57.1 $\pm$ 1.2 (55.9 to 58.3)
Reference Fluid	ARV [ $P_{\text{tot}}$ ] $\pm$ Uncertainty, (psi)	Recommended Instrument Manufacturer Tolerance, (psi)	Acceptable Testing Range for Reference Fuel [ $P_{\text{tot}}$ ], (psi)
Pentane	16.36 $\pm$ 0.03	$\pm$ 0.14	16.36 $\pm$ 0.17 (16.19 to 16.53)
2,2 Dimethylbutane	10.75 $\pm$ 0.03	$\pm$ 0.14	10.75 $\pm$ 0.17 (10.58 to 10.92)
2,3 Dimethylbutane	8.28 $\pm$ 0.03	$\pm$ 0.14	8.28 $\pm$ 0.17 (8.11 to 8.45)

<sup>A</sup> Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1619. Contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org).

6.3 *Syringe*, (optional, depending on sample introduction mechanism employed with each instrument) gas-tight, 1 mL to 20 mL capacity with a  $\pm 1\%$  or better accuracy and a  $\pm 1\%$  or better precision. If a syringe is used to measure the sample volume, the capacity of the syringe should not exceed two times the volume of the test specimen being dispensed.

6.4 *Iced Water Bath, Refrigerator, or Air Bath*, for chilling the samples and syringe to temperatures between 0 °C to 1 °C (32 °F to 34 °F).

6.5 *Pressure Measuring Device*, capable of measuring local station pressure with an accuracy of 0.20 kPa (0.03 psi), or better, at the same elevation relative to sea level as the apparatus in the laboratory.

6.5.1 When a mercury barometer is not used as the pressure measuring device, the calibration of the pressure measuring device employed shall be periodically checked (with traceability to a nationally recognized standard) to ensure that the device remains within the required accuracy specified in 6.5.

6.6 *McLeod Vacuum Gage or Calibrated Electronic Vacuum Measuring Device for Calibration*, to cover at least the range from 0.01 kPa to 0.67 kPa (0.1 mm Hg to 5 mm Hg). The calibration of the electronic vacuum measuring device shall be regularly verified in accordance with the annex section on Vacuum Sensors (A6.3) of Test Method D2892.

## 7. Reagents and Materials

7.1 *Purity of Reagents*—Use chemicals of at least 99 % purity for verification of instrument performance (see Section 11). Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.<sup>6</sup> Lower purities can be used, provided it is first ascertained that the reagent is of sufficient purity to permit its use without lessening the accuracy of the determination.

<sup>6</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

7.1.1 The chemicals in sections 7.2, 7.3, 7.4, 7.7, and 7.8 (blended by mass with pentane) are suggested for verification of instrument performance (see Section 11), based on the reference fuels analyzed in the 2003 interlaboratory study (ILS)<sup>2</sup> (see Table 1) and 2014 interlaboratory study (ILS)<sup>7</sup> (see Table 2). Such reference fuels are not to be used for instrument calibration. Table 1 and Table 2 identify the accepted reference value (ARV) and uncertainty limits, as well as the acceptable testing range for each of the reference fuels listed.

NOTE 6—Verification fluids reported by 28 of the 29 D5191 data set participants in the 2003 ILS<sup>2</sup> included the following (with number of data sets identified in parenthesis): 2,2-dimethylbutane (18), cyclopentane (5), pentane (2), 2,3-dimethylbutane (1), 3-methylpentane (1), and methanol (1).

7.2 *Cyclopentane*, (**Warning**—Cyclopentane is flammable and a health hazard).

7.3 *2,2-Dimethylbutane*, (**Warning**—2,2-dimethylbutane is flammable and a health hazard).

7.4 *2,3-Dimethylbutane*, (**Warning**—2,3-dimethylbutane is flammable and a health hazard).

7.5 *Methanol*, (**Warning**—Methanol is flammable and a health hazard).

7.6 *2-Methylpentane*, (**Warning**—2-methylpentane is flammable and a health hazard).

7.7 *Pentane*, (**Warning**—Pentane is flammable and a health hazard).

7.8 *Toluene*, (**Warning**—Toluene is flammable and a health hazard).

## 8. Sampling

### 8.1 General Requirements:

8.1.1 The extreme sensitivity of vapor pressure measurements to losses through evaporation and the resulting changes in the composition is such as to require the utmost precaution and the most meticulous care in the drawing and handling of samples.

<sup>7</sup> Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1805. Contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org).



**TABLE 2 Accepted Reference Value (ARV) and Acceptable Tolerance Range for Reference Fluids Analyzed in the 2014 ILS<sup>A</sup>**

Reference Fluid	ARV [ $P_{tot}$ ] $\pm$ Expanded Uncertainty, (kPa)	Reference Fluid Standard Deviation (kPa) Determined in ILS	Acceptable Tolerance Range for Single Result on Reference Fluid [ $P_{tot}$ ], (kPa)
Cyclopentane	72.97 $\pm$ 0.21	0.58	72.97 $\pm$ 1.95 (71.02 to 74.92)
22.0/78.0 (m/m) Blend of Pentane/Toluene	46.45 $\pm$ 0.13	0.36	46.45 $\pm$ 1.21 (45.24 to 47.66)
44.0/56.0 (m/m) Blend of Pentane/Toluene	68.78 $\pm$ 0.23	0.63	68.78 $\pm$ 2.10 (66.68 to 70.88)
68.5/31.5 (m/m) Blend of Pentane/Toluene	88.58 $\pm$ 0.21	0.58	88.58 $\pm$ 1.95 (86.63 to 90.53)

  

Reference Fluid	ARV [ $P_{tot}$ ] $\pm$ Expanded Uncertainty, (psi)	Reference Fluid Standard Deviation (psi) Determined in ILS	Acceptable Tolerance Range for Single Result on Reference Fluid [ $P_{tot}$ ], (psi)
Cyclopentane	10.58 $\pm$ 0.03	0.08	10.58 $\pm$ 0.28 (10.30 to 10.86)
22.0/78.0 (m/m) Blend of Pentane/Toluene	6.74 $\pm$ 0.02	0.05	6.74 $\pm$ 0.18 (6.56 to 6.92)
44.0/56.0 (m/m) Blend of Pentane/Toluene	9.98 $\pm$ 0.03	0.09	9.98 $\pm$ 0.30 (9.68 to 10.28)
68.5/31.5 (m/m) Blend of Pentane/Toluene	12.85 $\pm$ 0.03	0.08	12.85 $\pm$ 0.28 (12.57 to 13.13)

<sup>A</sup> Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1805. Contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org).

8.1.2 Obtain samples and test specimens in accordance with Practice **D4057**, except do not use the “Sampling by Water Displacement” section for fuels containing oxygenates. Use either 250 mL or 1 L (1 qt) sized containers filled between 70 % and 80 % with sample.

8.1.2.1 Samples in containers of other sizes, as prescribed in Practice **D4057**, may be used with the same ullage requirement but precision can be affected.

NOTE 7—The 2003 ILS<sup>2</sup> indicated that there was no statistically observable bias detected between 250 mL and 1 L containers.

8.1.3 To determine conformance with specifications, or regulations, or both, it may be necessary to prepare laboratory hand blends of gasoline blendstocks and denatured fuel ethanol for testing purposes, including vapor pressure. If necessary, then, a hand blend sample, prepared carefully in accordance with a procedure such as that described in Practice **D7717**, shall be considered suitable for the performance of this test. This hand blend, once prepared, shall then be considered equivalent to the ‘sample or samples,’ referred to in **8.1.4** through **8.5** of this Sampling section, as well as in subsequent sections of this test method.

8.1.4 Perform the vapor pressure determination on the first test specimen withdrawn from a sample container. Do not use the remaining sample in the container for a second vapor pressure determination. If a second determination is necessary, obtain a new sample.

NOTE 8—The effect of taking more than one test specimen from the same sample container was evaluated as part of the 2003 ILS.<sup>2</sup> A precision effect was observed between the first and second replicates taken from both the 1 L and 250 mL containers evaluated. The current precision statements were derived using the first test specimen withdrawn from 250 mL or 1 L containers.

8.1.5 Protect samples from excessive temperatures prior to testing. This can be accomplished by storage in an appropriate ice bath or refrigerator.

8.1.6 Do not test samples stored in leaky containers. Discard and obtain a new sample if leaks are detected.

8.2 *Sampling Handling Temperature*—Place the sample container and contents in an ice bath or refrigerator to the 0 °C to 1 °C (32 °F to 34 °F) range prior to opening the sample container. Allow sufficient time to reach this temperature.

NOTE 9—One way to verify the sample temperature is by direct measurement of the temperature of a similar liquid in a similar container placed in the cooling bath or refrigerator at the same time as the sample. Alternatively, temperature-monitoring studies conducted by laboratories have determined the minimum amount of time necessary to achieve the required temperature requirements stated in **8.2**, based upon typical sample receipt temperatures and cooling capacities of the instrumentation employed by the laboratory on samples included in such studies. The results of such studies have subsequently been applied to additional samples submitted for analysis. Typical minimum time durations reported by laboratories have ranged between approximately 30 min and 45 min, however, laboratories choosing this option need to determine the minimum cooling time required for their specific operation by conducting their own temperature-monitoring study.

8.3 *Verification of Sample Container Filling*—With the sample at a temperature of 0 °C to 1 °C, take the container from the cooling bath or refrigerator, and wipe dry with absorbent material. If the container is not transparent, unseal it and using a suitable gage, confirm that the sample volume equals 70 % to 80 % of the container capacity (see **Note 10**). If the sample is contained in a transparent glass container, verify that the container is 70 % to 80 % full by suitable means (see **Note 10**).

8.3.1 Discard the sample if the container is filled to less than 70 %, by volume, of the container capacity.

8.3.2 If the container is more than 80 % by volume full, pour out enough sample to bring the container contents within the 70 % to 80 % by volume range. Do not return any sample to the container once it has been withdrawn.

8.3.3 Reseal the container if necessary, and return the sample container to the cooling bath or refrigerator.

NOTE 10—For non-transparent containers, one way to confirm that the sample volume equals 70 % to 80 % of the container capacity is to use a dipstick that has been pre-marked to indicate the 70 % and 80 % container capacities. The dipstick should be of such material that it shows wetting after being immersed and withdrawn from the sample. To confirm the sample volume, insert the dipstick into the sample container so that it touches the bottom of the container at a perpendicular angle, before removing the dipstick. For transparent containers, using a marked ruler or by comparing the sample container to a like container which has the 70 % and 80 % levels clearly marked, has been found suitable.

8.4 *Air Saturation of the Sample in the Sample Container:*

8.4.1 *Transparent Containers Only*—Since **8.3** does not require that the sample container be opened to verify the