



Designation: D4055 – 04 (Reapproved 2019)

Standard Test Method for Pentane Insolubles by Membrane Filtration¹

This standard is issued under the fixed designation D4055; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of pentane insolubles for particles exceeding 0.8 μm in new and used lubricating oils.

NOTE 1—Pentane insolubles with particle sizes less than 0.8 μm may be studied with appropriate size membrane filters. Particle sizes above or below 0.8 μm can be studied. The precision of this test method has been determined only at 0.8 μm .

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* For specific warning statements, see 7.1, 8.2.1, and Annex A1.

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

D4057 Practice for Manual Sampling of Petroleum and Petroleum Products

D4175 Terminology Relating to Petroleum Products, Liquid Fuels, and Lubricants

D4177 Practice for Automatic Sampling of Petroleum and Petroleum Products

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.06 on Analysis of Liquid Fuels and Lubricants.

Current edition approved Jan. 1, 2019. Published February 2019. Originally approved in 1992. Last previous edition approved in 2013 as D4055 – 04 (2013). DOI: 10.1520/D4055-04R19.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3. Terminology

3.1 *Definitions:*

3.1.1 *pentane insolubles, in used oil analysis, n*—separated matter resulting when a used oil is dissolved in pentane.

3.1.1.1 *Discussion*—In this test method, the separation is effected by filtration through a membrane.

3.1.2 *used oil, n*—any oil that has been in a piece of equipment (for example, an engine, gearbox, transformer, or turbine) whether operated or not. **D4175**

4. Summary of Test Method

4.1 A sample of new or used lubricating oil is mixed with pentane in a volumetric flask. The oil solution is filtered through an 0.8 μm membrane filter. The flask, funnel, and filter are washed with additional pentane to effect a complete transfer of particulates onto the filter. The filter and its particulates are dried and weighed to give the pentane insolubles.

5. Significance and Use

5.1 Pentane insolubles above 0.8 μm in size may lead to increased wear. This increased wear can lead to premature equipment failure in critical applications.

6. Apparatus

6.1 In the development of this test method, it was noted that variations, particularly with respect to glassware and filter media, can affect the test result significantly.

6.2 When the user of this test method uses an alternate membrane filter, it is incumbent upon them to establish that the alternate filter will give equal results.

6.3 Precision data were established using the apparatus listed and a 0.8 μm filter pore size.

6.4 *Membrane Filter.*

6.5 *Borosilicate Filter Holder.*

6.6 *Borosilicate Filtering Flask.*

6.7 *Forceps* (plain flat tips, not serrated),

6.8 *Wash Bottle* equipped with 0.8 μm membranes.

6.9 *Analytical Balance*, capable of weighing with an accuracy of ± 0.1 mg.