

TECHNICAL REPORT

Summary of requirements and tests to products in the scope of IEC 60601-2-66
(standards.iteh.ai)

[IEC TR 62809:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fdcf8/iec-tr-62809-2019>



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

[IEC TR 62809:2019](#)

<https://standards.iec.ch/catalog/standards/sis/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019>

TECHNICAL REPORT

Summary of requirements and tests to products in the scope of IEC 60601-2-66
(standards.iteh.ai)

[IEC TR 62809:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.180.15; 17.140.50

ISBN 978-2-8322-7523-8

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD	3
INTRODUCTION	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	6
4 Summary of requirements and tests	7
Bibliography	18
Table 1 – Summary of requirements and tests	8

iTeh STANDARD PREVIEW **(standards.iteh.ai)**

[IEC TR 62809:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fdcf8/iec-tr-62809-2019>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SUMMARY OF REQUIREMENTS AND TESTS FOR
PRODUCTS IN THE SCOPE OF IEC 60601-2-66**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a Technical Report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 62809, which is a Technical Report, has been prepared by IEC technical committee 29: Electroacoustics.

This second edition cancels and replaces the first edition published in 2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) introduction of the term FITTED OSPL90 (FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL) (201.3.206 of IEC 60601-2-66:2019);
- b) the allowable maximum output sound pressure level is now based on FITTED MAXIMUM OUTPUT SOUND PRESSURE LEVEL (201.9.6 of IEC 60601-2-66:2019).
- c) ESSENTIAL PERFORMANCE is based on risk analysis (201.4.3).

The text of this Technical Report is based on the following documents:

Draft TR	Report on voting
29/1015/DTR	29/1019/RVDTR

Full information on the voting for the approval of this Technical Report can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

ITeH STANDARD PREVIEW
(standards.iteh.ai)

A bilingual version of this publication may be issued at a later date.

[IEC TR 62809:2019](#)

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019>

INTRODUCTION

During the preparation of IEC 60601-2-66, members of the involved technical committee and working group voiced concerns about the complexity of the document and its structure as part of the IEC 60601 series. Members felt distracted from the technical content by this complexity during reviews of the document stages. There was also concern that groups in the hearing aid community would have problems to understand and apply the standard and that this could be an issue with its acceptability.

In order to have a broad consensus for the new standard, it was agreed that the standard should be supported by this Technical Report, which should enable members of the community and the industry to have a basic understanding of the requirements of the standard, without the need to study the complete standard document and the documents that are referenced in it.

IEC 60601-2-66 was published to address the specific requirements for safety of hearing aids, and it is entitled “Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems”. It was published because IEC 60601-1 is a general standard intended to address a wide range of medical electrical equipment – including large scale facilities such as MRI machines, for example – and thus has large sections that are not relevant to low-voltage, low power, subminiature hearing aids.

If IEC 60601-2-66 was not published, test and regulatory organizations would probably have difficulty applying IEC 60601-1, because it does not contain specific guidance for hearing aids. This Technical Report contains all the requirements from IEC 60601-2-66 which relate to hearing aids and reduces discussion with those that do not relate to hearing aids.

It includes specific references to the applicable requirements within IEC 60601-1, and it is suggested that hearing aid designers and manufacturers along with test and regulatory organizations read this Technical Report as an overview of IEC 60601-2-66.

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019>

SUMMARY OF REQUIREMENTS AND TESTS FOR PRODUCTS IN THE SCOPE OF IEC 60601-2-66

1 Scope

This document, which is a Technical Report, provides an overview of the requirements and tests of IEC 60601-2-66 in combination with the applicable sections of IEC 60601-1, and the collateral standards of the IEC 60601 series.

NOTE The IEC 60601 series consists of three levels of standards: IEC 60601-1, known as the general standard, several IEC 60601-1-X documents, known as the collateral standards, and a series of particular standards covering requirements for specific types of equipment (IEC 60601-2-X).

It is intended to assist various groups involved in the product lifecycles process – like designers and suppliers – to get an overview of the basic requirements without studying all involved standard documents in detail. The table includes not all but just the more common requirements and tests.

It is crucial to understand that the summary in this document cannot serve as an input for a product requirement specification or as a test plan without consulting IEC 60601-2-66 itself. This document alone cannot be used to establish or assess compliance to IEC 60601-2-66.

The summary in Table 1 below does not preclude the user from reading the referenced standards in their entirety for a thorough knowledge of the basic safety of hearing aids and hearing aid systems.

[IEC TR 62809:2019](https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019)

2 Normative references

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019>

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-66:2019, *Medical electrical equipment – Part 2-66: Particular requirements for the basic safety and essential performance of hearing aids and hearing aid systems*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Summary of requirements and tests

The reference column in Table 1 shows the clause number of IEC 60601-2-66:2019 and, if applicable, the reference to IEC 60601-1:2005, or other documents. References to the particular standard IEC 60601-2-66:2019 start with the number 201, while references to the general standard, IEC 60601-1:2005, start directly with the clause or subclause number.

Other documents will be referred to explicitly. Some detailed references, for example describing tools, are placed in the text instead of in the reference column.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC TR 62809:2019](https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019)

<https://standards.iteh.ai/catalog/standards/sist/e93603cd-bdb8-47b9-bd05-5e8e03fddcf8/iec-tr-62809-2019>

Table 1 – Summary of requirements and tests

Reference	Subject	Requirements and tests
201.1.1 201.8	ACCESSORY	<p>Requirements:</p> <p>ACCESSORIES TO HEARING AIDS (e.g. remote-control units, audio streamers, battery chargers, power supplies) need to have documented proof of compliance to IEC 60065, IEC 60950-1, IEC 62368-1 or other applicable IEC safety standards. They form a HEARING AID SYSTEM when connected to the HEARING AID. Wherever this connection has an influence on the compliance to the requirements of IEC 60601-2-66, the HEARING AID has to pass the requirements while being connected to the ACCESSORY. If this connection results in additional requirements to the ACCESSORY, these requirements shall be fulfilled beyond the applicable IEC standards.</p> <p>Programming interfaces or ACCESSORIES in a clinical application are covered by IEC 60601 (all parts).</p> <p>For HEARING AIDS that are supplied by an external power source: If a particular separate power supply is specified, then the relevant tests are performed with the HEARING AIDS connected to it. If a generic separate power supply is specified, then the specification in the ACCOMPANYING DOCUMENTS is inspected.</p> <p>Tests:</p> <p>Inspection of ACCESSORY documentation, test configuration</p>
201.4.1 201.4 Clause 4 Clause 5	Type tests	<p>TYPE TESTS are performed on a representative sample of the item being tested. If multiple products are under consideration, which have a similar mechanical and electrical architecture, then an engineering analysis by the MANUFACTURER may justify a single representative sample for a family of products.</p> <p>Testing conditions shall consider not just NORMAL USE but also reasonably foreseeable misuse. Misuse and faults shall be the subject and results of the RISK ANALYSIS. The instructions for use shall be considered in testing conditions.</p> <p>The equipment is tested under the least favorable working conditions.</p>
4.2 4.5 ISO 14971	Risk management	<p>Requirements:</p> <p>A RISK MANAGEMENT complying with ISO 14971 shall be performed.</p> <p>Where IEC 60601-2-66 specifies requirements addressing particular RISKS, alternative means of addressing these RISKS are acceptable provided that the MANUFACTURER can justify that the RESIDUAL RISKS are the same or lower.</p> <p>Tests:</p> <p>Inspection of the RISK MANAGEMENT FILE.</p>
201.4.3 4.3	Essential performance	<p>Requirements:</p> <p>After a careful consideration of the clauses within this document, it was decided that they all deal with BASIC SAFETY as defined in the general standard. Manufacturers have the ability to identify functions of HEARING AIDS which are considered ESSENTIAL PERFORMANCE in accordance with their RISK MANAGEMENT PROCESS.</p>
4.4	Expected service life	<p>Requirements:</p> <p>The MANUFACTURER shall state the EXPECTED SERVICE LIFE of the HEARING AID in the RISK MANAGEMENT FILE.</p> <p>Tests:</p> <p>Inspection of the RISK MANAGEMENT FILE.</p>
4.8 4.9	Components	<p>Requirements:</p> <p>Components the failure of which could result in a HAZARDOUS SITUATION shall be used in accordance with their specified ratings. The reliability of components that are used as MEANS OF PROTECTION shall be assessed. They shall comply with the applicable safety requirements of a relevant IEC or ISO standard (options see 4.8).</p> <p>Tests:</p> <p>Inspection and, where necessary, test.</p>

Reference	Subject	Requirements and tests
7.1.3	Durability of markings	<p>Requirements:</p> <p>Markings shall be removable only with a TOOL or by appreciable force and shall be sufficiently durable to remain CLEARLY LEGIBLE during EXPECTED SERVICE LIFE.</p> <p>Tests:</p> <p>Markings are rubbed by hand, without undue pressure, first for 15 s with a cloth rag soaked with distilled water, then for 15 s with a cloth rag soaked with methylated spirit and then for 15 s with a cloth rag soaked with isopropyl alcohol.</p>
201.7.1.2 201.7.2.1	Legibility of markings	<p>Requirements:</p> <p>Markings shall be CLEARLY LEGIBLE when it is placed in the hand.</p> <p>The serial number and other markings shall be legible utilizing an optical aid if necessary.</p> <p>If the size of the HEARING AID does not allow affixation of all required markings, the markings shall be recorded in full in the ACCOMPANYING DOCUMENTS or instructions for use (IFU).</p> <p>Tests:</p> <p>Inspection and verification of the applicable requirements.</p>
201.7.2.2 7.3.3 7.4.2 7.8.2	Markings	<p>Requirements:</p> <p>HEARING AIDS shall be marked with:</p> <ul style="list-style-type: none"> – the name or trademark of the MANUFACTURER; – a MODEL OR TYPE REFERENCE. <p>HEARING AIDS shall be marked visibly on the outside or at a user accessible location (e.g. battery drawer), with:</p> <ul style="list-style-type: none"> – if needed: identification of right and left HEARING AID. Right = red. Left = blue; – serial number. <p>HEARING AIDS worn in the ear: The marking on the HEARING AID may be reduced to the serial number and the identification of right and left.</p> <p>The type of battery and the mode of insertion shall be marked on HEARING AIDS unless the design of the battery compartment prevents incorrect replacement of a battery</p> <p>Different positions of control devices and switches shall be indicated by figures, letters or other visual means.</p> <p>If the change of setting of a control could result in a RISK, such controls shall be provided with an indicating device or an indication of the direction in which the magnitude of the function changes.</p> <p>The color red shall be used only for emergency controls.</p> <p>Tests:</p> <p>Inspection and verification of the applicable requirements.</p>
201.7.2.17	Protective packaging	<p>Requirements:</p> <p>If special handling measures have to be taken during transport or storage, the packaging shall be marked accordingly.</p> <p>Tests:</p> <p>Inspection and verification of the applicable requirements.</p>
201.7.8.1 201.7.9 7.6.1	ACCOMPANYING DOCUMENTS	<p>Requirements:</p> <p>HEARING AIDS shall be accompanied by instructions containing at least:</p> <ul style="list-style-type: none"> – the purpose and INTENDED USE of the HEARING AID; – instructions for use, operating functions and a technical description; – easily understood diagrams, illustrations, or photographs of the fully assembled and ready-to-operate HEARING AID including all controls, visual information signals, and indicators, proper