

Edition 3.0 2020-09 REDLINE VERSION

INTERNATIONAL STANDARD



Medical electrical equipment — Standards

Part 2-50: Particular requirements for the basic safety and essential performance of infant phototherapy equipment and a standards.

Document Preview

IEC 60601-2-50:2020

https://standards.iteh.ai/catalog/standards/iec/995fdb90-ac99-4463-b0ce-c2be4b0cd7b0/iec-60601-2-50-2020





THIS PUBLICATION IS COPYRIGHT PROTECTED Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office 3, rue de Varembé CH-1211 Geneva 20 Switzerland

Tel.: +41 22 919 02 11

info@iec.ch www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished
Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.





Edition 3.0 2020-09 REDLINE VERSION

INTERNATIONAL STANDARD



Medical electrical equipment – Standards
Part 2-50: Particular requirements for the basic safety and essential performance of infant phototherapy equipment

Document Preview

IEC 60601-2-50:2020

https://standards.iteh.ai/catalog/standards/iec/995fdb90-ac99-4463-b0ce-c2be4b0cd7b0/iec-60601-2-50-2020

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ICS 11.040.60 ISBN 978-2-8322-8841-2

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

	FOREWO	RD	3	
	INTRODUCTION			
	201.1	Scope, object and related standards	7	
	201.2	Normative references	9	
	201.3	Terms and definitions	9	
	201.4	General requirements	11	
	201.5	General requirements for testing-of ME EQUIPMENT	11	
	201.6	Classification of ME EQUIPMENT and ME SYSTEMS	11	
	201.7	ME EQUIPMENT identification, marking and documents	12	
	201.8	Protection against electrical HAZARDS from ME EQUIPMENT	14	
	201.9	Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	14	
	201.10	Protection against unwanted and excessive radiation HAZARDS	16	
	201.11	Protection against excessive temperatures and other HAZARDS	17	
	201.12 outpu	Accuracy of controls and instruments and protection against hazardous	17	
	201.13	HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	20	
	201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	20	
	201.15	Construction of ME EQUIPMENT	20	
	201.16	ME SYSTEMS	20	
	201.17 ME S	Electromagnetic-COMPATIBILITY disturbances of ME EQUIPMENT and YSTEMS	20	
	202 Elect	romagnetic disturbances – Requirements and tests	20	
	Annexes	<u>1EC 60601-2-50:2020</u>	22	
	Annex AA	iteh a/catalog/standards/iec/995tdb90-ac99-4463-b0ce-c2be4b0cd7b0/iec-6060 (informative) Particular guidance and rationale	23	
Bibliography				
	Index of d	lefined terms used in this document	30	
	Figure 20	1.101 – Example of a measuring grid	18	
	_	1.102 – Layout of weight test devices		
Table 201.101 – List of symbols, abbreviations and acronyms				
	Table AA.1 – UV radiation exposure limits and spectral weighting function			
		·		

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT -

Part 2-50: Particular requirements for the basic safety and essential performance of infant phototherapy equipment

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicy Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

International standard IEC 60601-2-50 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2009 and Amendment 1:2016. This edition constitutes a technical revision.

This edition includes the following significant technical change with respect to the previous edition: re-dating of normative references.

The text of this International Standard d is based on the following documents:

FDIS	Report on voting
62D/1767/FDIS	62D/1775/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- test specifications: italic type;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type.
 Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

(https://standards.iteh.ai)

IEC 60601-2-50:2020

https://standards.iteh.ai/catalog/standards/iec/995fdb90-ac99-4463-b0ce-c2be4b0cd7b0/iec-60601-2-50-2020

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of INFANT PHOTOTHERAPY EQUIPMENT.

This particular standard amends and supplements IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance, hereinafter referred to as the general standard.

The requirements are followed by specifications for the relevant tests.

A general guidance and rationale for the requirements of this particular standard are given in Annex AA.

It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of this particular standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

iTeh Standards (https://standards.iteh.ai) Document Preview

IEC 60601-2-50:2020

https://standards.iteh.ai/catalog/standards/iec/995fdb90-ac99-4463-b0ce-c2be4b0cd7b0/iec-60601-2-50-2020

MEDICAL ELECTRICAL EQUIPMENT -

Part 2-50: Particular requirements for the basic safety and essential performance of infant phototherapy equipment

201.1 Scope, object and related standards

Clause 1 of the general standard applies, except as follows:

201.1.1 * Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of INFANT PHOTOTHERAPY EQUIPMENT, as defined in 201.3.203, also referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document, except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This particular standard specifies safety requirements for INFANT PHOTOTHERAPY EQUIPMENT, but alternate methods of compliance with a specific clause by demonstrating equivalent safety will not be judged as non-compliant if the MANUFACTURER has demonstrated in his RISK MANAGEMENT FILE that the RISK presented by the HAZARD has been found to be of an acceptable level when weighed against the benefit of treatment from the device.

This document does not apply to:

- devices supplying heat via BLANKETS, PADS or MATTRESSES in medical use; for information, see-IEC 80601-2-35 IEC 60601-2-35 [1]2;
- INFANT INCUBATORS; for information, see IEC 60601-2-19 [2];
- INFANT TRANSPORT INCUBATORS; for information, see IEC 60601-2-20 [3];
- INFANT RADIANT WARMERS; for information, see IEC 60601-2-21 [4].

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for INFANT PHOTOTHERAPY EQUIPMENT (as defined in 201.3.203), which reduce the safety HAZARDS to PATIENTS and OPERATORS as much as possible and to specify tests for demonstrating compliance with these requirements.

The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.

² The figures between brackets refer to the Bibliography.

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

– 8 –

IEC 60601-1-2:2014 applies as modified in Clause 202. IEC 60601-1-3 and IEC 60601-1-10³⁾ do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 is are referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standards corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard and applicable collateral standards are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through—3.139 3.147, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

³⁾ IEC 60601-1-10:2007, Medical electrical equipment — Part 1-10: General requirements for basic safety and essential performance — Collateral Standard: Requirements for the development of physiologic closed-loop controllers

The term "this document" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of the general standard applies, except as follows:

Amendment

Addition:

IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance IEC 60601-1:2005/AMD1:2012

Replacement:

IEC 60601-1-2:20072014, Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility disturbances – Requirements and tests

201.3 Terms and definitions

For the purposes of this document, the terms and definitions—given specified in IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012,—apply, except as follows and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at http://www.electropedia.org/
- ISO Online browsing platform: available at http://www.iso.org/obp

NOTE An index of defined terms is found on page 30. A list of symbols, abbreviations and acronyms used in this particular standard is given in Table 201.101.

Replacement:

201.3.76

PATIENT

INFANT, as specified under 201.3.202, who is being treated by means of visible radiation from INFANT PHOTOTHERAPY EQUIPMENT, as specified under 201.3.203

Addition:

201.3.201

EFFECTIVE IRRADIATED AREA

surface on which the PATIENT rests according to the intended position and which is irradiated by the INFANT PHOTOTHERAPY EQUIPMENT

Note 1 to entry: The EFFECTIVE IRRADIATED AREA is the intended treatment surface which is illuminated by the phototherapy light. The area of 60 cm × 30 cm is used as a standard-sized surface unless specified differently in the ACCOMPANYING DOCUMENTS.

201.3.202

INFANT

PATIENT up to the age of three months and a weight less than 10 kg

201.3.203

* INFANT PHOTOTHERAPY EQUIPMENT

ME EQUIPMENT which emits in the main radiation spectrum in the range between 400 nm and 550 nm for reducing the concentration of bilirubin in the body of INFANTS

201.3.204

TOTAL IRRADIANCE FOR BILIRUBIN

 E_{b}

irradiance equal to the total of all irradiance in the range between 400 nm and 550 nm

Table 201.101 - List of symbols, abbreviations and acronyms

Abbreviation	Term
AAP	American Academy of Pediatrics
°C	degrees Celsius (unit of temperature)
dB(A)	decibel A-weighted to human frequency response (a logarithmic measure of sound intensity)
Δ_{λ}	bandwidth (in nanometres)
E	irradiance (radiant power incidence per unit area on a surface)
E _{bi}	irradiance for bilirubin (total irradiance for 400 nm to 550 nm)
E _{eff}	effective irradiance
E_{λ}	spectral irradiance
EL	exposure limit
G ₂	uniformity of irradiance (unitless)
GHz	gigahertz (unit of frequency)
h	hour (unit of time)
IR	infrared radiation (with wavelengths between 700 nm and 1 mm)
IR-A	A region of infrared radiation (with wavelengths between 700 nm and 1 400 nm)
IR-B	B region of infrared radiation (with wavelengths between 1,4 μm and 3 $\mu m)$
IR-C	C region of infrared radiation (with wavelengths between 3 μm and 8 μm)
kg	kilograms (unit of mass)
λ	lambda (unit of wavelength)
m	meter (unit of length)
MHz	megahertz (unit of frequency)
min	minute (unit of time)
μW/cm ²	microwatts per square centimetre (unit of irradiance)
nm	nanometre (unit of length)
N	newton (unit of force)
s	second (unit of time)
S_{λ}	relative spectral effectiveness (unitless)
UV	ultraviolet radiation (with wavelength shorter than visible light)
UV-A	near-ultraviolet region (with wavelengths between 315 nm and 400 nm)
V/m	volts per meter (unit of electric field intensity)
W/cm ²	watts per square centimetre (unit of irradiance)

https

Abbreviation	Term
W/m ²	watts per square meter (unit of irradiance)

201.4 General requirements

Clause 4 of the general standard applies, except as follows:

201.4.3 * ESSENTIAL PERFORMANCE

Replacement:

There are no additional ESSENTIAL PERFORMANCE requirements for INFANT PHOTOTHERAPY EQUIPMENT.

201.5 General requirements for testing of ME EQUIPMENT

Clause 5 of the general standard applies, except as follows:

201.5.4 Other conditions

Additional subclauses:

201.5.4.101 * Pre-ageing

The following general operating conditions shall be taken into account for radiation measurements of INFANT PHOTOTHERAPY EQUIPMENT.

After 5 h of pre-ageing of the radiator source, or after the pre-ageing time specified by the MANUFACTURER, if the MANUFACTURER has specified a different pre-ageing time in the ACCOMPANYING DOCUMENTS, the initial values of TOTAL IRRADIANCE FOR BILIRUBIN $E_{\rm bi}$ for the INFANT PHOTOTHERAPY EQUIPMENT shall be measured at the normal operating conditions for the different irradiance settings defined by the MANUFACTURER.

201.5.4.102 Position of measurements

The radiation measurements shall be taken in the operating position of the lamp of the INFANT PHOTOTHERAPY EQUIPMENT at a distance specified by the MANUFACTURER disclosed in the instructions for use (see 201.7.9.2.9).

201.5.4.103 Stabilization period

The radiation measurements shall be taken when all important parameters for measurements have reached stable conditions. The stabilization period shall be at least 0,5 h, or longer, unless the MANUFACTURER states a different time in the ACCOMPANYING DOCUMENTS.

201.5.4.104 * Arrangement in space

The INFANT PHOTOTHERAPY EQUIPMENT shall be oriented as specified by the MANUFACTURER in the instructions for use (see 201.7.9.2.9).

201.6 Classification of ME EQUIPMENT and ME SYSTEMS

Clause 6 of the general standard applies, except as follows:

201.6.3 Protection against harmful ingress of water or particular particulate matter

Addition:

201.6.3.101 INFANT PHOTOTHERAPY EQUIPMENT located under the PATIENT

If INFANT PHOTOTHERAPY EQUIPMENT is located under the PATIENT, it shall at least comply with IPX3 specified in IEC 60529.

201.7 ME EQUIPMENT identification, marking and documents

Clause 7 of the general standard applies, except as follows:

201.7.2 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT PARTS (see also Table C.1 of the general standard)

Additional subclause:

201.7.2.101 * Safety sign for PATIENT eye shield

A safety sign indicating the requirement for requiring eye shields for the PATIENT shall be used if the PATIENT'S eyes can be exposed to the INFANT PHOTOTHERAPY EQUIPMENT'S radiation. See symbol number Safety 02 ISO 7010-M025 in IEC TR 60878:2015.

201.7.3.1 Heating elements or lamp holders

Addition:

The types of lamps specified or recommended by the MANUFACTURER shall be indicated.

201.7.9.2.2 Warning and safety notices

/ Addition:s, iteh.ai/catalog/standards/jec/995fdb90-ac99-4463-b0ce-c2be4b0cd7b0/jec-60601-2-50-2020

The instructions for use shall also include the following:

- a) a statement that the INFANT PHOTOTHERAPY EQUIPMENT should be used only by appropriately trained personnel and under the direction of qualified medical personnel familiar with currently known RISKS and benefits of INFANT PHOTOTHERAPY EQUIPMENT use;
- b) a statement by the MANUFACTURER explaining the effect of varying ambient conditions on the PATIENT, for example varying ambient temperatures, different radiation sources (sunlight), etc.;
- c) if necessary, a notice giving information about the filter and the protective barrier required for NORMAL USE;
- d) a notice that some PATIENTS' water balance may be disturbed;
- e) a notice that PATIENTS adjacent to the INFANT PHOTOTHERAPY EQUIPMENT may need to be protected, and a notice and details about additional protective measures (e.g. shields, protective glasses);
- f) a notice that the PATIENT'S bilirubin values shall be measured regularly;
- g) a notice that the use of reflective foils may cause hazardous body temperatures, if relevant to the type of INFANT PHOTOTHERAPY EQUIPMENT;
- h) advice to supply the PATIENT with an eye shield, whenever the PATIENT'S eye can be exposed to the INFANT PHOTOTHERAPY EQUIPMENT'S radiation;
- *i) the warning notice that the OPERATOR may experience some effects during prolonged exposure to the area irradiated by the INFANT PHOTOTHERAPY EQUIPMENT;