

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Electrical equipment for measurement, control and laboratory use –  
EMC requirements –  
Part 2-3: Particular requirements – Test configuration, operational conditions  
and performance criteria for transducers with integrated or remote signal  
conditioning**

<https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020>

**Matériel électrique de mesure, de commande et de laboratoire –  
Exigences relatives à la CEM –  
Partie 2-3: Exigences particulières – Configurations d'essai, conditions de  
fonctionnement et critères de performance des transducteurs avec un système  
de conditionnement du signal intégré ou à distance**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22,000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67,000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Electrical equipment for measurement, control and laboratory use –  
EMC requirements –  
Part 2-3: Particular requirements – Test configuration, operational conditions  
and performance criteria for transducers with integrated or remote signal  
conditioning** <https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020>

**Matériel électrique de mesure, de commande et de laboratoire –  
Exigences relatives à la CEM –  
Partie 2-3: Exigences particulières – Configurations d’essai, conditions de  
fonctionnement et critères de performance des transducteurs avec un système  
de conditionnement du signal intégré ou à distance**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 17.220.20; 25.040.40; 33.100.20

ISBN 978-2-8322-8980-8

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD .....	4
1 Scope .....	6
2 Normative references .....	8
3 Terms and definitions .....	8
4 General .....	9
5 EMC test plan .....	9
5.1 General .....	9
5.2 Configuration of EUT during testing .....	9
5.3 Operation conditions of EUT during testing .....	10
5.4 Specification of FUNCTIONAL PERFORMANCE .....	10
5.5 Test description .....	10
6 Immunity requirements .....	10
6.1 Conditions during the tests .....	10
6.2 Immunity test requirements .....	11
6.3 Random aspects .....	11
6.4 Performance criteria .....	11
7 Emission requirements .....	12
7.1 Conditions during measurements .....	12
7.2 Emission limits .....	12
8 Test results and test report .....	12
9 Instructions for use .....	12
Annex A (normative) Immunity test requirements for PORTABLE TEST AND MEASUREMENT EQUIPMENT powered by battery or from the circuit being measured .....	13
Annex AA (normative) Additional requirements and exceptions for specific types of transducers – Transducers for measurement of tension and compressive forces (force transducers) .....	14
AA.1 General considerations .....	14
AA.2 Test configuration .....	14
AA.3 Operation conditions .....	15
Annex BB (normative) Additional requirements and exceptions for specific types of transducers – Transducers for measurement of pressure (pressure transducers) .....	17
BB.1 General considerations .....	17
BB.2 Test configuration .....	17
BB.3 Operation conditions .....	18
Annex CC (normative) Additional requirements and exceptions for specific types of transducers – Transducers for measurement of temperature (temperature transducer) .....	19
CC.1 General considerations .....	19
CC.2 Test configuration .....	19
CC.3 Operation conditions .....	20
Bibliography .....	22
Figure 101 – Example of a TRANSDUCER WITH INTEGRATED SIGNAL CONDITIONING .....	7
Figure 102 – Example of a TRANSDUCER WITH REMOTE SIGNAL CONDITIONING .....	8
Figure AA.1 – Example of the configuration of a force transducer with remote signal conditioning .....	15

Figure BB.1 – Example of the configuration of a pressure transducer .....	18
Figure CC.1 – Example of the configuration of a temperature transducer with sensor and signal conditioning in the same housing .....	20
Figure CC.2 – Example of the configuration of a temperature transducer with remote signal conditioning .....	20
Table 101 – Performance criteria for the different functions .....	12
Table AA.1 – Circuitry actions for generating an output signal for simulation of a mechanical load on the transducer.....	16

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[IEC 61326-2-3:2020](https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020)

<https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020>

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ELECTRICAL EQUIPMENT FOR MEASUREMENT,  
CONTROL AND LABORATORY USE –  
EMC REQUIREMENTS –****Part 2-3: Particular requirements –  
Test configuration, operational conditions and performance  
criteria for transducers with integrated or remote signal conditioning**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This International Standard IEC 61326-2-3 has been prepared by subcommittee 65A: System aspects, of IEC technical committee 65: Industrial-process measurement, control and automation.

This third edition cancels and replaces the second edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- update of the document with respect to IEC 61326-1:2020.

The text of this International Standard is based on the following documents:

FDIS	Report on voting
65A/980/FDIS	65A/991/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document the following print types are used:

- Terms used throughout this document which have been defined in Clause 3 of this document and of IEC 61326-1:2020: SMALL CAPITALS.

This part of the IEC 61326 series is to be used in conjunction with IEC 61326-1:2020 and follows the same numbering of clauses, subclauses, tables and figures.

When a particular subclause of IEC 61326-1 is not mentioned in this part, that subclause applies as far as is reasonable. When this standard states “addition”, “modification” or “replacement”, the relevant text in IEC 61326-1 is to be adapted accordingly.

NOTE The following numbering system is used:

- subclauses, tables and figures that are numbered starting from 101 are additional to those in IEC 61326-1;
- unless notes are in a new subclause or involve notes in IEC 61326-1, they are numbered starting from 101 including those in a replaced clause or subclause;
- additional annexes are lettered AA, BB, etc.

A list of all parts of the IEC 61326 series, under the general title *Electrical equipment for measurement, control and laboratory use – EMC requirements*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

# ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL AND LABORATORY USE – EMC REQUIREMENTS –

## Part 2-3: Particular requirements – Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning

### 1 Scope

In addition to the requirements of IEC 61326-1, this part of IEC 61326 specifies more detailed test configurations, operational conditions and performance criteria for transducers with integrated or remote signal conditioning.

This document applies only to transducers characterized by their ability to transform, with the aid of an auxiliary energy source, a non-electric quantity to a process-relevant electrical signal, and to output the signal at one or more PORTS. This document includes transducers for electro-chemical and biological measured quantities.

The transducers covered by this document can be powered by AC or DC voltage and/or by battery or with internal power supply.

Transducers referred to by this document comprise at least the following items (see Figure 101 and Figure 102):

- one or more elements for transforming a non-electrical input quantity to an electrical quantity;
- a TRANSMISSION LINK for transferral of the electrical quantity to a component for signal conditioning;
- a unit for signal conditioning that converts the electrical quantity to a process-relevant electrical signal;
- an enclosure for enclosing the above-stated components fully or in parts.


Transducers referred to by this document can also have the following items (see Figure 101 and Figure 102):

- a communication and control unit;
- a display unit;
- control elements such as keys, buttons, switches, etc.;
- transducer output signals (for example, switch outputs, alarm outputs) which are clearly assigned to the input signal(s);
- transducers with signal conditioning which may be integrated or remote.

The manufacturer specifies the environment for which the product is intended to be used and utilizes the corresponding test levels of IEC 61326-1.

Additional requirements and exceptions for specific types of transducers are given in Annex AA, Annex BB and Annex CC to this document.





IEC


**Key**

- 1 non-electrical quantity
- 2 electrical quantity
- 3 TRANSMISSION LINK
- 4 signal conditioning
- 5 communication and control unit
- 6 input/output PORTS
- 7 power supply
- 8 signal PORT
- 9 AC/DC POWER PORT
- 10 enclosure

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

<http://standards.iteh.ai/catalog/standards/siv/694d3310-1347-4f7-8c66-5687ff898845/iec-61326-2-3-2020>

**Figure 101 – Example of a TRANSDUCER WITH INTEGRATED SIGNAL CONDITIONING**



IEC

**Key**

- 1 non-electrical quantity
- 2 electrical quantity
- 3 TRANSMISSION LINK
- 4 signal conditioning
- 5 communication and control unit
- 6 input/output PORTS
- 7 power supply
- 8 signal PORT
- 9 AC/DC POWER PORT
- 10 enclosure

**iTeh STANDARD PREVIEW**  
(standards.iteh.ai)

[IEC 61326-2-3:2020](#)

**Figure 102 – Example of a TRANSDUCER WITH REMOTE SIGNAL CONDITIONING**

**2 Normative references**

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

Clause 2 of IEC 61326-1:2020 applies, except as follows:

*Addition:*

IEC 61326-1:2020, *Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements*

**3 Terms and definitions**

For the purposes of this document, the terms and definitions of IEC 61326-1:2020 apply, except as follows.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

*Addition:*

### **3.101**

#### **transducer with integrated signal conditioning**

transducer in which all components for signal conditioning are integrated in the enclosure

Note 1 to entry: See Figure 101.

### **3.102**

#### **transducer with remote signal conditioning**

transducer whose components for signal conditioning are installed in separate enclosures

Note 1 to entry: See Figure 102.

### **3.104**

#### **transmission link**

connection between the individual components of a TRANSDUCER WITH REMOTE SIGNAL CONDITIONING

### **3.105**

#### **range**

range of indications obtainable with a particular setting of the controls of a measuring instrument

Note 1 to entry: The (nominal) RANGE is normally stated in terms of its lower and upper limits. Where the lower limit is zero, the (nominal) RANGE is commonly stated solely in terms of its upper limit.

[SOURCE: IEC 60050-300:2001, 311-03-14]

### **3.106**

#### **measuring range (of a transducer)** [IEC 61326-2-3:2020](#)

range defined by two values of the measured quantity within which the relationship between the output and input signals complies with the accuracy requirements

Note 1 to entry: For a 4 mA to 20 mA system, the output current 4 mA represents the lower limit for the measured quantity and 20 mA represent the upper limit.

[SOURCE: IEC 60050-300:2001, 314-04-04, modified – "output signal" has been changed to "measured quantity" and the note to entry has been added]

### **3.107**

#### **span**

algebraic difference between the values of the upper and lower limits of the MEASURING RANGE

[SOURCE: IEC 60050-300:2001, 311-03-13]

## **4 General**

Clause 4 of IEC 61326-1:2020 applies.

## **5 EMC test plan**

### **5.1 General**

Subclause 5.1 of IEC 61326-1:2020 applies.

### **5.2 Configuration of EUT during testing**

Subclause 5.2 of IEC 61326-1:2020 applies, except as follows.

### 5.2.1 General

Subclause 5.2.1 of IEC 61326-1:2020 applies, except as follows:

*Addition:*

A system for monitoring the behaviour of the EUT and for registering the output values shall be designed in such a way that the electromagnetic compatibility characteristics of the EUT are not impaired. The monitoring system shall also be designed such that its response is not affected by the immunity tests. The input impedance of the monitoring system shall correspond to the terminating impedance of the transducer, specified by the manufacturer. The distance between the monitoring system and the EUT should be at least 1,5 m.

The measurement uncertainty and the bandwidth of the monitoring system shall be adapted to the characteristics of the transducer.

TRANSMISSION LINKS are considered as separate input and output lines.

The tests shall be conducted in compliance with the environmental conditions for the transducer specified by the manufacturer and using the specified supply voltage.

In the case of battery-operated transducers that can also be used when connected with a power supply, both operating modes (stand-alone and externally supplied) shall be tested.

In cases in which the manufacturer's installation instructions stipulate the use of external protective equipment or particular protective measures that are explicitly stated in the operating manual, the test requirements given in this part of the document shall be applied for use together with the external protective equipment or measures.

### 5.3 Operation conditions of EUT during testing

Subclause 5.3 of IEC 61326-1:2020 applies.

### 5.4 Specification of FUNCTIONAL PERFORMANCE

Subclause 5.4 of IEC 61326-1:2020 applies.

### 5.5 Test description

Subclause 5.5 of IEC 61326-1:2020 applies.

## 6 Immunity requirements

### 6.1 Conditions during the tests

Subclause 6.1 of IEC 61326-1:2020 applies except as follows:

*Addition:*

Transducers shall be operated during the test with all lines connected, provided the PORTS do not have functions that disable the transducer's normal operation.

Configurations with alternative PORTS shall be tested separately.

Transducers shall be set to the most sensitive ranges or combination of ranges unless other ranges are known to provide worst-case immunity results within normal application.

Only operational functions compliant with the specified use under the nominal conditions are permitted. Defined functions that cannot be set under electromagnetic compatibility test conditions shall be simulated by appropriate measures. This shall be done in such a way that the electromagnetic compatibility behaviour of the transducer is not affected.

Measurement and supply circuits shall be grounded in accordance with the manufacturer's specifications. If no such specifications are given, the tests shall be carried out with the circuits grounded and with the circuits ungrounded.

## 6.2 Immunity test requirements

Subclause 6.2 of IEC 61326-1:2020 applies except as follows:

*Addition:*

After or during each test, the function of the transducer shall be tested.

Power inputs for voltages up to 75 V DC or voltages up to 50 V AC that are fed in a single cable together with the input and output PORTS are tested as input and output PORTS.

Power inputs for voltages up to 75 V DC or voltages up to 50 V AC with superimposed output signals (for example, 4 mA to 20 mA current loop with two-wire technology) are also tested as input/output PORTS.

The TRANSMISSION LINK of a TRANSDUCER WITH REMOTE SIGNAL CONDITIONING is tested as an input/output PORT.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

If insulation resistance is specified, it shall be verified after ESD, fast transient and surge immunity tests.

<https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020>

If the insulation's specifications are not satisfied, the transducer is deemed to have failed the EMC tests.

## 6.3 Random aspects

Subclause 6.3 of IEC 61326-1:2020 applies.

## 6.4 Performance criteria

Subclause 6.4 of IEC 61326-1:2020 applies except as follows:

*Addition:*

The performance criteria are used to assess the defined functions of a transducer under the effects of external electromagnetic disturbances. Since a transducer is often part of a chain of functions in a large process, effects on the overall process due to malfunctions of a transducer caused by external interference factors cannot be predicted without great difficulty. For this reason, it is particularly important that the behaviour of transducers under the influence of electromagnetic disturbances is described with performance criteria.

Table 101 describes the permissible effects of a disturbance on the different functions of a transducer with regard to the required performance criteria.

**Table 101 – Performance criteria for the different functions**

Function	Additional particular performance criteria		
	for performance criterion A	for performance criterion B	for performance criterion C
<b>Main function</b> <sup>a</sup>	See 6.4.2 of IEC 61326-1:2020	See 6.4.3 of IEC 61326-1:2020	The deviations during the test may be outside the limit values specified and documented. After the test, the measured values are within the specified range.  The time needed for the device to return to ready state is given in the documentation.
<b>Process communication</b>	Communication works as intended	Temporary interference of the communication is permitted during the test	Interference of the communication is permitted during the test.  The time needed for the device to return to ready state is given in the documentation.
<b>Alarm function</b>	No malfunctions are permitted	Temporary malfunctions are permitted during the test	Malfunctions are permitted.  The time needed for the device to return to ready state is given in the documentation.
<sup>a</sup> The main function of a measuring transducer is to transform a non-electrical quantity into a process-relevant signal as shown in Figure 101 and Figure 102.			

iTech STANDARD PREVIEW  
(standards.itech.ai)

## 7 Emission requirements

[IEC 61326-2-3:2020](https://standards.itech.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020)

<https://standards.itech.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020>

### 7.1 Conditions during measurements

Subclause 7.1 of IEC 61326-1:2020 applies except as follows:

*Addition:*

The additions made in Clauses 5 and 6 shall be taken into account.

### 7.2 Emission limits

Subclause 7.2 of IEC 61326-1:2020 applies.

## 8 Test results and test report

Clause 8 of IEC 61326-1:2020 applies.

## 9 Instructions for use

Clause 9 of IEC 61326-1:2020 applies.

**Annex A**  
(normative)

**Immunity test requirements for PORTABLE TEST AND MEASUREMENT  
EQUIPMENT powered by battery or from the circuit being measured**

Annex A of IEC 61326-1:2020 does not apply.

*Additional annexes:*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[IEC 61326-2-3:2020](https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020)

<https://standards.iteh.ai/catalog/standards/sist/694d33d0-b347-4fc7-8e66-5687ff898845/iec-61326-2-3-2020>