

TECHNICAL REPORT

Household and similar electrical appliances – Method of measuring
performance – Assessment of repeatability, reproducibility and uncertainty
(standards.iteh.ai)

[IEC TR 63250:2021](https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021)

[https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-
b0c307cb8091/iec-tr-63250-2021](https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021)



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2021 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC online collection - oc.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 18 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

[IEC TR 63250:2021](#)

<https://standards.iteh.ai/catalog/standards/sis/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021>

TECHNICAL REPORT

Household and similar electrical appliances – Method of measuring performance – Assessment of repeatability, reproducibility and uncertainty

[IEC TR 63250:2021](https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021)

<https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 97.030

ISBN 978-2-8322-9939-5

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms and definitions	7
4 Determination of standard deviations.....	10
4.1 General.....	10
4.2 Repeatability standard deviation	10
4.3 Reproducibility standard deviation	10
5 Assessment of repeatability, reproducibility, and uncertainty of a measurement method.....	11
5.1 Purpose.....	11
5.2 Requirements	11
5.3 Expression of repeatability and reproducibility	12
5.4.1 The importance of the uncertainty.....	12
5.4.2 Methods to estimate uncertainty	12
5.4.3 Expanded uncertainty calculation	13
6 Scrutiny of results for consistency and outliers.....	14
6.1 Purpose.....	14
6.2 Graphical consistency technique (Mandel's h and k statistics).....	14
6.2.1 Inter-laboratory consistency statistic h	14
6.2.2 Intra-laboratory consistency statistic k	14
6.2.3 Evaluation.....	14
6.3 Numerical outlier technique.....	15
6.3.1 Cochran's C test.....	15
6.3.2 Grubbs' test.....	15
6.3.3 Evaluation	15
7 Data to be reported for assessing the repeatability, reproducibility and uncertainty of a test method.....	16
Annex A (informative) Example of bottom-up analysis.....	17
A.1 General.....	17
A.2 Temperature measurement system	17
A.2.1 General	17
A.2.2 Calibration of thermocouples	17
A.2.3 Calibration of the DAQ system.....	17
A.3 Uncertainty temperature measurement.....	17
A.4 Analysis of each component in the uncertainty formulation, example thermocouple simulator.....	19
Annex B (informative) Guidance on how to conduct round robin tests for household and similar electrical appliances	21
B.1 General.....	21
B.2 Scope	21
B.3 Process and responsibilities.....	22
B.3.1 Process	22
B.3.2 Responsibilities	23
B.4 Testing laboratories	23

B.4.1	Potential laboratories.....	23
B.4.2	Announcement.....	23
B.4.3	Selection of laboratories	24
B.4.4	Final list of laboratories	24
B.5	Transportation of the product.....	24
B.5.1	Logistics	24
B.5.2	Packaging.....	24
B.6	Test.....	24
B.6.1	Performance of test	24
B.6.2	Laboratory visit.....	25
B.6.3	Transmission of result	25
B.7	Analysis, report and termination.....	25
B.7.1	Analysis.....	25
B.7.2	Report	25
B.7.3	Termination and publication of final external report.....	26
Annex C (informative)	Example of a round robin test and its analysis.....	27
C.1	General.....	27
C.2	Standard deviations and assessment of repeatability and reproducibility.....	27
C.3	Scrutiny of results for consistency and outliers.....	29
C.3.1	Example of Mandel's h and k statistics	29
C.3.2	Example of numerical outlier test.....	29
Annex D (informative)	Example of expression of results.....	32
Bibliography	33
	https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021	
Figure 1	– Visualisation of straggler and outlier of Cochran value	16
Figure C.1	– Mandel's h statistics.....	30
Figure C.2	– Mandel's k statistics.....	31
Table A.1	– Description of uncertainty parameters	18
Table A.2	– Measuring a temperature of -23 °C at a climate room temperature between 16 °C to 32 °C	19
Table A.3	– Expanded uncertainty in measured temperature ($U(T)$, $k = 2$) at a room temperature in the range of 16 °C to 32 °C and a lab temperature between 14 °C and 32 °C	19
Table A.4	– Calibration results of the simulator	20
Table C.1	– Measurement results	27
Table C.2	– Standard deviations, repeatability and reproducibility.....	28
Table C.3	– Mandel's h and k statistics	29
Table C.4	– Cochran's test C and Grubbs' test G	29
Table C.5	– Cochran's test C and Grubbs' test G summary	30

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**HOUSEHOLD AND SIMILAR ELECTRICAL APPLIANCES –
METHOD OF MEASURING PERFORMANCE – ASSESSMENT OF
REPEATABILITY, REPRODUCIBILITY AND UNCERTAINTY**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 63250 has been prepared by IEC technical committee 59: Performance of household and similar electrical appliances. It is a Technical Report.

The text of this Technical Report is based on the following documents

Draft	Report on voting
59/752/DTR	59/765/RVDTR

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this Technical Report is English.

Words **in bold** in the text are defined in Clause 3.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

iTeh STANDARD PREVIEW (standards.iteh.ai)

[IEC TR 63250:2021](#)

<https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021>

INTRODUCTION

To encourage the efficient use of energy and other resources, national governments and regional authorities have issued regulations that mandate the provision of information to consumers regarding the energy and water consumption of household appliances and associated performance characteristics.

Therefore, methods for measuring performance characteristics must be of sufficient **accuracy** to provide confidence to governments, consumers and manufacturers.

The **accuracy** of a test method is expressed in terms of **bias** and **precision**. **Precision**, when evaluating test methods, is expressed in terms of two measurement concepts: **repeatability** (intra-laboratory variability) and **reproducibility** (inter-laboratory variability). Therefore, standard procedures are required for determining the **repeatability** and the **reproducibility** of test methods. The determination of levels of **repeatability** and **reproducibility** is frequently done by carrying out **round robin tests** (RRT). The **repeatability** of a test method must be sufficiently accurate for comparative testing. The **reproducibility** of a test method must be sufficiently accurate for the determination of values that are declared, and for checking these declared values. Other ways to assess the uncertainty are possible.

Uncertainty reporting is essential to ensure measured data are interpreted correctly. Especially when data of measurements are to be compared between laboratories or when normative requirements are set up, it is necessary to know the uncertainty with which data can be measured.

In conformity assessment using a binary decision rule, a property of an item is measured, and the item is accepted as conforming if the measured value of the property lies within a defined acceptance interval. A measured value outside the acceptance interval leads to rejection of the item as non-conforming.

The objective of this technical report is to give guidelines for household and similar electrical appliances within TC 59, but it can also be used for assessing other types of appliances outside the technical committee 59 and its subcommittees' environment.

It is intended to collate and summarise the information needed for assessing the **repeatability**, **reproducibility** and uncertainty of measurements of performance of household and similar electrical appliances present in previous IEC publications¹.

¹ IEC TR 61923, IEC TR 62617 and IEC TR 62970

HOUSEHOLD AND SIMILAR ELECTRICAL APPLIANCES – METHOD OF MEASURING PERFORMANCE – ASSESSMENT OF REPEATABILITY, REPRODUCIBILITY AND UNCERTAINTY

1 Scope

This Technical Report deals with the determination of **repeatability** and **reproducibility** of test methods used for assessing the performance characteristics of household and similar electrical appliances. It also provides guidance for carrying out **round robin tests** (RRT).

It also specifies the uncertainty reporting of measurements of household and similar electrical appliances.

It describes methods to estimate the uncertainty of a measured result and to predict the range of measured values when the same appliance is measured in another laboratory applying the same measurement method.

It does not cover the development of measurement methods. It also does not deal with:

- the production variability of the appliance;
- how closely the measurement method reflects the normal use of appliances in households.

NOTE 1 Although this technical report does not cover the development of test methods, it can be taken into consideration for this purpose.

NOTE 2 For the purpose of this technical report, production variability includes the variation of the individual appliances of the same type and model manufactured on the same production line.

NOTE 3 For noise standardisation, some deviating definitions are used (see. IEC 60704-3:2019).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5725-2:2019, *Accuracy (trueness and precision) of measurement methods and results – Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

ISO 80000-1:2009, *Quantities and units – Part 1: General*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

accuracy

closeness of agreement between a test result or measurement result and the true value

Note 1 to entry: In practice, the accepted reference value is substituted for the true value.

Note 2 to entry: The term "**accuracy**", when applied to a set of test or measurement results, involves a combination of random components and a common systematic error or **bias** component.

Note 3 to entry: **Accuracy** refers to a combination of **trueness** and **precision**.

[SOURCE: ISO 3534-2:2006, 3.3.1, modified – Cross-references have been deleted]

3.2

precision

closeness of agreement between independent test/measurement results obtained under stipulated conditions

Note 1 to entry: **Precision** depends only on the distribution of random errors and does not relate to the true value or the specified value.

Note 2 to entry: The measure of **precision** is usually expressed in terms of imprecision and computed as a standard deviation of the test results or measurement. Less **precision** is reflected by a larger standard deviation.

Note 3 to entry: Quantitative measures of **precision** depend critically on the stipulated conditions. **Repeatability conditions** and **reproducibility conditions** are particular sets of extreme stipulated conditions.

[SOURCE: ISO 3534-2:2006, 3.3.4, modified – Cross-references have been deleted]

3.3

repeatability

precision under **repeatability conditions**.

Note 1 to entry: **Repeatability** can be expressed quantitatively in terms of the dispersion characteristics of the results.

[SOURCE: ISO 3534-2:2006, 3.3.5, modified – Cross-references have been deleted]

3.4

repeatability conditions

observation conditions where independent test/measurement results are obtained with the same method on identical test/measurement items in the same test or measuring facility by the same operator using the same equipment within short intervals of time

Note 1 to entry: **Repeatability** conditions include:

- the same measurement procedure or test procedure;
- the same operator;
- the same measuring or test equipment used under
- the same conditions;
- the same location;
- repetition over a short period of time.

[SOURCE: ISO 3534-2:2006, 3.3.6, modified – Cross-references have been deleted]

3.5

repeatability standard deviation

standard deviation of test results or measurement results obtained under **repeatability conditions**

Note 1 to entry: It is a measure of the dispersion of the distribution of test results or measurement results under **repeatability conditions**.

Note 2 to entry: Similarly, "repeatability variance" and "repeatability coefficient of variation" can be defined and used as measures of the dispersion of test or measurement results under **repeatability conditions**.

[SOURCE: ISO 3534-2:2006, 3.3.7, modified – Cross-references have been deleted]

3.6

reproducibility

precision under **reproducibility conditions**

Note 1 to entry: **Reproducibility** can be expressed quantitatively in terms of the dispersion characteristics of the results.

Note 2 to entry: Results are usually understood to be corrected results

[SOURCE: ISO 3534-2:2006, 3.3.10, modified – Cross-references have been deleted]

3.7 reproducibility conditions

observation conditions where independent test/measurement results are obtained with the same method on identical test/measurement items in different test or measurement facilities with different operators using different equipment

[SOURCE: ISO 3534-2:2006, 3.3.11, modified – Cross-references have been deleted]

3.8 reproducibility standard deviation:

standard deviation of test results or measurement results obtained under **reproducibility conditions**

Note 1 to entry: It is a measure of the dispersion of the distribution of test or measurement results under **reproducibility conditions**.

Note 2 to entry: Similarly, "reproducibility variance" and "reproducibility coefficient of variation" can be defined and used as measures of the dispersion of test or measurement results under **reproducibility conditions**.

[SOURCE: ISO 3534-2:2006, 3.3.12 modified – Cross-references have been deleted]

3.9 outlier

member of a small subset of observations that appears to be inconsistent with the remainder of a given sample

Note 1 to entry: The classification of an observation or a subset of observations as **outlier(s)** is relative to the chosen model for the population from which the data set originates. This or these observations are not to be considered as genuine members of the main population.

Note 2 to entry: An **outlier** may originate from a different underlying population, or be the result of incorrect recording or gross measurement error.

Note 3 to entry: The subset may contain one or more observations.

[SOURCE: ISO 16269-4:2010, 2.2 modified – Cross-references have been deleted]

3.10 statistical uncertainty

repeatability standard deviation obtained in one laboratory under **repeatability conditions**

3.11 expanded uncertainty

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand

Note 1 to entry: The fraction may be viewed as the coverage probability or level of confidence of the interval.

Note 2 to entry: To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterized by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.

Note 3 to entry: **Expanded uncertainty** is termed *overall uncertainty* in paragraph 5 of Recommendation INC-1 (1980)

[SOURCE: ISO/IEC Guide 98-3:2008, 2.3.5]

3.12 bias

difference between the expectation of a test result or measurement result and a true value

[SOURCE: ISO 3534-2:2006, 3.3.2, modified – Cross-references and Notes have been deleted]

3.13 round robin testing

RRT

ringtest

process in which one or more items are tested according to a specific protocol by a number of different laboratories.

4 Determination of standard deviations

4.1 General

Repeatability and **reproducibility standard deviations** serve as parameters for assessing:

- the suitability of a measurement method;
- the **accuracy** of a measurement result;
- conformance of measured values to acceptance interval.

Rounding is only applied to reported values in Clause 7. If numbers are rounded, they are rounded to the nearest number in accordance with ISO 80000-1:2009, Annex B, Rule B. If the rounding takes place to the right of the comma, the omitted places are not filled with zeros.

4.2 Repeatability standard deviation

The **repeatability standard deviation** $s_{L,i}$ of a measurement method within laboratory i is calculated from Equation (1):

iTeh STANDARD PREVIEW
(standards.iteh.ai)

$$s_{L,i} = \sqrt{\frac{1}{n_i} \sum_{k_i=1}^{n_i} (x_{ki} - \bar{x}_i)^2} \tag{1}$$

[IEC TR 63250:2021](https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021)

<https://standards.iteh.ai/catalog/standards/sist/5d643546-6360-403e-a620-b0c307cb8091/iec-tr-63250-2021>

where

n_i is the number of measurement results;

x_{ki} is the particular measurement result;

\bar{x}_i is the arithmetic mean value of n_i measurement results x_k of laboratory i .

The average **repeatability standard deviation** s_r of a measurement method within p laboratories is calculated from Equation (2):

$$s_r = \sqrt{\frac{1}{p} \sum_{i=1}^p s_{L,i}^2} \tag{2}$$

where

$s_{L,i}$ is the **repeatability standard deviation** within laboratory i ;

p is the number of laboratories participating in the inter-laboratory test.

4.3 Reproducibility standard deviation

The **reproducibility standard deviation** s_R of a measurement method is calculated from Equations (3), (4), and (5):