

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Safety requirements for electrical equipment for measurement, control, and laboratory use –**

**Part 2-040: Particular requirements for sterilizers and washer-disinfectors used to treat medical materials**

[IEC 61010-2-040:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/26400dbf-5541-430e-95a5->

**Exigences de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire –**

**Partie 2-040: Exigences particulières pour stérilisateurs et laveurs désinfecteurs utilisés pour traiter le matériel médical**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22.000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67.000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Safety requirements for electrical equipment for measurement, control, and laboratory use –  
Part 2-040: Particular requirements for sterilizers and washer-disinfectors used to treat medical materials**

[IEC 61010-2-040:2020](#)

<https://standards.iteh.ai/catalog/standards/sist/26400dbf-5541-430e-95a5-18205a2020>

**Exigences de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire –**

**Partie 2-040: Exigences particulières pour stérilisateurs et laveurs désinfecteurs utilisés pour traiter le matériel médical**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 19.080; 71.040.10

ISBN 978-2-8322-8301-1

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD .....	3
1 Scope and object .....	5
2 Normative references .....	6
3 Terms and definitions .....	6
4 Tests .....	7
5 Marking and documentation .....	9
6 Protection against electric shock .....	14
7 Protection against mechanical HAZARDS and against HAZARDS related to mechanical functions .....	15
8 Resistance to mechanical stresses .....	19
9 Protection against the spread of fire .....	19
10 Equipment temperature limits and resistance to heat .....	20
11 Protection against HAZARDS from fluids and solid foreign objects .....	21
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure .....	24
13 Protection against liberated gases, substances, explosion and implosion .....	25
14 Components and subassemblies .....	30
15 Protection by interlocks .....	32
16 HAZARDS resulting from application .....	32
17 RISK assessment .....	32
Annexes .....	33
Annex G (informative) Leakage and rupture from fluids under pressure .....	33
Annex L (informative) Index of defined terms .....	34
Bibliography .....	35

ITIH STANDARD PREVIEW  
(standards.iteh.ai)

[IEC 61010-2-040:2020](https://standards.iteh.ai/catalog/standards/sist/26400dbf-5541-430e-95a5-8d890968b2d7/iec-61010-2-040-2020)

<https://standards.iteh.ai/catalog/standards/sist/26400dbf-5541-430e-95a5-8d890968b2d7/iec-61010-2-040-2020>

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR  
MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 2-040: Particular requirements for STERILIZERS and  
WASHER-DISINFECTORS used to treat medical materials**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-040 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

This third edition cancels and replaces the second edition published in 2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) it is established on the basis of the third edition (2010) of IEC 61010-1 and its Amendment 1 (2016);
- b) added tolerance for stability of a.c. voltage test equipment to 6.8.3.1;

- c) the status of a Group Safety Publication has been removed (this does not change the technical requirements in the document).

The text of this International Standard is based on the following documents:

CDV	Report on voting
66/699/CDV	66/716/RVC

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

The reader's attention is drawn to the fact that Annex G lists all of the "in-some-country" clauses on differing practices of a less permanent nature relating to the subject of this standard.

A list of all parts in the IEC 61010 series, published under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This Part 2-040 is to be used in conjunction with IEC 61010-1. It was established on the basis of the third edition (2010) of IEC 61010-1 and its Amendment 1 (2016), hereinafter referred to as Part 1.

This Part 2-040 supplements or modifies the corresponding clauses in Part 1 so as to convert that publication into the IEC standard: *Particular requirements for STERILIZERS and WASHER-DISINFECTORS used to treat medical materials*.

<https://standards.iteh.ai/catalog/standards/sist/26400dbf-5541-430e-95a5-618286821875/iec-61010-2-040-2020>

Where a particular subclause of Part 1 is not mentioned in this Part 2-040, that subclause applies as far as is reasonable. Where this Part 2-040 states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in Part 1 shall be adapted accordingly.

In this standard:

- 1) the following print types are used:
  - requirements: in roman type;
  - NOTES: in small roman type;
  - conformity and tests: *in italic type*;
  - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS.
- 2) subclauses, figures, and tables which are additional to those in Part 1 are numbered starting from 101; additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "http://webstore.iec.ch" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

## SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

### Part 2-040: Particular requirements for STERILIZERS and WASHER-DISINFECTORS used to treat medical materials

#### 1 Scope and object

This clause of Part 1 is applicable except as follows:

##### 1.1.1 Equipment included in scope

*Replacement:*

*Replace the existing text with the following:*

This part of IEC 61010 specifies safety requirements for electrical equipment intended for sterilization, washing, and disinfection of medical materials in the medical, veterinary, pharmaceutical and laboratory fields, when used under the environmental conditions of 1.4.

Examples of such equipment include the following:

- a) STERILIZERS and disinfectors using steam and/or hot water as the sterilant;
- b) STERILIZERS and disinfectors using toxic gas, toxic aerosol or toxic vapour as the sterilant;
- c) STERILIZERS and disinfectors using hot air or hot inert gas as the sterilant; and
- d) WASHER-DISINFECTORS

##### 1.1.2 Equipment excluded from scope

*Addition:*

*Add the following note to item f):*

NOTE IEC 60601-1:2005, 3.63, defines "medical electrical equipment" as follows (notes to entry are omitted):

Electrical equipment, having an applied part or transferring energy to or from the patient or detecting such energy transfer to or from the patient and which is:

- a) provided with not more than one connection to a particular supply MAINS; and
- b) intended by its manufacturer to be used:
  - 1) in the diagnosis, treatment, or monitoring of a patient; or
  - 2) for compensation or alleviation of disease, injury or disability.

*Addition:*

*Add the following new second paragraph after the lettered list:*

This document does not apply to the following types of equipment:

- aa) equipment for use in hazardous atmospheres (see IEC 60079); however this document does apply to an atmosphere created inside equipment by a flammable sterilizing agent (see 13.2.101 and 13.2.102);
- bb) laboratory equipment for the heating of materials for purposes other than sterilization or disinfection (see IEC 61010-2-010);

- cc) laundry equipment (see IEC 60335-2-4, IEC 60335-2-7, IEC 60335-2-11, and ISO 10472 (all parts)), unless designed for disinfecting medical materials;
- dd) dishwashers (see IEC 60335-2-5 and IEC 60335-2-58).

### 1.2.1 Aspects included in scope

*Replacement:*

*Replace item g) with the following new text:*

- g) liberated gases (including the non-intentional escape of toxic gas), pathogenic substances, explosion and implosion (see Clause 13).

### 1.2.2 Aspects excluded from scope

*Addition:*

*Add the following two new items:*

- aa) special requirements for protection against chemical and high-risk micro-biological HAZARDS associated with the LOAD;
- bb) general requirements for the design of calorifiers, shell boilers and PRESSURE VESSELS.

NOTE National and other regulations or codes apply for the safety of calorifiers, shell boilers and PRESSURE VESSELS (see 14,101).

iTech STANDARD PREVIEW  
(standards.iteh.ai)

## 2 Normative references

This clause of Part 1 is applicable except as follows:

*Addition:*

<https://standards.iteh.ai/catalog/standards/sist/26400dbf-5541-430e-95a5-8d890968b2d7/iec-61010-2-040-2020>

*Add the following new references:*

IEC 61770, *Electric appliances connected to the water mains – Avoidance of backsiphonage and failure of hose-sets*

ISO 3585, *Borosilicate glass 3.3 – Properties*

ISO 4126-1, *Safety devices for protection against excessive pressure – Part 1: Safety valves*

ISO 4126-2, *Safety devices for protection against excessive pressure – Part 2: Bursting disc safety devices*

## 3 Terms and definitions

This clause of Part 1 is applicable except as follows:

### 3.2 Parts and accessories

*Addition:*

*Add the following new terms and definitions:*

#### 3.2.101

**CHAMBER**

part of the equipment which receives the LOAD



**3.2.102****LOAD**

equipment or materials put into a CHAMBER to be processed through an OPERATING CYCLE

**3.2.103****STERILIZER**

equipment designed to achieve sterilization which comprises a series of actions or operations needed to achieve the specified requirements for sterility

**3.2.104****PRESSURE VESSEL**

assembly comprising the CHAMBER, the jacket (if fitted), doors, and all other components in permanent open connection with the CHAMBER

Note 1 to entry: The PRESSURE VESSEL does not include parts from which it can be isolated, such as steam generators, pipework, and fittings.

**3.2.105****OPERATING CYCLE**

complete set of stages of the process that is carried out in a specified sequence

Note 1 to entry: Loading and unloading are not part of the OPERATING CYCLE.

**3.2.106****WASHER-DISINFECTOR**

equipment intended to clean and disinfect medical devices and other articles used in the context for example of medical, dental, pharmaceutical and veterinary practice

**3.5 Safety terms****3.5.2****HAZARD**

*Addition:*

*Add the following new Note 1 to entry:*

Note 1 to entry: In the context of this document, the term HAZARD relates only to potential sources of harm to the OPERATOR and surroundings (see 1.2.1), and does not include potential sources of harm related to the efficacy of the process.

**3.5.11****OPERATOR**

*Addition:*

*Add the following Note 1 to entry:*

Note 1 to entry: An OPERATOR includes persons installing, operating, adjusting, maintaining, cleaning, repairing or moving equipment.

**4 Tests**

This clause of Part 1 is applicable except as follows:

**4.3.2.4 Covers and removable parts**

*Addition:*

*Add the following new second paragraph:*

Covers including panels and control box ENCLOSURES which do not require the use of a TOOL for removal need not be removed if they have interlocks which meet the requirements of Clause 15, and which automatically de-activate all parts which would otherwise present a HAZARD when the cover is opened.

#### 4.3.2.12 Duty cycle

*Addition:*

*Add the following new second paragraph:*

Equipment which can be operated continuously shall also be tested without any interval between consecutive OPERATING CYCLES.

*Addition:*

*Add the following new subclause:*

#### 4.3.2.101 Non-electrical supplies and services

Non-electrical supplies and services shall be set to the least favourable RATED settings.

#### 4.4.2.1 General

*Replacement:*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

*Replace the first sentence of the first paragraph with the following sentence:*

Fault conditions shall include those specified in 4.4.2.2 to 4.4.2.14 and 4.4.2.101 to 4.4.2.103.

#### 4.4.2.5 Motors

*Addition:*

*Add the following new second paragraph:*

If it is impracticable to test a motor when installed, a separate identical motor can be tested but it shall be tested in the same conditions that exist inside the equipment.

#### 4.4.2.13 Interlocks

*Addition:*

*Add the following new second paragraph:*

If an interlock provides protection against accidental contact with a hazardous substance, the interlock is tested using a non-hazardous substance.

*Addition:*

*Add the following three new subclauses:*

#### 4.4.2.101 Pressure controllers

Pressure controllers, except for overpressure safety devices meeting the requirements of 11.7.4, shall be overridden to supply the service continuously.

#### 4.4.2.102 Failure, or partial failure, of the MAINS supply

The equipment shall be operated at 90 % and 110 % of the RATED voltage for one cycle. The voltage shall then be set to 90 % of the RATED voltage for 5 min. The voltage shall then be reduced gradually at a rate of approximately 10 V per min until the equipment fails to operate normally. The voltage shall then be reset to the RATED voltage with the equipment still switched on.

#### 4.4.2.103 Failure, or partial failure, of other supplies and services

In turn, each non-electrical supply and service shall be interrupted, or partially interrupted, whichever is less favourable.

NOTE Examples include air, steam, feedwater, sterilant gas, detergent, disinfectant, and systems for drainage, exhaust, and ventilation.

## 5 Marking and documentation

This clause of Part 1 is applicable except as follows:

### 5.1.2 Identification

*Replacement:*

Replace the existing text by the following:

The equipment shall as a minimum be marked with the following:

- the name and address of the manufacturer;
- any additional markings required by national and local regulations, including the name and address of the manufacturer's authorized representative in the country of intended use;
- a marking that uniquely identifies the individual unit of manufacture such as a serial number;
- year and place of manufacture, if different from manufacturer's address;
- model identification;
- designated purpose of the equipment.

*Conformity is checked by inspection*

*Addition:*

*Add the following two new subclauses:*

#### 5.1.101 Overpressure protective device

The device (see 11.7.4) shall be marked with the name of the manufacturer of the device, the model number, and the pressure to which it is set. If a bursting disc is located between the CHAMBER and the overpressure safety device, the disc shall be marked with its specified bursting pressure and associated temperature.

NOTE National, local regulations and other codes can apply.

#### 5.1.102 PRESSURE VESSELS and shell boilers

Attention is drawn to the existence of national and local regulations that can require additional markings.

## 5.2 Warning markings

*Addition:*

*Add the following new second paragraph:*

Warning and caution symbols shall be at least 10 mm high.

### 5.4.1 General

*Replacement:*

*Replace the first paragraph (excluding lettered list) by the following new paragraph:*

The following documentation necessary for safety purposes, as needed by the OPERATOR or RESPONSIBLE BODY, shall be marked with its date of issue or revision status and be provided with the equipment:

*Add the following new paragraph after item h) of the lettered list:*

If NORMAL USE involves the handling of a hazardous substance, documentation shall include information on constituents, correct storage, use, and safe disposal.

*Replace Note 2 by the following new Note 2:*

NOTE 2 Attention is drawn to the existence of national and local regulations that can apply to the documentation.

*Add a new paragraph directly before the conformity statement:*

Marking, information, and language shall

- 1) comply with regulations applying in the country of intended use;

NOTE 3 ISO 15223-2 offers guidance for equipment classified as a medical device.

- 2) include instructions for the disposal of the equipment, its accessories and its packaging;
- 3) give due consideration to the technical knowledge, education and training of different OPERATOR categories;
- 4) not contradict information contained in documentation provided to describe the equipment.

### 5.4.2 Equipment RATINGS

*Addition:*

*Add the following new item to the lettered list after item f):*

- aa) for each non-electrical service, if applicable, the RATED ranges of temperature, pressure and flow-rate.

### 5.4.3 Equipment installation

*Replacement:*

*Replace items a) to g) with the following:*

- a) location and mounting instructions;
- b) space required for safe and efficient maintenance;
- c) individual weights of principal heavy subassemblies;

- d) overall weight and floor loading requirements;
  - e) unpacking and assembly instructions (see also 7.108);
  - f) MAINS supply requirements and connections, including the temperature rating of any cable required to meet the requirements in 5.1.8;
  - g) for PERMANENTLY CONNECTED EQUIPMENT:
    - 1) supply wiring requirements;
    - 2) requirements for any external switch or circuit-breaker (see 6.11.3.1) and external overcurrent protection devices (see 9.6.1) and a recommendation that the switch or circuit-breaker be near the equipment;
  - h) ventilation requirements (see 11.101, 13.1.103.1, and 13.1.101);
  - i) drainage requirements (see 11.101);
  - j) instructions for protective earthing;
  - k) instructions relating to sound level (see 12.5.1);
  - l) requirements for special services, for example air, feedwater, cooling liquid;
  - m) requirements related to hazardous gas atmospheres (see Clause 13);
  - n) instructions to position the equipment so that it is not difficult to operate the disconnecting device;
  - o) instructions relating to the handling and containment of hazardous substances, including any need for additional equipment that can be required to control emissions (see 13.1);
  - p) instructions relating to HAZARDS caused by liquids or hot items falling from the equipment (see 9.1);
  - q) requirements for material used in the installation of the equipment and which can come into contact with sterilant (see 13.1.103.4 and 13.2.101);
  - r) instructions for ambient illumination (see also 11.102);
- NOTE ISO 12100 and EN 1837 give guidance on lighting
- s) instructions relating to heat emission.

*Addition:*

*Add the following new subclause:*

#### **5.4.3.101 Special systems**

Installation instructions shall include details of the following special systems, if needed to protect against possible HAZARDS:

- a) non-recirculating ventilation system for the room in which the equipment is installed (also see 13.1.103.3);

Such a ventilation system shall normally give a minimum of 10 air changes per hour, but for large installations this may need to be increased.
- b) for equipment using toxic sterilant, means to protect against HAZARDS arising from failure of the room ventilation system (see 13.1.103.3);
- c) a non-recirculating local exhaust system to remove fugitive emissions (see 13.1.101.4);
- d) a drainage system (see 13.1.101.3);
- e) a venting system for the drain (see 13.1.101.3);
- f) a CHAMBER exhaust system (see 13.1.101.2);
- g) a system used to control escaping biological emissions (see 13.1.104);
- h) any other supply, for example sterilant, steam, compressed air, hot or cold water (including instructions on the prevention of backsiphonage (see 11.104)).

Conformity is checked by inspection.

#### 5.4.4 Equipment operation

*Replacement:*

*Replace items a) to j) with the following:*

- a) identification of operating controls and their use in all operating modes;
- b) an instruction not to position the equipment so that it is difficult to operate the disconnecting device;
- c) instructions for interconnection to accessories and other equipment, including details of suitable accessories, detachable parts and any special materials;
- d) specification of limits for intermittent operation;
- e) an explanation of symbols related to safety which are used on the equipment (see 5.2);
- f) instructions for cleaning (see 11.2);
- g) instructions for making the equipment safe after an incomplete OPERATING CYCLE;
- h) instructions for the correct use of the lockable door closure prevention device (see 7.102 b));
- i) instructions to the RESPONSIBLE BODY for safe access to the LOAD in the CHAMBER in the event of a fault (see 13.1.102);
- j) instructions for action in case of a malfunction, including fault diagnosis;

NOTE 1 These instructions can include any special methods of interpreting data recorded or noted during the OPERATING CYCLE, to detect failure or trends that can lead to failure, for example the use of a temperature recorder.

- k) loading procedure;
- l) instructions for safe disposal of parts such as detergent containers, sterilant containers and parts contaminated by pathogenic material;

NOTE 2 Additional requirements on methods of disposal can be specified by national or local authorities.

- m) instructions for testing the function of critical safety devices in a safe manner, for example overpressure safety devices (see 11.7.4);
- n) if NORMAL USE involves the handling of substances, instructions on correct use and safety provisions. In addition, instructions shall be given on methods of safe handling before disposal, and recommendations on disposal (also see Note 2 above);
- o) details of methods of reducing burn HAZARDS from surfaces permitted to exceed the temperature limits specified in Table 19;
- p) guidelines to be followed in cases of emergency in which eye or skin contact or inhalation could occur, such as release of toxic material or pathogenic material, or leakage from a sterilizing agent container or disinfectant container or enzymatic, alkaline or acidic detergent container;

These guidelines shall also be prominently displayed on or near the equipment

- q) instructions for safely replenishing containers of dosing chemicals (see 13.102);
- r) if a HAZARD could result from the use of equipment with a type of LOAD other than those for which it is intended, there shall be an appropriate warning in the instructions, and a warning marking (see 5.2) shall state the types of LOAD which can be used. If small equipment has insufficient space for this warning marking, symbol 14 of Table 1 shall be marked;
- s) instructions for inspection, replenishment, and storage of consumable materials which could cause a HAZARD, including details of HAZARDS which could arise from the introduction of incorrect quantities of recommended consumable materials, also procedures and details of the protection needed to minimize such HAZARDS;

- t) identification of residual RISKS and instructions on necessary protective procedures (see Clause 17).

#### 5.4.5 Equipment maintenance and service

*Replacement:*

*Replace the existing text with the following new text:*

Instructions shall be provided to the RESPONSIBLE BODY in sufficient detail to permit safe maintenance, inspection and testing of the equipment and to ensure continued safety of the equipment after the maintenance, inspection and test procedure.

Instructions shall include:

- a) details of maintenance required on parts subject to wear and tear if failure could lead to a HAZARD;
- b) inspection and replacement, if necessary, of any hoses/pipes or other parts containing fluids, if their failure could cause a HAZARD;
- c) details of safety devices fitted together with their settings and replacement procedures;
- d) procedures for making the equipment safe prior to maintenance;
- e) maintenance schedules and repair procedures, including ambient lighting level (see 11.102) and any special precautions necessary to protect against HAZARDS during maintenance;
- f) methods of safe handling for repair or disposal of any part containing or contaminated by toxic and/or pathogenic material;

NOTE 1 Requirements on methods of disposal can be specified by national or local authorities.

NOTE 2 Aspects of environmental impact are addressed in ISO 14971 and in applicable parts of IEC 61508.

- g) battery types for equipment using replaceable batteries;
- h) RATINGS and characteristics of replaceable fuses;
- i) a list of parts (if any), restricted to examination and/or supply by the manufacturer or the manufacturer's agent;
- j) residual RISKS (see Clause 17) and protective measures for these RISKS;
- k) verification of the safe state of the equipment after repair.

*Conformity is checked by inspection.*

*Addition:*

*Add the following two new subclauses:*

#### 5.4.101 OPERATOR training

##### 5.4.101.1 General

In order that OPERATORS be adequately trained in the safe use of the equipment, the manufacturer's instructions shall state that the RESPONSIBLE BODY should ensure:

- a) that all personnel who operate or maintain the equipment are trained in its operation and in its safe use;
- b) that, if exposure limits (i.e. short-term exposure limit (STEL) or long-term exposure limit (LTEL)) or permissible working environmental concentration limit (see the notes to 13.1) could be exceeded during NORMAL USE, personnel working with toxic chemicals, gases, and vapours are given comprehensive instructions in the process. These instructions