



Designation: E287 – 02 (Reapproved 2019)

# Standard Specification for Laboratory Glass Graduated Burets<sup>1</sup>

This standard is issued under the fixed designation E287; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers requirements for glass burets of precision and general-purpose grades suitable for laboratory use.

1.1.1 *Class A*—Each buret of precision grade may be marked with the letter “A” to signify compliance with applicable construction and accuracy requirements in [Table 1](#) and [Table 2](#). Each buret may be marked with a permanent serial number at the option of the manufacturer.

1.1.2 *Class B*—General purpose burets are of the same basic design as Class A burets. However, volumetric tolerances for Class B burets shall be within twice the specified range allowed for Class A burets. These burets need not be marked with their class designation or serial number. See [Table 1](#) and [Table 3](#) for tolerances and dimensional requirements.

NOTE 1—The Twelfth General (International) Conference on Weights and Measures redefined the litre as a “special name for the cubic decimetre,” but agreed to permit continuance of the terms litre, millilitre, and mL, except in association with measurements of the highest precision. For volumetric glassware the difference between the old and new meanings of litre is negligible. Therefore, either mL or cm<sup>3</sup> may be marked on ware covered by this standard.

1.1.3 Product with a stated capacity not listed in this standard may be specified Class A Tolerance when product conforms to the tolerance range of the next smaller volumetric standard product listed in [Table 1](#).

1.2 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Laboratory Ware and Supplies.

Current edition approved July 1, 2019. Published July 2019. Originally approved in 1965. Last previous edition in 2012 as E287 – 02(2012). DOI: 10.1520/E0287-02R19.

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>2</sup>

D664 Test Method for Acid Number of Petroleum Products by Potentiometric Titration

E675 Specification for Interchangeable Taper-Ground Stopcocks And Stoppers

E911 Specification for Glass Stopcocks with Polytetrafluoroethylene (PTFE) Plugs

E920 Specification for Commercially Packaged Laboratory Apparatus

E921 Specification for Export Packaged Laboratory Apparatus

E1133 Practice for Performance Testing of Packaged Laboratory Apparatus for United States Government Procurements

E1157 Specification for Sampling and Testing of Reusable Laboratory Glassware

## 3. General Requirements

3.1 *Calibration*—Burets shall be calibrated to deliver the intended volume at 20 °C. No after-drainage period is allowed. Volumetric tolerances shall be within the limits specified in [Table 1](#).

## 4. Design

4.1 *Shape*—This specification covers 10, 25, 50, and 100-mL straight burets with single-bore stopcock; 10, 25, 50, and 100-mL straight burets with 3-way stopcock; 100-mL burets with 50-mL bulb (intermediate bulb style); and 10, 25, 50, and 100-mL self-zeroing burets.

4.1.1 Self-zeroing burets shall have an ungraduated, tapered, interior extension that terminates with an overflow orifice. The exterior of the buret shall extend reasonably beyond the terminal point of the interior overflow orifice. An exterior outlet tube positioned below the overflow orifice shall

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.