



Designation: E1162 – 11 (Reapproved 2019)

Standard Practice for Reporting Sputter Depth Profile Data in Secondary Ion Mass Spectrometry (SIMS)¹

This standard is issued under the fixed designation E1162; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers the information needed to describe and report instrumentation, specimen parameters, experimental conditions, and data reduction procedures. SIMS sputter depth profiles can be obtained using a wide variety of primary beam excitation conditions, mass analysis, data acquisition, and processing techniques **(1-4)**.²

1.2 *Limitations*—This practice is limited to conventional sputter depth profiles in which information is averaged over the analyzed area in the plane of the specimen. Ion microprobe or microscope techniques permitting lateral spatial resolution of secondary ions within the analyzed area, for example, image depth profiling, are excluded.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards*:³

¹ This practice is under the jurisdiction of ASTM Committee E42 on Surface Analysis and is the direct responsibility of Subcommittee E42.06 on SIMS.

Current edition approved Nov. 1, 2019. Published November 2019. Originally approved in 1987. Last previous edition approved in 2011 as E1162 – 11. DOI: 10.1520/E1162-11R19.

² The boldface numbers in parentheses refer to the references at the end of this standard.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

[E673 Terminology Relating to Surface Analysis](#) (Withdrawn 2012)⁴

3. Terminology

3.1 For definitions of terms used in this practice, see Terminology [E673](#).

4. Summary of Practice

4.1 Experimental conditions and variables that affect SIMS sputter depth profiles **(1-4)** and tabulated raw data (where feasible) are reported to facilitate comparisons to other laboratories or specimens, and to results of other analytical techniques.

5. Significance and Use

5.1 This practice is used for reporting the experimental conditions as specified in Section 6 in the “Methods” or “Experimental” sections of other publications (subject to editorial restrictions).

5.2 The report would include specific conditions for each data set, particularly, if any parameters are changed for different sputter depth profile data sets in a publication. For example, footnotes of tables or figure captions would be used to specify differing conditions.

6. Information to Be Reported

6.1 Instrumentation:

6.1.1 If a standard commercial SIMS system is used, specify the manufacturer and instrument model number and type of analyzer, such as, magnetic sector, quadrupole, time-of-flight, and so forth. Specify, the model numbers and manufacturer of any accessory or auxiliary equipment relevant to the depth profiling study (for example, special specimen stage, primary mass filter, primary ion source, electron flood gun, vacuum pumps, data acquisition system, and source of software, etc.).

6.1.2 If a nonstandard commercial SIMS system is used, specify the manufacturer and model numbers of components (for example, primary ion source, mass analyzer, data system, and accessory equipment).

⁴ The last approved version of this historical standard is referenced on www.astm.org.