

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

**Medical electrical equipment –  
Part 2-71: Particular requirements for the basic safety and essential performance  
of functional near-infrared spectroscopy (NIRS) equipment**

**Appareils électromédicaux –  
Partie 2-71: Exigences particulières pour la sécurité de base et les performances  
essentiels des appareils de spectroscopie dans le proche infrarouge (NIRS)  
fonctionnelle**





**THIS PUBLICATION IS COPYRIGHT PROTECTED**  
**Copyright © 2025 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

#### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

[IEC 80601-2-71:2025](https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025)

<https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025>

#### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

#### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)



IEC 80601-2-71

Edition 2.0 2025-01

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Medical electrical equipment –  
Part 2-71: Particular requirements for the basic safety and essential performance  
of functional near-infrared spectroscopy (NIRS) equipment**

**Appareils électromédicaux –  
Partie 2-71: Exigences particulières pour la sécurité de base et les performances  
essentielle des appareils de spectroscopie dans le proche infrarouge (NIRS)  
fonctionnelle**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.55

ISBN 978-2-8327-0096-9

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards.....	8
201.2 Normative references .....	10
201.3 Terms and definitions .....	10
201.4 General requirements.....	14
201.5 General requirements for testing ME EQUIPMENT .....	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	14
201.7 ME EQUIPMENT identification, MARKING and documents .....	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	16
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	16
201.10 Protection against unwanted and excessive radiation HAZARDS .....	16
201.11 Protection against excessive temperatures and other HAZARDS .....	16
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	18
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	29
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	29
201.15 Construction of ME EQUIPMENT .....	29
201.16 ME SYSTEMS .....	30
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	30
201.101 Requirements for the FUNCTIONAL NIRS EQUIPMENT ACCESSORIES .....	30
202 Electromagnetic disturbances – Requirements and tests.....	30
206 Usability.....	31
Annexes .....	32
Annex C (informative) Guide to MARKING and labelling requirements for ME EQUIPMENT and ME SYSTEMS .....	33
Annex AA (informative) Particular guidance and rationale .....	35
ANNEX BB (normative) Evaluating ME EQUIPMENT performance using the FUNCTIONAL NIRS PHANTOM.....	39
ANNEX CC (informative) Skin temperature at the FUNCTIONAL NIRS PROBE.....	48
Annex DD (informative) Reference to the IMDRF essential principles and labelling guidance .....	49
Bibliography.....	51
Index of defined terms .....	55
Figure 201.101 – FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION .....	11
Figure 201.102 – Measurement of the AVERAGE OPTICAL POWER.....	19
Figure 201.103 – Measurement of PEAK WAVELENGTH and FWHM .....	20
Figure 201.104 – Measurement of the signal stability.....	22
Figure 201.105 – Measurement of the RESPONSE TIME .....	23
Figure 201.106 – Rise time and fall time in the RESPONSE TIME.....	24
Figure 201.107 – Measurement of the signal-to-noise ratio of the detected light intensity .....	25

Figure 201.108 – Measurement of signal-to-noise ratio of the PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE.....	27
Figure 201.109 – Measurement of SIGNAL CROSS-TALK.....	28
Figure BB.1 – The FUNCTIONAL NIRS PHANTOM in two states with different detected light intensities .....	42
Figure BB.2 – FUNCTIONAL NIRS PHANTOM measurement using the reference system.....	43
Figure BB.3 – FUNCTIONAL NIRS PHANTOM measurement using the ME EQUIPMENT to be evaluated.....	43
Figure BB.4 – Schematic for measurement of OPTICAL LOSS .....	47
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements.....	14
Table 201.102 – Performance tests employing the FUNCTIONAL NIRS EQUIPMENT or attenuator and the required OPTICAL LOSS.....	19
Table 201.C.101 – MARKING on the outside of FUNCTIONAL NIRS EQUIPMENT or their parts.....	33
Table 201.C.102 – ACCOMPANYING DOCUMENTS general .....	33
Table 201.C.103 – INSTRUCTIONS FOR USE .....	34
Table 201.C.104 – TECHNICAL DESCRIPTION .....	34
Table DD.1 – Correspondence between this document and the IMDRF essential principles.....	49
Table DD.2 – Correspondence between this document and the IMDRF labelling principles.....	50

<https://standards.iteh.ai>  
Document Preview

[IEC 80601-2-71:2025](https://standards.iteh.ai)

<https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025>

# INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-71: Particular requirements for the basic safety and essential performance of functional near-infrared spectroscopy (NIRS) equipment

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC and ISO draw attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC and ISO take no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC and ISO had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch> and [www.iso.org/patents](http://www.iso.org/patents). IEC and ISO shall not be held responsible for identifying any or all such patent rights.

IEC 80601-2-71 has been prepared by a Joint Working Group of IEC subcommittee 62D: Particular medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems, and ISO subcommittee SC3: Respiratory devices and related equipment used for patient care, of ISO technical committee 121: Anaesthetic and respiratory equipment. It is an International Standard.

This second edition cancels and replaces the first edition published in 2015. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-8:2006, IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-8:2006/AMD2:2020, IEC 60601-1-2:2014, IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-6:2010, IEC 60601-1-6:2010/AMD1:2013 and IEC 60601-1-6:2010/AMD2:2020;
- b) added requirements for ESSENTIAL PERFORMANCE;
- c) added requirements for PRIMARY OPERATING FUNCTIONS;
- d) added requirements for protection against excessive temperatures;
- e) added requirements for the display legibility for OPERATORS wearing personal protective equipment;
- f) harmonization with ISO 20417, where appropriate.

This publication is published as a double logo standard.

The text of this International Standard is based on the following documents of IEC:

Draft	Report on voting
62D/2169/FDIS	62D/2196/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

In this document, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS defined in Clause 3 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

NOTE The attention of the users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

[IEC 80601-2-71:2025](https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025)

<https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025>



## INTRODUCTION

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of FUNCTIONAL NIRS EQUIPMENT.

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" text giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this document.

**iTeh Standards**  
**(<https://standards.iteh.ai>)**  
**Document Preview**

[IEC 80601-2-71:2025](https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025)

<https://standards.iteh.ai/catalog/standards/iec/cc769223-dbc4-4f56-941e-8ffcab54be29/iec-80601-2-71-2025>

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-71: Particular requirements for the basic safety and essential performance of functional near-infrared spectroscopy (NIRS) equipment

#### 201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

##### 201.1.1 Scope

###### *Replacement:*

This part of IEC 80601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of FUNCTIONAL NIRS EQUIPMENT, as defined in 201.3.205, intended to be used by itself, or as a part of an ME SYSTEM hereinafter referred to as ME EQUIPMENT.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 7.2.13 and 8.4.1.

NOTE Additional information can be found in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, 4.2.

This document is not applicable to

- equipment for the measurement of oxygen saturation of the haemoglobin in the micro vessels (capillaries, arterioles and venules), i.e. tissue oximeters;
- frequency-domain and time-domain equipment for functional near-infrared spectroscopy;
- equipment for the measurement of changes in the concentration of chromophores other than oxy- and deoxy-haemoglobin;
- equipment for the measurement of changes in the concentration of oxy- and deoxy-haemoglobin in tissues other than the brain.

This document does not specify the requirements for:

- cerebral tissue oximeter equipment, which are given in ISO 80601-2-85 [1]<sup>1</sup>; and
- pulse oximeter equipment, which are given in ISO 80601-2-61 [2].

##### 201.1.2 Object

###### *Replacement:*

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for FUNCTIONAL NIRS EQUIPMENT as defined in 201.3.205.

NOTE This document has been prepared to address the relevant essential principles [3] and labelling principles [4] of the International Medical Devices Regulators Forum (IMDRF) as indicated in Annex DD.

<sup>1</sup> Numbers in square brackets refer to the Bibliography.

### 201.1.3 Collateral standards

*Addition:*

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and Clause 201.2 of this document.

IEC 60601-1-3 and IEC 60601-1-10 [5] do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

### 201.1.4 Particular standards

*Addition:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020, including the collateral standards, as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY or ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and applicable collateral standards.

The numbering of clauses and subclauses of this particular standard corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix "201" (e.g. 201.1 in this document addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.6 in this document addresses the content of Clause 6 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

**"Replacement"** means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is replaced completely by the text of this document.

**"Addition"** means that the text of this document additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard.

**"Amendment"** means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However due to the fact that definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

Where there is no corresponding clause or subclause in this document, the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

## 201.2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

*Addition:*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*  
IEC 60601-1:2005/AMD1:2012  
IEC 60601-1:2005/AMD2:2020

IEC 60825-1:2014, *Safety of laser products – Part 1: Equipment classification and requirements*

IEC 62471:2006, *Photobiological safety of lamps and lamp systems*

IEC 62570:2014, *Standard practice for marking medical devices and other items for safety in the magnetic resonance environment*

[IEC 80601-2-71:2025](https://standards.iteh.ai/standards/IEC/80601-2-71/2025)

<https://standards.iteh.ai/standards/ISO/17664-1/2021>, *Processing of health care products – Information to be provided by the medical device manufacturer for the processing of medical devices – Part 1: Critical and semi-critical medical devices*

ISO 17664-2:2021, *Processing of health care products – Information to be provided by the medical device manufacturer for the processing of medical devices – Part 2: Non-critical medical devices*

ISO 20417:2021, *Medical devices – Information to be supplied by the manufacturer*

## 201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, ISO 17664-2:2021, ISO 20417:2021 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

**201.3.201****AVERAGE OPTICAL POWER**

temporal average power of continuous light or repeated light pulses from each discrete wavelength, from the EMITTER OPTODE connected to the FUNCTIONAL NIRS MONITOR

**201.3.202****DETECTOR OPTODE**

part of the FUNCTIONAL NIRS PROBE which detects light from the living tissue that forms part of the APPLIED PART

**201.3.203****EMITTER OPTODE**


part of the FUNCTIONAL NIRS PROBE which emits light to the living tissue that forms part of the APPLIED PART

Note 1 to entry: An EMITTER OPTODE can contain two or more light sources (e.g. laser diodes or light-emitting diodes) operating at different NOMINAL wavelengths.

**201.3.204****FWHM****FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION**

difference of the wavelength between the two points whose corresponding power values are equal and 3 dB lower than the values at each PEAK WAVELENGTH

Note 1 to entry: FWHM is a measure of the width of the spectral power distribution emitted by the EMITTER OPTODE connected to the FUNCTIONAL NIRS MONITOR. Figure 201.101 provides a visual representation. If there are more than two wavelengths where power value is 3 dB lower than the values at each PEAK WAVELENGTH, FWHM shall be calculated from the difference between minimum and maximum wavelengths.



**Figure 201.101 – FULL WIDTH AT HALF MAXIMUM OF SPECTRAL POWER DISTRIBUTION**

**201.3.205****FUNCTIONAL NIRS EQUIPMENT****FUNCTIONAL NEAR-INFRARED SPECTROSCOPY EQUIPMENT**

ME EQUIPMENT that estimates PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGES in living tissue by illuminating tissue with continuous light and detecting changes in the infrared and visible light intensity diffusively reflected from the tissue

**201.3.206****FUNCTIONAL NIRS MONITOR  
FUNCTIONAL NEAR-INFRARED SPECTROSCOPY MONITOR  
MONITOR**

part of the FUNCTIONAL NIRS EQUIPMENT that encompasses the electronics, display and operator-equipment interface excluding the EMITTER OPTODES and DETECTOR OPTODES

Note 1 to entry: The FUNCTIONAL NIRS MONITOR typically encompasses the electronics, display and OPERATOR-EQUIPMENT INTERFACE. It can consist of a separate control unit, a computer and a display. Electronics or a control unit integrated in the APPLIED PART are not part of the FUNCTIONAL NIRS MONITOR.

**201.3.207****FUNCTIONAL NIRS PROBE  
FUNCTIONAL NEAR-INFRARED SPECTROSCOPY PROBE  
PROBE**

part of the functional NIRS EQUIPMENT that includes the APPLIED PART

Note 1 to entry: This definition of PROBE is consistent with the definition of PROBE for cerebral tissue oximeters as given in ISO 80601-2-85.

Note 2 to entry: The PROBE contains single or multiple EMITTER OPTODES and DETECTOR OPTODES. A reflectance PROBE design is the typical configuration.

**201.3.208****FUNCTIONAL NIRS PHANTOM**

apparatus that simulates a PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE by giving the ME EQUIPMENT a specified known change in attenuation to evaluate the difference between the measured value of the pseudo PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE obtained from the measurement on the phantom and the reference value calculated from the attenuation change

Note 1 to entry: The FUNCTIONAL NIRS PHANTOM plays a role in determining the performance of FUNCTIONAL NIRS EQUIPMENT, especially PATHLENGTH-DEPENDENT HAEMOGLOBIN CHANGE measurement. A description of the function and specifications regarding the manufacturing of the FUNCTIONAL NIRS PHANTOM is found in Annex BB.

Note 2 to entry: A FUNCTIONAL NIRS PHANTOM is developed during design and is used at the time of inspection in manufacturing or after being placed into service.

Note 3 to entry: The change in attenuation corresponds to a change in OPTICAL LOSS of the FUNCTIONAL NIRS PHANTOM of the same magnitude.

**201.3.209****MEASUREMENT CHANNEL  
CHANNEL**

combination of an EMITTER OPTODE and a DETECTOR OPTODE that provide an output

**201.3.210****OPTICAL LOSS**

ratio of the total optical power exiting the FUNCTIONAL NIRS PHANTOM or attenuator through a specified aperture, to the optical power emitted by the EMITTER OPTODE connected to the FUNCTIONAL NIRS MONITOR and placed on the entrance side of the FUNCTIONAL NIRS PHANTOM or attenuator

Note 1 to entry: The OPTICAL LOSS is denoted in dB.  $X$  dB OPTICAL LOSS is equivalent to  $10^{-X/10}$ .

Note 2 to entry: The optical power exiting the EMITTER OPTODE and the FUNCTIONAL NIRS PHANTOM can be measured with an optical power meter.

Note 3 to entry: For details of the measurement of the OPTICAL LOSS, refer to BB.3.3.

**201.3.211****PATHLENGTH-DEPENDENT DEOXYHAEMOGLOBIN CHANGE**

value calculated from the changes in detected light intensities, equal to the product of the apparent change in the concentration of deoxyhaemoglobin and the mean optical pathlength