



Designation: D3325 – 90 (Reapproved 2020)

# Standard Practice for Preservation of Waterborne Oil Samples<sup>1</sup>

This standard is issued under the fixed designation D3325; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This practice covers the preservation of waterborne oil samples from the time of collection to the time of analysis. Information is provided to ensure sample integrity and to avoid contamination and to minimize microbial degradation.

1.2 The practice is for controlled field or laboratory conditions and specifies thorough preparation of equipment and precise operation. Where these details must be compromised in a field emergency, nonstandard simplifications are recommended that will minimize or eliminate consequent errors.

NOTE 1—Procedures for the analysis of oil spill samples are Practices D3326, D3415, and D4489, and Test Methods D3650, D3327, D3328, and D3414. A guide to the use of ASTM test methods for the analysis of oil spill samples is found in Practice D3415.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

D1129 Terminology Relating to Water

D3326 Practice for Preparation of Samples for Identification of Waterborne Oils

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.06 on Methods for Analysis for Organic Substances in Water.

Current edition approved Jan. 1, 2020. Published January 2020. Originally approved in 1974. Last previous edition approved in 2013 as D3325 – 90 (2013). DOI: 10.1520/D3325-90R20.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

D3327 Test Method for Analysis of Selected Elements in Waterborne Oils (Withdrawn 1993)<sup>3</sup>

D3328 Test Methods for Comparison of Waterborne Petroleum Oils by Gas Chromatography

D3414 Test Method for Comparison of Waterborne Petroleum Oils by Infrared Spectroscopy (Withdrawn 2018)<sup>3</sup>

D3415 Practice for Identification of Waterborne Oils

D3650 Test Method for Comparison of Waterborne Petroleum Oils By Fluorescence Analysis (Withdrawn 2018)<sup>3</sup>

D4489 Practices for Sampling of Waterborne Oils

## 3. Terminology

3.1 *Definitions:*

3.1.1 For definitions of terms used in this standard, refer to Terminology D1129.

## 4. Summary of Practice

4.1 Special types of sample containers and shipping containers are recommended. Samples may be of several types: tar balls, collected oil, oil-water mixtures, emulsions, and oil and water on collecting devices such as silanized glass cloth, TFE-fluorocarbon polymer, or other materials. Instructions are given for the care of samples to minimize changes due to autoxidation and microbial attack between the time of sampling and the time of analysis. Services available for transportation of samples are described.

## 5. Apparatus

5.1 *Sample Containers*—Borosilicate glass containers that have been thoroughly cleaned are preferable. All glass containers, new or used, must be thoroughly cleaned and washed prior to use. The cleaning steps consist of an initial wash with a warm aqueous detergent mixture followed by six hot tap water rinses, two rinses with reagent water, a rinse with reagent-grade acetone, and a final rinse with a solvent such as pentane, hexane, cyclohexane, dichloromethane, or chloroform followed by drying in a clean oven at 105°C or hotter for 30 min. If the glassware requires cleaning under field conditions, it should be washed with warm aqueous detergent followed by extensive water rinsing. A solvent rinse with acetone should be made, if possible, followed by lengthy air drying to remove

<sup>3</sup> The last approved version of this historical standard is referenced on www.astm.org.

\*A Summary of Changes section appears at the end of this standard