



Designation: D5454 – 11 (Reapproved 2020)

Standard Test Method for Water Vapor Content of Gaseous Fuels Using Electronic Moisture Analyzers¹

This standard is issued under the fixed designation D5454; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of the water vapor content of gaseous fuels by the use of electronic moisture analyzers. Such analyzers commonly use sensing cells based on phosphorus pentoxide, P_2O_5 , aluminum oxide, Al_2O_3 , or silicon sensors piezoelectric-type cells and laser based technologies.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.3 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[D1142 Test Method for Water Vapor Content of Gaseous Fuels by Measurement of Dew-Point Temperature](#)

[D1145 Test Method for Sampling Natural Gas](#) (Withdrawn 1986)³

[D4178 Practice for Calibrating Moisture Analyzers](#)

3. Terminology

3.1 *Definitions of Terms Specific to This Standard:*

3.1.1 *capacitance-type cell*—this cell uses aluminum coated with Al_2O_3 as part of a capacitor. The dielectric Al_2O_3 film

changes the capacity of the capacitor in relation to the water vapor present. Silicone cells also operate on this principal by reporting a capacitance change when adsorbing or desorbing water vapor.

3.1.2 *electrolytic-type cell*—this cell is composed of two noble metal electrode wires coated with P_2O_5 . A bias voltage is applied to the electrodes, and water vapor chemically reacts, generating a current between the electrodes proportional to the water vapor present.

3.1.3 *piezoelectric-type cell*—sensor consists of a pair of electrodes which support a quartz crystal (QCM) transducer. When voltage is applied to the sensor a very stable oscillation occurs. The faces of the sensor are coated with a hygroscopic polymer. As the amount of moisture absorbed onto the polymer varies, a proportional change in the oscillation frequency is produced.

3.1.4 *laser-type cell*—consists of a sample cell with an optical head mounted on one end and a mirror mounted on the other; however, some models will not need a mirror to reflect the light wavelength emitted from the laser. The optical head contains a NIR laser, which emits light at a wavelength known to be absorbed by the water molecule. Mounted, the laser is a detector sensitive to NIR wavelength light. Light from the laser passes through the far end and returns to the detector in the optical head. A portion of the emitted light, proportional to the water molecules present, is absorbed as the light transits the sample cell and returns to the detector.

3.1.5 *water content*—water content is customarily expressed in terms of dewpoint, °F or °C, at atmospheric pressure, or the nonmetric term of pounds per million standard cubic feet, lb/MMSCF. The latter term will be used in this test method because it is the usual readout unit for electronic analyzers. One lb/MMSCF = 21.1 ppm by volume or 16.1 mgm/m³ of water vapor. Analyzers must cover the range 0.1 to 50 lb/MMSCF.

3.1.6 *water dewpoint*—the temperature (at a specified pressure) at which liquid water will start to condense from the water vapor present. Charts of dewpoints versus pressure and water content are found in Test Method [D1142](#).

¹ This test method is under the jurisdiction of ASTM Committee D03 on Gaseous Fuels and is the direct responsibility of Subcommittee D03.06.04 on Analysis by Colorimetric Techniques.

Current edition approved Jan. 1, 2020. Published January 2020. Originally approved in 1993. Last previous edition approved in 2011 as D5454 – 11^{e1}. DOI: 10.1520/D5454-11R20.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.