

ISO/ASTM 51631:2020(E)



Standard Practice for Use of Calorimetric Dosimetry Systems for Dose Measurements and Routine Dosimetry System Calibration in Electron Beams¹

This standard is issued under the fixed designation ISO/ASTM 51631; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

1. Scope

1.1 This practice covers the preparation and use of semi-adiabatic calorimetric dosimetry systems for measurement of absorbed dose and for calibration of routine dosimetry systems when irradiated with electrons for radiation processing applications. The calorimeters are either transported by a conveyor past a scanned electron beam or are stationary in a broadened beam.

1.2 This document is one of a set of standards that provides recommendations for properly implementing dosimetry in radiation processing, and describes a means of achieving compliance with the requirements of ISO/ASTM Practice 52628 for a calorimetric dosimetry system. It is intended to be read in conjunction with ISO/ASTM Practice 52628.

1.3 The calorimeters described in this practice are classified as Type II dosimeters on the basis of the complex effect of influence quantities. See ISO/ASTM Practice 52628.

1.4 This practice applies to electron beams in the energy range from 1.5 to 12 MeV.

1.5 The absorbed dose range depends on the calorimetric absorbing material and the irradiation and measurement conditions. Minimum dose is approximately 100 Gy and maximum dose is approximately 50 kGy.

1.6 The average absorbed-dose rate range shall generally be greater than 10 Gy·s⁻¹.

1.7 The temperature range for use of these calorimetric dosimetry systems depends on the thermal resistance of the calorimetric materials, on the calibration range of the temperature sensor, and on the sensitivity of the measurement device.

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate*

appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

E666 Practice for Calculating Absorbed Dose From Gamma or X Radiation

E668 Practice for Application of Thermoluminescence-Dosimetry (TLD) Systems for Determining Absorbed Dose in Radiation-Hardness Testing of Electronic Devices

E3083 Terminology Relating to Radiation Processing: Dosimetry and Applications

2.2 ISO/ASTM Standards:²

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51649 Practice for Dosimetry in an Electron Beam Facility for Radiation Processing at Energies Between 300 keV and 25 MeV

51707 Guide for Estimating Uncertainties in Dosimetry for Radiation Processing

52628 Practice for Dosimetry in Radiation Processing

2.3 International Commission on Radiation Units and Measurements (ICRU) Reports:³

ICRU Report 34 The Dosimetry of Pulsed Radiation

ICRU Report 35 Radiation Dosimetry: Electron Beams with Energies Between 1 and 50 MeV

ICRU Report 80 Dosimetry Systems for use in Radiation Processing

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

¹ This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.02 on Dosimetry Systems, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved Jan. 15, 2020. Published February 2020. Originally published as E 1631 – 94. The present Fourth Edition of International Standard ISO/ASTM 51631:2020(E) is a minor revision of the Third Edition of ISO/ASTM 51631–2013(E).

² For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from the Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, U.S.A.

2.4 *Joint Committee for Guides in Metrology (JCGM) Reports:*

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement⁴

JCGM 200:2012, VIM International vocabulary of metrology – Basic general concepts and general terms⁵

3. Terminology

3.1 Definitions:

3.1.1 *primary-standard dosimetry system*—dosimetry system that is designated or widely acknowledged as having the highest metrological qualities and whose value is accepted without reference to other standards of the same quantity.

3.1.2 *reference standard dosimetry system*—dosimetry system, generally having the highest metrological quality available at a given location or in a given organization, from which measurements made there are derived.

3.1.3 *transfer standard dosimetry system*—dosimetry system used as an intermediary to calibrate other dosimetry systems.

3.1.4 *type II dosimeter*—dosimeter, the response of which is affected by influence quantities in a complex way that cannot practically be expressed in terms of independent correction factors.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *adiabatic*—no heat exchange with the surroundings.

3.2.2 *calorimeter*—assembly consisting of calorimetric body (absorber), thermal insulation, and temperature sensor with wiring that, when irradiated, exhibits increase in the absorber temperature that can be related to absorbed dose. This language parallels that of dosimeter.

3.2.3 *calorimetric body*—mass of material absorbing radiation energy and whose temperature is measured.

3.2.4 *calorimetric dosimetry system*—dosimetry system consisting of calorimeter, measurement instruments and their associated reference standards, and procedures for the system's use.

3.2.5 *endothermic reaction*—chemical reaction that consumes energy.

3.2.6 *exothermic reaction*—chemical reaction that releases energy.

3.2.7 *heat defect (thermal defect)*—amount of energy released or consumed by chemical reactions caused by the absorption of radiation energy.

3.2.8 *specific heat capacity*—amount of energy required to raise 1 kg of material by the temperature of 1 K.

⁴ Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1). Available free of charge at the BIPM website (<http://www.bipm.org>).

⁵ Document produced by Working Group 2 of the Joint Committee for Guides in Metrology (JCGM/WG 2). Available free of charge at the BIPM website (<http://www.bipm.org>).

3.2.9 *thermistor*—electrical resistor with a well-defined relationship between resistance and temperature of the thermistor.

3.2.10 *thermocouple*—junction of two metals producing an electrical voltage with a well-defined relationship to junction temperature.

3.3 Definitions of other terms used in this standard that pertain to radiation measurement and dosimetry may be found in ASTM Terminology E3083. Definitions in E3083 are compatible with ICRU Report 85a; that document, therefore, may be used as an alternative reference.

4. Significance and use

4.1 This practice is applicable to the use of calorimetric dosimetry systems for the measurement of absorbed dose in electron beams, the qualification of electron irradiation facilities, periodic checks of operating parameters of electron irradiation facilities, and calibration of other dosimetry systems in electron beams. Calorimetric dosimetry systems are most suitable for dose measurement at electron irradiation facilities utilizing conveyor systems for transport of product during irradiation.

NOTE 1—For additional information on calorimetric dosimetry system operation and use, see ICRU Report 80. For additional information on the use of dosimetry in electron accelerator facilities, see ISO/ASTM 51649, and ICRU Reports 34 and 35, and Refs (1-3).⁶

4.2 The calorimetric dosimetry systems described in this practice are not primary standard dosimetry systems. The calorimeters are classified as Type II dosimeters (ISO/ASTM 52628). They might be used as internal standards at an electron beam irradiation facility, including being used as transfer standard dosimetry systems for calibration of other dosimetry systems, or they might be used as routine dosimeters. The calorimetric dosimetry systems are calibrated by comparison with transfer standard dosimeters.

4.3 The dose measurement is based on the measurement of the temperature rise (dosimeter response) in an absorber (calorimetric body) irradiated by an electron beam. Different absorbing materials are used, but the response is usually defined in terms of dose to water.

NOTE 2—The calorimetric bodies of the calorimeters described in this practice are made from low atomic number materials. The electron fluences within these calorimetric bodies are almost independent of energy when irradiated with electron beams of 1.5 MeV or higher, and the mass collision stopping powers are approximately the same for these materials.

4.4 The absorbed dose in other materials irradiated under equivalent conditions can be calculated. Procedures for making such calculations are given in ASTM Practices E666 and E668, and Ref (1).

4.4.1 Calorimeters for use at industrial electron accelerators have been constructed using graphite, polystyrene or a Petri dish filled with water as the calorimetric body (4-10). The thickness of the calorimetric body should be less than the range of the incident electrons.

4.4.2 Polymeric materials other than polystyrene might also be used for calorimetric measurements. Polystyrene is used

⁶ The boldface numbers in parentheses refer to the bibliography at the end of this practice.

because it is known to be resistant to radiation (11) and because almost no exo- or endothermic reactions take place (12).

5. Interferences

5.1 *Extrapolation*—The calorimetric dosimetry systems described in this practice are not adiabatic, because of the exchange of heat with the surroundings or within the calorimeter assembly. The maximum temperature reached by the calorimetric body is different from the temperature that would have been reached in the absence of that heat exchange. The temperature drifts before and after irradiation should be extrapolated to the midpoint of the irradiation period in order to determine the true temperature increase due to the absorbed dose.

5.2 *Heat Defect*—Chemical reactions in irradiated material (resulting in what is called the heat defect or thermal defect) might be endo- or exothermic and might lead to measurable temperature changes (3).

5.3 *Specific Heat Capacity*—The specific heat capacity of some materials used as a calorimetric body might change with accumulated absorbed dose, thereby affecting the response of the calorimeters. This is notably the case for polymers, such as polystyrene, and it will therefore be necessary to recalibrate calorimetric dosimetry systems at intervals that will depend on the total accumulated dose.

NOTE 3—For calorimeters using polystyrene as material for the calorimetric body, the change in specific heat capacity might be in the order of 1 % per accumulated dose of 1 MGy. It can therefore be useful to track accumulated dose for polystyrene calorimeters.

5.4 *Influence Quantities*—The response of the calorimetric dosimetry systems to absorbed dose does not depend on ambient relative humidity and temperature.

5.5 *Temperature Effects from Accelerator Structure*—The calorimeters are often irradiated on a conveyor used for passing products and samples through the irradiation zone. Recognizing that the thermal insulation around the calorimetric body is not perfect, there is possibility that, for example, radiated heat from the mechanical structures of the irradiation facility and from the conveyor might contribute to the measured temperature increase in the calorimeters.

5.6 *Thermal Equilibrium*—The most reproducible results are obtained when the calorimeters are in thermal equilibrium with their surroundings before irradiation.

5.7 *Foreign Materials*—The temperature sensors, wires, etc. of the calorimeter represent foreign materials, which might influence the temperature rise of the calorimetric body. These components should be as small as possible.

5.8 *Dose Gradients*—Dose gradients will exist within the calorimetric body when it is irradiated with electrons. These gradients must be taken into account, for example, when other dosimeters are calibrated by comparison with calorimetric dosimetry systems.

6. Apparatus

6.1 *A Typical Graphite Calorimeter* is a disc of graphite placed in a thermally insulating material such as foamed plastic (4-6). A calibrated thermistor or thermocouple is embedded

inside the disc. Some typical examples of graphite disc thicknesses and masses are listed in Annex A2 (2).

6.2 *A Typical Water Calorimeter* is a sealed polystyrene Petri dish filled with water and placed in thermally insulating foamed plastic (4). A calibrated temperature sensor (thermistor) is placed through the side of the dish into the water. The shape and size of the water calorimeter can be similar to the shape and size of the polystyrene calorimeter (see 6.3).

6.3 *A Typical Polystyrene Calorimeter* is a polystyrene disc placed in thermally insulating foamed plastic. A calibrated thermistor or thermocouple is imbedded inside the disc. The dimension of the polystyrene disc might be similar to that of the graphite and water calorimeters (9). See Fig. 1 as an example of a polystyrene calorimeter designed for use at 10 MeV electron irradiation (13).

6.4 The thickness of the calorimetric body should be less than the range of the irradiating electrons, typically not exceeding $\frac{1}{3}$ of the range of the incident electrons. This will limit the effects of variation of the dose gradients within the calorimetric body.

6.5 Radiation-resistant components should be used for the parts of the calorimeter that are exposed to the electron beam. This also applies to insulation of electrical wires.

6.6 Good thermal contact must exist between the temperature sensor and the calorimetric body. For graphite and polystyrene calorimeters, this can be assured by adding a small amount of heat-conducting compound when mounting the temperature sensor.

6.7 *Measurement*—The response of the calorimeters is the temperature rise of the calorimetric body. This temperature rise is usually registered by thermistors or thermocouples.

6.7.1 *Thermistor*—A high-precision ohm-meter can be used for measurement of thermistor resistance. The meter should have a reproducibility of better than $\pm 0.1\%$ ($k=1$) and a combined uncertainty of better than $\pm 0.2\%$ ($k=1$). It should preferably be equipped for four-wire type resistance



Courtesy of Risø High Dose Reference Laboratory.

FIG. 1 Example of a polystyrene calorimeter used for routine measurements at a 10-MeV industrial electron accelerator

