



Designation: **F1611—00 (Reapproved 2018) F1611 – 20**

Standard Specification for Intramedullary Reamers¹

This standard is issued under the fixed designation F1611; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification provides requirements for material, dimensions and tolerances, finish and marking, and care and handling for reamers intended to cut a cylindrical path along the medullary canal of diaphyseal bone.

1.2 Intramedullary reamers are commonly used to prepare the medullary canal for the insertion of intramedullary fixation devices (~~IMFD~~)-(IMFDs). As such, the relationship between the intramedullary reamer diameter and the IMFD's diameter are considered.

1.3 *Units*—The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

[A564/A564M Specification for Hot-Rolled and Cold-Finished Age-Hardening Stainless Steel Bars and Shapes](#)

[A693 Specification for Precipitation-Hardening Stainless and Heat-Resisting Steel Plate, Sheet, and Strip](#)

[A705/A705M Specification for Age-Hardening Stainless Steel Forgings](#)

[F86 Practice for Surface Preparation and Marking of Metallic Surgical Implants](#)

[F565 Practice for Care and Handling of Orthopedic Implants and Instruments](#)

[F899 Specification for Wrought Stainless Steels for Surgical Instruments](#)

[F983 Practice for Permanent Marking of Orthopaedic Implant Components](#)

[F1264 Specification and Test Methods for Intramedullary Fixation Devices](#)

<https://standards.iteh.ai/catalog/standards/sist/2513e9fc-150f-44f7-b08e-f91a59bd1d7c/astm-f1611-20>

3. Terminology

3.1 *Definitions of Terms Specific to This Standard:*

3.1.1 *cutting head, n*—the portion of the reamer, which consists of flutes, or edges, which ~~cut~~cuts the bone.

3.1.2 *reamer diameter, n*—the diameter of the circumscribed circle of the cutting head's cross-section (shown in Fig. 1).

3.1.3 *reamer shaft diameter, n*—the diameter of the circumscribed circle of the long portion of the reamer, which connects the cutting portion of the reamer to the drill.

4. Classification

4.1 In general, intramedullary reamers consist of two types:

4.1.1 *One-piece reamer*—*Reamer*—A design where the reamer shaft and cutting head are permanently attached to each other.

4.1.2 *Modular Reamer*—A design where the reamer shaft and cutting head are two separate components, fixed to each other temporarily at the time of use via a geometric connection, for example, dovetail joint.

¹ This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.21 on Osteosynthesis.

Current edition approved Feb. 1, 2018; Feb. 1, 2020. Published April 2018; April 2020. Originally approved in 1995. Last previous edition approved in 2013 as F1611 – 00(2013); (2018). DOI: 10.1520/F1611-00R18-10.1520/F1611-20.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.