

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-54: Particular requirements for the basic safety and essential
performance of X-ray equipment for radiography and radioscopy**

IEC 60601-2-54:2022

<https://standards.iteh.ai/catalog/standards/sist/ef58c40c-25e3-4ea9-9102-667de4713615/iec-60601-2-54-2022>



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC 60601-2-54:2022

<https://standards.iteh.ai/catalog/standards/sist/ef58c40c-25e3-4ea9-9102-667de4713615/iec-60601-2-54-2022>



IEC 60601-2-54

Edition 2.0 2022-09
REDLINE VERSION

INTERNATIONAL STANDARD



**Medical electrical equipment –
Part 2-54: Particular requirements for the basic safety and essential
performance of X-ray equipment for radiography and radioscopy**

[IEC 60601-2-54:2022](https://standards.iteh.ai/catalog/standards/sist/ef58c40c-25e3-4ea9-9102-667de4713615/iec-60601-2-54-2022)

<https://standards.iteh.ai/catalog/standards/sist/ef58c40c-25e3-4ea9-9102-667de4713615/iec-60601-2-54-2022>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.50

ISBN 978-2-8322-5784-5

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD	4
INTRODUCTION	2
INTRODUCTION TO AMENDMENT 1	2
INTRODUCTION TO AMENDMENT 2	2
201.1 Scope, object and related standards	9
201.2 Normative references	11
201.3 Terms and definitions	11
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	15
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	15
201.7 ME EQUIPMENT identification, marking and documents	15
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	19
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	22
201.10 Protection against unwanted and excessive radiation HAZARDS	26
201.11 Protection against excessive temperatures and other HAZARDS	26
201.12 Accuracy of controls and instruments and protection against hazardous outputs	27
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	27
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	27
201.15 Construction of ME EQUIPMENT	28
201.16 ME SYSTEMS	28
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	28
202 Electromagnetic compatibility disturbances – Requirements and tests	28
202.101 * Immunity testing of ESSENTIAL PERFORMANCE	28
203 RADIATION PROTECTION in diagnostic X-RAY EQUIPMENT	29
203.4 General requirements	29
203.5 ME EQUIPMENT identification, marking and documents	30
203.6 RADIATION management	33
203.7 RADIATION QUALITY	48
203.8 Limitation of the extent of the X-RAY BEAM and relationship between X-RAY FIELD and IMAGE RECEPTION AREA	49
203.9 FOCAL SPOT TO SKIN DISTANCE	57
203.10 ATTENUATION of the X-RAY BEAM between the PATIENT and the X-RAY IMAGE RECEPTOR	58
203.11 Protection against RESIDUAL RADIATION	59
203.12 Protection against LEAKAGE RADIATION	61
203.13 Protection against STRAY RADIATION	61
Annexes	67
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	68
Annex AA (informative) Particular guidance and rationale	70
Bibliography	75

Index of defined terms used in this document	78
Figure 203.101 – Zone of EXTRA-FOCAL RADIATION	50
Figure 203.102 – Discrepancies in covering the IMAGE RECEPTION AREA.....	52
Figure 203.103 – Discrepancies in visual indication of the X-RAY FIELD	56
Figure 203.104 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY SOURCE ASSEMBLY below the PATIENT SUPPORT)	64
Figure 203.105 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE ASSEMBLY below the PATIENT SUPPORT)	65
Figure 203.106 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY SOURCE ASSEMBLY above the PATIENT SUPPORT)	65
Figure 203.107 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE ASSEMBLY above the PATIENT SUPPORT).....	66
Table 201.101 – Distributed potential ESSENTIAL PERFORMANCE requirements	15
Table 203.101 – Tests for verifying reproducibility and linearity	37
Table 203.102 – LOADINGS for testing automatic exposure controls	38
Table 203.103 – ATTENUATION for the measurement of AIR KERMA.....	40
Table 203.104 – ATTENUATION EQUIVALENT of items in the X-RAY BEAM.....	58
Table 203.105 – Application categories.....	60
Table 203.106 – Requirements for PRIMARY PROTECTIVE SHIELDING	61
Table 203.107 – STRAY RADIATION in SIGNIFICANT ZONES OF OCCUPANCY	63
Table 201.C.101 – Marking on the outside of ME EQUIPMENT or its parts	68
Table 201.C.102 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS	68

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This redline version of the official IEC Standard allows the user to identify the changes made to the previous edition IEC 60601-2-54:2009+AMD1:2015+AMD2:2018 CSV. A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text.

IEC 60601-2-54 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice. It is an International Standard.

This second edition cancels and replaces the first edition published in 2009, Amendment 1:2015 and Amendment 2:2018. This edition constitutes a technical revision.

This edition includes editorial and technical changes to reflect the IEC 60601-1:2005/AMD2:2020. It also contains corrections and technical improvements. Significant technical changes with respect to the previous edition are as follows:

- a) a new specific term DOSIMETER is introduced to replace the general term DOSEMETER;
- b) terms and definitions taken exclusively from IEC TR 60788:2004 and which are specifically applicable in this document have been moved to 201.3;
- c) the collateral standards IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 are applicable if MANUFACTURER so declares;
- d) the subclause 201.11.101 “Protection against excessive temperatures of X-ray tube assemblies” has been removed from this document as its requirements are sufficiently and clearly covered by IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020 and IEC 60601-2-28:2017;
- e) to adopt changes which are introduced with respect to indicator lights in 7.8.1 of the IEC 60601-1:2005/AMD2:2020 clarification of requirements is provided to avoid conflicts with requirements of indicator lights stipulated for X-RAY EQUIPMENT;
- f) explanation of the term ESSENTIAL PERFORMANCE is provided in Annex AA to emphasize the performance of the clinical function under NORMAL and SINGLE FAULT CONDITIONS.

The text of this document is based on the following documents:

IEC 60601-2-54:2022	
https://standards.iech.org/catalog/standards/sist/c8c49101-9101-667de4713615/iec-60601-2-54-2022/	
Draft	Report on voting
62B/1285/FDIS	62B/1293/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type*;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);

- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 and IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The “colour inside” logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This document has been prepared to provide, based on IEC 60601-1:2005 (third edition) and its collaterals, a complete set of safety requirements for ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. The purpose of this second edition is to introduce changes to reference the second amendment (2020) to IEC 60601-1:2005 and associated collateral standards. Moreover, in Annex AA a clarification of the term for ESSENTIAL PERFORMANCE is provided. ~~While the previously existing standards for such equipment were dedicated to components and subsystems,~~ This document addresses the system level of X-RAY EQUIPMENT, which consists of a combination of an X-RAY GENERATOR, ASSOCIATED EQUIPMENT and ACCESSORIES. Component functions are addressed as far as necessary.

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. Requirements for additional provisions for ME EQUIPMENT for interventional applications are covered by IEC 60601-2-43.

iTeh STANDARD PREVIEW
(standards.iteh.ai)

IEC 60601-2-54:2022

<https://standards.iteh.ai/catalog/standards/sist/ef58c40c-25e3-4ea9-9102-667de4713615/iec-60601-2-54-2022>

~~INTRODUCTION TO AMENDMENT 1~~

~~The purpose of this first amendment to IEC 60601-2-54:2009 is to introduce changes to reference the first amendment (2012) to IEC 60601-1:2005. As neither IEC 60601-2-54:2009 nor this amendment refers to specific elements of IEC 60601-1-2, the introduction of a dated reference to the latter document has been removed. In addition, a number of technical errors have been corrected.~~

~~INTRODUCTION TO AMENDMENT 2~~

~~The purpose of this second amendment to IEC 60601-2-54:2009 is to introduce changes which take the current state of the art into account. Therefore, X-RAY EQUIPMENT specified for DIRECT RADIOSCOPY is no longer in the scope of this document. The normative references were also updated in this amendment, and editorial clarifications and new terms and definitions were added. Provisions for QUALITY CONTROL PROCEDURES to be recommended by the MANUFACTURER are emphasized. Specific attention is paid to EXAMINATION PROTOCOLS in a new subclause which differentiate between adult and paediatric applications, in particular for X-RAY EQUIPMENT without an AUTOMATIC CONTROL SYSTEM. In addition, fixed periods for termination of LOADING after release of the RADIATION control by the OPERATOR are stipulated for RADIOSCOPY.~~

~~A new subclause on electronic documentation of EXAMINATION PROTOCOLS is introduced. It recommends providing access to electronic documentation containing relevant parameters of the PRE-PROGRAMMED EXAMINATION PROTOCOL. In another new subclause, the creation of basic documentation of the RADIATION DOSE STRUCTURED REPORT (RDSR) according to IEC 61910-1 is recommended. Furthermore, the subclause describing the LAST IMAGE HOLD RADIOGRAM has been revised and requires that the last image in RADIOSCOPY be displayed rather than provide just a means to display it.~~

~~This amendment recommends providing a graphical DISPLAY of the position of the BEAM LIMITING DEVICE blades on the IMAGE DISPLAY DEVICE in the subclause "Indication on the X-RAY EQUIPMENT".~~

~~Finally, the requirement for providing means to limit the FOCAL SPOT TO SKIN DISTANCES for radiosopic X-RAY EQUIPMENT differentiates between MOBILE and FIXED EQUIPMENT and extends, in the latter case, the minimum distance in possible clinical applications.~~

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy

201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of ME EQUIPMENT and ME SYSTEMS intended to be used for projection RADIOGRAPHY and INDIRECT RADIOSCOPY. IEC 60601-2-43 applies to ME EQUIPMENT and ME SYSTEMS intended to be used for interventional applications and refers to applicable requirements in this document.

ME EQUIPMENT and ME SYSTEMS intended to be used for bone or tissue absorption densitometry, computed tomography, mammography or dental or radiotherapy applications are excluded from the scope of this document. The scope of this document also excludes radiotherapy simulators.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

201.1.2 Object

Replacement:

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ME EQUIPMENT and ME SYSTEMS for RADIOGRAPHY and RADIOSCOPY.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 as modified in 201.2.

IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-3:2008 and IEC 60601-1-3:2008/AMD1:2013 and IEC 60601-1-3:2008/AMD2:2021 apply, as modified in Clauses 202 and 203 respectively. If the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in a HOME HEALTHCARE ENVIRONMENT, then IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020 apply and if the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in an EMERGENCY MEDICAL SERVICES ENVIRONMENT, then IEC 60601-1-12:2014 and IEC 60601-1-12:2015/AMD1:2020 apply. IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10, ~~IEC 60601-1-11 and IEC 60601-1-12~~ do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE 1 OPERATORS of X-RAY EQUIPMENT are used to audible signals as ~~required~~ specified in this document rather than to the concepts of IEC 60601-1-8. Therefore IEC 60601-1-8 does not apply.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020.

~~For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.~~

The numbering of clauses and subclauses of this document corresponds to that of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 with the prefix “201” (e.g. 201.1 in this document addresses the content of Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020) or applicable collateral standard with the prefix “20x” where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this document addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this document addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are specified by the use of the following words:

“(Replacement)” means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard is replaced completely by the text of this document.

<https://standards.iteh.ai/catalog/standards/sist/ef58c40c-25e3-4ea9-9102-667de4713615/iec-60601-1-2-54-2022>
“(Addition)” means that the text of this document is additional to the requirements of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard.

“(Amendment)” means that the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard is amended as indicated by the text of this document.

Subclauses, figures or tables which are additional to those of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered starting from 201.101. However, due to the fact that definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 are numbered 3.1 through 3.139154, additional definitions in this document are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where “x” is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

~~The term “this standard” is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.~~

Where there is no corresponding clause or subclause in this document, the clause or subclause of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 or the applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC

60601-1:2005/AMD2:2020 or the applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this document.

201.2 Normative references

NOTE Informative references are listed in the Bibliography.

Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

Addition:

IEC 60336:2020, *Medical electrical equipment – X-ray tube assemblies for medical diagnosis – ~~Characteristics of focal spots~~ Focal spot dimensions and related characteristics*

IEC 60580:20002019, *Medical electrical equipment – Dose area product meters*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 60806, *Determination of the maximum symmetrical radiation field ~~from a rotating anode~~ of X-ray tube assemblies and X-ray source assemblies for medical diagnosis*

~~IEC 62220-1-1:2015, Medical electrical equipment – Characteristics of digital X-ray imaging devices – Part 1-1: Determination of the detective quantum efficiency – Detectors used in radiographic imaging~~

IEC 61910-1:2014, *Medical electrical equipment – Radiation dose documentation – Part 1: Radiation dose structured reports for radiography and radioscopy*

IEC 62494-1:2008, *Medical electrical equipment – Exposure index of digital X-ray imaging systems – Part 1: Definitions and requirements for general radiography*

Amendment:

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment*

IEC 60601-1-3:2008/AMD1:2013

IEC 60601-1-3:2008/AMD2:2021

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, ~~applicable collateral standards~~ and IEC 60601-1:2005/AMD2:2020, IEC TR 60788:2004 and the following apply, ~~except as follows~~.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE An Index of defined terms is found in the last part of this document.

Addition:

201.3.201

APPARENT RESISTANCE OF SUPPLY MAINS

for diagnostic X-RAY GENERATOR, resistance of the SUPPLY MAINS determined under specific load conditions

201.3.202

AUTOMATIC INTENSITY CONTROL

in an X-RAY GENERATOR, mode of operation in which one or more LOADING FACTORS are controlled automatically in order to obtain at a pre-selected location a desired rate of a RADIATION QUANTITY

201.3.203

DIRECT RADIOGRAPHY

RADIOGRAPHY in which the permanent recording is effected at an IMAGE RECEPTION AREA

Example: Film-screen or film RADIOGRAPHY.

201.3.204

DIRECT RADIOSCOPY

RADIOSCOPY in which the visible images are presented at the IMAGE RECEPTION AREA, or close to it, in the RADIATION BEAM

201.3.205

DOSE AREA PRODUCT

product of the area of the cross-section of an X-RAY BEAM and the averaged AIR KERMA over that cross-section. The unit is the gray square metre ($\text{Gy}\cdot\text{m}^2$)

Note 1 to entry: This definition is equivalent to AIR KERMA area product.

201.3.206

DOSIMETER

EQUIPMENT which uses ionization chambers or semiconductor detectors for the measurement of AIR KERMA or AIR KERMA RATE in the beam of an X-RAY EQUIPMENT used for diagnostic medical RADIOLOGICAL examinations

201.3.207

ENTRANCE FIELD SIZE

dimensions of the field in the entrance plane of an X-RAY IMAGE RECEPTOR that can be used for the transmission of an X-RAY PATTERN under specific conditions

201.3.208

EXAMINATION PROTOCOL

full set of any programmed technical factors, control functions and settings, including image processing settings, designed to optimize the image acquisition and DISPLAY

201.3.209

EXAMINATION PROTOCOL SELECTION CONTROL

control to select a PRE-PROGRAMMED EXAMINATION PROTOCOL

201.3.210

HIGH-VOLTAGE GENERATOR

in an X-RAY GENERATOR, combination of all components for control and production of the electrical energy to be supplied to an X-RAY TUBE, usually consisting of a high-voltage transformer assembly and a control assembly

201.3.211

IMAGE RECEPTION PLANE

plane containing the greatest dimensions of the IMAGE RECEPTION AREA

201.3.212

INDIRECT RADIOGRAPHY

RADIOGRAPHY in which the permanent recording is effected after TRANSFER of the information obtained at an IMAGE RECEPTION AREA

Examples: CR systems, digital detector systems, image intensifier systems.

201.3.213

INDIRECT RADIOSCOPY

RADIOSCOPY in which the images are presented at a location outside the RADIATION BEAM after TRANSFER of the information

201.3.214

INTERLOCK

means preventing the start or the continued operation of ME EQUIPMENT unless certain predetermined conditions prevail

201.3.215

ISOCENTRE

in RADIOLOGICAL equipment with several modes of movement of the REFERENCE AXIS around a common centre, centre of the smallest sphere through which the X-RAY BEAM AXIS passes

201.3.216

LAST IMAGE HOLD RADIOGRAM

LIH RADIOGRAM

single image obtained by sampling or temporal processing of one or more images from the end of a radiosopic IRRADIATION

Note 1 to entry: This note applies to the French language only.

201.3.217

NOMINAL ELECTRIC POWER

for a HIGH-VOLTAGE GENERATOR, highest constant electric power which can be delivered for a single X-RAY TUBE load in a specific LOADING TIME

201.3.218

NOMINAL SHORTEST IRRADIATION TIME

shortest LOADING TIME for which a required constancy of the controlled radiation quantity is maintained

Note 1 to entry: The IRRADIATION TIME is controlled by a HIGH-VOLTAGE GENERATOR with AUTOMATIC CONTROL SYSTEMS.

201.3.219

PRE-PROGRAMMED EXAMINATION PROTOCOL

single hardware or software setting, or both, which is associated with an EXAMINATION PROTOCOL

201.3.220

QUALITY CONTROL

operational techniques and activities that are used to fulfil requirements for quality

201.3.221

RADIATION OUTPUT

AIR KERMA per CURRENT TIME PRODUCT (mGy/mAs) at a given distance from the FOCAL SPOT in the primary X-RAY BEAM