



Designation: D1217 – 20

Standard Test Method for Density and Relative Density (Specific Gravity) of Liquids by Bingham Pycnometer¹

This standard is issued under the fixed designation D1217; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope*

1.1 This test method covers the measurement of the density of pure hydrocarbons or petroleum distillates boiling between 90 °C and 110 °C that can be handled in a normal fashion as a liquid at the specified test temperatures of 20 °C and 25 °C.

1.2 This test method provides a calculation procedure for the conversion of density to relative density (specific gravity).

1.3 **WARNING**—Mercury has been designated by many regulatory agencies as a hazardous substance that can cause serious medical issues. Mercury, or its vapor, has been demonstrated to be hazardous to health and corrosive to materials. Use Caution when handling mercury and mercury-containing products. See the applicable product Safety Data Sheet (SDS) for additional information. The potential exists that selling mercury or mercury-containing products, or both, is prohibited by local or national law. Users must determine legality of sales in their location.

1.4 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* Specific warning statements are given in Section 7.

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.04.0D on Physical and Chemical Methods.

Current edition approved May 1, 2020. Published May 2020. Originally approved in 1952. Last previous edition approved in 2015 as D1217 – 15. DOI: 10.1520/D1217-20.

2. Referenced Documents

2.1 *ASTM Standards:*²

D4052 Test Method for Density, Relative Density, and API Gravity of Liquids by Digital Density Meter

E1 Specification for ASTM Liquid-in-Glass Thermometers

3. Terminology

3.1 *Definitions:*

3.1.1 *density, n* —mass per unit volume at a specified temperature. **D4052**

3.1.2 *relative density, n* —the ratio of the density of a material at a stated temperature to the density of water at a stated temperature. **D4052**

4. Summary of Test Method

4.1 The liquid sample is introduced into a pycnometer, equilibrated to the desired temperature, and weighed. The relative density (specific gravity) or density is then calculated from this weight and the previously determined weight of water that is required to fill the pycnometer at the same temperature, both weights being corrected for the buoyancy of air.

5. Significance and Use

5.1 Density is a fundamental physical property which can be used in conjunction with other properties to characterize pure hydrocarbons and their mixtures.

5.2 This test method was originally developed for the determination of the density of the ASTM Knock Test Reference Fuels *n*-heptane and *isooctane*, with an accuracy of 0.00003 g/mL. Although it is no longer employed extensively for this purpose, this test method is useful whenever accurate densities of pure hydrocarbons or petroleum fractions with boiling points between 90 °C and 110 °C are required.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard




FIG. 1 Bingham-Type Pycnometer, 25 mL

6. Apparatus

6.1 *Pycnometer*, Bingham-type, conforming to the dimensions given in Fig. 1, constructed of borosilicate glass, and having a total weight not exceeding 30 g.

6.2 *Constant-Temperature Bath*, provided with suitable pycnometer holders or clips and means for maintaining temperatures constant to ± 0.01 °C in the desired range.

6.3 *Bath Thermometer*, graduated in 0.1 °C subdivisions and standardized for the ice point and the range of use to the nearest 0.01 °C. ASTM Saybolt Viscosity Thermometer 17C as prescribed in Specification E1, designed for tests at 21.1 °C and 25 °C, is recommended. A standardized platinum resistance thermometer may also be used, and offers the best means for observing minute temperature changes in the bath. Whichever means are available, it must be realized that for most hydrocarbons the density coefficient is about 0.0008 units/°C, and therefore an error of ± 0.013 °C would cause an error of ± 0.00001 in density.

6.4 *Hypodermic Syringe*, 30 mL capacity, of chemically resistant glass, equipped with a 152 mm needle made of stainless steel tubing as shown in Fig. 2.

6.5 *Draw-Off Needle*, made of stainless steel tubing as shown in Fig. 2.

6.6 *Solvent-Cleaning Assembly*, as shown in Fig. 3.

6.7 *Chromic Acid Cleaning Apparatus*, similar to that shown in Fig. 4.

6.8 *Balance*, capable of reproducing weighings within 0.1 mg. Mechanical balances should have sensitivity which causes the pointer to be deflected 2 or 3 scale divisions per 1 mg when carrying a load of 30 g or less on each pan. The balance should be located in a room shielded from drafts and fumes and in which the temperature changes between related weighings (empty and filled pycnometer) do not cause a significant change in the ratio of the balance arms. Otherwise weighings shall be made by the method of substitution, in which the calibrated weights and pycnometer are alternately weighed on the same balance pan. The same balance shall be used for all related weighings.

6.9 *Weights*, whose relative values are known to the nearest 0.05 mg or better. The same set of weights shall be used for the calibration of the pycnometer and the determination of densities.

7. Reagents and Materials

7.1 *Acetone*—(Warning—Extremely flammable. Use adequate ventilation.)

7.2 *Isopentane*—(Warning—Extremely flammable. Avoid buildup of vapors and remove all sources of ignition, especially non-explosion proof electrical apparatus.)

7.3 *Chromic Acid (Potassium Dichromate/Conc. Sulfuric Acid)*—(Warning—Causes severe burns. A recognized carcinogen. Do not get in eyes, or on skin or clothing.)

8. Preparation of Apparatus

8.1 Thoroughly clean the pycnometer with hot chromic acid cleaning solution by means of the assembly shown in Fig. 4. Chromic acid solution (Warning—See 7.3) is the most effective cleaning agent. However, surfactant cleaning fluids have also been used successfully. Mount the apparatus firmly and connect the trap to the vacuum. Warm the necessary amount of cleaning acid in the beaker, place the pycnometer on the ground joint, and evacuate by opening the stopcock to vacuum. Fill the pycnometer with acid by turning the stopcock, repeat several times or remove the filled pycnometer, and allow it to stand for several hours at 50 °C to 60 °C. Remove the acid from the pycnometer by evacuation, empty the acid from the trap, and flush the pycnometer with water. Cleaning should be made in this manner whenever the pycnometer is to be calibrated or whenever liquid fails to drain cleanly from the walls of the pycnometer or its capillary. Ordinarily, the pycnometer may be cleaned between determinations by washing with a suitable solvent, rinsing with pure, dry acetone, followed by isopentane, and vacuum drying.

8.2 Transfer the pycnometer to the cleaner assembly shown in Fig. 3, with vacuum line and trap attached to the side tube as indicated. Place the pycnometer on the cleaner with the upper hypodermic needle extending upward into the pycnometer, and press the edge of the ground joint on the rubber stopper until the vacuum holds it in place. Draw out all the liquid or sample. Immerse the lower end of the hypodermic tube in a suitable solvent and draw 20 mL to 25 mL through the




FIG. 2 Accessories for Bingham-Type Pycnometer





FIG. 3 Cleaner Assembly for Bingham-Type Pycnometer

pycnometer. Leaving the pycnometer in place, draw air through it until it is dry. Clean the hypodermic syringe with the same apparatus.

9. Calibration of Pycnometer

9.1 Proceeding as directed in Section 10, determine the weight of freshly-boiled and cooled distilled water (distilled from alkaline permanganate through a tin condenser) held by the pycnometer when equilibrated to volume at the bath

temperature to be used in the determination. Repeat until at least three values agree to ± 0.2 mg.

10. Procedure

10.1 Using another 25 mL pycnometer as a tare (Note 1), weigh the clean, dry pycnometer to 0.1 mg and record the weight.

NOTE 1—It is convenient to use the lightest of a set of pycnometers as a tare. For best results the treatment and environment of both pycnometer




FIG. 4 All-Glass Pycnometer Cleaner Assembly for Use with Hot Chromic Acid Cleaning Solution

and tare should be identical for some time prior to weighing.

10.2 Cool the sample to 5 °C to 10 °C below the test temperature, and fill the clean 30 mL hypodermic syringe. Transfer the sample to the pycnometer through the filling needle; avoid trapping air bubbles (**Warning**—Extremely flammable. Avoid buildup of vapors and remove all sources of ignition, especially non-explosion proof electrical apparatus) in the bulb or capillary of the pycnometer. If any are present, draw them into the syringe where possible. Also remove with the syringe or draw-off needle any liquid above the calibration mark in the capillary or overflow reservoir. Dry the remainder with a cotton fiber pipe cleaner or cotton swab which has been dampened slightly with acetone.

NOTE 2—For work of highest accuracy on pure compounds, dissolved air may be removed from the sample by repeated freezing and remelting of the sample under vacuum in the pycnometer.

10.3 Close the pycnometer with the glass stopper and immerse it to a point above the calibration mark in the constant-temperature bath adjusted to a constancy of ±0.01 °C at the desired temperature. Periodically, or before the liquid expands into the overflow chamber, remove the stopper, raise the pycnometer sufficiently to expose the calibration mark to view, and readjust the liquid level to the mark by withdrawing liquid through the steel draw-off needle until expansion has stopped, indicating that the liquid has reached the temperature of the thermostat. Do not allow the liquid to expand more than 10 mm above the calibration mark at any time, to minimize errors caused by faulty drainage. Allow the contents to equilibrate an additional 10 min and draw the level down exactly to the calibration line, avoiding parallax and using a magnifier, if necessary, to obtain good visibility. Remove any liquid adhering to the walls above the calibration mark, with the draw-off needle or pipe cleaner, depending upon the volatility of the sample. Portions in the overflow bulb may be removed with a cotton swab moistened with acetone.

10.4 Replace the glass stopper, remove the pycnometer from the bath, wash the outside surface with acetone, and dry thoroughly with a chemically clean, lint-free, slightly damp cloth. Place the pycnometer in or near the balance case for 20 min and weigh to the nearest 0.1 mg. In atmospheres of low humidity (60 % or lower), drying the pycnometer by rubbing

with a dry cotton cloth will induce static charges equivalent to a loss of about 1 mg in the weight of the pycnometer. This charge need not be completely dissipated in less than 30 min. The use of about 0.1 mg radium bromide- or polonium-coated foil in the balance case, or maintaining the relative humidity at 60 % or higher, aids in reducing weighing difficulties due to static charges.

10.5 Record temperature of the balance, barometric pressure, and relative humidity.

11. Calculation

11.1 Calculate the true density of the sample as follows:

$$\text{Density, g/mL at } ^\circ\text{C} = \frac{W_s(1 + (d_a/d_s) - (d_a/d_{wt}))d_w}{W_w(1 + (d_a/d_w) - (d_a/d_w))} \quad (1)$$

where:

- W_s = weight in air of sample contained in the pycnometer at the test temperature, g,
- W_w = weight in air of the water contained in the pycnometer at the calibration temperature, g,
- d_w = density of water at the calibration temperature, as obtained from Table 1,
- d_a = density of air in balance case at the time of weighing, as calculated from 10.3,
- d_{wt} = density of weights used in weighing the sample and water (brass = 10.4 g/mL, stainless steel = 7.75 g/mL), and
- d_s = approximate density of sample or $(W_s \times d)/W_w$ (2)

11.2 The equation assumes that the weighings of the pycnometer empty and filled are made in such a short time interval that the air density has not changed. If significant change should occur, the calculated apparent weight of the sample, W_s , in this equation, must be corrected for the difference in air buoyancy exerted on the pycnometer as follows:

$$W_s = W_{PS}^2 - W_p(1 + (d'_a/2.2) - (d'_a/d_{wt}))/((1 + (d_a^2/2.2) - (d_a^2/d_{wt}))) \quad (3)$$

where: