



Designation: D6561 – 20

Standard Test Method for Determination of Aerosol Monomeric and Oligomeric Hexamethylene Diisocyanate (HDI) in Air with (Methoxy-2-phenyl-1) Piperazine (MOPIP) in the Workplace¹

This standard is issued under the fixed designation D6561; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of aerosol hexamethylene diisocyanate (HDI) in air samples collected from workplace and ambient atmospheres. The method described in this test method collects separate fractions. One fraction will be dominated by vapor, and the other fraction will be dominated by aerosol. The results obtained from the analysis of the separate fractions do not necessarily represent the true partition of the measured HDI physical phases, and should only be considered a representation of the general trend in the physical phase partition within samples. The analyses of the two fractions are different, and are provided in separate, linked, standards to avoid confusion. This test method is principally used to determine short term exposure (15 min) of HDI in workplace environments for personal monitoring or in ambient air. The analysis of the vapor fraction is performed separately, as described in Test Method [D6562](#).

1.2 Differential air sampling is performed with a segregating device. The aerosol fraction is collected on a polytetrafluoroethylene (PTFE) filter.

1.3 The analysis of the aerosol fraction is performed by using a high performance liquid chromatograph (HPLC) equipped with an ultraviolet (UV) detector. An ultra high performance liquid chromatograph (UPLC) can also be used, provided that its performance is equivalent to what is stated in this standard. The range of application of the test method has been validated from 0.052 to 1.04 μg of monomeric HDI/mL, which corresponds, based on a 15 L air sample, to concentrations from 0.004 to 0.070 mg/m^3 of HDI. Those concentrations correspond to a range of aerosol phase concentrations from 0.5 ppb (V) to 10 ppb (V) and cover the established threshold limit value (TLV) value of 5 ppb (V).

1.4 The quantification limit for the monomeric HDI is 0.041 μg per mL, which corresponds to 0.003 mg/m^3 for a 15 L

¹ This test method is under the jurisdiction of ASTM Committee [D22](#) on Air Quality and is the direct responsibility of Subcommittee [D22.04](#) on Workplace Air Quality.

Current edition approved March 1, 2020. Published May 2020. Originally approved in 2000. Last previous edition approved in 2016 as D6561 – 06 (2016). DOI: 10.1520/D6561-20.

sampled air volume. This value is equivalent to ten times the standard deviation obtained from ten measurements carried out on a standard solution in contact with the PTFE filter whose concentration of 0.1 $\mu\text{g}/\text{mL}$ is close to the expected detection limit.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* See Section 9 for additional hazards.

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

- [D1193 Specification for Reagent Water](#)
- [D1356 Terminology Relating to Sampling and Analysis of Atmospheres](#)
- [D1357 Practice for Planning the Sampling of the Ambient Atmosphere](#)
- [D4840 Guide for Sample Chain-of-Custody Procedures](#)
- [D5337 Practice for Flow Rate Adjustment of Personal Sampling Pumps](#)
- [D6562 Test Method for Determination of Gaseous Hexamethylene Diisocyanate \(HDI\) in Air with 9-\(N-methylaminomethyl\) Anthracene Method \(MAMA\) in the Workplace](#)

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.




FIG. 1 MOPIP Solution

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this test method, refer to Terminology **D1356**.

4. Summary of Test Method

4.1 Vapor and aerosol fractions are sampled simultaneously by using a segregating sampling device. The aerosols are collected on PTFE filter while the gaseous fraction is being adsorbed on a second filter made of glass fiber, impregnated with a 9-(N-methylaminomethyl) anthracene (MAMA).

4.2 The analysis of the monomer in the gaseous fraction is performed separately in accordance with the procedure described in Test Method **D6562**.

4.3 Immediately after sampling, the PTFE filter is transferred into a jar containing a (methoxy-2 phenyl-1) piperazine (MOPIP) solution in toluene (**1, 2**).³ See **Fig. 1**.

4.3.1 The solution is then evaporated to dryness and redissolved, using an acetic anhydride solution (see **8.11**). Monomeric and oligomeric HDI are separated by using a reversed phase HPLC column, and detection is made by using an HPLC equipped with UV detection. An ultra high performance liquid chromatograph (UPLC) can also be used, provided that its performance is equivalent to what is stated in this standard.

4.4 Concentration of monomeric and oligomeric diisocyanates contained in a sample is calculated by using an external standard of the monomeric HDI.

5. Significance and Use

5.1 HDI is mostly used in the preparation of paints. The use of isocyanates and their industrial needs have been in constant growth.

5.2 Diisocyanates and polyisocyanates are irritants to skin, eyes, and mucous membranes. They are recognized to cause respiratory allergic sensitization, asthmatic bronchitis, and acute respiratory intoxication (**3-6**).

5.3 The American Conference of Governmental Industrial Hygienists (ACGIH) has adopted a threshold limit value–time weighted average (TLV-TWA) of 0.005 ppm (V) or 0.034 mg/m³ for monomeric HDI (**7**). The Occupational Safety & Health Administration of the U.S. Department of Labor (OSHA) has not listed a permissible exposure limit (PEL) for HDI (**8**).

5.4 In any case, there are not separate exposure standards for vapor and aerosol. Therefore, in comparing the results for

isocyanate against a standard, results from the two fractions should be combined to give a single total value.

5.5 Due to its low LOD and low required volume (15 L), this test method is well suited for monitoring of respiratory and other problems related to diisocyanates and polyisocyanates. Its short sampling times are compatible with the duration of many industrial processes, and its low detection limit with the concentrations often found in the working area.

6. Interferences

6.1 Any substance, including strong oxidizing agents, that can be deposited on the PTFE filter and react with MOPIP reagent can affect the analysis efficiency.

6.2 Any compound that has the same retention time as the HDI-MOPIP derivative and contributes to UV response is an interference. Chromatographic conditions can sometimes be changed to eliminate an interference.

7. Apparatus

7.1 Sampling Equipment:

7.1.1 *Personal Sampling Pump*, equipped with a flow-monitoring device (rotameter, critical orifice) or a constant-flow device capable of drawing 1.0 L/min of air through the sampling device for a period of at least 4 h.

7.1.2 *Double Filter Sampling Device*, 37 mm in diameter, three-piece personal monitor, plastic holder loaded with a PTFE filter close to the mouth, followed by a glass fiber filter (GFF) impregnated with MAMA and a plastic back-up pad. The GFF is impregnated with an amount of MAMA in the range from 0.07 to 0.25 mg.

7.1.3 *Flow Measuring Device*, used in accordance with Practice **D5337**.

7.2 Analytical Equipment:

7.2.1 *Liquid Chromatograph*, HPLC, equipped with a UV detector (242 nm wavelength), connected in series with a diode detector, and equipped with an automatic or manual sampling port injection. An ultra high performance liquid chromatograph (UPLC) can also be used provided that the other equipment mentioned below are adapted and offer at least the same performance stated in this standard.

7.2.2 *Liquid Chromatographic Column*, an HPLC stainless steel column, capable of separating the urea derivatives. This test method recommends a 150 by 3.2 mm internal diameter stainless steel column packed with 3 μm C-18, or an equivalent column.

7.2.3 *Electronic Integrator*, or any other effective method for determining peak areas.

7.2.4 *Analytical Balance*, with a precision of ±0.0001 g.

7.2.5 *Microsyringes and Pipets*—Microsyringes are used in the preparation of urea derivatives and standards. An automatic pipet, or any equivalent equipment, is required for sample preparation.

7.2.6 *pH Meter*, or any equivalent device capable of assaying a pH range between 2.5 and 7.

7.2.7 *Culture Tubes*, 16 mm × 100 mm, disposable, in borosilicate glass for evaporation of derivatized samples.

7.2.8 *Glass Jars*, 30 mL, and lids, capable of receiving 37 mm filters, used for derivatization of samples.

³ The boldface numbers in parentheses refer to the list of references at the end of this standard.

7.2.9 *Vacuum Filtration System*, filter 47 mm, with 0.22 μm pore size polyamide filters, or any equivalent method.

7.2.10 *Syringe Operated Filter Unit*, syringes with 4 mm, polyvinylidene fluoride 0.22- μm pore size filter unit, or any equivalent device.

7.2.11 *Injection Vials*, 1.5 mL vials with PTFE-coated septums.

7.2.12 *Bottle*—Amber colored bottle with cap and PTFE-coated septum for conservation of stock and diluted standard solutions of HDI.

7.2.13 *Vacuum Evaporator*, capable of heating to 55°C, or any equivalent device.

7.2.14 *Vortex Movement Mixer*, or any equivalent device.

8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. All reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—Unless otherwise indicated, water shall be reagent water as defined by Type 2 of Specification D1193, HPLC grade.

8.3 *Acetic Acid*, glacial (CH_3COOH), HPLC grade.

8.4 *Acetic Anhydride* ($(\text{CH}_3\text{CO})_2\text{O}$), certified by American Chemical Society (ACS).

8.5 *Acetonitrile*, HPLC grade.

8.6 *Buffer*—In a 1-L volumetric flask, dissolve 1 g sodium acetate ($\text{NaC}_2\text{H}_3\text{O}_2$) (see 8.12) in water and dilute to volume. Add glacial acetic acid (CH_3COOH) (see 8.11) to acidify to pH = 6.0. Under vacuum, filter the buffer with a 0.22- μm pore size filter.

8.7 *Derivatization Solution*—Weigh 50 mg of MOPIP (see 8.9), and dilute to 500 mL in a volumetric flask with toluene (see 8.13). This solution is equivalent to 0.1 mg MOPIP/mL.

8.8 *Hexamethylene Diisocyanate (HDI)*, (F.W. 168), 98 % purity.

8.9 (*Methoxy-2-phenyl-1) Piperazine (MOPIP)*, (F.W. 192.2). 98 % purity.

8.10 *Mobile Phase*, a solvent mixture of 60 % (v/v) acetonitrile (see 8.5) and 40 % (v/v) buffer (see 8.6).

8.11 *Redissolution Solution*—Dilute 500 μL of acetic anhydride ($(\text{CH}_3\text{CO})_2\text{O}$) (see 8.4) to 100 mL with acetonitrile (see 8.5).

8.12 *Sodium Acetate* ($\text{NaC}_2\text{H}_3\text{O}_2$).

8.13 *Toluene*, HPLC grade.

⁴ ACS Reagent Chemicals, Specifications and Procedures for Reagents and Standard-Grade Reference Materials, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

9. Hazards

9.1 **Warning**—Diisocyanates are potentially hazardous chemicals and are extremely reactive. Refer to material safety data sheets for reagents.

9.2 **Warning**—Avoid exposure to diisocyanate and solvents. Sample and standard preparations should be done in an efficient operating hood. For remedial statement, see Ref (9).

9.3 **Warning**—Wear safety glasses at all times and other laboratory protective equipment if necessary.

9.4 **Warning**—Avoid skin contact with isocyanates and all solvents.

10. Sampling

10.1 Refer to Practice D1357 for general information on sampling.

10.2 This test method recommends sampling in accordance with the method described in Refs (9-11).

10.3 Adjust the personal sampling pumps (see 7.1.1) to the recommended flow rate with an assembled cassette (see 7.1.2) between the pump and the flow-measuring device in accordance to Practice D5337. After the sampling, perform a post sampling flow rate verification. If the post sampling flow rate varies more than $\pm 5\%$ from the recommended flow rate, invalidate the sample.

10.4 Equip the worker, whose exposure is to be evaluated, with a filter holder connected to a belt-supported sampling pump. Place the filter holder pointing downward, if possible, at an optimum angle of 45° from horizontal in the breathing zone of the worker. Draw air through the sampling device, and collect 15 L at a rate of approximately 1.0 L/min.

10.5 For stationary monitoring, use a tripod or any other support to locate the sampler in a general room area at a height equivalent to the breathing zone.

10.6 A field blank is used to monitor contamination during the combined sampling, transportation, and storage process. Open the field blanks in the environment to be sampled and immediately close them. Process field blanks in the same manner as samples. Submit at least one field blank for every ten samples.

10.7 Immediately after sampling, open the cassette, withdraw the PTFE filter, place it in a glass jar containing 5 mL of MOPIP derivatization solution (see 8.7), and close the jar. This filter is used to analyze the aerosol fraction of diisocyanates.

10.8 Close the cassette leaving the GFF and the plastic pad support. The GFF is used to analyze the gaseous fraction of diisocyanates (see Test Method D6562).

10.9 Send the jars and the cassettes to be analyzed to the laboratory. Keep away from light.

10.10 The samples are stable for 14 days at room temperature and 42 days at 4°C.

10.11 Follow sampling chain of custody procedures in accordance with Guide D4840 to ensure sample traceability. Ensure that the documentation that accompanies the samples is suitable for a “chain of custody” to be established.