



Designation: E2409 – 20a

Standard Test Method for Glycol Impurities in Mono-, Di-, Tri- and Tetraethylene Glycol and in Mono- and Dipropylene Glycol (Gas Chromatographic Method)¹

This standard is issued under the fixed designation E2409; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method describes the gas chromatographic determination of glycol impurities in Mono-, Di-, Tri-, and Tetraethylene Glycol (MEG, DEG, TEG, and TetraEG), and in Mono- and Dipropylene Glycol (MPG and DPG).

1.2 This test method is applicable to MEG, DEG, TEG, and TetraEG with impurities to 3000 mg/kg. The limit of detection (LOD) is 22 mg/kg and the limit of quantitation (LOQ) is 73 mg/kg.

NOTE 1—LOD and LOQ were calculated using the lowest level sample in the ILS.

1.3 This test method is applicable to MPG and DPG to 2.5 %.

1.4 The following applies for the purposes of determining the conformance of the test results using this test method to applicable specifications, results shall be rounded off in accordance with the rounding-off method of Practice E29.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Section 7.

1.7 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

¹ This test method is under the jurisdiction of ASTM Committee D16 on Aromatic, Industrial, Specialty and Related Chemicals and is the direct responsibility of Subcommittee D16.14 on Alcohols & Glycols.

Current edition approved Oct. 1, 2020. Published October 2020. Originally approved in 2004. Last previous edition approved in 2020 as E2409 – 20. DOI: 10.1520/E2409-20A.

2. Referenced Documents

2.1 *ASTM Standards:*²

- D1193 Specification for Reagent Water
- D6809 Guide for Quality Control and Quality Assurance Procedures for Aromatic Hydrocarbons and Related Materials
- E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications
- E180 Practice for Determining the Precision of ASTM Methods for Analysis and Testing of Industrial and Specialty Chemicals (Withdrawn 2009)³
- E300 Practice for Sampling Industrial Chemicals
- E1064 Test Method for Water in Organic Liquids by Coulometric Karl Fischer Titration

2.2 *Other Document:*

- Manufacturers' instruction manuals of gas chromatograph and digital integration system used

3. Summary of Test Method

3.1 A portion of the test sample is analyzed by temperature-programmed, capillary gas chromatography over a polyethylene glycol column, using flame ionization detection. For quantification, the External Standard Technique or the Internal Standard (Marker) Technique are applied. When applying the Internal Standard Technique, the standard addition technique is used to eliminate the effect of other impurities present in the glycols. For this purpose, a blank glycol is used, as 100 % pure glycol samples are not available.

4. Significance and Use

4.1 Knowledge of the impurities is required to establish whether the product meets the requirements of its specifications.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

*A Summary of Changes section appears at the end of this standard

5. Apparatus

5.1 Autoinjectors are required for all gas chromatograph standards using an external standard to calculate results.

5.2 *Gas Chromatograph(s)*, provided with a sample splitter or on-column injection, flame ionization detector and temperature-programming facilities. Optional are pressure programming and auto sampler facilities. The instrument must be suitable for analysis according to the operating instructions given in [Table 1](#) or [Table 2](#).

5.3 *Columns*—The analytical column (chemically bonded cross-linked polyethylene glycol) used must completely separate.

MEG, DEG, TEG, TetraEG, PentaEG (Penta-ethylene Glycol), and 1,4-butanediol, or

MPG, DPG, TPG, and TetraPG (Tetrapropylene Glycol).

[Fig. A1.1](#) through [Fig. A1.5](#) show examples of chromatograms conforming to the requirements.

5.4 *Chromatographic data systems are preferred but electronic integration may be used if the user can demonstrate that the results are consistent with the precision statement.*

5.5 *Analytical Balance*, readability 0.1 mg, calibrated. Recalibrate or verify at regular intervals.

5.6 *Crimp Top Vials*, 1 mL and 5 mL.

5.7 *Crimper/De-capper*, for capping and de-capping the vials.

5.8 *Micro Syringes*, 10 μL .

5.9 *Bottles*, 50 mL, with screw cap.

6. Reagents and Materials

6.1 *Purity of Reagents*—Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁴ Other

⁴ *ACS Reagent Chemicals, Specifications and Procedures for Reagents and Standard-Grade Reference Materials*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

TABLE 1 Recommended Operating Parameters for the GC Analysis of Glycol Impurities in MEG, DEG, TEG, and TetraG

| | |
|----------------------|--|
| Column ⁴ | |
| Type | Capillary wide-bore |
| Material | Fused silica |
| Length \times I.D. | 15 m \times 0.53 mm |
| Stationary Phase | Polyethylene glycol, for example, DB-Wax |
| Film Thickness | 1 μm |
| Detector System | |
| Type | FID |
| Sensitivity | The ratio of the signal to the noise level must be at least 2:1 at a concentration of 5 mg/kg DEG in MEG |
| Temperatures | |
| Column Oven | 0.05 min at 70°C Programmed from 70 to 230°C at 25°C/min 10 min at 230°C |
| Detector | 250°C |
| Carrier Gas | Helium, nitrogen, or hydrogen. Warning! Helium carrier gas was used to develop this standard. Use of nitrogen or hydrogen requires different conditions. The user must conduct the necessary evaluation to determine that equivalent results are obtained. |
| Calibration | see Section 10 |
| Injected Volume | 0.2 μL (on-column injection), or 0.5 μL up to 1 μL (using split injection technique) |
| Split Ratio | 1:10 or appropriate split ratio to allow adequate sensitivity as defined under Detector System (only if split injection technique is used) |

⁴ The choice of column is based on resolution requirements. Any column may be used that is capable of resolving all significant impurities from the major component. The column and conditions described in [Table 1](#) have been used successfully and shall be used as a referee in cases of dispute. However, the chromatogram obtained must be equivalent, with regard to separation of the glycol components and 1,4-butanediol, to those illustrated in [Fig. A1.1](#), [Fig. A1.2](#), and [Fig. A1.3](#), or [Fig. A1.4](#) and [Fig. A1.5](#).

TABLE 2 Typical Operating Parameters for the GC Analysis of Glycol Impurities in MPG or DPG

| | |
|---------------------|--|
| Column ^A | |
| Type | Capillary wide-bore |
| Material | Fused silica |
| Length × I.D. | 30 m × 0.32 mm |
| Stationary Phase | Poly ethylene glycol, for example, DB-Wax |
| Film Thickness | 0.5 µm |
| Detector System | |
| Type | FID |
| Sensitivity | The ratio of the signal to the noise level must be at least 2 to 1 at a concentration of 0.01 % (m/m) DPG in MPG |
| Temperatures | |
| Column Oven | 5 min at 150°C Programmed from 150 to 180°C at 5°C/min 0 min at 180°C Programmed from 180 to 240°C at 30°C/min 10 min at 240°C |
| Detector | 300°C |
| Carrier Gas | Helium |
| Calibration | see Section 10 |
| Injected Volume | 0.1 µL or 0.5 µL (using split injection technique) |
| Split Ratio | 1 to 10 or appropriate split ratio to allow adequate sensitivity as defined under Detector System |

grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

6.2 Calibration Standards:

6.2.1 *Mono-ethylene Glycol* (MEG), minimum purity 99.5 mass %.

6.2.2 *Di-ethylene Glycol* (DEG), minimum purity 99.5 mass %.

6.2.3 *Tri-ethylene Glycol* (TEG), minimum purity 99.5 mass %.

6.2.4 *Tetra-ethylene Glycol* (TetraEG), of maximum purity available.

6.2.5 *Penta-ethylene Glycol* (PentaEG), of maximum purity available, or

6.2.6 *Mono-propylene Glycol* (MPG), minimum purity 99.5 mass %.

6.2.7 *Di-propylene Glycol* (DPG), minimum purity 99.5 mass %.

6.2.8 *Tri-propylene Glycol* (TPG), of maximum purity available.

6.2.9 *Tetra-propylene Glycol* (TetraPG), of maximum purity available.

6.3 Internal Standard:

6.3.1 *1,4-Butanediol* minimum purity 97 mass %, for ethylene glycols, if necessary.

6.3.2 *n-Octane* minimum purity 97 mass %, for propylene glycols, if necessary.

6.4 *Ethylene Glycol Quality Control Sample*, fiber grade MEG, DEG, TEG, or TeEG or *Propylene Glycol Quality Control Sample*, MPG or DPG (only required if maintaining a

control chart). Store nitrogen capped at a temperature between 0 and 5°C. Warm to ambient temperature before use. See Section 15.

6.5 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to type I of Specification D1193.

6.6 Solutions:

6.6.1 *Internal Standard Solution*—Weigh about 0.15 g 1,4-butanediol (m_1) to the nearest 0.1 mg into a 50 mL bottle. Add ultra-pure water up to a total mass of 50 g (m_2), weighing to the nearest 0.1 mg. Calculate the concentration of this solution to the nearest 0.1 mg/kg; or

6.6.2 *External Standard Solution*, of accurately known MEG, DEG, TEG, TetraEG, and PentaEG content; or MPG, DPG, TPG, and TetraPG content (see 10.4).

7. Hazards

7.1 Consult current OSHA regulations, suppliers' Safety Data Sheets, and local regulations for all materials used in this test method.

8. Sampling, Test Specimens, and Test Units

8.1 Follow the relevant instructions for sampling as given in Practice E300.

9. Preparation of Apparatus

9.1 *Gas Chromatograph(s) and Column(s)*—Check the performance of the gas chromatograph and column as described in Section 10.

10. Calibration and Standardization

10.1 Two methods of quantification may be employed: the Internal Standard (Marker) Technique or the External Standard Technique.

10.2 Internal Standard Technique for Ethylene Glycols:

10.2.1 Prepare calibration solutions, containing approximately 500, 1000, and 2000 mg/kg of each of the glycol components to be determined, by adding the relevant calibration standard (see 6.2) to a blank sample of the glycol being analyzed. Calculate the exact concentration of each glycol component (c_1) in the calibration solutions.

10.2.2 Weigh 0.5 g of each calibration solution (m_3) to the nearest 0.1 mg, into separate 5-mL vials. Add, also weighed to the nearest 0.1 mg, 0.5 g internal standard solution (see 6.6.1; m_4) and add water up to a total mass of approximately 5 g. Cap the vials and mix thoroughly.

10.2.3 Prepare a blank calibration solution by weighing 0.5 g blank sample of the glycol being analyzed (m_5), weighed to the nearest 0.1 mg, into a 5-mL vial. Add 0.5 g internal standard solution (see 6.6.1; m_6), also weighed to the nearest 0.1 mg, and add HPLC grade water up to a total mass of approximately 5 g. Cap the vial and mix thoroughly.

10.2.4 Calibrate separately for each impurity in MEG, DEG, TEG, or TeEG by using the Internal Standard (Marker) Technique.

10.2.5 Fill a 1-mL sample vial with the calibration solution from the 5-mL vial (see 10.2.2 and 10.2.3). Close the vial by means of an aluminum crimp cap.

10.2.6 Analyze each calibration solution and the blank solution using the operating parameters given in Table 1. Inject each solution at least twice and calculate the average peak areas for each of the calibration solutions. Apply digital integration equipment for measuring the peak areas.

10.2.7 For each chromatogram, calculate the system response factor (f) of each of the components as described in 10.2.8 through 10.2.10.

10.2.8 Calculate the amount of internal standard (1,4-butanediol) added to the calibration solution:

$$\text{Mass of Internal Standard } (m_7), \text{ g} = \frac{m_4 \times m_1}{m_2} \quad (1)$$

where:

m_1 = mass of 1,4-butanediol in internal standard solution (6.6.1), g,

m_2 = total mass of internal standard solution (6.6.1), g, and

m_4 = mass of internal standard solution added, g.

10.2.9 Calculate the amount of internal standard (1,4-butanediol) added to the blank solution:

$$\text{Mass of Internal Standard } (m_8), \text{ g} = \frac{m_6 \times m_1}{m_2} \quad (2)$$

where:

m_6 = mass of internal standard solution added (10.2.3), g.

10.2.10 Calculate the response factor of each component of interest in the calibration solutions by means of the following equation:

$$f = \frac{c_1 \times 10^{-6}}{\left(\frac{m_7 \times A_1}{m_3 \times A_2}\right) - \left(\frac{m_8 \times A_3}{m_5 \times A_4}\right)} \quad (3)$$

where:

c_1 = added concentration of glycol compound in the calibration solution, (10.2.1), mg/kg,

A_1 = peak area of component in calibration solution, arbitrary units,

A_2 = peak area of internal standard in calibration solution, same arbitrary units,

A_3 = peak area of component in blank solution, same arbitrary units,

A_4 = peak area of internal standard in blank solution, same arbitrary units,

m_3 = mass of calibration solution (10.2.2), g,

m_5 = mass of blank solution (10.2.3), g,

m_7 = mass of internal standard in calibration solution, as obtained in 10.2.8, g, and

m_8 = mass of internal standard in blank solution, as obtained in 10.2.9, g.

10.2.11 Calculate the mean of the response factors. If the individual factors differ by more than 5 % from the mean response factor, repeat the measurement of the respective calibration solution.

10.3 Internal Standard Technique for Propylene Glycols—Calibrate by determining the calibration factor for each component of interest relative to the internal standard on the basis of peak area versus mass as follows:

10.3.1 Prepare a calibration solution by accurately weighing 0.5 g of each of the components of interest and of the internal standard, to the nearest 0.1 mg into a previously tarred, 50 mL bottle. Fill the bottle with a suitable solvent (for example, acetone/cyclohexane), close, and reweigh to the nearest 0.1 mg. Homogenize the calibration solution.

10.3.2 Analyze the calibration solution following the operating parameters given in Table 2. Introduce the calibration solution at least twice. Determine the area of the components of interest and the reference component.

10.3.3 Calculate the mean peak area of the components of interest for the calibration solution. If the two single peak areas differ by more than 3 % relative, repeat the analysis. If no satisfactory results can be obtained, stabilize the conditions and repeat 10.3.1 and 10.3.2.

10.3.4 Calculate the calibration factor (f_i) for all individual compounds, relative to the internal standard, by means of the following equation:

$$f_i = \frac{m_i \times A_m}{m_m \times A_i} \quad (4)$$

TABLE 3 External Standard Recommended Weights

| Standard # | Target Weight of Glycol Component, ± 0.0001 g | Target Weight of High Purity Blank Glycol, ± 0.1 g |
|------------|---|--|
| 200 mg/kg | 0.010 | 50 |
| 500 mg/kg | 0.025 | 50 |
| 1000 mg/kg | 0.050 | 50 |

where:

m_i = mass of component i in calibration solution (10.3.1), g.
 m_m = mass of internal standard in calibration solution (10.3.1), g.
 A_i = peak area of component i (10.3.3), arbitrary units.
 A_m = peak area of internal standard (10.3.3), same arbitrary units.

NOTE 2—An alternative for the empirical calibration factors as described in 10.2 and 10.3 is the use of theoretical factors, based on the molecular structure of the compounds of interest. Theoretical factors calculated are as follows: For MPG 3.045, for all DPG isomers 2.512, for all TPG isomers 2.244, all relative to octane. See Footnote 6.⁵

10.4 External Standard Technique Ethylene Glycols, similar for Propylene Glycols:

10.4.1 Prepare at least three calibration solutions, for example, containing 200, 500, and 1000 mg/kg of each of the glycol components to be determined, by adding the relevant calibration standard (see 6.2) to a blank sample of the glycol being analyzed and mix thoroughly. Weigh each glycol component to the nearest 0.1 mg and the blank glycol to the nearest 0.1 g. (See Table 3 for recommended weights.)

10.4.1.1 Calculate the exact concentration of each glycol component (C_i) in the calibration solutions. The calibration range can be adjusted if needed.

$$C_i = \frac{W_{(Comp;i)}}{W_{(Comp;i)} + W_{(Blank;i)}} \times \frac{10^6 \mu\text{g}}{\text{g}} \quad (5)$$

where:

C_i = the concentration of each glycol component in the calibration standard of interest,
 i = glycol component standard of interest,
 $W_{(Comp;i)}$ = mass (g) of glycol component added to the calibration standard of interest, and
 $W_{(Blank;i)}$ = mass (g) of blank glycol added to the calibration standard of interest.

10.4.2 Analyze each calibration solution and the blank solution using the operating parameters given in Table 1 or Table 2. Inject each solution at least twice.

10.4.2.1 Prepare a plot of area counts of the glycol component (y -axis) versus the concentration of the glycol component (mg/kg) added to the standard of interest (x -axis). Using a computer program, determine the best-fit line through the data using linear regression analysis. The relationship between concentration and peak area will be linear. Record the intercept value (concentration; mg/kg) where the resulting line crosses the x -axis ($y = 0$). Apply digital integration equipment for measuring the peak areas.

10.4.2.2 Calculate the corrected concentration (mg/kg) of the glycol component in each calibration standard as follows:

$$\text{Corrected } C_i = C_i + Y \quad (6)$$

where:

Corrected C_i = the corrected concentration (mg/kg) of the glycol component in each calibration standard of interest,

C_i = the concentration (mg/kg) of the glycol component added to the calibration standard of interest, and
 Y = absolute value of the concentration of blank glycol determined from the linear regression graph (intercept value) for each calibration standard.

10.4.3 For each chromatogram, calculate the system response factor (f) of each of the glycol components by means of the following equation:

$$f = \frac{\text{Corrected } C_i}{A_i} \quad (7)$$

where:

Corrected C_i = concentration of component in external standard solution, mg/kg, and
 A_i = peak area of component, arbitrary units.

10.4.3.1 Calculate the mean of the response factors for each of the glycol components. If the individual factors differ by more than 5% from the mean response factor, repeat the measurement of the respective calibration solution.

NOTE 3—Many gas chromatographs have the ability to calculate a calibration graph automatically after measuring the calibration solutions and subsequently to show the concentration of the component being measured directly on a display. In such cases, no calibration graphs need to be constructed. It is, however, recommended to verify the calibration procedure of the instrument and to establish the characteristics of the calibration graph according to suitable regression analysis software.

11. Procedure

11.1 Internal Standard Technique for Ethylene Glycols:

11.1.1 Weigh a test portion of 0.5 g (m_0), weighed to the nearest 0.1 mg, into a 5-mL vial.

NOTE 4—This method is for the determination of glycol impurities in the range of 5 to 3000 mg/kg. Higher levels of glycol impurities (<5000 mg/kg) can be determined, if the intake is adjusted as follows:

$$\text{Mass Intake Sample, g} = \frac{2000}{c} \times 0.5 \quad (8)$$

where:

c = the expected maximum concentration of component in the sample, mg/kg.

11.1.2 Add 0.5 g internal standard solution (m_{10}), weighed to the nearest 0.1 mg, and add HPLC grade water up to a total mass of approximately 5 g and weigh. Close the vial and mix thoroughly.

11.1.3 Fill a 1-mL sample vial with the test solution. Close the vial by means of an aluminum crimp-cap.

11.1.4 Analyze the test solution using the operating parameters given in Table 1. Examples of the chromatograms are shown in Fig. A1.1, Fig. A1.2, and Fig. A1.3. Apply digital integration equipment for measuring the area of the peaks.

11.2 Internal Standard Technique for Propylene Glycols:

11.2.1 Prepare a test solution by weighing 100 μL of internal standard to the nearest 0.1 mg into a previously tarred, 50 mL bottle. Fill the bottle with test sample, close and reweigh to the nearest 0.1 mg. Homogenize the test solution. Calculate the concentration of the internal standard in the test solution.

⁵ Sternberg, J.C. *Gas Chromatography*, Academic Press, New York, 1962; pp. 231-267.

11.2.2 Analyze the test solution following the operating parameters given in **Table 2**. Apply the chromatography data system for measuring the areas of the peaks of interest. An example of a chromatogram is given in **Fig. A1.4**.

11.3 *External Standard Technique*—When applying the external standard technique, analyze the test sample using the operating parameters given in **Table 1** or **Table 2**. Sample dilution, with the sample glycol, may be necessary if the component of interest is beyond the range of the method. Apply digital integration equipment for measuring the area of the peaks.

11.4 *Determination of Water Content*—If it is required to calculate and report the purity of the sample, determine the water content in % mass (m/m) according to Test Method **E1064**.

12. Calculation

12.1 *Internal Standard Technique (Ethylene Glycol)*:

12.1.1 Calculate the amount (m_{11}) of internal standard (1,4-butanediol) added to the test sample by means of the following equation:

$$\text{Mass of Internal Standard } (m_{11}), \text{ g} = \frac{m_{10} \times m_1}{m_2} \quad (9)$$

where:

- m_{10} = mass of internal standard solution added, (11.1.2), g,
- m_1 = mass of 1,4-butanediol in internal standard solution (6.6.1), g, and
- m_2 = total mass of internal standard solution (6.6.1), g.

12.1.2 Calculate the concentration of each component of interest in the sample by means of the following equation:

$$\text{Glycol Component, mg/kg} = \frac{f \times m_{11} \times A_5}{m_9 \times A_6} \times 10^6 \quad (10)$$

where:

- A_5 = peak area of the component, arbitrary units,
- A_6 = peak area of internal standard, same arbitrary units,
- f = relative response factor of component of interest, as obtained in 10.2.11,
- m_9 = mass of sample, that is, without internal standard, (11.1.1), g, and
- m_{11} = mass of internal standard in test solution, as obtained in 12.1.1, g.

NOTE 5—If the concentration of the calculated glycol component is

required to be expressed in % mass (m/m), divide the result obtained above by a factor of 10^4 .

12.2 *Internal Standard Technique (Propylene Glycol)*:

12.2.1 Calculate the concentration of each component of interest by means of the following equation:

$$\text{Component, mg/kg or \% (m/m)} = f_i \times \frac{A_i}{A_{is}} \times C_{is} \quad (11)$$

where:

- f_i = calibration factor for relevant component as obtained in 10.3.4, or theoretical factor (see **Note 2**).
- A_i = peak area of relevant component peak in test solution (11.2.2), arbitrary units.
- A_{is} = peak area of internal standard in test solution (11.2.2), same arbitrary units.
- C_{is} = concentration of internal standard in test solution (11.2.1), mg/kg or mass %, whichever is relevant.

12.3 *External Standard Technique*:

12.3.1 Obtain the concentrations of the glycol impurities in the test sample in mg/kg, as presented by the software of the applied gas chromatograph (see **Note 3**). If an automated system is not being applied, read the concentration of the glycol impurities, in mg/kg, from the respective calibration graph, or

12.3.2 Calculate the concentration of the glycol components, in mg/kg, in the test sample by means of the following equation (see also **Note 4**):

$$\text{Glycol Component, mg/kg} = A_i \times f \quad (12)$$

where:

- A_i = peak area of relevant glycol component, arbitrary units, and
- f = the average response factor of component of interest, as obtained in 10.4.3.

12.4 *Purity*:

12.4.1 Calculate the purity of the sample by means of the following equation:

$$\text{Glycol of Interest Purity, mass \%} = 100 - O - W \quad (13)$$

where:

- O = other glycols, sum content as calculated in 12.1, 12.2, or 12.3, mass % of each of the minor glycol components, and

TABLE 4 E2409 Glycol Impurities by GC

| Test Result mg/kg | Sample | Average over all Laboratories | Repeatability Standard Deviation | Intermediate Standard Deviation | Reproducibility Standard Deviation | Repeatability Limit | Intermediate Limit | Reproducibility Limit |
|-------------------|--------|-------------------------------|----------------------------------|---------------------------------|------------------------------------|---------------------|--------------------|-----------------------|
| DEG | MEG | 374.59 | 7.3 | 7.3 | 34.0 | 20.6 | 20.6 | 95.3 |
| MEG | DEG | 1479.73 | 46.3 | 76.0 | 215.1 | 129.7 | 212.9 | 602.4 |
| TEG | DEG | 3499.69 | 92.8 | 143.2 | 306.5 | 260.0 | 401.0 | 858.3 |
| DEG | TEG | 489.32 | 56.8 | 70.9 | 201.7 | 159.1 | 198.5 | 564.9 |
| TTEG | TEG | 1020.00 | 96.3 | 96.3 | 244.1 | 269.8 | 269.8 | 683.5 |
| DEG | TeEG | 1646.25 | 55.4 | 55.4 | 95.4 | 155.1 | 155.1 | 267.1 |
| TEG | TeEG | 7908.35 | 221.9 | 221.9 | 1350.7 | 621.2 | 621.2 | 3782.0 |
| PentaEG | TeEG | 2084.93 | 58.7 | 72.9 | 156.3 | 164.5 | 204.1 | 437.5 |

W = water content of the sample, determined by Test Method **E1064** (11.4), mass %.

13. Report

13.1 Report the concentrations of DEG in MEG and MEG in DEG to the nearest mg/kg and all other impurities to the nearest 10 mg/kg. Results <22 mg/kg should be reported as <22 mg/kg.

13.2 Report the concentrations of DPG and TPG in MPG, and MPG, TPG, and TetraPG in DPG to the nearest 0.01 mass %.

13.3 Report the purity of the sample to the nearest 0.01 mass %.

14. Precision and Bias

14.1 An ILS was conducted which included 17 labs:

14.1.1 14 laboratories analyzed four samples of DEG for MEG in duplicate on two different days.

14.1.2 16 laboratories analyzed four samples of MEG for DEG in duplicate on two different days.

14.1.3 9 laboratories analyzed four samples of TEG for DEG in duplicate on two different days.

14.1.4 5 laboratories analyzed four samples of TeEG for DEG in duplicate on two different days.

14.1.5 13 laboratories analyzed four samples of DEG for TEG in duplicate on two different days.

14.1.6 5 laboratories analyzed four samples of TEG for DEG in duplicate on two different days.

14.1.7 10 laboratories analyzed four samples of TEG for TTEG in duplicate on two different days.

14.1.8 4 laboratories analyzed four samples of TeEG for PentaEG in duplicate on two different days.

14.1.9 Practice **E180** was followed for the design and analysis of the data: the details of this study are given in ASTM Research Report No. RR:E15-1063.⁶

14.2 Helium carrier gas was used to develop this standard. In cases of dispute, the carrier gas used to develop this standard must be used.

14.3 *Repeatability*—Two test results obtained within one laboratory shall be judged not equivalent if they differ by more than the r value for that material; r is the interval representing the critical difference between two test results for the same material, obtained by the same operator using the same equipment on the same day in the same laboratory.

14.4 *Reproducibility*—Two test results shall be judged not equivalent if they differ by more than the R value for that material; R is the interval representing the difference between two test results for the same material, obtained by different operators using different equipment in different laboratories.

14.5 *Intermediate Precision*—The day-to-day standard deviation within a laboratory for results produced by the same operator, determined through statistical analysis following Practice **E180**. Practice **E180** was used to conform to this particular study design which required an estimate of intermediate precision. The statistical analysis was conducted using the SAS statistical analysis software, Version 8.0.

14.5.1 The Practice **E180** analysis considers the two test results from each day as being run under repeatability conditions and estimates the repeatability, intermediate, and reproducibility precision for each assay. The repeatability precision would be estimated from the two sets of duplicate test results within each day, and the intermediate precision would be estimated from the agreement between the two days, all pooled over laboratories. Caveat: Since two days is a short time period, the intermediate precision would probably be underestimated by the Practice **E180**, Test Method **E1064** analysis.

14.5.2 Any judgment in accordance with these two statements would have an approximate 95 % probability of being correct.

14.6 *Bias*—At the time of the study, there was no accepted reference material suitable for determining the bias for this test method, therefore no statement on bias is being made.

14.7 The precision statement was determined through statistical examination of qualified results, from seventeen laboratories, on four materials. These four materials were described as the following: Fluid 1: Monoethylene Glycol, Fluid 2: Diethylene Glycol, Fluid 3: Triethylene Glycol, and Fluid 4: Tetraethylene glycol. To judge the equivalency of two test results, it is recommended to choose the material closest in characteristics to the test material. See **Table 5**.

14.8 Precision for MPG and DPG: to be evaluated.

⁶ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:E15-1063. Contact ASTM Customer Service at service@astm.org.

TABLE 5 E2409 Glycol Impurities by GC (mg/kg)

| Test Result mg/kg | Sample | Average over all Laboratories | Repeatability Standard Deviation | Intermediate Standard Deviation | Reproducibility Standard Deviation | Repeatability Limit | Intermediate Limit | Reproducibility Limit |
|-------------------|---------|-------------------------------|----------------------------------|---------------------------------|------------------------------------|---------------------|--------------------|-----------------------|
| DEG | MEG | 374.59 | 7.3 | 7.3 | 34.0 | 20.6 | 20.6 | 95.3 |
| MEG | DEG | 1479.73 | 46.3 | 76.0 | 215.1 | 129.7 | 212.9 | 602.4 |
| TEG | DEG | 3499.69 | 92.8 | 143.2 | 306.5 | 260.0 | 401.0 | 858.3 |
| DEG | TEG | 489.32 | 56.8 | 70.9 | 201.7 | 159.1 | 198.5 | 564.9 |
| TetraEG | TEG | 1020.00 | 96.3 | 96.3 | 244.1 | 269.8 | 269.8 | 683.5 |
| DEG | TetraEG | 1646.25 | 55.4 | 55.4 | 95.4 | 155.1 | 155.1 | 267.1 |
| TEG | TetraEG | 7908.35 | 221.9 | 221.9 | 1350.7 | 621.2 | 621.2 | 3782.0 |
| PentaEG | TetraEG | 2084.93 | 58.7 | 72.9 | 156.3 | 164.5 | 204.1 | 437.5 |