

Designation: D3663 - 03 (Reapproved 2015) D3663 - 20

# Standard Test Method for Surface Area of Catalysts and Catalyst Carriers<sup>1</sup>

This standard is issued under the fixed designation D3663; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\varepsilon$ ) indicates an editorial change since the last revision or reapproval.

# 1. Scope

- 1.1 This test method covers the determination of surface areas of catalyst and catalyst carriers that have Type II or IV nitrogen adsorption isotherms, and at least 1 m<sup>2</sup>/g of area. A volumetric measuring system is used to obtain at least four data points which fit on the linear BET<sup>2</sup> equation line.
- 1.2 <u>Units</u>—The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only and are not considered standard.
- 1.3 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety safety, health, and health environmental practices and determine the applicability of regulatory limitations prior to use.
- 1.4 This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

# 2. Referenced Documents

2.1 ASTM Standards:<sup>3</sup>

#### ASTM D3663-20

D3766 Terminology Relating to Catalysts and Catalysis 2e2ab-d2fc-417a-9462-094be995324e/astm-d3663-20

E177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods

E456 Terminology Relating to Quality and Statistics

E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

#### 3. Terminology

- 3.1 Consult Terminology D3766 for definitions of other terms used.
  - 3.2 Definitions:
- 3.2.1 *surface area of a <del>catalyst</del>* <u>catalyst</u>, <u>n</u> the total <u>probe accessible</u> surface of the <del>catalyst</del>. <u>It catalyst</u>; it is expressed in square metres per gram.
  - 3.3 Symbols:

<sup>&</sup>lt;sup>1</sup> This test method is under the jurisdiction of ASTM Committee D32 on Catalysts and is the direct responsibility of Subcommittee D32.01 on Physical-Chemical Properties.

Current edition approved April 1, 2015Oct. 1, 2020. Published June 2015November 2020. Originally approved in 1978. Last previous edition approved in 20082015 as D3663 – 03 (2008), (2015). DOI: 10.1520/D3663-03R15.10.1520/D3663-20.

<sup>&</sup>lt;sup>2</sup> Brunauer, Emmett, Teller, Journal of American Chemical Society, JACS, No. 60, 1938, p. 309.

<sup>&</sup>lt;sup>3</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

```
P_{\mathrm{H_1}}
          = initial helium pressure, torr
          = helium pressure after equilibration, torr
          = temperature of manifold at initial helium pressure, °C
          = temperature of manifold after equilibration, °C
          = initial N_2 pressure, torr
T_1
          = manifold temperature at initial N_2 pressure, K
T_1
          = manifold temperature at initial N_2 pressure, °C
             pressure after equilibration, torr
P_0^2
T_s
X
          = liquid nitrogen vapor pressure, torr
          = liquid nitrogen temperature, K
          = relative pressure, P_2/P_0
         = volume of manifold, cm<sup>3</sup>
          = extra volume bulb, cm<sup>3</sup>
          = dead-space volume, cm<sup>3</sup>
W
          = mass of sample, g
          = tare mass of sample tube, g
W
          = sample + tare mass of tube, g
          = volume of nitrogen in the dead-space, cm<sup>3</sup>
          = see 10.4.4
          = see 10.4.6
          = see 10.4.7
          = see 10.4.9
          = see 10.8
          = initial extra-volume bulb temperature, K
   x'(i) = initial extra-volume bulb temperature, °C
          = extra-volume bulb temperature after equilibrium, K
T_{2x}'(i) = extra-volume bulb temperature after equilibrium, °C
```

# 4. Summary of Test Method

- 4.1 The surface area of a catalyst or catalyst carrier is determined by measuring the volumeamount of nitrogen gas adsorbed at various low-pressure levels by the catalyst sample. Pressure differentials caused by introducing the catalyst surface area to a fixed volume of nitrogen in the test apparatus are measured and used to calculate BET surface area.
- 5. Apparatus and ards. iteh. ai/catalog/standards/sist/b2c2e2ab-d2fc-417a-9462-094be995324e/astm-d3663-20
- 5.1 A schematic diagram of the <u>a suitable</u> apparatus is shown in Fig. 1. It may be constructed of glass or of metal. <u>Commercial metal instruments are available</u>. It has the following features:
  - 5.1.1 Distribution Manifold, having a volume between 20 and 35 cm<sup>3</sup>,  $(V_d)$ , known to the nearest 0.05 cm<sup>3</sup>. This volume is defined as the volume between the stopcocks or valves and includes the pressure gauge.




FIG. 1 Schematic Diagram of Surface Area Apparatus

<sup>&</sup>lt;sup>4</sup> Automated equipment is commercially available.

- 5.1.2 *Vacuum System*, capable of attaining pressures below  $10^{-4}$  torr (1 torr = 133.3 Pa). This will<u>can</u> include a vacuum gauge (not shown in Fig. 1). Access to the distribution manifold is through the valve V.
- 5.1.3 Constant-Volume Gauge or Mercury Manometer, Pressure-Measuring Device, that operates at constant volume and is capable of measurements to the nearest 0.1 torr, in the range from 0 to 1000 torr (1 torr = 133.3 Pa).

Note 1—See, for example, the article by Joy<sup>5</sup> for a description of a constant-volume manometer.

- 5.1.4 *Valve* (*H*), from the helium supply to the distribution manifold.
- 5.1.5 Valve (N), from the nitrogen supply to the distribution manifold.
- 5.1.6 The connection between the sample tube and the S valve can be a standard-taper glass joint, a glass-to-glass seal, or a compression fitting.
- 5.1.7 Extra Volume Bulb,  $(V_x)$ , should be 100 to 150 cm<sup>3</sup>, known to the nearest 0.05 cm<sup>3</sup>.  $V_x$  includes the volume of the stopcock bore in the glass apparatus.
- 5.2 Sample Tubes, with volumes from 5 to 100 cm<sup>3</sup> depending on the application. Markings Use sample tubes recommended by the instrument manufacturer. If necessary, markings should be placed on the sample tubes about 30 to 50 mm-50 mm below the connectors to indicate the desired liquid nitrogen level.
- 5.3 Heating Mantles or Small Furnaces.
  - (https://standards.itah.a
- 5.5 Laboratory Balance, with 0.1 mg  $(10^{-7} \text{ kg})$  sensitivity.
- 5.6 *Thermometer or Thermocouple*, for measuring the temperature of the distribution manifold  $[T_1'(i)]$  or  $T_2'(i)$  in degrees Celsius or kelvins.
- 5.6.1 It is preferred that the manifold be thermostated at a particular temperature, a few degrees above ambient, to obviate the necessity of recording this temperature at each reading.
- 5.7 Thermometer, for measuring the temperature of the liquid nitrogen bath  $[T_s(i)]$  in kelvins. This will preferably be a nitrogen vapor-pressure thermometer, often referred to in a commercial instrument as a pressure saturation tube, from which  $P_0$  values may be derived.

# 6. Reagents

5.4 Dewar Flasks.

- 6.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.<sup>6</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.
- 6.2 Helium Gas—A cylinder of helium gas at least 99.9 % 99.999 % pure.
  - 6.3 Liquid Nitrogen, of such purity that  $P_0$  is not more than 20 torr above barometric pressure. A fresh daily supply is recommended.

<sup>&</sup>lt;sup>5</sup> Joy, A. S., Vacuum, Vol 3, 1953, p. 254.

6.4 Nitrogen Gas-A cylinder of nitrogen gas at least 99.999 % pure.

# 7. Procedure—Sample Preparation and Degassing

- 7.1 Select a sample tube of the desired size. A 5 cm<sup>3</sup> sample tube is preferred for samples not exceeding about 1 g, to minimize the dead-space. However, a 25 cm<sup>3</sup> sample tube may be preferred for finely powdered catalysts, to avoid "boiling" when degassing is started.
- 7.2 Fill the sample tube with nitrogen or helium, at atmospheric pressure, after removing air by evacuation. This may be done on the surface area unit, or on a separate piece of equipment.
- 7.3 Remove the sample tube from the system, cap, and weigh. Record the mass as  $W_1$ .
- 7.4 Place the catalyst sample, whose mass is known approximately, into the sample tube. Choose the sample size to provide an estimated total sample surface area of 20 to 100 m<sup>2</sup>.
- 7.5 Attach the sample tube to the apparatus. If other samples are to be run, attach them at this time to the other ports.
- 7.6 Open the S valves where there are samples.
- 7.7 It may be necessary to close the V valve system periodically to protect the diffusion pump fluid from exposure to pressures above 0.1 torr for periods of more than 30 s. Close the valve off for 2 min.
- 7.8 Install a heating mantle or furnace around each sample and raise the temperature to about 300°C300°C (573 K).
- Note 2—Take special precautions if the moisture content exceeds approximately 5 % to avoid "bumping" of powdered catalyst, and to avoid surface area loss by self-steaming. It is recommended that the heating rate not exceed 100 KA stepped or pressure-controlled heating routine is recommended that these circumstances.
  - Note 3—Not all catalysts or their supports are stable to 300 °C. This should be noted and the appropriate temperature should be used based on thermal properties of the material.
- 7.9 Continue degassing at about 300°C 300 °C (573 K) for a minimum of 3 h, 3 h, at a pressure not to exceed 10<sup>-3</sup> torr. Overnight degassing is permissible.
  - Note 3—Certain materials will decompose at 300°C (for example, alumina hydrates) or will sinter (for example, platinum black). Lower degassing temperatures are permissible for such materials; however, the degassing temperature should be specified when reporting the results.
  - 7.10 Remove the heating mantles, and allow the samples to cool.
  - 7.11 Close the EV valve, if open.
  - 7.12 Close the S valve.
  - 7.13 It is permissible to exercise the option of preliminary degassing on an external unit. In such a case, follow the procedures of 7.4 7.11 and then repeat on the surface area unit, except that the degassing time in 7.9 should not exceed  $\frac{1 1.1}{1.1}$  h.
  - 7.14 If it is desired to weigh the sample after preliminary degassing on an external unit, backfill with the same gas used in 7.2 to above atmospheric pressure. Close the S valve.
  - 7.15 Detach the sample tube from the apparatus, recap with the stopper used previously, and weigh. Record the mass as  $W_2$ .



7.16 Remove the backfilled gas by evacuation to less than  $10^{-4}$  torr at room temperature.

# 8. Procedure—Dead-Space Determination

- 8.1 From this point on, each sample being tested for surface area must be run on an individual basis. Thus each Step (8.2 9.17) must be carried out separately for each tube in test.
  - 8.2 The "dead-space" is the void volume of the charged sample tube, including the S valve, when the tube is immersed in liquid nitrogen to the proper depth (see 5.2).
  - Note 4—The dead-space may be determined after the nitrogen adsorption, if more convenient, as long as adequate degassing precedes it. In that case, replace the liquid nitrogen bath after Step 9.14 before proceeding with Steps 8.3 8.9.
  - Note 5—Some commercial equipment allows for helium-free determination of the dead-space by calibration of a blank cell and accounting for the skeletal volume of the sample when the cell is charged. This can be considered equivalent to determination of dead-space by helium and can be used if helium is not available. Follow the recommendations of the instrument manufacturer for these measurements.
  - 8.3 Place a Dewar flask of liquid nitrogen around the sample and adjust the liquid level to a fixed point on the sample tube. Maintain this level throughout the test.
  - 8.4 Zero the pressure gauge.
  - 8.5 Admit the helium gas into the manifold to a pressure of 600 to 900 torr by carefully opening the H valve. Record this pressure as  $P_{H1}$ , and the manifold temperature,  $T_{H1}$ .
  - 8.6 Open the S valve to admit helium to the sample.
  - 8.7 After about 5 min of equilibration, readjust the liquid nitrogen level, and record the pressure as  $P_{H2}$ , and manifold temperature as  $T_{H2}$ .
  - 8.8 Repeat 8.5 8.7 for each sample on the manifold. <u>ASTM D3663-20</u> https://standards.iteh.ai/catalog/standards/sist/b2c2e2ab-d2fc-417a-9462-094be995324e/astm-d3663-20
  - 8.9 Open all S valves; then slowly open the V valve to remove the helium gas.
  - 8.10 When a pressure less than 0.01 torr has been attained, close the S valve. This operation should take 5 to 10 min.

# 9. Procedure—Nitrogen Adsorption

- 9.1 Close the V valve and open the EV\_valve if the extra-volume bulb is to be used, when the surface area is known to be high.
- 9.2 Recheck the zero setting of the pressure gauge.
- 9.3 Admit nitrogen gas, and record the pressure as  $P_1(1)$  (torr) and the temperature as  $T_1'(1)$  (degrees Celsius) and read the temperature of the extra-volume bulb and record it as  $T_{1x}(1)$ . It is desirable, but not necessary, to choose  $P_1(1)$  such that the first equilibrium adsorption pressure,  $P_2(1)$ , will be about 40 torr equivalent to  $P_2(1)/P_0(1)$  of about 0.05. If the surface area is small, it may be desirable to eliminate use of the extra-volume bulb by closing the EV valve.
- 9.4 Open the S valve to admit nitrogen to the catalyst.
- 9.5 Allow sufficient time for equilibration, readjusting the liquid nitrogen level when necessary. Equilibrium shall be considered as attained when the pressure change is no more than 0.02 torr/min.
- 9.6 Record the equilibrium pressure as  $P_2(1)$ , manifold temperature  $T_2'(1)$ , and the extra volume bulb temperature  $T_{2x}(1)$ .