



Designation: D4198 – 20

Standard Test Methods for Evaluating Absorbent Pads Used with Membrane Filters for Bacteriological Analysis and Growth¹

This standard is issued under the fixed designation D4198; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 These test methods cover the determination of the nutrient-holding capacity and the toxic or nutritive effect on bacterial growth of organisms retained on a membrane filter, when the absorbent pad being tested is used as a nutrient reservoir and medium supply source for the retained bacteria.

1.2 The tests described are conducted on 47 mm diameter disks, although other size disks may be employed for bacterial culture techniques.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[D1129 Terminology Relating to Water](#)

[D1193 Specification for Reagent Water](#)

[D3508 Method for Evaluating Water Testing Membrane Filters for Fecal Coliform Recovery](#) (Withdrawn 1995)³

3. Terminology

3.1 *Definitions:*

3.1.1 For definitions of terms used in this standard, refer to Terminology [D1129](#).

4. Summary of Test Methods

4.1 Test Method A involves saturating a 47 mm absorbent pad with water and determining the volume of water held by the pad by weighing the pad dry and when fully saturated.

4.2 Test Method B involves culturing micro-organisms from suspensions of pure cultures on a 0.45 μm membrane filter, which is placed on the test absorbent pad saturated with the appropriate growth medium. The resultant cultures are compared to cultures grown on spread plates and to membrane filters placed directly on agar with no absorbent pad.

5. Significance and Use

5.1 These test methods are appropriate for qualifying absorbent pads used with membrane filters for bacteriological enumeration.

5.1.1 The test methods described are applicable to quality control testing of absorbent pads by the suppliers and users of these pads and to specification testing of absorbent pads intended for use with membrane filters in bacteriological enumeration.

5.2 Other pure culture organisms and their appropriate culture medium may be substituted for the *E. coli* and M-FC media for specification testing, as required.

6. Apparatus

6.1 *Filtration Units* for membrane filters with side-arm flask and tubing.

6.2 *Vacuum Source*.

6.3 *Vortex Mixer* or similar mixer.

¹ These test methods are under the jurisdiction of ASTM Committee D19 on Water and are the direct responsibilities of Subcommittee D19.08 on Membranes and Ion Exchange Materials.

Current edition approved Nov. 1, 2020. Published November 2020. Originally approved in 1982. Last previous edition approved in 2011 as D4198 – 82 (2011) which was withdrawn July 2020 and reinstated in November 2020. DOI: 10.1520/D4198-20.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

6.4 *Forceps*, flat-bladed.

6.5 *Incubator* capable of maintaining temperatures of 44.5 ± 0.2 °C.

6.6 *Stereoscopic Microscope and Illuminator*.

6.7 *Illuminated Magnifying Stand* for counting colonies on agar spread plates.

6.8 *Hand Tally Counter*.

6.9 *Autoclave*.

6.10 *Analytical Balance* readable to the nearest 1 mg.

6.11 *Petri Dish*, 50 mm, nonsterile.

6.12 *Expendable Equipment*:

6.12.1 Filters (gridded, 0.45 µm, 47 mm) sterile, for water testing.

6.12.2 Absorbent pads (47 mm), sterile for the growth test.

6.12.3 Petri dishes, sterile 50 mm and 100 mm.

6.12.4 Pipets, sterile, 10 mL, 0.1 mL graduations, accuracy of ±5 %.

6.12.5 Test tubes, sterile, 20 mL, with screw caps.

6.12.6 Bent glass rod, sterile, for spreading bacterial cultures.

6.12.7 Burner, for flame sterilization.

7. Reagents and Materials

7.1 *Purity of Water*—Unless otherwise indicated, reference to water shall be understood to mean reagent water conforming to Specification **D1193**, Type II, with 0.2 µm membrane filtration. In addition, suitability tests for determining the bactericidal properties of the reagent grade water should be performed.

7.2 *M-FC Agar with Rosolic Acid*—Fecal coliform medium specific for the membrane filter technique.

7.3 *M-FC Broth with Rosolic Acid*—Broth nutrient for bacterial growth.

7.4 *Peptone Water*, 0.1 %.

7.5 *E. coli* (ATCC 11229).

8. Preparation of Equipment and Materials

8.1 *Washing and Cleaning*—Clean all glassware and filtration equipment thoroughly, using a suitable detergent in hot water, rinse with hot water, and then rinse in reagent grade water. Dry the equipment thoroughly prior to sterilization.

8.2 *Sterilization*—Follow standard microbiological laboratory practices for preparing glassware and filtration equipment prior to placing in the autoclave. Autoclave at 121 °C for 15 min. Refer to Method **D3508** for details.

8.3 *Incubator*—Set incubator at 44.5 ± 0.2 °C.

9. Media Preparation

9.1 *M-FC Broth with Rosolic Acid*—Dissolve in reagent grade water in accordance with the manufacturer's instructions.

9.2 *M-FC Agar with Rosolic Acid*—To a solution of M-FC, add agar (15 g per 1000 mL), mix while heating in accordance with the manufacturer's instructions, cool, and dispense into 100 mm petri dishes.

10. Culture Preparation

10.1 Resuspend culture in accordance with the supplier's instructions.

10.2 Using a sterile loop, streak an agar plate with culture of *E. coli*.

10.3 Incubate 24 h at 44.5 °C.

10.4 Using a sterile needle, inoculate M-FC agar with organisms from a single colony on the streak plate.

10.5 Let the culture incubate for 5 to 6 h at 35 °C.

10.6 Plate out a series of dilutions and store the remainder of the culture in the refrigerator. Incubate the plates for 18 ± 2 h at 44.5 °C.

10.7 Based on the 24-h plate count, dilute a portion of the culture to give a solution with 200 to 800 microorganisms per millilitre.

11. Procedure

11.1 *Test Method A—Water Retention*:

11.1.1 Weigh three dry 50 mm plastic petri dishes to the nearest 1 mg, for each lot of pads to be tested.

11.1.2 Randomly select three absorbent pads from each lot, place a dry pad in each of the dishes, cover, and weigh again.

11.1.3 To each pad, add an excess of water.

11.1.4 After the pads are fully saturated (20 to 30 s), pour off the excess water and shake out any remaining excess.

11.1.5 Cover the dishes and weigh again.

11.2 *Test Method B—Culture Technique*:

11.2.1 Prepare a set of ten 100 mm sterile petri dishes with 16 ± 1 mL of M-FC agar. Make sure the agar plates are at room temperature and that the surfaces are dry before using.

11.2.2 Prepare a set of five 50 mm sterile petri dishes with the sterile pads. To each pad add 1.8 mL of sterile M-FC broth and pour off the excess.

11.2.3 Test five replicate sets of three aliquots. Each replicate shall include (a) two membrane-filtered samples (one on agar, one on a pad), and (b) one spread plate.

11.2.4 Add 0.1 mL of the diluted culture solution from **10.7** to the agar plate from **11.2.1** and using a sterile bent glass rod, spread over the surface of the agar. Cover the plate.

11.2.5 Set up two sterile filter funnels with flasks so that the two membrane samples in the set can be run concurrently with the spread plate.

11.2.6 Place a sterile gridded 0.45 µm membrane onto each of the two filtration bases and assemble the funnels.

11.2.7 Add 0.1 mL of the diluted culture (20 to 80 organisms per 0.1 mL) from **10.7** to each of the two tubes, which contain 20 mL of sterile 0.1 % peptone.

11.2.8 Cap the tubes and mix on a vortex mixer.

11.2.9 Add the contents of one tube to each funnel and turn on the vacuum.