

---

---

## Laboratory glassware — Filter funnels

*Verrerie de laboratoire — Entonnoirs à filtrer*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 4798:1997

<https://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-a8a90d8d6e0f/iso-4798-1997>



## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

This International Standard ISO 4798 has been prepared by Technical Committee ISO/TC 48, *Laboratory glassware and related apparatus*, Subcommittee SC 2, *General laboratory glassware (other than measuring apparatus)*.

Annex A of this International Standard is for information only.

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

ISO 4798:1997

<https://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-a8a90d8d6e0f/iso-4798-1997>

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet central@iso.ch  
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland

# Laboratory glassware — Filter funnels

## 1 Scope

This International Standard specifies requirements and dimensions for glass filter funnels suitable for general laboratory purposes, fitting together with other general-use glassware such as boiling flasks and volumetric flasks.

NOTE — Annex A lists additional International Standards for other general-purpose laboratory glassware.

## 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

<https://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-asa70d8d6c0f/iso-4798-1997>

ISO 719:1985, *Glass — Hydrolytic resistance of glass grains at 98 °C — Method of test and classification.*

ISO 3585:1991, *Borosilicate glass 3.3 — Properties.*

## 3 Types and sizes

Two types of filter funnels are specified, in sizes as follows (given in millimetres):

### 3.1 Plain filter funnels of bowl diameter:

35 – 55 – 75 – 100 – 125 – 150 – 200

### 3.2 Analytical filter funnels of bowl diameter:

55 – 75 – 100

## 4 Material

4.1 When tested in accordance with the procedure and classification given in ISO 719, the glass shall comply with the requirements of class HGB3 or better.

The glass shall be free from visible defects and shall be free from internal stress which would impair the performance of the filter funnels.

**4.2** At the manufacturer's discretion, filter funnels may be manufactured from borosilicate glass 3.3 in accordance with ISO 3585.

## **5 Construction**

### **5.1 General appearance**

The funnel shall consist of a bowl, having the shape of a right circular cone, and a stem which shall be coaxial with the bowl.

### **5.2 Bowl**

**5.2.1** The wall of the bowl shall diverge from the axis of the funnel so as to give an included angle of  $60^{\circ} \begin{smallmatrix} 0 \\ -3^{\circ} \end{smallmatrix}$ .

**5.2.2** In plain filter funnels, the bowl shall be plain or with a beaded rim.

**5.2.3** In analytical funnels, the bowl shall either be plain or ribbed internally or externally. Where the bowl is ribbed externally, there shall be at least three such ribs evenly spaced round the bowl.

### **5.3 Rim**

ISO 4798:1997

<https://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-a8a90d8d6c0f/iso-4798-1997>

The rim of the bowl shall be fire-polished, beaded or ground to a plane surface with the outside slightly bevelled. Any rim bead should not project internally so as to interfere with the lie of the filter paper.

### **5.4 Stem**


**5.4.1** In plain filter funnels, the stem should consist of a separate piece of tubing joined neatly and securely to the bowl, or be drawn out from the bowl. The stem may be slightly tapered from the upper to the lower end.

**5.4.2** In analytical filter funnels, the stem shall consist of a separate piece of tubing joined neatly and securely to the bowl.

**5.4.3** The end of the stem shall be ground at  $30^{\circ}$  minimum to the axis, and the outer edge of the ground part shall be slightly bevelled or fire-polished.

## **6 Dimensions**

The dimensions of the filter funnels shall be as shown in table 1 and table 2.



**Figure 1 — General design of glass filter funnel  
(plain or analytical)**

## iTeh STANDARD PREVIEW (standards.iteh.ai)

**Table 1 — Dimensions for plain filter funnels**

ISO 4798:1997

Dimensions in millimetres

<https://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-a8a90d8d6e0f/iso-4798-1997>

Bowl		Stem		
Internal diameter	Wall thickness min.	Nominal		Wall thickness min.
		External diameter	Length	
$35 \pm 5$	1	6	35	0,8
$55 \pm 5$	1	8	55	1,3
$75 \pm 5$	1	8	75	1,3
$100 \pm 5$	1,2	10	100	1,3
$125 \pm 5$	1,2	16	125	1,3
$150 \pm 10$	1,2	16	150	1,6
$200 \pm 10$	1,5	24	150	1,6

Table 2 — Dimensions for analytical filter funnels

Dimensions in millimetres

Bowl	Stem			
Internal diameter	Diameter		Wall thickness	Length
$\pm 5$	external max.	internal min.	min	$\pm 5$
55	9,4	2,5	1,5	150
75				
100				

## 7 Marking

The following inscriptions shall be marked on each filter funnel:

- the manufacturer's and/or supplier's name or mark;
- the type of glass, if not otherwise identifiable.

[ISO 4798:1997](http://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-a8a90d8d6e0f/iso-4798-1997)

The following are optional:

- the internal diameter of the bowl without tolerances;
- the number of this International Standard.

## Annex A

(informative)

### Bibliography

- [1] ISO 383:1976, *Laboratory glassware — Interchangeable conical ground joints.*
  - [2] ISO 384:1978, *Laboratory glassware — Principles of design and construction of volumetric glassware.*
  - [3] ISO 641:1975, *Laboratory glassware — Interchangeable spherical ground joints.*
  - [4] ISO 1773:1997, *Laboratory glassware — Narrow-necked boiling flasks.*
  - [5] ISO 3819:1985, *Laboratory glassware — Beakers.*
  - [6] ISO 4142:1997, *Laboratory glassware — Test tubes and culture tubes.*
  - [7] ISO 4785:1997, *Laboratory glassware — General-purpose glass stopcocks.*
  - [8] ISO 4796:1977, *Laboratory glassware — Bottles.*
  - [9] ISO 4797:1981, *Laboratory glassware — Flasks with conical ground joints.*
  - [10] ISO 4799:1978, *Laboratory glassware — Condensers.*
  - [11] ISO 4800:1977, *Laboratory glassware — Separating funnels and dropping funnels.*
  - [12] ISO 6556:1981, *Laboratory glassware — Filter flasks.*
- ISO 4798:1997  
<https://standards.iteh.ai/catalog/standards/sist/c7cece5-626c-4a14-983a-a8a90d8d6e0f/iso-4798-1997>

## iTeh STANDARD PREVIEW (standards.iteh.ai)

ISO 4798:1997

<https://standards.iteh.ai/catalog/standards/sist/7c7cece5-626c-4a14-983a-a8a90d8d6e0f/iso-4798-1997>

---

---

### ICS 71.040.20

**Descriptors:** laboratory equipment, laboratory glassware, funnels, specifications, dimensions, marking.

Price based on 5 pages

---

---