

INTERNATIONAL STANDARD



**Radiation protection instrumentation – Security screening of humans –
Measuring the imaging performance of X-ray systems**

(<https://standards.iteh.ai>)

Document Preview

[IEC 62709:2024](https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024)

<https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024>



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

International
Standards
Document Preview
standards.iteh.ai

[IEC 62709:2024](https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024)

<https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024>



IEC 62709

Edition 2.0 2024-09

INTERNATIONAL STANDARD



**Radiation protection instrumentation – Security screening of humans –
Measuring the imaging performance of X-ray systems**

<https://standards.iteh.ai>
Document Preview

[IEC 62709:2024](https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024)

<https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 13.280

ISBN 978-2-8322-9243-3

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	8
2 Normative references.....	8
3 Terms, definitions, abbreviated terms, quantities and units.....	9
3.1 Terms and definitions.....	9
3.2 Abbreviated terms.....	12
3.3 Quantities and units.....	12
4 Imaging performance evaluation procedures.....	12
4.1 General characteristics and test procedures.....	12
4.2 Location of testing.....	13
4.3 Body phantom and test objects.....	14
4.4 Pentalith resolution test.....	15
4.4.1 Purpose.....	15
4.4.2 Test object description.....	16
4.4.3 Procedure.....	16
4.4.4 Evaluation and record.....	16
4.5 Wire detection test.....	16
4.5.1 Purpose.....	16
4.5.2 Test object description.....	17
4.5.3 Procedure.....	17
4.5.4 Evaluation and record.....	17
4.6 Materials detection on body test.....	18
4.6.1 General.....	18
4.6.2 Purpose.....	18
4.6.3 Test object description.....	18
4.6.4 Procedure.....	18
4.6.5 Evaluation and record.....	18
4.7 Materials detection in air test.....	18
4.7.1 General.....	18
4.7.2 Purpose.....	19
4.7.3 Test object description.....	19
4.7.4 Procedure.....	19
4.7.5 Evaluation and record.....	19
4.8 Penetration test.....	19
4.8.1 General.....	19
4.8.2 Purpose.....	19
4.8.3 Test object description.....	20
4.8.4 Procedure.....	20
4.8.5 Evaluation and record.....	20
5 Minimum acceptable imaging performance.....	20
6 Environmental requirements.....	21
Annex A (normative) Mechanical drawings of the test objects.....	22
Annex B (informative) Example of reporting form.....	35
Annex C (informative) Image resolution measurement using the pentalith.....	37
C.1 General.....	37

C.2	Strategy	37
C.3	Pentalith description	37
C.4	Pass/fail criterion	40
C.5	Repeatability	41
Annex D (informative)	Comparison of whole body imaging systems	42
Bibliography	43

Figure 1	– Generic illustration of the testing configuration showing a HDPE body phantom with a test object on one end supported 1 m off the ground	13
Figure 2	– Body phantom and test objects	15
Figure A.1	– Components of the test phantom	22
Figure A.2	– Material detection in air phantom	23
Figure A.3	– Subassembly of the material detection in air phantom (Figure A.2), metal comb, three teeth	23
Figure A.4	– Subassembly of the material detection in air phantom (Figure A.2), metal comb, two teeth	24
Figure A.5	– Subassembly of the material detection in air phantom (Figure A.2), metal comb, one tooth	24
Figure A.6	– Subassembly of the material detection in air phantom (Figure A.2), plastic comb	25
Figure A.7	– Subassembly of the material detection in air phantom (Figure A.2), mounting sheet	25
Figure A.8	– Material detection on body 1	26
Figure A.9	– Material detection on body 2	26
Figure A.10	– Wire detection phantom	27
Figure A.11	– Subassembly of the wire detection phantom (Figure A.10), mounting base	27
Figure A.12	– Subassembly of the wire detection phantom (Figure A.10), cover	28
Figure A.13	– Pentalith resolution phantom	29
Figure A.14	– Subassembly of the pentalith resolution phantom (Figure A.13), mounting base	30
Figure A.15	– Subassembly of the pentalith resolution phantom (Figure A.13); hole placement in mounting base	31
Figure A.16	– Subassembly of the pentalith resolution phantom (Figure A.13), cover	32
Figure A.17	– Body phantom, 55 mm thick	32
Figure A.18	– Body phantom, 75 mm thick	33
Figure A.19	– Body phantom, 50 mm thick	33
Figure A.20	– Storing space	34
Figure C.1	– Dimensional design of the pentalith pattern	38
Figure C.2	– Example of a pentalith overlying a pixel grid	38
Figure C.3	– Example of a pentalith test phantom suitable for optical measurements	39
Figure C.4	– Example of a pentalith test phantom suitable for X-ray imaging	39
Figure C.5	– Example of using image thresholding as an objective pass/fail criterion	41
Table 1	– Wire sizes for the wire detection test	17
Table 2	– Minimum acceptable imaging performance at the reference location	20

Table 3 – Standard test conditions 21
Table D.1 – Comparison of whole body imaging systems for security screening 42

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[IEC 62709:2024](https://standards.itih.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024)

<https://standards.itih.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**RADIATION PROTECTION INSTRUMENTATION –
SECURITY SCREENING OF HUMANS –
MEASURING THE IMAGING PERFORMANCE OF X-RAY SYSTEMS****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 62709 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This second edition cancels and replaces the first edition published in 2014. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) Clarified the test procedures to maintain consistency with IEC 62463.
- b) Changed the term "spatial resolution" to "pentalith resolution".
- c) Modified some standard test conditions.
- d) Modified some terms and definitions.
- e) Changed the imaging requirements for transmission general-use systems.

The text of this International Standard is based on the following documents:

Draft	Report on voting
45B/1059/FDIS	45B/1069/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

iTeh Standards

<https://standards.itih.ai>

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

[IEC 62709:2024](#)

<https://standards.itih.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024>

INTRODUCTION

This document establishes standard test methods and test objects for measuring the imaging performance of X-ray systems for security screening of humans. For each image quality test, this document also sets minimum acceptable levels of performance. These procedures and minimum acceptable requirements should not be construed as an all-inclusive measure of performance for any situation. Depending on the circumstances and detection needs, user institutions will continue to generate their own requirements and are encouraged to do so. Rather, it is hoped that this document will provide a starting point for evaluating systems, provide a uniform set of readily available information to compare equipment, and offer a standard procedure for periodic quality control testing.

Four annexes are included. Annex A (normative) provides mechanical drawings of the imaging test objects. Sample test report forms are given in Annex B (informative). Annex C (informative) provides a generic description of the pentolith resolution test object. Annex D (informative) seeks to describe the different types of security systems presently being used for whole-body imaging.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC 62709:2024](https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024)

<https://standards.iteh.ai/catalog/standards/iec/df53a3cc-104a-40aa-bfe7-2a4ae8c13c89/iec-62709-2024>

RADIATION PROTECTION INSTRUMENTATION – SECURITY SCREENING OF HUMANS – MEASURING THE IMAGING PERFORMANCE OF X-RAY SYSTEMS

1 Scope

This document applies to security screening systems that utilize X-ray radiation and are used to inspect people who are not inside vehicles, containers, or enclosures. Specifically, this document applies to systems used to detect objects carried on or within the body of the individual being inspected.

The following types of systems are included in the scope of this document:

- Systems designated as mobile or fixed.
- Systems employing detection of primary radiation, backscattered radiation, forward-scattered radiation, (see Annex D) or some combination of these modalities to form two-dimensional X-ray images.
- Systems that are primarily imaging but that also may have complementary features such as material discrimination, automatic active or passive detection alerts. This document does not address how to test these complementary features.

The objective is to provide standard methods of measuring and reporting imaging quality characteristics that enable system manufacturers, potential system users and other interested parties to:

- a) Establish a consistent indicator of the expected technical performance of screening systems used for the inspection of individuals. Such technical performance testing complements explicit detection testing and evaluation. In this document "detection" refers to items in an image.
- b) Provide repeatable and verifiable imaging performance data that can be used to compare systems from different vendors.
- c) Establish a baseline that can be used over time to calibrate the system or detect any performance degradation. (It is not intended that the entire test method be employed for daily quality assurance testing.)
- d) Establish minimum acceptable performance requirements for the systems described above.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050-395:2014, *International Electrotechnical Vocabulary (IEV) – Part 395: Nuclear instrumentation – Physical phenomena, basic concepts, instruments, systems, equipment and detectors*

IEC 60050-395:2014/AMD1:2016

IEC 60050-395:2014/AMD2:2020

IEC 60050-881:1983, *International Electrotechnical Vocabulary (IEV) – Part 881: Radiology and radiological physics*

IEC 60050-881:1983/AMD1:2014

IEC 60050-881:1983/AMD2:2019

IEC 60050-881:1983/AMD3:2020

IEC 62463:2024, *Radiation protection instrumentation – X-ray systems for the security screening of persons*

ISO 683-17:2023, *Heat-treatable steels, alloy steels and free-cutting steels – Part 17: Ball and roller bearing steels*

3 Terms, definitions, abbreviated terms, quantities and units

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

3.1 Terms and definitions

For the purposes of this document, the following terms and definitions apply. The general terminology concerning X-ray systems and radiological physics is given in IEC 60050-395:2014 and IEC 60050-881:1983.

3.1.1 backscattered radiation backscatter

scattering of photons by material through angles greater than 90° with respect to their initial direction

3.1.2 backscatter system

security screening system that makes use of backscattered radiation to form an image

3.1.3 body phantom

object whose absorption and scattering effects on ionizing radiation are equivalent to a human

3.1.4 contrast sensitivity

ability to distinguish a small difference of intensity in an area of an X-ray image from a surrounding uniform background

[SOURCE: IEC 62523:2010, 3.11]

3.1.5 edge detection

ability to discern edges of objects or anomalies even when the bulk of the objects or anomalies may appear with the same brightness as the background

3.1.6 effective dose

dose quantity intended to reflect the stochastic health risk to the whole body due to radiation exposure. It is calculated based on the sum of the equivalent doses in various organs multiplied by the appropriate tissue weighting factors

3.1.7**forward-scattered radiation****forward-scatter**

scattering of photons by material through angles less than 90° with respect to their initial direction

3.1.8**forward-scatter system**

security screening system that makes use of forward-scattered radiation to form an image

3.1.9**floor of the scanner**

surface that individuals stand on when scanned

3.1.10**general-use system**

X-ray screening system that is configured to deliver an effective dose of less than 0,25 µSv per screening (using the dose estimation methods defined in IEC 62463) and operating using the administrative controls specified in IEC 62463. Given proper justification and certain restrictions, general-use systems may be operated without specific controls that would limit the number of individuals scanned or the number of scans per individual in a year

Note 1 to entry: This definition was reproduced, with the permission of the Health Physics Society (HPS), from ANSI/HPS N43.17-2009 (R2018)

[SOURCE: IEC 62463:2024, 3.4]

3.1.11**influence quantity**

quantity that is not the measurand but that affects the result of the measurement

3.1.12**limited-use system**

personnel screening system that is configured to deliver an effective dose that does not exceed 10 µSv per screening (using the dose estimation methods defined in IEC 62463) which does not meet the definition of a general-use system. Limited-use systems require additional controls and documentation to ensure that annual individual dose limits are not exceeded.

Note 1 to entry: This definition was reproduced, with the permission of the Health Physics Society (HPS), from ANSI/HPS N43.17.

[SOURCE: IEC 62463:2024, 3.7]

3.1.13**materials detection**

test of the ability to detect materials on or off the body phantom

3.1.14**operator**

person that controls one or more aspects of the screening procedure. An operator is authorized to perform their duties, appropriately trained, and performs their duties according to the standard operating procedure

[SOURCE: IEC 62463:2024, 3.9]

3.1.15**partial body field of view**

field of view of systems designed to scan parts of the body, such as cast and prostheses scanners or shoe scanners