

SLOVENSKI STANDARD SIST EN 25814:1996

01-oktober-1996

Kakovost vode - Ugotavljanje raztopljenega kisika - Elektrokemijska metoda (ISO 5814:1990)

Water quality - Determination of dissolved oxygen - Electrotechnical probe method (ISO 5814:1990)

Wasserbeschaffenheit - Bestimmung des gelösten Sauerstoffs - Elektrochemisches Verfahren (ISO 5814:1990) h STANDARD PREVIEW

Qualité de l'eau - Dosage de l'oxygene dissous - Méthode électrochimique a la sonde (ISO 5814:1990) <u>SIST EN 25814:1996</u>

https://standards.iteh.ai/catalog/standards/sist/43b9439d-9da0-4bcf-b8d2-

Ta slovenski standard je istoveten z: EN 25814-1996

<u>ICS:</u>

13.060.50 Preiskava vode na kemične snovi Examination of water for chemical substances

SIST EN 25814:1996

en

iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN 25814:1996 https://standards.iteh.ai/catalog/standards/sist/43b9439d-9da0-4bcf-b8d2e7dd20fl f3c0/sist-en-25814-1996

EUROPEAN STANDARD

EN 25814:1992

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 1992

UDC 628.1/3:620.1:543.37:543.25:546.21

Descriptors: Water, quality, water testing, chemical analysis, determination of content, oxygen, dissolved gases, electrochemical methods

English version

Water quality - Determination of dissolved oxygen - Electrochemical probe method (ISO 5814:1990)



<u>SIST EN 25814:1996</u> https://standards.iteh.ai/catalog/standards/sist/43b9439d-9da0-4bcf-b8d2e7dd20f1f3c0/sist-en-25814-1996

This European Standard was approved by CEN on 1992-10-05. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization Comité Européen de Normalisation Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

1992 Copyright reserved to CEN members

Ref. No. EN 25814:1992 E

Page 2 EN 25814:1992

Foreword

This European Standard is the endorsement of ISO 5814. Endorsement of ISO 5814 was recommended by CEN/Technical Committee 230 "Water analysis" under whose competence this European Standard will henceforth fall.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 1993, and conflicting national standards shall be withdrawn at the latest by April 1993.

The Standard was approved and in accordance with the CEN/CENELEC Internal Regulations, the following countries are bound to implement this European Standard : Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, United Kingdom, DARD PREVIEW

(standards.iteh.ai)

Endorsement notice SIST EN 25814:1996

The text of the international standard ISO 5814:1990 was approved by CEN as a European Standard Without any modification.

INTERNATIONAL STANDARD

ISO 5814

Second edition 1990-04-01

Water quality – Determination of dissolved oxygen – Electrochemical Qualité de l'eau – Dosage de l'oxygene dissous électrochimique à la sonde



Reference number ISO 5814:1990(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75% of the member bodies casting a vote.

International Standard ISO 5814 was prepared by Technical Committee ISO/TC 147, Water quality.

This second edition cancels and replaces Sthe Efirst 8 edition (ISO 5814:1984), of which it constitutes a minor revision g/standards/sist/43b9439d-9da0-4bcf-b8d2-

Annexes A and B of this International Standard are for information only.

© ISO 1990

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization

Case Postale 56 • CH-1211 Genève 20 • Switzerland Printed in Switzerland

ii

Water quality — Determination of dissolved oxygen — **Electrochemical probe method**

1 Scope

This International Standard specifies an electrochemical method for the determination of dissolved oxygen in water by means of an electrochemical cell which is isolated from the sample by a gas permeable membrane.

Depending on the type of probe employed, measurement can be made either as concentration of RD solved oxygen - lodometric method. oxygen in milligrams per litre, percentage saturation (% dissolved oxygen) or both. The method meas-in the method meas-

100 % saturation. However, most instruments permit measurement of values higher than 1003% ENe25814:19 mmersion of a probe, consisting of a cell enclosed https://standards.iteh.ai/catalog/standards/sist/byb9439selective:f-membrane and containing the supersaturation.

The method is suitable for measurements made in the field and for continuous monitoring of dissolved oxygen as well as measurements made in the laboratory. It is the preferred method for highly coloured and turbid waters, and also for waters containing iron and iodine fixing substances, all of which may interfere in the iodometric method specified in ISO 5813. Gases and vapours such as chlorine, sulfur dioxide, hydrogen sulfide, amines, ammonia, carbon dioxide, bromine and iodine which diffuse through the membrane, may interfere, if present, by affecting the measured current. Other substances present in the sample may interfere with the measured current by causing obstruction, or deterioration of the membrane or corrosion of the electrodes. These include solvents, oils, sulfides, carbonates and algae.

The method is suitable for natural, waste and saline waters. If used for saline waters such as sea waters, or estuarine waters, a correction for salinity is essential.

2 Normative reference

The following standard contains provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the edition indicated was valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent edition of the standard indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 5813:1983. Water quality — Determination of dis-

e7dd20flf3c0/sist-en-25 electrolyte and two metallic electrodes, in the water to be analysed. (The membrane is practically impermeable to water and ionic dissolved matter, but is permeable to oxygen and a certain number of the other gases and lyophylic substances.)

> Because of the potential difference between the electrodes, caused by galvanic action or an external voltage, oxygen passing through the membrane is reduced at the cathode, while metal ions pass into solution at the anode.

> The current so produced is directly proportional to the rate of transport of oxygen through the membrane and the layer of electrolyte and hence to the partial pressure of the oxygen in the sample at a given temperature.

> The permeability of the membrane to gases varies greatly with temperature, and compensation is reguired for readings taken at different temperatures of the sample. This can be done mathematically, for example, by the use of a suitable nomograph or computer program. The majority of modern instruments automatically compensate for temperature variation by inclusion of temperature-sensitive elements in the electronic circuitry. However, instruments reading directly in percentage solubility, unless provided with a pressure transducer in the circuitry to compensate for pressure differences, will

display an observed percentage reading. This is a reading with respect to atmospheric pressure and not the true reading if the pressure differs from atmospheric pressure.

4 Reagents

During the analysis, use only reagents of recognized analytical grade and only distilled water or water of equivalent purity.

4.1 Sodium sulfite, anhydrous (Na_2SO_3) or heptahydrate, $(Na_2SO_3.7H_2O)$.

4.2 Cobalt(II) salt, for example cobalt(II) chloride hexahydrate ($CoCl_2.6H_2O$).

5 Apparatus

5.1 Measuring instrument, comprising the following components.

5.1.1 Measuring probe, either of the galvanic type The procedure is (for example lead/silver) or the polarographic type (for example silver/gold) with on required, and instructions temperature-sensitive compensating device.

5.1.2 Meter, graduated to show the concentration

of dissolved oxygen directly, and/or the percentager EN 25Adjust/the electrical zero of the instrument, where saturation with oxygen, or the current in microam gstandarapplicable 39d-9da0-4bcf-b8d2peres. e7dd20flf3c0/sist-en-25814-1996 NOTE 1 Some instruments are zero compensated and

5.2 Thermometer, graduated in divisions of 0,5 °C.

5.3 Barometer, graduated in divisions of 10 Pa.

6 Procedure

When using the measuring instrument, the manufacturer's instructions should be followed.

6.1 Measuring technique and precautions to be taken

6.1.1 Never touch the active surface of the membrane with the fingers.

6.1.2 After changing the electrolyte and the membrane, or if the membrane has been allowed to dry out, wet the membrane and allow the reading to become stable before carrying out the calibration (see 6.2). The time required depends on that necessary for consumption of the oxygen dissolved in the electrolyte.

6.1.3 Ensure that air bubbles are not trapped in the probe when immersing it in the sample.

6.1.4 It is essential that the sample should flow past the membrane of the probe to prevent the occurrence of false readings due to depletion of the oxygen in the part of the sample in immediate contact with the membrane. Ensure that the flow rate is such that variations in readings are not produced, and consult the instrument manufacturer's instructions on this matter.

6.1.5 In the case of a discrete sample, carry out the determination in a vessel filled to overflowing, sealed to exclude air and containing a stirrer, for example a magnetic bar. Adjust the stirring rate such that the reading remains stable after reaching equilibrium, and there is no entrapment of air.

6.1.6 In the case of a flowing sample, such as a water course, check the flow rate to ensure that it is sufficient. If not, either move the probe about in the sample or take a discrete sample and treat it as described in 6.1.5.

6.2 Calibration

The procedure is described in 6.2.1 to 6.2.3, but it is necessary to consult the instrument manufacturer's instructions.

need no adjustment.

6.2.2 Checking the zero

Check and, if possible, adjust the zero setting of the instrument by immersing the probe in 1 litre of water to which about 1 g of sodium sulfite (4.1) and about 1 mg of the cobalt(II) salt (4.2) have been added to render the water free from oxygen.

NOTE 2 Modern probes typically achieve a stable response in 2 min to 3 min. However, different probes can have different response rates and the manufacturer's instructions should be consulted.

6.2.3 Calibration at a value near saturation

Bubble air through water at a constant temperature so that its oxygen content is brought to saturation or near saturation. Leave for about 15 min at this temperature and determine the dissolved oxygen concentration, for example by the iodometric method specified in ISO 5813.

Regulate the instrument.

Immerse the probe in a bottle completely filled with the sample, prepared and standardized as de-

scribed. After allowing the probe to stabilize in the stirred solution for 10 min (see the note to 6.2.2), adjust the instrument reading to the known oxygen concentration of the sample, if necessary.

Replace the electrolyte and the membrane when the instrument can no longer be calibrated or when the response becomes unstable or slow (see the manufacturer's instructions).

NOTES

3 If previous experience has shown that the bubbling time and the air flow rate used provide a sample saturated with air, the iodometric determination may be replaced by consultation of table A.1 and table A.2.

Many instruments allow calibration in air. 4

6.2.4 Linearity curve

Check the linearity before using the instrument for routine purposes and carry out checks at regular intervals.

Carry out these checks by preparing a range of distilled water samples of varying intermediate oxygen concentration. Completely fill three or four narrownecked 250 ml bottles with distil(ed water and s.iteh.ai) callbration. deoxygenate to varying extents by gently bubbling oxygen-free argon or nitrogen through each for an appropriate period. Take readings with the probe under test, until a required concentration of disso reached. Then determine the disso tent by the probe and immediatel termine the dissolved oxygen con iodometric method in ISO 5813.

If there is agreement between the across the range of dissolved trations, the probe can be said to sponse. Check any departure from necessary, consult the manufacturer of the probe.

The reading of the instrument is usually linear NOTE 5 for dissolved oxygen saturations up to 100 %.

6.3 Determination

Carry out the determination on the water to be analysed following the instrument manufacturer's instructions.

After immersion of the probe in the sample, allow sufficient time for the probe to attain the water temperature and reach a stable reading. If necessary, because of the type of instrument used and the result required, check the water temperature and/or the atmospheric pressure.

Calculation and expression of results 7

7.1 Dissolved oxygen concentration

Express the dissolved oxygen concentration, in milligrams of oxygen per litre, and report the result to the first decimal place.

If the reading on the sample was made at a temperature different from that at which the instrument was calibrated, correct the value given by the instrument to allow for this. Some instruments make this correction automatically. This correction makes allowance for the difference in the solubility of oxygen at the two temperatures. Calculate the real value by multiplying the value read at the temperature at which the measurement was made by the ratio

$$\frac{\rho(0)_{m}}{\rho(0)_{c}}$$

where

 $ho(O)_m$ is the solubility at the temperature of measurement;

 $\rho(0)_{c}$ is the solubility at the temperature of

Example:

from time to time pproximately time lved ⁷ doxygen 0/ist-en-258 lved oxygen con- y afterwards de- centration by the	43 Jemperature of calib ration
	814-1996 Solubility at 25 °C 8,3 mg/l
	Temperature at the time of the measurement
e two procedures oxygen concen- show a linear re- n linearity and, if	Reading of the instrument 7 mg/l
	Solubility at 10 °C 11,3 mg/l
	Real value at 10 °C 11,3/8,3 × 7,0 = 9,5 mg/l

NOTES

6 The values for $\rho(O)_m$ and $\rho(O)_c$, expressed in milligrams per litre, in the given example are obtainable from table A.1, column 2.

7 For information on how the solubility of oxidation is dependent on the temperature, pressure and salinity, see annex A.

7.2 Dissolved oxygen concentration expressed as a percentage saturation

If required, calculate the percentage saturation of dissolved oxygen in water from:

$$\frac{\rho(O)}{\rho(O)_{s}} \times 100$$