

## SLOVENSKI STANDARD SIST EN 26777:1996

01-junij-1996

## Kakovost vode - Določanje nitrita - Molekularna absorpcijska spektometrijska metoda (ISO 6777:1984)

Water quality - Determination of nitrite - Molecular absorption spectrometric method (ISO 6777:1984)

Wasserbeschaffenheit - Bestimmung von Nitrit - Spektrometrisches Verfahren (ISO 6777:1984) **Teh STANDARD PREVIEW** 

Qualité de l'eau - Dosage des nitrites - Méthode par spectrométrie d'absorption moléculaire (ISO 6777:1984)

SIST EN 26777:1996

https://standards.iteh.ai/catalog/standards/sist/bf9a297e-bb0b-43ac-931a-

Ta slovenski standard je istoveten z: EN 26777-1996

ICS:

13.060.30 Odpadna voda

Sewage water

SIST EN 26777:1996

en



## iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 26777:1996</u> https://standards.iteh.ai/catalog/standards/sist/bf9a297e-bb0b-43ac-931ae456fbf3168f/sist-en-26777-1996

#### SIST EN 26777:1996

#### EUROPEAN STANDARD

#### EN 26777:1993

#### NORME EUROPÉENNE

### EUROPÄISCHE NORM

January 1993

UDC 628.1/.3:620.1:543.3:546.173

Descriptors:

Water, quality, water testing, chemical analysis, determination of content, nitrites, molecular absorption spectrophotometry

English version

#### Water quality - Determination of nitrite - Molecular absorption spectrometric method (ISO 6777:1984)

## iTeh STANDARD PREVIEW

Qualité de l'eau - Dosage des nitrites -Méthode par spectrométrie d'apsorption dards.iteh.aspektrometrisches Verfahren (ISO 6777:1984) moléculaire (ISO 6777:1984)

#### SIST EN 26777:1996

https://standards.iteh.ai/catalog/standards/sist/bf9a297e-bb0b-43ac-931ae456fbf3168f/sist-en-26777-1996

This European Standard was approved by CEN on 1993-01-20. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

### CEN

European Committee for Standardization Comité Européen de Normalisation Europäisches Komitee für Normung

Central Secretariat: rue de Stassart,36 B-1050 Brussels



Page 2 EN 26777:1993

1

#### Foreword

This European Standard is the endorsement of ISO 6777. Endorsement of ISO 6777 was recommended by Technical Committee CEN/TC 230 "Water analysis" under whose competence this European Standard will henceforth fall.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 1993, and conflicting national standards shall be withdrawn at the latest by July 1993.

The Standard was approved and in accordance with the CEN/CENELEC Internal Regulations, the following countries are bound to implement this European Standard : Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, United Kingdom.

## **iTeh STANDARD PREVIEW**

(St Endorsement interceai)

The text of the International Standard TSO 6777: 1984 was approved by CEN as he European Standardawithout fany modification.

e456fbf3168f/sist-en-26777-1996

International Standard



INTERNATIONAL ORGANIZATION FOR STANDARDIZATION MEXA HAPODHAR OPPAHUSALUR TO CTAHDAPTUSALUU ORGANISATION INTERNATIONALE DE NORMALISATION

# Water quality – Determination of nitrite – Molecular absorption spectrometric method

Qualité de l'eau – Dosage des nitrites – Méthode par spectrométrie d'absorption moléculaire

## First edition – 1984-08-01 iTeh STANDARD PREVIEW (standards.iteh.ai)

<u>SIST EN 26777:1996</u> https://standards.iteh.ai/catalog/standards/sist/bf9a297e-bb0b-43ac-931ae456fbf3168f/sist-en-26777-1996

UDC 543.344 : 543.42

Ref. No. ISO 6777-1984 (E)

**Descriptors** : water, chemical analysis, determination of content, nitrites, spectrometric analysis.

#### SIST EN 26777:1996

#### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of developing International Standards is carried out through ISO technical committees. Every member body interested in a subject for which a technical committee has been authorized has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council.

International Standard ISO 6777 was developed by Technical Committee ISO/TC 147, EVIEW Water quality, and was circulated to the member bodies in December 1982.

It has been approved by the member bodies of the following countries:

Australia	Hungary	ST EN 26777:1996
Austria	India	Romania
Belgium	Iran e456fbf3	South Africa, Rep. of
Brazil	Iraq	Spain
Canada	Italy	Sweden
Czechoslovakia	Japan	Switzerland
Denmark	Korea, Dem. P. Rep. of	Thailand
Egypt, Arab Rep. of	Mexico	United Kingdom
China	Netherlands	USSR
France	New Zealand	
Germany, F.R.	Norway	

No member body expressed disapproval of the document.

© International Organization for Standardization, 1984 •

Printed in Switzerland

## Water quality — Determination of nitrite — Molecular absorption spectrometric method

#### 1 Scope

This International Standard specifies a molecular absorption spectrometric method for the determination of nitrite in potable, raw and waste water.

#### 2 Field of application

#### 2.1 Range

A nitrite nitrogen concentration,  $\varrho_{\rm N}$ , of up to 0,25 mg/l can be determined when using the maximum volume (40 ml) of test II en SIA portion.

#### 2.2 Limit of detection 1)

## 4.2 Orthophosphoric acid, approximately 1,5 mol/l solu-

When using cells of optical path length 40 mm and a test por 26777:1996 tion of 40 ml, the limit of detectionthas been determined storlierds/sist/bf9a297e-bb0b-43ac-931a-

#### 2.3 Sensitivity 1)

Using a 40 ml test portion and a cell of optical path length 40 mm,  $\rho_{\rm N} = 0.062$  mg/l gives an absorbance of about 0,66 units.

Using a 40 ml test portion and a cell of optical path length 10 mm,  $\rho_{\rm N}$  = 0,25 mg/l gives an absorbance of about 0,67 units.

#### 2.4 Interferences

If the alkalinity of the sample is high, some interference may be encountered (see clause 9).

A range of substances often encountered in water samples has been tested for possible interference. Full details are given in the annex. Of the substances tested, only chloramine, chlorine. thiosulfate, sodium polyphosphate and iron(III) interfere significantly.

within the range  $\rho_{\rm N} = 0,001$  to 0,002 mg/l. e456fb[3]68f/sist-en-26Add, by means of a pipette, 25 ml of the orthophosphoric acid (4.1) to 150  $\pm$  25 ml of water. Mix and cool to room temperature. Transfer the solution to a 250 ml one-mark volumetric flask and dilute to the mark with water.

> Store in an amber glass bottle. The solution is stable for at least 6 months.

#### 4.3 Colour reagent.

WARNING - This reagent is hazardous. Skin contact or ingestion of it or its ingredients must be avoided.

Dissolve 40,0  $\pm$  0,5 g of 4-aminobenzene sulfonamide  $(NH_2C_6H_4SO_2NH_2)$  in a mixture of 100  $\pm$  1 ml of the orthophosphoric acid (4.1) and 500  $\pm$  50 ml of water in a beaker.

Dissolve 2,00  $\pm$  0,02 g of N-(1-naphthyl)-1,2-diaminoethane dihydrochloride (C<sub>10</sub>H<sub>7</sub>-NH-CH<sub>2</sub>-CH<sub>2</sub>-NH<sub>2</sub>·2HCI) in the resulting solution. Transfer to a 1 000 ml one-mark volumetric flask and dilute to the mark with water. Mix well,

Store in an amber glass bottle. The solution is stable for 1 month if stored at 2 to 5 °C.

#### Principle 3

Reaction of nitrite in the test portion with 4-aminobenzene sulfonamide reagent in the presence of orthophosphoric acid at pH 1,9 to form a diazonium salt which forms a pink-coloured dye with N-(1-naphthyl)-1,2-diaminoethane dihydrochloride (added with the 4-aminobenzene sulfonamide reagent). Measurement of the absorbance at 540 nm.

#### 4 Reagents

4.1

(standards.iten1.70 g/ml).

During the analysis, use only reagents of recognized analytical grade and only distilled water or water of equivalent purity.

Orthophosphoric acid, 15 mol/l solution,

<sup>1)</sup> Information derived from a United Kingdom interlaboratory trial involving five participants.

 $1,9 \pm 0,1.$  (See clause 9.)

7.3 Correction for colour

determinations.

7.4 Blank test

#### **4.4** Nitrite, standard solution, $\rho_N = 100 \text{ mg/l}$ .

Dissolve 0,492 2  $\pm$  0,000 2 g of sodium nitrite (dried at 105 °C for at least 2 h) in about 750 ml of water. Transfer quantitatively to a 1 000 ml one-mark volumetric flask and dilute to the mark with water.

Store in a stoppered amber glass bottle at 2 to 5 °C. This solution is stable for at least 1 month. (See clause 10.)

**4.5** Nitrite, standard solution,  $\rho_N = 1,00 \text{ mg/l}$ .

Transfer, by means of a pipette, 10 ml of the standard nitrite solution (4.4) to a 1 000 ml one-mark volumetric flask and dilute to the mark with water.

Prepare this solution each day as required, and discard after use.

#### **5** Apparatus

All glassware shall be carefully cleaned using approximately 2 mol/l hydrochloric acid and then rinsed thoroughly with water.

Ordinary laboratory apparatus, and

**Spectrometer**, suitable for measurements at a wavelength of 540 nm, together with cells of optical path length between 10 and 50 mm.

iTeh STAND

Gal mark volumetric flasks, the volumes of the standard nitrite solution (4.5) shown in table 1.
<u>SIST EN 26777:1996</u>
Dilute the contents of each flask with water to give a volume of

Place, by means of a burette, into a series of nine 50 ml one-

Add, by means of a pipette, 1,0 ml of the colour reagent (4.3). Mix immediately by swirling and dilute to the mark with water.

Mix and allow to stand. The pH at this stage should be

At least 20 min after addition of the reagent, measure the ab-

sorbance of the solution at the wavelength of maximum absor-

bance, approximately 540 nm, in a cell of suitable optical path

NOTE - The wavelength of maximum absorbance should be checked

when this method is first used, and should be used in all subsequent

If the colour of the test portion is such that it may interfere with

the measurement of absorbance, treat a duplicate test portion

as described in 7.2, but replacing the colour reagent (4.3) with

Carry out a blank test by proceeding as described in 7.2, but

1,0 ml of the orthophosphoric acid solution (4.2).

replacing the test portion with 40  $\pm$  2 ml of water.

7.5 Preparation of the calibration graph

length, using water as the reference liquid.

# https://standards.iteh.ai/catalog/standards

Subtract the absorbance of the zero term from the absorbances obtained for the other standard solutions and plot a graph of absorbance against the mass of nitrite, as nitrogen, for each optical path length. The graph should be linear and should pass through the origin.

#### 8 Expression of results

#### 8.1 Method of calculation

The corrected absorbance,  $A_{\rm r}$ , of the test solution is given by the equation

$$A_{\rm r} = A_{\rm s} - A_{\rm b}$$

or, if correction for colour was made, by means of the equation

$$A_{\rm r} = A_{\rm s} - A_{\rm b} - A_{\rm c}$$

where

 $A_{\rm s}$  is the absorbance, as measured, of the test solution;

 $A_{\rm b}$  is the absorbance of the blank test solution;

 $A_{\rm c}$   $\,$  is the absorbance of the solution prepared for the correction for colour.

#### 6 Sampling and samples

Laboratory samples should be collected in glass bottles and should be analysed as soon as possible within 24 h of collection. Storage of the samples at 2 to 5 °C may preserve many types of sample, but this should be verified.

#### 7 Procedure

#### 7.1 Test portion

The maximum volume of test portion is 40 ml. This is suitable for the determination of nitrite concentrations of up to  $\rho_{\rm N} = 0.25$  mg/l. Smaller test portions may be used as appropriate in order to accommodate much higher nitrite concentrations. If the laboratory sample contains suspended matter, this should be allowed to settle, or the sample should be filtered through a glass fibre paper before taking the test portion.

#### 7.2 Determination

Transfer, by means of a pipette, the selected volume of test portion to a 50 ml one-mark volumetric flask, and, if necessary, dilute to 40  $\pm$  2 ml with water.

NOTE — It is always essential to adjust the volume to 40  $\pm$  2 ml to ensure that the correct pH is obtained (after addition of the reagent) for the reaction.

NOTE – It is essential that the values of  $A_s$ ,  $A_b$  and  $A_c$  are measured in cells of the same optical path length for a particular sample.

From the corrected absorbance  $A_r$ , determine from the calibration graph (7.5), for the appropriate optical path length of the cell, the corresponding mass of nitrite, as nitrogen, in micrograms.

The nitrite content, expressed in milligrams of nitrogen per litre, is given by the formula

$$\frac{m_{\rm N}}{V}$$

where

 $m_{\rm N}$  is the mass, in micrograms, of nitrite nitrogen corresponding to the corrected absorbance ( $A_r$ );

V is the volume, in millilitres, of the test portion.

The result may be expressed as the mass concentration of nitrogen,  $\varrho_{\rm N}$ , or nitrite,  $\varrho_{\rm NO2^-}$ , in milligrams per litre, or as the

amount of substance concentration of nitrite ion,  $c(NO_2)$ , in micromoles per litre. The appropriate conversion factors are given in table 2.

Table 2

	e <sub>N</sub>	₽ <sub>NO2</sub>	$c(NO_2)$
	mg/l	mg/l	µmol/l
$\varrho_{\rm N} = 1  {\rm mg/l}$	1	3,29	71,4
$\varrho_{\rm NO_2^-} = 1  \rm mg/l$	0,304	1	21,7
$c(NO_2) = 1  \mu mol/l$	0,014	0,046	1

#### Example:

A nitrogen concentration of 1 mg/l corresponds to a nitrite concentration of 3,29 mg/l.

#### 8.2 Precision

Repeatability and reproducibility standard deviations have been determined as indicated in table 3.

nitrite solution (4.5)	Mass of nitrite nitrogen, m	Optical path length of cell		
ml	μg	mm		
0,00	SIST EN 2607001996	10 and 40*		
http9:59tandards.iteh.	ai/catalog/standar@s59st/bf9a297e-b	b0b-43ac-931a- 40		
1,00 <sub>e4</sub>	156fbf3168f/sist-ep-96777-1996	10 and 40		
1,50	1,50	40		
2,00	2,00	40		
2,50	2,50	10 and 40		
5,00	5,00	10		
7,50	7,50	10		
10,00	10,00	10		

## iTeh STANDATED PREVIEW

\* 50 mm cells may also be used.