



SLOVENSKI STANDARD
SIST EN 872:1997

01-avgust-1997

Kakovost vode - Določanje suspendiranih snovi - Metoda s filtracijo skozi filtre iz steklenih vlaken

Water quality - Determination of suspended solids - Method by filtration through glass fibre filters

Wasserbeschaffenheit - Bestimmung suspendierter Feststoffe - Verfahren durch Abtrennung mittels Glasfaserfilter

Qualité de l'eau - Dosage des matières en suspension - Méthode par filtration sur filtre en fibres de verre

PREVIEW
(standards.iteh.ai)
<https://standards.iteh.ai/catalog/standards/sist/9561e374-b83e-4f75-ba73-73265166e3cb/sist-en-872-1997>

Ta slovenski standard je istoveten z: EN 872:1996

ICS:

13.060.30 Odpadna voda Sewage water

SIST EN 872:1997 **en**

iTeh STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 872:1997

<https://standards.iteh.ai/catalog/standards/sist/9561e374-b83e-4f75-ba73-73265166e3cb/sist-en-872-1997>

EUROPEAN STANDARD

EN 872

NORME EUROPÉENNE

EUROPÄISCHE NORM

February 1996

ICS 13.060.40

Descriptors: water tests, water, effluents, sewage, quality, chemical analysis, determination of content, suspended matter, filtration, filters, glass fibres

English version

**Water quality - Determination of suspended solids
- Method by filtration through glass fibre filters**

Qualité de l'eau - Dosage des matières en suspension - Méthode par filtration sur filtre en fibres de verre

Wasserbeschaffenheit - Bestimmung suspendierter Feststoffe - Verfahren durch Abtrennung mittels Glasfaserfilter

STANDARD PREVIEW
(standards.iteh.ai)

SIST EN 872:1997
<https://standards.iteh.ai/catalog/standards/sist/9561e374-b83e-4f75-ba73-7326516683c6/sist-en-872-1997>

This European Standard was approved by CEN on 1995-12-07. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Foreword

This European Standard has been prepared by the Technical Committee CEN/TC 230 "Water analysis", of which the secretariat is held by DIN.

This European Standard shall be given the status of a National Standard, either by publication of an identical text or by endorsement, at the latest by August 1996, and conflicting national standards shall be withdrawn at the latest by August 1996.

According to the CEN/CENELEC Internal Regulations, the following countries are bound to implement this European Standard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

Annex A is informative.

1 Scope

This European Standard describes a method for the determination of suspended solids in raw waters, waste waters and effluents by filtration through glass fibre filters. The lower limit of the determination is about 2 mg/l. No upper limit has been established.

Water samples are not always stable which means that the content of suspended solids depends on storage time, means of transportation, pH value and other circumstances. Results obtained with unstable samples should to be interpreted with caution.

Floating oil and other immiscible organic liquids will interfere (see Annex A).

Samples containing more than about 1 000 mg/l of dissolved solids can require special treatment, see 8.6.

NOTE 1: The result of the determination depends to some extent on the type of filter used, see 5.2. It is therefore recommended that the type of filter is specified.

NOTE 2: The size distribution of particles in different samples can vary widely. Therefore there is no correlation between results obtained with filters of different pore width and no conversion factor can be given for the conversion of results obtained with one type of filter to another.

2 Normative References

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

- EN 25667-2 : 1993 Water Quality — Sampling — Part 2 : Guidance on Sampling Techniques (ISO 5667-2 : 1991)
- ISO 6107-2 : 1989 Water quality — Vocabulary — Part 2

3 Definitions

For the purposes of this European Standard the following definitions apply:

3.1 Suspended solids

Solids removed by filtration or centrifuging under specified conditions [4.24.3 of ISO 6107 : 1989].

3.2 Dissolved solids

The substances remaining, after filtration and evaporation to dryness of a sample, under specified conditions [4.24.1 of ISO 6107-2 : 1989].



4 Principle

Using a vacuum or pressure filtration apparatus the sample is filtered through a glass fibre filter. The filter is then dried at 105 °C and the mass of the residue retained on the filter is determined by weighing.

5 Apparatus

5.1 Equipment for vacuum or pressure filtration, to accommodate the selected filters (see 5.2).

Equipment for membrane filtration can in most cases be used for other types of filters. The plate supporting the filter needs to have sufficient permeance to allow the water to pass freely.

5.2 Borosilicate glass fibre filters which contain no binders. The filters shall be circular and of the appropriate diameter to fit the filtering device (5.1). The loss of mass in a blank test shall be less than 0,3 mg per filter. Preferably the mass per unit area shall be between 50 g/m² and 100 g/m².

Check the loss of mass during filtration by running the procedure in clause 8 but using 150 ml of distilled water instead of the sample. Check each box or batch separately. Use 3 filters, selected at random, to increase the sensitivity of the test.

NOTE: To remove water-soluble constituents, the filters may be prewashed. Individual or a small number of filters (less than 10) are prewashed by filtering 150 ml of distilled water through the filter(s) and then drying at 105 °C for at least 1 h.

It is recommended to bulk-wash filters by soaking in distilled water for several hours. The wash water is drained off and the filter dried at 105 °C for at least 1 h or preferably overnight before use.

Glass fibre filters from different manufacturers can have somewhat different filtering characteristics. State the type of filter used and its manufacturer in the report (see clause 10).

5.3 Drying oven, capable of maintaining a temperature of 105 °C ± 2 °C.

5.4 Analytical balance, capable of weighing to an accuracy of at least 0,1 mg.

5.5 Drying support of suitably surfaced material, to support the filters in the drying oven (see 5.3).

<https://standards.iteh.ai/catalog/standards/sist/9561e374-b83e-4f75-ba73-73265166e3cb/sist-en-872-1997>

6 Reagents

6.1 Reference suspension, microcrystalline cellulose, 500 mg/l. Weigh 0,500 g (oven dry basis) of microcrystalline cellulose (C₆H₁₀O₅)_n, the grade used for thin layer chromatography (TLC), or equivalent, and transfer it quantitatively into a 1 000 ml volumetric flask and make up to the mark with distilled water. Shake the suspension well before use.
It may be stored for at least three months.

NOTE: The dry matter content of the microcrystalline cellulose can be determined by drying a separate sample in an oven at 105 °C ± 2 °C.

6.2 Working reference suspension, ρ = 50 mg/l.

Shake the reference suspension (see 6.1) until it is completely uniform. With a minimum delay measure out (100 ml ± 1 ml) into a 100 ml volumetric flask. Transfer the measured volume quantitatively into a 1 000 ml volumetric flask and make up to the mark with distilled water. Shake the suspension well before use. Prepare a fresh reagent daily.

7 Sampling and sample handling

Obtain samples as described in the relevant sampling guidelines, such as EN 25667-2. Samples shall preferably be taken in bottles of transparent material. Avoid filling the bottles completely to allow efficient mixing by shaking the bottle.

Analyse samples for the determination of suspended solids as soon as possible after sampling, preferably within 4 h. Store samples which cannot be analysed within 4 h in the dark at below 8 °C, but do not allow the sample to freeze. Interpret results obtained for samples that have been stored more than 24 h with caution. Samples for the determination of suspended solids shall not be preserved by any additives.
If the time period from sampling to analysis exceeds 4 h, this shall be stated in the report as well as the conditions of storage.



8 Procedure

8.1 Allow the samples to attain room temperature.

8.2 Ensure that the filters fulfil the requirement that the mass loss shall be less than 0,3 mg per filter.

8.3 Allow a filter to attain moisture equilibrium with the air near the balance and weigh it to the nearest 0,1 mg using the balance (see 5.4). Take care to avoid dust contaminating the filter for example by using a desiccator.

8.4 Place the filter, the smooth side down, in the funnel of the the filtering device (see 5.1) and connect the device to a vacuum (or pressure) line.

WARNING: The evacuation of large glass vessels can cause dangerous implosions if the vessel is damaged by scratches etc. It should be ensured that the relevant safety precautions have been observed.

8.5 Shake the sample bottle vigorously and immediately transfer in one stroke a suitable volume of sample to a measuring cylinder.

If the sample is obtained in a completely filled bottle, mix the sample by transferring it "back and forth" between two bottles. Check that the second bottle is dry and free from contaminants before use.

Select the sample volume so that the dry residue on the filter will be in the optimum range of the determination, which is between 5 mg and 50 mg. However, avoid sample volumes exceeding 1 l. To be valid, the result shall be based on a dry residue of at least 2 mg. Read the sample volume with an accuracy of 2 % or better. Sample volumes of less than 25 ml shall be determined by weighing.

8.6 Filter the sample and rinse the measuring cylinder with about 20 ml of distilled water and use this portion to wash the filter. Rinse the inner sides of the funnel with another 20 ml portion of distilled water.

If the sample contains more than 1000 mg/l of dissolved solids, repeat the washing of the filter with 3 portions each consisting of 50 ml of distilled water. Take care to wash the rim of the filter.

NOTE: The filtering normally is complete within less than 1 min. However, some types of waters contain materials that block the filter pores or reduce their width. This increases the filtering time and the results can become a function of the sample volume. If such blocking of the filter is observed, the determination should be repeated with smaller volumes. The results should be interpreted with caution.

73265166e3cb/sist-en-872-1997

Release the vacuum (or pressure) when the filter is almost dry. Carefully remove it from the funnel with a pair of forceps having flat ends. The filter may be folded if desired. Place the filter on the drying support (see 5.5) and dry it in the oven (see 5.3) at $105\text{ °C} \pm 2\text{ °C}$ for 1 h to 2 h. Remove from the oven and allow the filter to attain equilibrium with the air surrounding the balance and weigh it as before.

9 Check control

Repeat the test procedure (see clause 8) using 200 ml of the working reference suspension (see 6.2) as the sample. The recovery shall be between 90 % and 110 %.

10 Calculation and report

10.1 Calculate the content of suspended solids from the expression

$$\rho = \frac{1000 \cdot (b - a)}{V}$$

where:

ρ is the content of suspended solids, in milligrams per litre;

b is the mass of the filter after the filtration, in milligrams;

a is the mass of the filter before the filtration, in milligrams;

V is the volume of the sample, in millilitres. If the sample has been weighed, consider 1 g as equivalent to 1 ml.

10.2 Report results below 2 mg/l as "below 2 mg/l" and other results in milligrams per litre with two significant figures.

The test report shall include reference to this European Standard and the following particulars:

- a) date and place of testing;
- b) identification mark of the sample tested;
- c) the manufacturer and the designation of the filter used;
- d) the result;
- e) any departure from the procedure (see clause 8) described in this Standard or any other circumstances that can have affected the results, for example blocking of filters (see the note to 8.6) and storage time before analysis.

11 Precision

The precision of data for the content of suspended matter, determined as specified in this standard, depends mainly on the nature of the sample rather than on the method itself. Furthermore, some influence from the particular make of filter used cannot be excluded.

No generally valid data for the reproducibility of the results can be given since it is virtually impossible to perform an interlaboratory study using authentic waters of relevant types with a guarantee that the subsamples are identical at the arrival in different laboratories. Samples containing living organisms or slimy material (for example carbohydrate polymers) that block filters are particularly sensitive to transport and testing conditions.

In an interlaboratory study (Swedish Environmental Protection Agency, 1992) 175 laboratories analysed two samples by a method, essentially the same as in this standard. Synthetic samples prepared with kaolin and microcrystalline cellulose were used. The results summarized in table 1 illustrate the repeatability:

Table 1: Data from a Swedish interlaboratory trial

	Sample 1	Sample 2
Number of accepted results	171	172
Mean	961 mg/l	790 mg/l
Standard deviation	41 mg/l	36 mg/l
Coefficient of variation	4,29 %	4,72 %
Rejected results	4	3

In a German study, reported in 1994, the results indicated in table 2 were obtained:

Table 2: Data from a German interlaboratory trial

Sample	Number of laboratories	Accepted results	Rejected results	Mean, mg/l	Coefficient of variation, in %
Microcrystalline cellulose, 10 mg/l	8	32	0	9,95	9,0
Microcrystalline cellulose, 100 mg/l	8	32	0	96,6	7,8
Microcrystalline cellulose, 500 mg/l	7	26	6	499,5	3,5
Kaolin, 10 mg/l	8	30	2	9,59	7,8
Kaolin, 100 mg/l	8	32	0	92,4	7,5
Kaolin, 500 mg/l	8	31	1	463,4	6,0
Fibrous cellulose (pulp), 10 mg/l	7	28	0	10,3	8,3
Fibrous cellulose (pulp), 100 mg/l	6	24	4	102,3	9,2
Fibrous cellulose (pulp), 500 mg/l	7	28	0	474,1	13,6
Microcrystalline/kaolin, 1/1, 10 mg/l	8	31	1	9,6	4,4
Microcrystalline/kaolin, 1/1, 50 mg/l	8	30	2	49,9	1,7
Microcrystalline/kaolin, 1/1, 200 mg/l	8	32	0	195,4	3,3

The results illustrate the reproducibility that can be expected under ideal conditions, i.e. with samples containing inert particles of well defined sizes. Ideal conditions can, however, rarely be expected with authentic samples.