



Designation: ~~E2808--21~~ E2808 – 21a

## Standard Guide for Microspectrophotometry in Forensic Paint Analysis<sup>1</sup>

This standard is issued under the fixed designation E2808; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### INTRODUCTION

Color is one of the most important comparative characteristics of paints. The comparison of color is one of the first steps taken in a forensic paint comparison. Subjective terms such as “blue,” “violet,” or “purple” are descriptors of color but are inadequate for clear communication of color as terms could suggest different colors to different people. It is essential to note that this guide does not propose the use of instrumental color comparison for objects that are distinguishable to the eye. Since the 1940s, analytical instruments have been able to discriminate colors that the average human eye cannot distinguish. Microspectrophotometers (MSPs), in particular, allow for an objective measurement of the color of small, millimetre or submillimetre samples and are generally more sensitive than the more subjective results of visual microscopical color comparisons.

Suitable instruments with appropriate optics, sensitivity, resolution, and dynamic range can measure spectral curves from small samples as that light is transmitted, absorbed, reflected, or emitted (by means of fluorescence) by the sample. The spectral limits of different instruments can vary and can extend from the ultraviolet (UV) (~190 to 380 nm) through the visible spectral region (~380 to 780 nm) to the near infrared region (NIR) (~780 to 2500 nm). MSPs should not be confused with broadband or absorption filter-based tristimulus systems.

### 1. Scope

1.1 This guide is intended to assist forensic analysts who conduct UV, visible, NIR, or fluorescence emission spectral analyses on small fragments of paint or use Guide **E1610**, as this guide is to be used in conjunction with a broader analytical scheme.

1.2 This guide deals primarily with color measurements within the visible spectral range but will also include some details concerning measurements in the UV and NIR spectral ranges. The particular method(s) employed by each analyst depends upon available equipment, examiner training (Practices **E2917**, **E3234**), sample suitability, and sample size.

1.3 This guide provides basic recommendations and information about microspectrophotometers.

1.4 This guide does not address other areas of color evaluation such as colorimetric values, paint surface texture or pigment particle size, shape, or dispersion within a paint film that are evaluated by other forms of microscopy.

1.5 This guide is directed at the color analysis of commercially prepared paints and coatings. It does not address the analysis or determination of provenance of artistic, historical, or restorative paints, but it could be useful in those fields.

<sup>1</sup> This guide is under the jurisdiction of ASTM Committee **E30** on Forensic Sciences and is the direct responsibility of Subcommittee **E30.01** on Criminalistics. Current edition approved March 1, 2021/Sept. 1, 2021. Published April 2021/September 2021. Originally approved in 2011. Last previous edition approved in 2019/2021 as E2808 – 1921<sup>ε1</sup>. DOI: [10.1520/E2808-21](https://doi.org/10.1520/E2808-21); [10.1520/E2808-21A](https://doi.org/10.1520/E2808-21A).

1.6 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.7 *This standard is intended for use by competent forensic science practitioners with the requisite formal education, discipline-specific training (see Practices E2917, E3234), and demonstrated proficiency to perform forensic casework.*

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

**iTeh Standards**  
**(<https://standards.itih.ai>)**  
**Document Preview**

[ASTM E2808-21a](#)

<https://standards.itih.ai/catalog/standards/sist/63186da6-cefc-4b29-b12e-66d5918e828f/astm-e2808-21a>

1.9 This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>2</sup>

- D16 Terminology for Paint, Related Coatings, Materials, and Applications
- E179 Guide for Selection of Geometric Conditions for Measurement of Reflection and Transmission Properties of Materials
- E275 Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers
- E284 Terminology of Appearance
- E1610 Guide for Forensic Paint Analysis and Comparison
- E1492 Practice for Receiving, Documenting, Storing, and Retrieving Evidence in a Forensic Science Laboratory
- E2917 Practice for Forensic Science Practitioner Training, Continuing Education, and Professional Development Programs
- E3234 Practice for Forensic Paint Analysis Training Program

## 3. Terminology

3.1 *Definitions*—For definitions of paint-associated terminology used in this guide, see Terminologies D16 and E284, Guide E1610, and Practice E3234.

### 3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *charge-coupled device (CCD), n*—a silicon-based semiconductor chip consisting of a linear or two-dimensional array of photo sensors or pixels that transfers an electrical charge and converts it into a digital value.

3.2.2 *effect pigment, n*—any paint pigment that is designed to produce a significant change in color attribute(s) in a paint film when the film is viewed or illuminated from varied angles.

3.2.3 *exclusionary difference, n*—a difference in a feature or property between compared items that is substantial enough to conclude that they did not originate from the same source.

#### 3.2.3.1 *Discussion*—

An exclusionary difference is statistically supported when an appropriate statistical analysis shows a result outside the range of what usually occurs when the items originate from the same source.

#### 3.2.3.2 *Discussion*—

When a statistical analysis is not suitable, an exclusionary difference can be determined by expert judgement.

3.2.4 *grating, n*—parallel set of linear, regularly repeating grooves that, when illuminated, produces dispersion of light into its requisite wavelengths with maxima and minima of light intensity as a consequence of interference.

#### 3.2.4.1 *Discussion*—

These maxima and minima vary in position with wavelength. This allows radiation of any given wavelength to be isolated from a mixture of wavelengths (for example, white light) and allows the grating to be used as part of a monochromator. The dispersion or ability to resolve separate wavelengths is expressed as the number of lines (or steps) in the grating per millimetre.

3.2.5 *measuring aperture, n*—element in the optical path of a MSP system that limits the area of illumination reaching the detector focal plane.

3.2.6 *metameric samples, n*—two or more samples that appear to have the same color under one type of illumination but can appear dissimilar under different lighting conditions, or two or more samples that appear to be the same color under all lighting conditions, yet their reflectance/transmittance spectral curves are different.

3.2.7 *microspectrophotometer (MSP), n*—a specialized spectrophotometer designed to measure the absorbance, reflectance and emission spectra of microscopic areas on samples.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3.2.8 *monochromator, n*—device designed to isolate narrow wavelength ranges of light from complex, broad-spectrum radiation.

3.2.9 *photomultiplier tube (PMT), n*—photosensitive vacuum tube device that quantitatively converts photons of light into electrical energy.

3.2.10 *pigment, n*—a finely ground, organic or inorganic, insoluble, and dispersed particle. Besides color, a pigment can provide many of the essential properties of paint, such as opacity, hardness, durability, and corrosion resistance (see also *effect pigment*).

3.2.11 *pixel binning, v*—the process of combining counts from adjacent pixels in a CCD detector during readout.

3.2.12 *spectral resolution, n*—measure of the ability to distinguish between adjacent peaks in a spectrum; it is usually determined by measuring peak width at half the maximum value of the peak height or full-width half-maximum (FWHM).

3.2.12.1 *Discussion*—

Spectral resolution is not to be confused with spatial resolution (the smallest features that can be resolved in the field of view of the MSP camera or eyepieces or can be used to refer to the smallest spectral sampling area of the MSP).

## 4. Summary of Guide

4.1 One of the most obvious properties of paint is its color. Pigments are used in paint to modify color or other properties. The interaction of pigments with light is very complex with light being scattered, absorbed, reflected, and emitted by a paint layer. A MSP can measure reflected, transmitted, or emitted radiation over a range of wavelengths.

4.2 MSP is an integrated instrument consisting of a microscope, a light source, a spectrophotometer, and a data-processing device. The microscope not only allows for analysis location selection but also reflects or transmits light energy efficiently, uniformly, and reproducibly and images light onto the spectrophotometer entrance aperture. The spectrophotometer contains a dispersive element, most commonly a diffraction grating, and a detector. The system is designed to measure the intensity of light energy with respect to its wavelength. All MSPs are single-beam instruments: a standard or a blank is measured, the result is stored, then a sample is measured and a ratio made to the standard to yield a transmittance or reflectance spectrum.

4.3 MSP analysis typically includes the visible spectral region (~380 to 780 nm), which captures information about the visual color of a sample. Most MSP systems are also sensitive to the NIR (~780 to 1100 nm). For UV-configured systems, the UV region (~190 to 380 nm) can provide additional information, for example, about UV absorbers in a clear coat layer. Furthermore, the spectrum of fluorescence emission (UV and visible excitation with UV to NIR emission) can also be captured.

4.4 MSP systems are generally used in forensic analysis because of the small sample sizes presented by paint film fragments. MSP is typically minimally destructive and microscopic samples can be analyzed. Instruments for color measurement from large samples can be used for color comparison, and can also be used for compiling color databases. Instruments for macroscopic color measurements produce results that are not necessarily directly comparable to MSP results.

4.5 Paint colors are usually measured in transmittance through thin sections. In transmittance measurements, a paint thin section is illuminated and the fraction of light transmitted or absorbed by the sample in the spectral range of interest (UV-Vis-NIR) is recorded relative to the light transmitted by a colorless background. Transmittance spectra can be plotted in either percent transmittance or absorbance.

4.6 In reflectance measurements of paint, a sample is illuminated and the fraction of light reflected from the sample in the spectral range of interest (UV-Vis-NIR) is recorded relative to the light reflected from a white reference standard. Reflectance spectral measurements are greatly affected by surface observation angle, surface texture, and the lack of microscopically reproducible diffuse reflectance standards. As a result, reflectance is seldom used for detailed color comparison, but it can be useful in exclusionary comparisons of bulk colors or for opaque samples. Reflectance spectra are plotted in percent reflectance.

## 5. Significance and Use

5.1 This guide is designed to assist an analyst in the selection of appropriate sample preparation methods and instrumental parameters for the analysis and comparison of paint pigments and colors. When used for comparison purposes, the goal is to determine whether any meaningful exclusionary differences exist between the samples.

5.2 Paint sample spectra can be measured by reflectance or transmittance spectroscopy for comparison purposes. Transmittance measurements are generally preferred and are required for the analysis of UV absorbers in clear coats and the detailed analysis of effect pigments that are not opaque. Emission comparison by means of fluorescence is also measurable.

5.3 It is not the intention of this guide to present comprehensive theories and methods of MSP. It is necessary that the analyst have an understanding of UV-Vis-NIR MSP and general concepts of specimen preparation before using this guide. This information is available from manufacturers' reference materials, training courses, and references such as Eyring (1),<sup>3</sup> Stoecklein (2), and Purcell (3).

## 6. Sample Preparation

6.1 The general collection, handling, and tracking of samples should meet or exceed the requirements of Practice E1492.

6.2 Verify that the work area and tools used for the preparation of samples are free of all extraneous materials that could transfer to the sample.

### 6.3 *Transmittance Measurements:*

6.3.1 Prepare hand-cut or microtome-cut thin sections for transmittance measurements. Although embedding followed by microtomy could be considered labor-intensive, this method (when questioned and known samples are mounted together) produces samples with an equivalent path length. Samples with equivalent path lengths are preferred for all transmission measurements and required for analysis of UV absorbers when the comparison relies upon differences in relative concentrations (4).

6.3.2 Paint sections of approximately 3  $\mu\text{m}$  thick are generally appropriate for measurements of pigmented layers. Lightly colored samples or clear coats could require thicker sections (for example, up to 20  $\mu\text{m}$  thick) (4, 5) for increased sensitivity of the instrument to low concentrations of UV absorbers or pigments.

6.3.3 Photometric reproducibility could be difficult to achieve with thin peels of paint layers or hand-cut cross sections due to uncontrolled path lengths within and between samples. This can lead to difficulties in the interpretation of shade (that is, intensity) differences when spectral curves do not exhibit exclusionary features.

### 6.4 *Reflectance Measurements:*

6.4.1 Conduct specular reflection measurements on clean and undamaged sample areas. Clean the surface of the sample carefully with deionized water, or alcohol if necessary. Reflectance measurements are usually conducted without further preparation of the sample surface so that surface features (for example, weathering) that might distinguish between samples are not altered. In some cases, it will be necessary to remove surface features. Surface weathering, for example, can be removed by polishing the paint using a diamond paste or another polishing medium so the measured properties of the surface are comparable to data for new, undamaged paint surfaces. However, caution is recommended before removal of features as the potential for other evidential loss increases. Retain a portion of the original unaltered sample whenever possible.

6.4.2 To obtain optimal results during spectral measurements, prepare the surface so it is free of scratches, dirt, and blemishes and mounted perpendicular to the optical axis of the microscope. Obtain a sharp focus on the paint sample surface.

6.4.3 Conduct cross section reflectance measurements on polished or microtome cut surfaces. Prepare samples for comparison in the same manner. Simplify and enhance reproducibility of the surface finish by mounting and polishing edge-mounted known and questioned samples side by side.

6.4.4 When the signal-to-noise ratio of a reflectance measurement is poor due to small sample size, prepare and analyze a thin cross section by transmittance.

6.5 For paint smears, different portions of the smear can be separated and mounted for analysis. These results can provide

---

<sup>3</sup> The boldface numbers in parentheses refer to a list of references at the end of this standard.

conclusive exclusionary data but could lack meaningful associative information. Smear paints are difficult for reflectance measurements; analyzing smears by transmittance methods is recommended.

## 6.6 Mounting Samples:

6.6.1 Mount and prepare questioned and known samples in the same manner.

6.6.2 For reflectance measurements, record the angle of illumination and collection.

6.6.3 For transmittance measurements, mount the thin section or smear particles on a microscope slide under a coverslip in an appropriate refractive index medium (typically one that is close to that of the sample).

6.6.3.1 Use a microscope slide with transmittance characteristics appropriate for the region of the spectrum being analyzed. Glass (borosilicate) slides and coverslips are suitable for measurements in the visible and NIR portions of the spectrum. Mount samples to be analyzed in the UV region of the spectrum on quartz or fused silica; however, not all slides have the same UV transmittance characteristics.

6.6.3.2 Add a mounting medium between the slide and coverslip for transmittance measurements. Select a mounting medium that is compatible with the sample (that is, will not dissolve the sample) and with the spectral range being investigated (that is, glycerol is transparent to the UV region of the spectrum). Mounting media include, but are not limited to, water, xylene, xylene substitutes, glycerol, refractive index oils ( $n = 1.52$  or  $1.66$  are common).

## 7. Performance Checks

7.1 Prior to use of the instrument, turn on the microscope, illumination sources, and spectrometer and allow them to stabilize. Allow the lamp to warm up and stabilize according to manufacturers' instructions or laboratory experience, whichever yields consistent results.

7.2 Checking instrument performance verifies that an instrument is operating within required standards and any errors that might affect data or analytical conclusions are known, limited, accounted for, or corrected. It is essential to demonstrate wavelength and absorbance/photometric accuracy through a performance check, such as that described in Practice E275.

7.2.1 Conduct a performance check daily, prior to analyses.

7.2.2 A performance check should be conducted prior to use after any maintenance.

7.2.3 Use a similar configuration each time a performance check is conducted on the system to ensure that historical performance check data are comparable.

7.3 Maintain a record of all performance checks. A historical record of this data provides a mechanism for monitoring system performance and provides an operator with an early warning of system trends and deterioration.

7.4 Performance check parameters include:

7.4.1 *Wavelength Accuracy*—Routinely check wavelength accuracy over the measured range with the aid of National Institute of Standards and Technology (NIST)-traceable holmium, erbium, or didymium oxide filters. The resolution used during the wavelength accuracy checks should be the same or higher than that used in casework and consistent for each wavelength accuracy check. Transmittance is used for these measurements.

7.4.2 *Photometric Accuracy*—Neutral density filters are used to demonstrate the photometric response of the system is linear. A typical set of neutral density calibration filters could include some or all of the following filters: 0.1, 0.5, 1.0, 2.0, 2.5, and 3.0 absorbance units.

## 8. Instrument and Scanning Parameters

8.1 Microscope parameters include:

8.1.1 *Illuminator*—Select an illuminator appropriate to the analysis being conducted. Select the emission spectrum of the illuminator that has sufficient intensity across the entire wavelength range of interest so as to provide a spectrum with an acceptable signal-to-noise ratio. Tungsten, halogen, and xenon are commonly used for visible and NIR analysis. Commonly, xenon lamps are used for UV analysis and mercury lamps are used for fluorescence excitation. While LED illuminators are available over much of the spectrum, they are of little utility for MSP due to their lower intensity and limited spectral range (3).

8.1.1.1 Instrument design with regard to lighting can affect the ease of collecting reflectance measurements. Epi-illumination results in specular reflectance of the illuminating light, causing a loss of detail in the viewed image, making layer distinction difficult to impossible at times. The illumination/measuring geometry in reflectance analysis of  $45^{\circ}/0^{\circ}$  is preferred such that the specularly reflected portion is not gathered by the microscope. This geometry can be best achieved with the help of a darkfield/brightfield illuminator fitted with darkfield objectives. However, such illumination excludes UV measurements as UV transmitting darkfield objectives are not available. A general discussion of geometric considerations for reflectance and transmittance measurements is found in Guide E179.

8.1.1.2 Background/system/reference transmittance spectra can be used to monitor illuminator performance and warn of unsuitable system alignment.

8.1.1.3 *Illumination Centration*—Slight adjustments to the position of the bulb can serve to increase or decrease the emission over specific regions of the spectrum. For example, it is possible to maximize UV illumination, often at the expense of some light in the visible wavelengths. Generally, the slight loss of intensity in the visible region is not problematic due to the high intensity of modern bulbs.

8.1.1.4 *Illumination Intensity*—For some illuminators, this can be a fixed parameter. When the voltage of an illuminator is adjustable, it should be held fixed following the photometric intensity performance check (7.4.2).

8.1.2 *Field Diaphragm*—To minimize stray light, with the specimen in focus, the edges of the field diaphragm should be brought into view and sharply focused with adjustment of the substage condenser. The field diaphragm is then opened so that its edges are either just outside the collection aperture or just beyond the field of view. Readjust the focus and size of the aperture when the objective is changed.

8.1.3 *Substage Aperture (that is, Condenser Iris)*—The substage aperture should be opened until the desired image contrast is obtained. As adjustment of this aperture impacts the amount of light reaching the detector, this aperture setting could be different from that used to produce an ideal image (6). In some instances, it is desirable to further increase the opening of the substage aperture to allow more signal to reach the detector. Aperture adjustment should occur prior to the collection of the background spectrum, and should remain in a fixed position between the background and sample spectra collection and between samples when they are being compared. The appropriate aperture level is typically that which produces an emission intensity for the most intense emission peak of no more than approximately 80 % of the detector saturation value. It is critical that the detector is not saturated anywhere over the region being measured.

8.1.4 *Objective*—Select an objective that permits imaging of the paint layer or inclusion to be analyzed. A balance between the objective magnification and size of the measuring aperture is selected by the analyst. Typically, measurements are made using objectives between 10× and 50×. Quartz optics are required for measurements made in the UV region. Once the appropriate objective is selected, measure all samples being compared at a fixed magnification.

8.1.5 *Measuring Aperture*—If the MSP system is equipped with variable or multiple collection apertures, select and use the same aperture for the measurement of all samples being compared. In general, select the largest aperture that will remain within the boundaries of the sample area to be measured. This is particularly important for paints with effect pigments, which show variation over larger areas. The edge of a layer or sample should be avoided due to edge effects that could impact the spectrum. For smears and thin peels, an oversized aperture (one that extends beyond the boundary of the sample) is undesirable in transmittance measurements as it increases the noise in the spectrum. In fluorescence emission, the spectrum is not the result of a ratio to a reference scan; the strength of the signal is determined by absolute counts. Therefore, an oversized aperture when analyzing smears or thin peels for fluorescence emission can be used to increase the signal reaching the detector. As the background is black, there is no significant increase in noise to detract from the quality of the collected data.