



**SLOVENSKI STANDARD**  
**SIST EN ISO 14402:2000**  
**01-december-2000**

---

Water quality - Determination of phenol index by flow analysis (FIA and CFA)(ISO 14402:1999)

Water quality - Determination of phenol index by flow analysis (FIA and CFA)(ISO 14402:1999)

Wasserbeschaffenheit - Bestimmung des Phenolindex mit der Fließanalytik (FIA und CFA) (ISO 14402:1999)

Qualité de l'eau - Détermination de l'indice phénol par analyse en flux (FIA et CFA) (ISO 14402:1999)

**ITeh STANDARD PREVIEW**  
**(standards.iteh.ai)**  
<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ca780cab8a8f/sist-en-iso-14402-2000>

**Ta slovenski standard je istoveten z: EN ISO 14402:1999**

---

**ICS:**

13.060.70	Preiskava bioloških lastnosti vode	Examination of biological properties of water
-----------	------------------------------------	---

**SIST EN ISO 14402:2000**

**en**

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN ISO 14402:2000

<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000>

EUROPEAN STANDARD

EN ISO 14402

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 1999

ICS

English version

## Water quality - Determination of phenol index by flow analysis (FIA and CFA) (ISO 14402:1999)

Qualité de l'eau - Détermination de l'indice phénol par  
analyse en flux (FIA et CFA) (ISO 14402:1999)

Wasserbeschaffenheit - Bestimmung des Phenolindex mit  
der Fließanalytik (FIA und CFA) (ISO 14402:1999)

This European Standard was approved by CEN on 23 July 1999.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

SIST EN ISO 14402:2000

<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

## Foreword

The text of the International Standard ISO 14402:1999 has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2000, and conflicting national standards shall be withdrawn at the latest by March 2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

**NOTE FROM CEN/CS:** The foreword is susceptible to be amended on reception of the German language version. The confirmed or amended foreword, and when appropriate, the normative annex ZA for the references to international publications with their relevant European publications will be circulated with the German version.

## Endorsement notice

The text of the International Standard ISO 14402:1999 was approved by CEN as a European Standard without any modification.

**(standards.iteh.ai)**

[SIST EN ISO 14402:2000](https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000)

<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000>

STANDARDS  
OPERATIONAL  
INTERNATIONAL

# INTERNATIONAL STANDARD

**ISO**  
**14402**

First edition  
1999-09-01

---

---

## Water quality — Determination of phenol index by flow analysis (FIA and CFA)

*Qualité de l'eau — Détermination de l'indice phénol par analyse en flux (FIA  
et CFA)*

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 14402:2000](https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000)

[https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-  
ea780cab8a8f/sist-en-iso-14402-2000](https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000)



Reference number  
ISO 14402:1999(E)

## ISO 14402:1999(E)

Contents	Page
1 Scope .....	1
2 Normative references .....	1
3 Determination of phenol index (without distillation) after extraction .....	1
3.1 Principle.....	1
3.2 Interferences .....	1
3.3 Reagents .....	2
3.4 Apparatus .....	4
3.5 Sampling.....	6
3.6 Procedure .....	7
3.7 Calculation of results .....	8
4 Determination of phenol index (without extraction) after distillation .....	9
4.1 Principle.....	9
4.2 Interferences .....	9
4.3 Reagents.....	9
4.4 Apparatus .....	10
4.5 Sampling.....	13
4.6 Procedure .....	13
4.7 Calculation of results .....	14
5 Expression of results .....	14
6 Precision and accuracy.....	14
7 Test report .....	14
Annex A (informative) Statistical data.....	15
Bibliography .....	17

iTeh STANDARD PREVIEW  
(standards.iteh.ai)

[SIST EN ISO 14402:2000](https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ca780cab8a81/sist-en-iso-14402-2000)

<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ca780cab8a81/sist-en-iso-14402-2000>

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet iso@iso.ch

Printed in Switzerland

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 14402 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical biochemical methods*.

Annex A of this International Standard is for information only.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

[SIST EN ISO 14402:2000](https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000)

<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000>

## Introduction

Methods for determination of water quality using flow analysis and automatic wet chemical procedures are particularly suitable for the processing of large sample series at a high analysis frequency.

Differentiation is needed between flow injection analysis (FIA) [1, 2] and continuous flow analysis (CFA) [3]. Both methods include automatic dosage of the sample into a flow system (manifold) where the analytes in the sample react with the reagent solutions on their way through the manifold. The sample preparation may be integrated in the manifold. The reaction product is measured in a flow detector.

Phenol index is an analytical convention. It represents a group of aromatic compounds which under the specific reaction conditions form coloured condensation products. The analytical result is expressed in terms of phenol concentration.

This International Standard describes two methods: the determination of phenol index (without distillation) after extraction, and the determination of phenol index (without extraction) after distillation.

It should be investigated whether and to what extent particular problems will require the specification of additional marginal conditions.

## iTeh STANDARD PREVIEW (standards.iteh.ai)

SIST EN ISO 14402:2000

<https://standards.iteh.ai/catalog/standards/sist/82f54436-6e9b-4d90-a7fd-ea780cab8a8f/sist-en-iso-14402-2000>



# Water quality — Determination of phenol index by flow analysis (FIA and CFA)

## 1 Scope

This International Standard specifies two methods for the determination of the phenol index in waters of different origin (such as ground waters, surface waters, seep waters, and waste waters) in mass concentrations of 0,01 mg/l to 1 mg/l (in the undiluted sample). In particular cases, the range of application may be adapted by varying the operating conditions. Clause 3 describes the determination of phenol index (without distillation) after extraction, and in clause 4 the determination of phenol index (without extraction) after distillation is given.

## 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this International Standard. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

SIST EN ISO 14402:2000

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*.

ISO 5667-3:1994, *Water quality — Sampling — Part 3: Guidance on sample handling and preservation*.

ISO 6439:1990, *Water quality — Determination of phenol index — 4-Aminoantipyrine spectrometric methods after distillation*.

## 3 Determination of phenol index (without distillation) after extraction

### 3.1 Principle

The sample is fed into a continuously flowing carrier stream and mixed with also continuously flowing solutions of 4-aminoantipyrine and potassium peroxydisulfate. Phenolic compounds in the sample are oxidized by potassium peroxydisulfate, and the resulting quinones react with 4-aminoantipyrine, forming coloured condensation products. These are extracted in a flow extraction unit from the aqueous phase into chloroform. The chloroform phase is separated by a suitable phase separator (e.g. a hydrophobic semipermeable membrane), and the absorbance of the organic phase is measured spectrometrically in a flow spectrometer at 470 nm to 475 nm. More information on this analytical technique is given in the references [6 to 9].

It is absolutely essential that the test described in this International Standard be carried out by suitably qualified staff.

### 3.2 Interferences

#### 3.2.1 Chemical interferences

Under the prevailing reaction conditions, aromatic amines will also form condensation products with 4-aminoantipyrine, leading to positive bias.

Interferences can occur when the sample, after the addition of the reagent solutions, does not reach a pH of 10,0 to 10,5. In particular this may occur in the cases of strongly acidic, strongly alkaline and buffered samples. In these cases, the sample is adjusted to a pH between 5 and 7 prior to addition of the reagent solutions.

Further information on interferences is given in [5].

### 3.2.2 Physical interferences arising from applying CFA and FIA

If the samples contain particulate matter, refer to 3.5 (last paragraph). Turbid samples do not cause interferences with the determination. In the event of coloured samples, check whether the colour can be extracted with chloroform, and determine the sample blank without the addition of reagents R1 and R2. The difference in response between the two measurements shall be taken into account with the evaluation (according to 3.7).

The interlaboratory trial (see clause 6 and annex A) has shown that detergents in waste water can strongly influence the determination, because the foam produced in the flow system can disturb on the one hand the steam distillation of volatile phenols (phenol index after distillation, see clause 4, and on the other hand the phase segmentation and phase separation procedures (phenol index after extraction, see clause 3). In general such interferences can easily be discovered.

In the case of significant detergent content, this International Standard is only applicable for phenol mass concentrations above 0,1 mg/l.

### 3.3 Reagents

Use only reagents of recognized analytical grade quality. The reagent blank value shall regularly be checked (see 3.6.3). The solutions used for the flow system shall be degassed. If not stated otherwise, it is recommended to degas the solutions under reduced pressure, because by this procedure the solutions are simultaneously purified.

**WARNING — Phenol is toxic and can easily be absorbed through the skin. Chloroform is toxic and cancerogenic. Waste containing these substances should be disposed of appropriately.**

**3.3.1 Water**, of grade 1 in accordance with ISO 3696

**3.3.2 Potassium hydroxide**, KOH

**3.3.3 Sodium hydrogencarbonate**, NaHCO<sub>3</sub>

**3.3.4 4-aminoantipyrine** (4-amino-2,3-dimethyl-1-phenyl-3-pyrazolin-5-one), C<sub>11</sub>H<sub>13</sub>N<sub>3</sub>O

**3.3.5 Potassium peroxydisulfate**, K<sub>2</sub>S<sub>2</sub>O<sub>8</sub>

**3.3.6 Phenol**, C<sub>6</sub>H<sub>5</sub>OH

**3.3.7 Boric acid**, H<sub>3</sub>BO<sub>3</sub>

**3.3.8 Ethanol**, C<sub>2</sub>H<sub>5</sub>OH, 96 % mass fraction

**3.3.9 2-Propanol**, C<sub>3</sub>H<sub>7</sub>OH, 100 % mass fraction

**3.3.10 Sulfuric acid**,  $\rho(\text{H}_2\text{SO}_4) = 1,84 \text{ g/ml}$

**3.3.11 Hydrochloric acid**, HCl, 50 % mass fraction

**3.3.12 Potassium hydroxide solution**,  $c(\text{KOH}) = 1,0 \text{ mol/l}$

**3.3.13 Buffer solution**

Dissolve in a 1 000 ml graduated flask in approximately 500 ml of water (3.3.1): 23 g of sodium hydrogencarbonate (3.3.3), 27 g of boric acid (3.3.7), and 35 g of potassium hydroxide (3.3.2) and make up to volume with water.

The pH of the buffer solution is approximately 10,3. The solution is stable for 1 month.

**3.3.14 Carrier solution** (symbol C in Figure 1)

Use water (3.3.1) degassed under reduced pressure.

**3.3.15 4-Aminoantipyrine solution I** (symbol R1 in Figures 1 and 2)

Dissolve in a 100 ml graduated flask 0,5 g of 4-aminoantipyrine (3.3.4) in approximately 50 ml of buffer solution (3.3.13), and make up to volume with buffer solution (3.3.13).

Degas the solution, e.g. by membrane filtration.

Prepare fresh solution every day.

**3.3.16 Potassium peroxodisulfate solution** (symbol R2 in Figures 1 and 2)

Dissolve in a 100 ml graduated flask 5 g of potassium peroxodisulfate (3.3.5) in approximately 90 ml of water (3.3.1), adjust to pH 11 with potassium hydroxide solution (3.3.12) and make up to volume with water.

Degas the solution, e.g. by membrane filtration.

Prepare fresh solution daily.

**3.3.17 Chloroform**,  $\text{CHCl}_3$  (symbol Org in Figures 1 and 2)

Degas the chloroform solution either by membrane filtration or for 3 min in an ultrasonic bath.

**3.3.18 Phenol stock solution**,  $\rho = 1\,000\text{ mg/l}$ 

Dissolve in a 1 000 ml graduated flask 1,000 g of phenol (3.3.6) in water (3.3.1) and make up to volume with water. Use only colourless phenol crystals.

The cooled solution (2 °C to 5 °C) is stable for one month.

**3.3.19 Phenol standard solution I**,  $\rho = 10\text{ mg/l}$ 

Pipette 1 ml of the stock solution (3.3.18) into a 100 ml graduated flask, and make up to volume with water (3.3.1).

The cooled solution (2 °C to 5 °C) is stable for one week.

**3.3.20 Phenol standard solution II**,  $\rho = 1\text{ mg/l}$ 

Pipette 10 ml of the standard solution I (3.3.19) into a 100 ml graduated flask, and make up to volume with water (3.3.1).

The cooled solution (2 °C to 5 °C) is stable for one week.

**3.3.21 Calibration solutions**

Prepare the calibration solutions according to the origin of the sample and the expected concentrations by diluting the phenol standard solution I or II respectively (3.3.19 or 3.3.20).

Prepare a minimum of at least five calibration solutions per working range.

Proceed as follows for working ranges I and II, if using e.g. six calibration solutions:

## a) Working range I, (0,1 mg/l to 1 mg/l):

Into each of a series of 100 ml graduated flasks pipette 1 ml, 3 ml, 5 ml, 6 ml, 8 ml and 10 ml respectively of the standard solution I (3.3.19), and make up to volume with water (3.3.1).

The concentration of phenol in these calibration solutions is 0,1 mg/l, 0,3 mg/l, 0,5 mg/l, 0,6 mg/l, 0,8 mg/l and 1,0 mg/l, respectively.