



Designation: D7707 – 21

# Standard Specification for Wipe Sampling Materials for Beryllium in Surface Dust<sup>1</sup>

This standard is issued under the fixed designation D7707; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers requirements for premoistened wipe materials that are used to collect settled dusts on surfaces for the subsequent determination of beryllium.

1.2 For wipe materials used for the determination of lead in surface dust, refer to Specification E1792. This is mentioned to insure that users of wipes recognize that there is some relationship between wipes and the analyte of interest.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 This specification contains notes that are explanatory and are not part of the mandatory requirements of the specification.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

D1356 Terminology Relating to Sampling and Analysis of Atmospheres

D6966 Practice for Collection of Settled Dust Samples Using Wipe Sampling Methods for Subsequent Determination of Metals

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee D22 on Air Quality and is the direct responsibility of Subcommittee D22.04 on Workplace Air Quality.

Current edition approved Sept. 1, 2021. Published October 2021. Originally approved in 2011. Last previous edition approved in 2016 as D7707 – 11 (2016). DOI: 10.1520/D7707-21.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

D7035 Test Method for Determination of Metals and Metalloids in Airborne Particulate Matter by Inductively Coupled Plasma Atomic Emission Spectrometry (ICP-AES)

D7202 Test Method for Determination of Beryllium in the Workplace by Extraction and Optical Fluorescence Detection

D7439 Test Method for Determination of Elements in Airborne Particulate Matter by Inductively Coupled Plasma–Mass Spectrometry

E105 Guide for Probability Sampling of Materials

E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

E1792 Specification for Wipe Sampling Materials for Lead in Surface Dust

## 3. Terminology

3.1 For definitions of pertinent terms not listed here, see Terminology D1356.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *lot, n*—a finite quantity of a given product manufactured under production conditions that are considered uniform.

3.2.2 *standard beryllium test method, n*—as used in this specification, any of the following test methods: D7035, D7202, D7439, the graphite furnace atomic absorption technique in NIOSH Method 7102,<sup>3</sup> or a equivalent test method issued by a national or international consensus standard organization.

3.2.3 *wipe, n*—a disposable towellette that may or may not be premoistened with an acceptable wetting agent, depending on the intended application for obtaining beryllium samples.

3.2.3.1 *Discussion*—The towellette is used to collect a sample of settled dust on a surface for subsequent beryllium analysis.

## 4. Significance and Use

4.1 This specification is intended for use by manufacturers and suppliers to evaluate the performance of wipe sampling materials for beryllium in surface dust.

<sup>3</sup> National Institute for Occupational Safety and Health, *NIOSH Manual of Analytical Methods*, 4th ed., NIOSH, Cincinnati, OH, 1994, available from www.cdc.gov/niosh/nmam.

4.2 This specification may also be employed by users of wipes to compare the performance of candidate wipes for the sampling of beryllium in surface dust.

## 5. Manufacture

5.1 The wipes shall be made from materials using methods that ensure compliance with the requirements of Sections 6 and 8, and shall be clean and free of imperfections that would affect their performance.

5.2 An expiration date shall be established based on date of manufacture, and shall be marked on each package (see 10.2). The expiration date should not be less than one year from date of manufacture.

## 6. General Requirements

6.1 Wipes shall conform to the requirements in 6.1.1 – 6.1.8. Test procedures for each requirement are found in Section 8.

6.1.1 *Background Beryllium*—The mean background beryllium content per wipe shall be less than 0.0005  $\mu\text{g}$ .

6.1.2 *Ruggedness*—Wipes shall be sufficiently rugged so as to be capable to be used to wipe a 2000  $\text{cm}^2$  area (large wipes; see 6.1.4.1(1)) or 500  $\text{cm}^2$  area (small wipes; see 6.1.4.1(2)) of a vinyl or urethane-coated vinyl tile surface without tearing.

6.1.3 *Moisture Content*—Each wipe, when examined, must be fully wetted upon removal from the package or container. Wipes shall have a moisture content such that the coefficient of variation for a random sampling of the lot of wipes be no greater than 25 % across individually packaged wipes, and 25 % within and between multi-packaged wipes.

6.1.4 *Dimensions*:

6.1.4.1 The mean area of the wipe shall be either:

(1) No smaller than 200  $\text{cm}^2$  and no larger than 400  $\text{cm}^2$  with the mean dimension of a side or diameter no smaller than 10 cm or larger than 20 cm (referred to as the “large wipe”); or,

(2) No smaller than 16  $\text{cm}^2$  and no larger than 38  $\text{cm}^2$  with the mean dimension of a side or diameter no smaller than 4.0 cm or larger than 8.0 cm (referred to as the “small wipe”).

NOTE 1—Large wipes are intended for use on larger areas (for example, 1000  $\text{cm}^2$ ). Small wipes are intended for use on smaller areas (for example, 100  $\text{cm}^2$ ).

6.1.5 *Thickness*—Dry wipes (wipes without the moistening agent) shall have a mean thickness of at least 0.005 cm, but no greater than 0.05 cm (see 8.6).

6.1.6 *Mass*—The coefficient of variation in mass of dry wipes in a lot shall not exceed 10 %.

6.1.7 *Beryllium Recoveries*—The mean beryllium recoveries from wipes spiked with Certified Reference Materials (CRMs) shall be 100  $\pm$  20 % at the 95 % confidence level in accordance with 8.2.

NOTE 2—It is not imperative that the wipe be completely dissolved when digested/extracted in accordance with a standard beryllium test method to meet the recovery criterion. However, the digestion/extraction solution that is to be analyzed should be free of suspended particulates and gelatinous material.

6.1.8 *Collection Efficiency*—Collection efficiency of an individual wipe, using an initial wipe on a given vinyl or urethane-coated vinyl tile test surface, shall be determined

using a solid beryllium-containing CRM in accordance with 8.2. The mass of beryllium-containing material (particulate or dust) loaded per surface area to be sampled shall be  $0.50 \pm 0.05$  g for the large wipe and  $0.050 \pm 0.005$  g for the small wipe. The minimum collection efficiency of at least 95 % of the individual wipes shall be 75 %, as measured against the known mass of beryllium loaded on the test surface prior to wiping (1000  $\text{cm}^2$  minimum surface area for the large wipe and 100  $\text{cm}^2$  for the small wipe).

## 7. Selection and Handling of Wipes for Testing

7.1 Tests described in Section 8 shall be conducted on wipes selected in accordance with the random sampling procedure described in Practice E105, using wipes selected after packaging, and representative of each lot.

7.2 Wipes shall not be removed from their packages until immediately prior to testing.

## 8. Procedure

8.1 *Background and Recovery*—Recoverability of beryllium from spiked wipes (see 6.1.7) shall be measured in accordance with a standard beryllium test method. Background beryllium in unspiked wipes (see 6.1.1) shall be measured in accordance with the same procedure. A minimum of seven samples per each concentration level (unspiked,  $0.020 \pm 0.002$   $\mu\text{g}$ ,  $0.10 \pm 0.02$   $\mu\text{g}$ ,  $0.20 \pm 0.02$   $\mu\text{g}$ ,  $1.5 \pm 0.2$   $\mu\text{g}$ ,  $3.0 \pm 0.2$   $\mu\text{g}$ , and  $6.0 \pm 0.2$   $\mu\text{g}$ ) shall be tested (see 6.1.1 and 6.1.7), using wipes randomly selected from each lot in accordance with Practice E105.

8.1.1 Compute the mean and coefficient of variation for each set of samples. See Practice E691 for details regarding statistical computations. Compare with the requirements of 6.1.1 and 6.1.7.

8.2 *Collection Efficiency*—Collection efficiency of beryllium shall be measured in the following manner:

8.2.1 A delineated area of the vinyl or urethane-coated tile test surface (minimum area 1000  $\text{cm}^2$  for large wipes and 100  $\text{cm}^2$  for small wipes as in 6.1.8) is loaded with a known mass ( $0.50 \pm 0.05$  g for large wipes and  $0.050 \pm 0.005$  g for small wipes) of CRM particulate or dust (see 6.1.7) and then wiped in accordance with Practice D6966.

8.2.2 An equivalent alternative procedure consists of manually distributing a known amount (mass) of beryllium-containing CRM uniformly onto the vinyl tile test surface of 1000  $\text{cm}^2$  (large wipe) or 100  $\text{cm}^2$  (small wipe) minimum area, and then wiping the applicable surface with the appropriately sized wipe in accordance with Practice D6966.

8.2.3 The collection efficiency is determined by comparing the amount of beryllium collected in the wipe against the total amount of beryllium loaded onto the area of interest on the test surface (in accordance with 8.2.1 or 8.2.2). A minimum of seven wipes, randomly selected from the lot in accordance with Practice E105, shall be tested for each beryllium level.

8.2.4 For the measurement of beryllium content, the wipes shall be digested/extracted and the beryllium content determined in accordance with a standard beryllium test method. Determine the collection efficiency for each wipe and compare with the requirements of 6.1.8.