



Edition 3.0 2024-07 COMMENTED VERSION

# INTERNATIONAL STANDARD



Medical electrical equipment – Dosimeters with ionization chambers and/or semiconductor detectors as used in X-ray diagnostic imaging

### **Document Preview**

IEC 61674:2024

https://standards.iteh.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16f36e519a13/iec-61674-2024





#### THIS PUBLICATION IS COPYRIGHT PROTECTED Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

**IEC** Secretariat 3, rue de Varembé CH-1211 Geneva 20 Switzerland

Tel.: +41 22 919 02 11 info@iec.ch www.iec.ch

#### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

#### IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

#### Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.





Edition 3.0 2024-07 COMMENTED VERSION

# INTERNATIONAL STANDARD



Medical electrical equipment – Dosimeters with ionization chambers and/or semiconductor detectors as used in X-ray diagnostic imaging

### **Document Preview**

IEC 61674:2024

https://standards.iteh.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16f36e519a13/iec-61674-2024

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ICS 11.040.50

ISBN 978-2-8322-9379-9

Warning! Make sure that you obtained this publication from an authorized distributor.

### CONTENTS

FC	OREWC	RD	4
IN	TRODU	ICTION	7
1	Scop	e and object	8
	1.1	Scope	8
	1.2	bject	8
2	Norm	native references	8
3	Terms and definitions		9
4	Gene	eral requirements	17
	4 1	Performance requirements	17
	4.1	REFERENCE VALUES and STANDARD TEST VALUES	17
	4.3	General test conditions	
	431	STANDARD TEST CONDITIONS	18
	4.3.2	Statistical fluctuations	
	4 3 3		19
	434	Adjustments during test	19
	435	Batteries	19
	4.0.0	Constructional requirements as related to performance	10
	 4 4 1	Components	10
	442	Display TIEII Standards	20
	4 A A	Indication of battery condition	20
	4.4.0 / / /	Indication of polarizing voltage failure	20
	4.4.4		20
	4.4.0	MEASURING ASSEMBLIES with multiple DETECTOR ASSEMBLIES	20
	4.4.0	Padioactive STABILITY CLECK DEVICE	2 I 21
	4.4.7	UNCERTAINTY of measurement C 61674:2024	2 I 21
ttps <del>;</del> //	standai	s of DEDEODMANCE CHARACTERISTICS	1673,202
± J			22
	5.1		22
	5.2	Repeatability	22
	5.2.1		22
	5.2.2	Repeatability in the ATTENUATED BEAM	22
	5.2.3	Repeatability in the UNATTENUATED BEAM	23
	5.3	RESOLUTION of reading	23
	5.4	STABILIZATION TIME	23
	5.5	Effect of pulsed radiation on AIR KERMA and AIR KERMA LENGTH PRODUCT	24
	56	Stability	24 25
	561	Long term stability	20
	5.6.2	Accumulated dose stability	20
	5.6		20
	5.7	Effects of LEAKAGE CURRENT	
	5.7		25
6		S OF VARIATION for effects of INFLUENCE OUANTITIES	20 26
0			20
	0.1 0.0		
	b.2	Energy aependence of RESPONSE	26
	0.3	AIR KERMA RATE GEPENGENCE OF AIR KERMA AND AIR KERMA LENGTH PRODUCT measurements	28
	<del>6.3.1</del>	MEASURING ASSEMBLY	·····

<del>6.3.</del>	2 IONIZATION CHAMBER Recombination losses	
6.4	Dependence of DETECTOR RESPONSE on angle of incidence of radiation	29
6.4.	1 Non-CT detectors	29
6.4.	2 CT DETECTORS	29
6.5	Operating voltage	29
6.5.	1 Mains-operated DOSIMETERS	29
6.5.	2 Battery-operated DOSIMETERS	29
6.5.	3 Mains rechargeable, battery-operated DOSIMETERS	30
6.6	Air pressure	30
6.7	Air pressure Equilibration time of the radiation detector	30
6.8	Temperature and humidity	31
6.9	Electromagnetic compatibility	31
6.9.	1 ELECTROSTATIC DISCHARGE	31
6.9.	2 Radiated electromagnetic fields	32
6.9.	3 CONDUCTED DISTURBANCES induced by bursts and radio frequencies	32
6.9.	4 Voltage dips, short interruptions and voltage VARIATIONS	32
6.10	Field size	33
6.11	EFFECTIVE LENGTH and spatial uniformity of RESPONSE of CT DOSIMETERS	33
7 Mar	7 Marking	
7.1	DETECTOR ASSEMBLY	33
7.2	MEASURING ASSEMBLY	33
7.3	Radioactive STABILITY CHECK DEVICE	34
8 Acc	OMPANYING DOCUMENTS	34
Annex A	(informative) Combined standard uncertainty for dosimeter performance	
Bibliogra	phy. Document preview	
Index of	defined terms	38
	<u>IEC 61674:2024</u>	
ttps://standa	niments. rds.iten.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16136e519a13/iec	-61674-20

Table 1 – REFERENCE and STANDARD TEST CONDITIONS	17
Table 2 – Number of readings required to detect true differences $\Delta$ (95 % confidence level) between two sets of instrument readings	19
Table 3 – Maximum values for the COEFFICIENT OF VARIATION, $v_{max}$ , for measurements	
in the attenuated beam	23
Table 4 – Maximum values for the COEFFICIENT OF VARIATION, $v_{max}$ , for measurements	
in the unattenuated beam and mammography	23
Table 5 – LIMITS OF VARIATION for the effects of INFLUENCE QUANTITIES	27
Table 6 – Climatic conditions	31
Table A.1 – Estimation of COMBINED STANDARD UNCERTAINTY for dosimeter performance	36

#### INTERNATIONAL ELECTROTECHNICAL COMMISSION

### MEDICAL ELECTRICAL EQUIPMENT – DOSIMETERS WITH IONIZATION CHAMBERS AND/OR SEMICONDUCTOR DETECTORS AS USED IN X-RAY DIAGNOSTIC IMAGING

#### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.

7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications

- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at https://patents.iec.ch. IEC shall not be held responsible for identifying any or all such patent rights.

This commented version (CMV) of the official standard IEC 61674:2024 edition 3.0 allows the user to identify the changes made to the previous IEC 61674:2012 edition 2.0. Furthermore, comments from IEC SC 62C experts are provided to explain the reasons of the most relevant changes, or to clarify any part of the content.

A vertical bar appears in the margin wherever a change has been made. Additions are in green text, deletions are in strikethrough red text. Experts' comments are identified by a blue-background number. Mouse over a number to display a pop-up note with the comment.

This publication contains the CMV and the official standard. The full list of comments is available at the end of the CMV.

IEC 61674 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This third edition cancels and replaces the second edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) for mammography, the manufacturer specifies the REFERENCE VALUE for the RADIATION QUALITY; **1**
- b) for mammography, the manufacturer provides the MINIMUM RATED RANGE of RADIATION QUALITIES for the compliance test on energy dependence of response;
- c) the compliance test for analogue displays was removed; 2
- d) the compliance tests for range reset, the effect of leakage and recombination losses were removed. These tests are already covered by the test on linearity and cannot be conducted for modern devices. The estimation of COMBINED STANDARD UNCERTAINTY was changed accordingly;
- e) the compliance test for mains rechargeable and battery-operated dosimeters were updated for modern devices.

The text of this International Standard is based on the following documents:

-		indards	
	Draft	Report on voting	•
(ľ	62C/909/FDIS	62C/913/RVD	

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

#### EC 61674:2024

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members\_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used.

- requirements and definitions: roman type.
- test specifications: italic type.
- informative material appearing outside of tables, such as notes, examples and references: in smaller type.
  Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2:2021. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

## iTeh Standards (https://standards.iteh.ai) Document Preview

IEC 61674:2024

https://standards.iteh.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16f36e519a13/iec-61674-2024

#### INTRODUCTION

Diagnostic radiology is the largest contributor to man-made IONIZING RADIATION to which the public is exposed. The reduction in the exposure received by PATIENTS undergoing medical radiological examinations or procedures has therefore become a central issue in recent years. The PATIENT dose will be minimized when the X-ray producing equipment is correctly adjusted for image quality and radiation output. These adjustments require that the routine measurement of AIR KERMA, AIR KERMA LENGTH PRODUCT and/or AIR KERMA RATE be made accurately. The equipment covered by this document plays an essential part in achieving the required accuracy. It is important that the DOSIMETERS used for adjustment and control measurements **must be** are of satisfactory quality and **must** therefore fulfil the special requirements laid down in this document.

## iTeh Standards (https://standards.iteh.ai) Document Preview

IEC 61674:2024

https://standards.iteh.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16f36e519a13/iec-61674-2024

#### MEDICAL ELECTRICAL EQUIPMENT – DOSIMETERS WITH IONIZATION CHAMBERS AND/OR SEMICONDUCTOR DETECTORS AS USED IN X-RAY DIAGNOSTIC IMAGING

#### 1 Scope and object

#### 1.1 Scope

This document specifies the performance and some related constructional requirements of DIAGNOSTIC DOSIMETERS intended for the measurement of AIR KERMA, AIR KERMA LENGTH PRODUCT or AIR KERMA RATE, in photon radiation fields used in medical X-ray imaging, such as RADIOGRAPHY, including mammography, RADIOSCOPY and COMPUTED TOMOGRAPHY (CT), for X-RADIATION with generating potentials not greater than in the range of 20 kV to 150 kV.

This document is applicable to the performance of DOSIMETERS with VENTED IONIZATION CHAMBERS and/or SEMICONDUCTOR DETECTORS as used in X-ray diagnostic imaging.

#### 1.2 Object

The object of this document is

- a) to establish requirements for a satisfactory level of performance for DIAGNOSTIC DOSIMETERS, and
- b) to standardize the methods for the determination of compliance with this level of performance.

This document is not concerned with the safety aspects of DOSIMETERS. The DIAGNOSTIC DOSIMETERS covered by this document are not intended for use in the PATIENT ENVIRONMENT and, therefore, the requirements for electrical safety applying to them are contained in IEC 61010-1.

#### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60050 (all parts), International Electrotechnical Vocabulary (available at <a href="http://www.electropedia.org">http://www.electropedia.org</a>)

IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance

IEC 60601-1-3:2008, Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment

IEC 60417, *Graphical symbols for use on equipment,* available at http://www.graphical-symbols.info/equipment

IEC 60731:2011, Medical electrical equipment — Dosimeters with ionization chambers as used in radiotherapy

IEC 61674:2024 CMV © IEC 2024 - 9 -

IEC TR 60788:2004, Medical electrical equipment – Glossary of defined terms

IEC 61000-4 (all parts), *Electromagnetic compatibility (EMC) – Part 4: Testing and measuring techniques* 

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test* 

IEC 61000-4-3:2020, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test* 

IEC 61000-4-4, Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test

IEC 61000-4-6, *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields* 

IEC 61000-4-11, Electromagnetic compatibility (EMC) – Part 4-11: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations immunity tests for equipment with input current up to 16 A per phase

IEC 61187, *Electrical and electronic measuring equipment – Documentation* 

IEC 61267:2005, Medical diagnostic X-ray equipment – Radiation conditions for use in the determination of characteristics

ISO/IEC GUIDE 98-3:2008, Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)

ISO/IEC Guide 99:2007, International vocabulary of metrology – Basic and general concepts and associated terms (VIM) IEC 61674:2024

tps://standards.iteh.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16f36e519a13/iec-61674-2024 ISO 3534-1:2006, Statistics Vocabulary and symbols Part 1: General statistical terms and terms used in probability

#### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC TR 60788:2004 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at https://www.electropedia.org/
- ISO Online browsing platform: available at https://www.iso.org/obp

#### 3.1

#### DIAGNOSTIC DOSIMETER

#### DOSIMETER

equipment which uses IONIZATION CHAMBERS and/or SEMICONDUCTOR DETECTORS for the measurement of AIR KERMA, AIR KERMA LENGTH PRODUCT and/or AIR KERMA RATE in the beam of an X-RAY EQUIPMENT used for diagnostic medical radiological examinations

Note 1 to entry: A DIAGNOSTIC DOSIMETER contains the following components:

- one or more DETECTOR ASSEMBLIES which may or may not be an integral part of the MEASURING ASSEMBLY;
- a MEASURING ASSEMBLY;

- one or more STABILITY CHECK DEVICES (optional).

#### 3.1.1

#### DETECTOR ASSEMBLY

RADIATION DETECTOR and all other parts to which the RADIATION DETECTOR is permanently attached, except the MEASURING ASSEMBLY

- 10 -

Note 1 to entry: The DETECTOR ASSEMBLY normally includes:

- the RADIATION DETECTOR and the stem (or body) on which the RADIATION DETECTOR is permanently mounted (or embedded);
- the electrical fitting and any permanently attached cable or pre-amplifier.

#### 3.1.1.1

#### RADIATION DETECTOR

element which transduces AIR KERMA, AIR KERMA LENGTH PRODUCT or AIR KERMA RATE into a measurable electrical signal

Note 1 to entry: A radiation detector may be either an ionization chamber or a semiconductor detector.

#### 3.1.1.1.1 IONIZATION CHAMBER

#### CHAMBER

ionizing RADIATION DETECTOR consisting of a CHAMBER filled with air, a suitable gas, or a gaseous mixture, in which an electric field insufficient to produce gas multiplication is provided for the total collection, at the electrodes, of charges associated with the ions and the electrons produced in the measuring sensitive volume of the detector by the ionizing radiation

Note 1 to entry: An ionization chamber can be sealed or vented.

Note 2 to entry: Vented ionization chambers are constructed in such a way as to allow the air inside the measuring volume to communicate freely with the atmosphere, so that corrections to the response for changes in air density need to be made.

Note 3 to entry: Sealed ionization chambers are not suitable, because the necessary wall thickness of a sealed chamber may cause an unacceptable energy dependence of the response and because the long-term stability of sealed chambers is not guaranteed.

[SOURCE: <u>IEC 60731:2011, 3.1.1.1, modified – three new notes to entry have replaced the two</u> original notes. IEC 60050-395:2014, 395-03-07, modified – Two new notes to entry were added.]

#### 3.1.1.1.2

#### VENTED IONIZATION CHAMBER

IONIZATION CHAMBER constructed in such a way as to allow the air inside the measuring volume to communicate freely with the atmosphere such that corrections to the RESPONSE for changes in air density need to be made

[SOURCE: IEC 60731:2011, 3.1.1.1.3, modified – The term has been changed from "vented chamber" to "VENTED IONIZATION CHAMBER".]

#### 3.1.1.1.3

#### SEMICONDUCTOR DETECTOR

semiconductor device that utilises the production and motion of electron-hole pairs in a charge carrier depleted region of the semiconductor for the detection and measurement of IONIZING RADIATION

Note 1 to entry: The production of electron-hole pairs is caused <u>either directly</u> by interaction of the IONIZING RADIATION with the semiconductor material, <del>or</del>. In the purview of this document **3**, detectors qualify as semiconductor detectors, even when the production of electron-hole pairs is caused indirectly by first converting the incident radiation energy to light in a scintillator material directly in front of and optically coupled to a semiconductor photodiode, which then produces the electrical signal.

#### 3.1.2

#### MEASURING ASSEMBLY

device to measure the charge (or current) electrical signal from the RADIATION DETECTOR and convert it into a form suitable for displaying the values of DOSE or KERMA or their corresponding rates

[SOURCE: IEC 60731:2011, 3.1.2, modified – The term IONIZATION CHAMBER in the original definition has been replaced by the term RADIATION DETECTOR. The words "measure the charge (or current) from the IONIZATION CHAMBER" have been replaced with "measure the electrical signal from the RADIATION DETECTOR".]

#### 3.1.3

#### STABILITY CHECK DEVICE

device which enables the stability of RESPONSE of the MEASURING ASSEMBLY and/or CHAMBER ASSEMBLY to be checked

Note 1 to entry: The STABILITY CHECK DEVICE may be a purely electrical device, or a radiation source, or it may include both.

[SOURCE: IEC 60731:2011, 3.1.3]

#### 3.1.4

#### **CT** DOSIMETER

DIAGNOSTIC DOSIMETER which uses long narrow IONIZATION CHAMBERS and/or SEMICONDUCTOR DETECTORS for the measurement of AIR KERMA integrated along the length of the DETECTOR when the DETECTOR is exposed to a cross-sectional X-ray scan of a computed tomograph

Note 1 to entry: A CT DOSIMETER contains the following components:

- one or more DETECTOR ASSEMBLIES;
- a MEASURING ASSEMBLY.

#### 3.1.5

#### IEC 61674:2024

TTP RADIATION DETECTOR which is used for CT dosimetry e89-4e66-a621-16f36e519a13/iec-61674-2024

#### 3.2

#### INDICATED VALUE

**CT** DETECTOR

value of a quantity derived from the reading of an instrument together with any scale factors indicated on the control panel of the instrument

[SOURCE: IEC 60731:2011, 3.2, modified – The note has been deleted.]

#### 3.3

#### TRUE VALUE

value of the physical quantity to be measured by an instrument

[SOURCE: IEC 60731:2011, 3.3, modified - The note has been deleted.]

#### 3.4

#### CONVENTIONAL TRUE VALUE

value used instead of the TRUE VALUE when calibrating or determining the performance of an instrument, since in practice the TRUE VALUE is unknown and unknowable

Note 1 to entry: The CONVENTIONAL TRUE VALUE will usually be the value determined by the WORKING STANDARD with which the instrument under test is being compared.

[SOURCE: IEC 60731:2011, 3.4, modified – The second note has been deleted.]

#### 3.5

#### MEASURED VALUE

best estimate of the TRUE VALUE of a quantity, being derived from the INDICATED VALUE of an instrument together with the application of all relevant CORRECTION FACTORS and the CALIBRATION FACTOR

Note 1 to entry: The MEASURED VALUE is sometimes also referred to as "result of a measurement".

[SOURCE: IEC 60731:2011, 3.5, modified – The existing note has been replaced with a new note to entry has been added.]

#### 3.5.1

#### ERROR OF MEASUREMENT

difference remaining between the  $\ensuremath{\mathsf{MEASURED}}$  value of a quantity and the  $\ensuremath{\mathsf{TRUE}}$  value of that quantity

[SOURCE: IEC 60731:2011, 3.5.1]

#### 3.5.2

#### **OVERALL UNCERTAINTY**

UNCERTAINTY associated with the MEASURED VALUE

Note 1 to entry: I.e. it represents the bounds within which the ERROR OF MEASUREMENT is estimated to lie (see also 4.5).

[SOURCE: IEC 60731:2011, 3.5.2, modified – The parenthesis has been added to the note to entry, and the second note has been deleted.]

#### 3.5.3

### EXPANDED UNCERTAINTY

quantity defining an interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand IEC 61674:2024

ttps://standards.iteh.ai/catalog/standards/iec/cce42257-4e89-4e66-a621-16f36e519a13/iec-61674-2024 [SOURCE: ISO/IEC GUIDE 98-3:2008, 2.3.5, modified – The three notes have been deleted.]

#### 3.6

#### CORRECTION FACTOR

dimensionless multiplier which corrects the INDICATED VALUE of an instrument from its value when operated under particular conditions to its value when operated under stated REFERENCE CONDITIONS

[SOURCE: IEC 60731:2011, 3.6]

#### 3.7

#### INFLUENCE QUANTITY

any external quantity that may affect the performance of an instrument

[SOURCE: IEC 60731:2011, 3.7]

#### 3.8

#### **INSTRUMENT PARAMETER**

any-internal property of an instrument that may affect the performance of this instrument

[SOURCE: IEC 60731:2011, 3.8]