

Designation: D6505 - 00 (Reapproved 2022)

Standard Test Method for Assay of *normal*-Propyl Bromide Content¹

This standard is issued under the fixed designation D6505; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

- 1.1 This test method provides a basis for the determination of the *normal*-propyl bromide (weight %) in the presence of stabilizers and impurities, in virgin or reclaimed *normal*-propyl bromide (*n*PB). The application range is from 50 wt % to 100 wt %.
- 1.2 This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

2. Summary of Test Method

2.1 This analytical test method describes the determination of the assay of *n*-propyl bromide by gas chromatography using an internal standard. The results are reported in weight percent.

3. Significance and Use

3.1 This test method provides an analysis in weight percent of the *normal*-propyl bromide concentration of virgin, formulated, or reclaimed *normal*-propyl bromide: compounds that co-elute with *normal*-propyl bromide or *normal*-heptane (internal standard) may interfere with this test method.

4. Apparatus

- 4.1 Gas Chromatograph—Hewlett Packard 5890 Series II equipped with a split/splitless injector, a flame ionization detector, and an autosampler, HP, or equivalent, if available.
- 4.2 *Column*—30 m by 0.25 mm, 1.0 μm film thickness AllTech AT-5, or equivalent, capillary column.
 - 4.3 Syringe—5 μL or 10 μL GC autosampler syringe.
- 4.4 Data Acquisition and Analysis Device—VAX MULTI-CHROM or equivalent chromatography data collection and processing system, or integrator.
- ¹ This test method is under the jurisdiction of ASTM Committee D26 on Halogenated Organic Solvents and Fire Extinguishing Agents and is the direct responsibility of Subcommittee D26.04 on Test Methods.
- Current edition approved May 1, 2022. Published May 2022. Originally approved in 2000. Last previous edition approved in 2017 as D6505–00 (2017). DOI: 10.1520/D6505-00R22.

- 4.5 *Crimp Top Vials*—Hewlett Packard 2 mL glass crimp top vials with TFE-fluorocarbon-lined silicone septa or equivalent.
 - 4.6 Volumetric Pipet—25 mL.
- 4.7 Glass Vials—32 mL glass vials (25 mm by 95 mm), with TFE-fluorocarbon-lined screw caps.
 - 4.8 Transfer Pipets—Borosilicate glass transfer pipets.
 - 4.9 Precision Balance, accurate to 0.1 mg.

5. Required Chemicals

- 5.1 *n-Propyl Bromide* (1-Bromopropane), 99+ %, Aldrich Chemical Company
- 5.2 *n-Heptane* (internal standard), Reagent Grade, 99.5+ %, J.T. Baker (or equivalent)
- 5.3 Methylene Chloride, 99+ %, Aldrich Chemical Company (or equivalent)

6. Suggested Analysis Conditions

6.1 *Chromatographic Conditions:*

Injector	150 °C
Detector temperature	250°C-d6505-002022
FID detector range	most sensitive setting
Temperature program	30 °C for 10 min 30 °C to 250 °C at 15 °C /min hold at 250 °C for 10 min
Carrier gas	He (or $\rm N_2$) at 1 mL/min split flow: 25 mL/min column head pressure: 7.5 psi
Make-up gas	He (or N ₂) at 30 mL/min

7. Analytical Procedure

Injection volume

7.1 Internal Standard Calibration Mixture—Carefully weigh about 0.25 g of *n*-heptane internal standard and about 0.5 g of *n*-propyl bromide into a 32 mL glass vial. Record the actual weights for each. Pipet 25 mL of methylene chloride into this vial and cap immediately. Mix well and then transfer approximately 1 mL of this standard solution into an autosampler vial. Cap the vial and analyze in accordance with the conditions specified in the chromatographic conditions section.

0.5 µL

Inject this standard at least three times and calibrate an internal standard method in the data system using the results from the standard analyses.

7.2 Preparation of Samples for Analysis—Weigh about 0.25 g of n-heptane and about 0.5 g of sample into a 32 mL glass vial. Record the actual weights for each. Pipet 25 mL of methylene chloride into this vial and cap immediately. Mix well and transfer approximately 1 mL of this solution to an autosampler vial. Cap the vial and analyze in accordance with the conditions specified in the chromatographic conditions section. The following table shows components of interest, their absolute retention time (ART) in minutes, their relative retention time (RRT) in relation to n-heptane, the internal standard, as well as their relative response factor (RRf).

Component	ART	RRT	RRf
n-propyl bromide	13.30	0.83	0.3495
n-heptane (internal standard)	16.08	1.00	1.0000

8. Calculation

- 8.1 A calibrated internal standard weight percent method in the data system will yield the required weight percent data. Alternatively, the weight percent calculations may be performed manually as described as follows:
- 8.2 Calculate the relative response factor (RRf) for n-propyl bromide in the standard relative to the n-heptane internal standard. Average the peak areas of the three standard injections:

$$RRf = \frac{Ay \times Wtl}{Wty \times Al}$$
 (1)

where:

Ay = average area of *n*-propyl bromide peak in standard calibration mixture,

Al = average area of *n*-heptane (internal standard) peak in standard calibration mixture,

Wty = amount of n-propyl bromide in standard calibration mixture, and

Wtl = amount of n-heptane (internal standard) in standard calibration mixture.

The assay of n-propyl bromide (weight % y) in the sample is calculated as follows from the sample injection:

weight %
$$y = \frac{Ay \times Wtl \times 100}{Al \times RRf \times WtS}$$
 (2)

where:

Ay =area of *n*-propyl bromide peak in sample mixture,

Al = area of n-heptane (internal standard) peak in sample

mixture,

Wtl = weight of internal standard in sample mixture,

WtS = weight of sample in sample mixture, and

RRf = relative response factor from above.

9. Precision and Bias

- 9.1 *Precision*—The following criteria should be used for judging the acceptability of results:
- 9.1.1 Repeatability (Single Analyst)—The standard deviation for a single determination has been estimated to be 0.037 % absolute at 22 df. The 95 % limit for the difference between 2 such runs is 0.10 % absolute.
- 9.1.2 Laboratory Precision (Within-Laboratories, Between Days Variability, formerly called Repeatability—The standard deviation of results (each the average of duplicates) obtained by the same analyst on different days, has been estimated to be 0.62 % absolute at 8 df. The 95 % limit for the difference between 2 such runs is 1.72 % absolute.
- 9.1.3 Reproducibility, Multilaboratory—The standard deviation of results (each the average of duplicates) obtained by analysts at different laboratories, has been estimated to be 0.93 % absolute at 7 df. The 95 % limit for the difference between 2 such runs is 2.39 % absolute.

10. Keywords

10.1 assay; brominated; gas chromatography; halogenated; *normal*-propyl bromide; purity; solvent

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/