



Designation: ~~D5423 – 14~~ D5423 – 22

Standard Specification for Forced-Convection Laboratory Ovens for Evaluation of Electrical Insulation¹

This standard is issued under the fixed designation D5423; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

1. Scope

1.1 This specification covers forced-convection ventilated electrically-heated ovens, operating over all or part of the temperature range from ~~20°C~~ 20 °C above the ambient temperature to ~~500°C~~ 500 °C, and used for thermal endurance evaluation of electrical insulating materials.

1.2 The specification requirements for Type I ovens are based on IEC Publication ~~216-4-1, 60216-4-1~~, and are technically identical to it. The requirements for Type II ovens are essentially identical to the requirements of Specification ~~D2436~~. This specification and an associated test method, ~~D5374~~, have replaced Specification ~~D2436~~.

1.3 While the ovens covered by this specification are intended primarily for thermal endurance evaluation, ~~they can also be used wherever their characteristics make them suitable for other applications.~~ applications as applicable.

1.4 ~~This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.~~ ://standards.iteh.ai/catalog/standards/sist/d501d894-d230-4013-9066-659395c96624/astm-d5423-22

1.5 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.

1.6 This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

2. Referenced Documents

2.1 ASTM Standards:²

[D2436 Specification for Forced-Convection Laboratory Ovens for Electrical Insulation \(Withdrawn 1994\)](#)³

[D5374 Test Methods for Forced-Convection Laboratory Ovens for Evaluation of Electrical Insulation](#)

¹ This specification is under the jurisdiction of ASTM Committee [D09](#) on Electrical and Electronic Insulating Materials and is the direct responsibility of Subcommittee [D09.17](#) on Fire and Thermal Characteristics/Properties.

Current edition approved ~~April 1, 2014~~ May 1, 2022. Published ~~May 2014~~ May 2022. Originally approved in 1993. Last previous edition approved in ~~2005~~ 2014 as ~~D5423 – 93 (2004)~~ D5423 – 14, which was ~~withdrawn in January 2014 and reinstated in April 2014.~~ (1999). DOI: [10.1520/D5423-14](#). DOI: [10.1520/D5423-22](#).

² For referenced ASTM standards, visit the ASTM website, [www.astm.org](#), or contact ASTM Customer Service at [service@astm.org](#). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on [www.astm.org](#).

2.2 Other Document:

~~IEC Publication 216-4-1~~~~60216-4-1~~ [Guide for the Determination of Thermal Endurance Properties of Electrical Insulating Materials, Part 4—Aging Ovens, Section Electrical Insulating Materials—Thermal Endurance Properties—Part 4-1: Ageing](#)
~~1—Single-Chamber Ovens—Single-Chamber Ovens~~⁴

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *rate of ventilation, n*—the number of air changes per hour in the oven chamber.

3.1.2 *set temperature, n*—the average of all of the measured temperatures within the oven, averaged over the period of any cyclic temperature variation that ~~may occur~~.occurs.

3.1.2.1 Discussion—

This is the actual operating temperature of the oven. In IEC 60216-4-1, this term is called the exposure temperature.

3.1.3 *temperature fluctuation, n*—~~temperature differences~~ maximum change in temperature at one point in the oven over a period of time.

3.1.3.1 Discussion—

This property depends upon the sensitivity and type (on/off or proportional) of control used and the heater mass in relation to surface area.

3.1.4 *temperature gradient, n*—the maximum temperature difference at one time between ~~different~~ any two points in the oven chamber.

3.1.4.1 Discussion—

This property depends on such factors as uniformity of heater temperature, heater distribution about the oven, and air flow patterns within the oven. In IEC 60216-4-1, this term is called the temperature difference.

3.1.5 *temperature variation, n*—~~temperature differences~~ maximum change in temperature with time and location due to the combination of temperature gradient and temperature fluctuation.

3.1.6 *thermal lag time, n*—the time required for a defined specimen to reach a specified temperature (or range of temperature).

3.1.6.1 Discussion—

This property is largely dependent upon the rate of air circulation within the oven. In IEC 216-4-1, 60216-4-1, this term is called “time—called the “time constant.”

3.1.7 *time constant, n*—See thermal lag time.

4. Classification

4.1 *Type I*—Ovens with rate of ventilation of 5 to 20 air changes per hour.

4.2 *Type II*—Ovens with rate of ventilation of 100 to 200 air changes per hour.

5. Test Methods

5.1 Determine the properties specified in Section 6 using Test Methods **D5374**.

6. Performance Requirements

6.1 *Rate of Ventilation*—The rate of ventilation at $\pm 00^{\circ}\text{C}$ 100 °C and at the maximum temperature at which the oven is used shall be:

6.1.1 For Type I ovens, 5 to 20 air changes per hour,

⁴ Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036.

6.1.2 For Type II ovens, 100 to 200 air changes per hour.

NOTE 1—Past experience in the United States has been to specify 100 to 200 air changes per hour, as required for Type II ovens. However, international practice, as defined in IEC Publication 216-4-1, has been to use current international practice is to prefer the lower rate of ventilation, which is ventilation specified for Type I ovens.

6.2 *Set Temperature*—The set temperature, measured over a five-day test period, shall not vary by more than the values given in **Table 1**.

6.3 *Temperature Variation*—The temperature shall not exceed that given in **Table 1** for any test temperature.

NOTE 2—Some test standards may require closer control of temperature variation or of set temperature. In such cases, reduction of the variations specified in **Table 1** is subject to agreement between purchaser and manufacturer.

6.4 *Thermal Lag Time*—The thermal lag time, measured with the oven operating at a temperature of $200 \pm 5^\circ\text{C}$, shall not exceed 660 s.

7. Construction Requirements

7.1 The oven shall be soundly constructed of suitable materials and all electrical and mechanical fittings shall be readily accessible for inspection, maintenance, and replacement purposes.

7.2 The interior of the oven shall be constructed of suitable corrosion-resistant, non-absorbent material, and fabricated so any joints are leak-proof and not subject to corrosion, and having interior surfaces that are easy to clean.

7.3 The door(s) shall be fitted with hinges and latches to provide sufficient pressure for an effective seal between the interior of the oven and the external atmosphere.

7.4 The oven shall have provisions for automatically switching off the heaters when the actual oven temperature substantially exceeds the set temperature.

7.5 Ports for admitting and exhausting air shall be provided, with fittings for attaching conduits for exhaust air. When specified, provisions shall be made for supplying inlet air or other gases from a controlled source.

7.6 Dampers for adjusting the rate of ventilation and the air circulation shall be provided to enable compliance with other requirements of this specification.

8. Certification

8.1 For each oven furnished under this specification, the purchaser shall be furnished certification that ovens of exactly the same design have been tested in accordance with this specification and Test Methods **D5374**, and that the requirements have been met.

8.2 Where specified, the purchaser shall be furnished with results of tests on the ovens furnished. In such cases, the positions of the dampers during the tests shall be clearly marked.

TABLE 1 Temperature Requirements

| Set Temperature Range °C | Temperature Variation °C, max | Variation in Set Temperature, ±°C, max |
|--------------------------|-------------------------------|--|
| ≤80 | 4 | 2 |
| >80 to ≤180 | 5 | 2.5 |
| >180 to ≤300 | 6 | 3 |
| >300 to ≤400 | 8 | 4 |
| >400 to ≤500 | 10 | 5 |