



Designation: E394 – 22

Standard Test Method for Iron in Trace Quantities Using the 1,10-Phenanthroline Method¹

This standard is issued under the fixed designation E394; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the determination of trace concentrations of iron in the range from 1 to 100 $\mu\text{g/g}$ in a wide variety of products.

1.2 In determining the conformance of the test results using this method to applicable specifications; results shall be rounded off in accordance with the rounding-off method of Practice E29.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 Consult current OSHA regulations, suppliers' Safety Data Sheets, and local regulations for all materials used in this specification.

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

D1193 Specification for Reagent Water

D6809 Guide for Quality Control and Quality Assurance Procedures for Aromatic Hydrocarbons and Related Materials

E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

E60 Practice for Analysis of Metals, Ores, and Related Materials by Spectrophotometry

¹ This test method is under the jurisdiction of ASTM Committee D16 on Aromatic, Industrial, Specialty and Related Chemicals and is the direct responsibility of Subcommittee D16.04 on Instrumental Analysis.

Current edition approved Sept. 1, 2022. Published September 2022. Originally approved in 1970. Last previous edition approved in 2015 as E394 – 15. DOI: 10.1520/E0394-22.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

E180 Practice for Determining the Precision of ASTM Methods for Analysis and Testing of Industrial and Specialty Chemicals (Withdrawn 2009)³

E275 Practice for Describing and Measuring Performance of Ultraviolet and Visible Spectrophotometers

E300 Practice for Sampling Industrial Chemicals

3. Summary of Test Method

3.1 This test method is based upon a photometric determination of the 1,10-phenanthroline complex with the iron(II) ion. The sample is dissolved in a suitable solvent and the iron is reduced to the divalent state by the addition of hydroxylamine hydrochloride. The color is then developed, by the addition of 1,10-phenanthroline. After a short reaction period, the absorbance of the solution is measured at approximately 510 nm using a suitable photometer. The absorbance of the solution, once the color is developed, is stable for at least several months.

4. Significance and Use

4.1 This test method is suitable for determining trace concentrations of iron in a wide variety of products, provided that appropriate sample preparation has rendered the iron and sample matrix soluble in water or other suitable solvent (see 10.1 and Note 4).

4.2 This test method assumes that the amount of color developed is proportional to the amount of iron in the test solution. The calibration curve is linear over the specified range. Possible interferences are described in Section 5.

5. Interferences

5.1 Fortune and Mellon⁴ have made a comprehensive study of the interferences of various inorganic ions in this determination. Table 1 and Table 2, taken from their report, show the effects of various cations and anions on the determination of 2.0 $\mu\text{g/g}$ (ppm) iron. If the maximum level of 500 $\mu\text{g/g}$ (ppm) does not interfere, it is very likely that the ion will not interfere

³ The last approved version of this historical standard is referenced on www.astm.org.

⁴ Fortune, W. B., and Mellon, M. G., *Industrial and Engineering Chemistry, Analytical Edition*, IENAA Vol 10, 1938, pp. 60–64.

*A Summary of Changes section appears at the end of this standard

TABLE 1 Effect of Cations on the Determination of 2 µg/g (ppm) Iron

Ion	Added As	Maximum Added Without Interference, µg/g (ppm)	Applicable pH Range
Aluminum	AlCl ₃	500	2.0–3.0
Ammonium	NH ₄ Cl	500	2.0–9.0
Antimony	SbCl ₃	30	3.0–9.0
Arsenic	As ₂ O ₅	500	3.0–9.0
Arsenic	As ₂ O ₃	500	3.0–9.0
Barium	BaCl ₂	500	3.0–9.0
Beryllium	Be(NO ₃) ₂	500	3.0–5.5
Bismuth	Bi(NO ₃) ₃	... ^A	... ^A
Cadmium	Cd(NO ₃) ₂	50	3.0–9.0
Calcium	Ca(NO ₃) ₂	500	2.0–9.0
Chromium	Cr ₂ (SO ₄) ₃	20	2.0–9.0
Cobalt	Co(NO ₃) ₂	10	3.0–5.0
Copper	Cu(NO ₃) ₂	10	2.5–4.0
Lead	Pb(C ₂ H ₃ O ₂) ₂	500	2.0–9.0
Lithium	LiCl	500	2.0–9.0
Magnesium	Mg(NO ₃) ₂	500	2.0–9.0
Manganese	MnSO ₄	500	2.0–9.0
Mercury	HgCl ₂	1	2.0–9.0
Mercury	Hg ₂ (NO ₃) ₂	10	3.2–9.0
Molybdenum	(NH ₄) ₆ Mo ₇ O ₂₄	100	5.5–9.0
Nickel	Ni(NO ₃) ₂	2	2.5–9.0
Potassium	KCl	1000	2.0–9.0
Silver	AgNO ₃	... ^A	... ^A
Sodium	NaCl	1000	2.0–9.0
Strontium	Sr(NO ₃) ₂	500	2.0–9.0
Thorium	Th(NO ₃) ₄	250	2.0–9.0
Tin	H ₂ SnCl ₆	20	3.0–6.0
Tin	H ₂ SnCl ₄	10	2.0–6.0
Tungsten	Na ₂ WO ₄	10	2.5–9.0
Uranium	UO ₂ (C ₂ H ₃ O ₂) ₂	100	2.0–6.0
Zinc	Zn(NO ₃) ₂	10	2.0–9.0
Zirconium	Zr(NO ₃) ₄	50	2.0–9.0

^A Must be completely absent because of precipitation.

TABLE 2 Effect of Anions on the Determination of 2 µg/g (ppm) Iron

Ion	Added As	Maximum Added Without Interference, µg/g (ppm)	Applicable pH Range
Acetate	NaC ₂ H ₃ O ₂	500	2.0–9.0
Tetraborate	Na ₂ B ₄ O ₇	500	3.0–9.0
Bromide	NaBr	500	2.0–9.0
Carbonate	Na ₂ CO ₃	500	3.0–9.0
Chlorate	KClO ₃	500	2.5–9.0
Chloride	NaCl	1000	2.0–9.0
Citrate	H ₃ C ₆ H ₅ O ₇	500	2.0–9.0
Cyanide	KCN	10	2.0–9.0
Dichromate	K ₂ Cr ₂ O ₇	20	2.5–9.0
Fluoride	NaF	500	4.0–9.0
Iodide	KI	500	2.0–9.0
Nitrate	KNO ₃	500	2.0–9.0
Nitrite	KNO ₂	500	2.5–9.0
Oxalate	(NH ₄) ₂ C ₂ O ₄	500	6.0–9.0
Perchlorate	KClO ₄	100	2.0–9.0
Phosphate	(NH ₄) ₂ HPO ₄	20	2.0–9.0
Pyrophosphate	Na ₄ P ₂ O ₇	50	6.0–9.0
Silicate	Na ₂ SiO ₃	100	2.0–4.5
Sulfate	(NH ₄) ₂ SO ₄	500	2.0–9.0
Sulfite	Na ₂ SO ₃	500	2.0–9.0
Tartrate	(NH ₄) ₂ C ₄ H ₃ O ₆	500	3.0–9.0
Thiocyanate	KCNS	500	2.0–9.0
Thiosulfate	Na ₂ S ₂ O ₃	500	3.0–9.0

in any quantity. The data were obtained under slightly different conditions than those specified in the present test method, but the interferences should be similar. For a more detailed description of interferences, the original literature should be consulted.

5.2 Aldehydes, ketones, and oxidizing agents interfere by consuming the hydroxylamine hydrochloride added as a reducing agent.

6. Apparatus

6.1 *Photometer*, capable of measuring light absorption at 510 nm and holding a 5-cm or 1-cm cell. Check the performance of the photometer at regular intervals according to the guidelines given in Practice E275 and the manufacturer's manual.

NOTE 1—If a filter photometer is used, a narrow band filter having its maximum transmission at 480 to 520 nm should be used. A discussion of photometers and photometric practice is given in Practice E60.

6.2 *Absorption Cells*, 5-cm or 1-cm light path.

7. Reagents and Materials

7.1 *Purity of Reagents*—Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁵ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean Type I or II reagent water as defined in Specification D1193.

7.3 *Hydroxylamine Hydrochloride Solution* (100 g/L)—Dissolve 10 g of hydroxylamine hydrochloride ($\text{HONH}_2 \cdot \text{HCl}$) in approximately 60 mL of water, filter, and dilute to 100 mL.

7.4 *Iron, Standard Solution* (1 mL = 0.01 mg Fe)⁶ may be prepared from one of the following options:

7.4.1 *Option 1*—Dissolve 0.1000 g of iron wire in 10 mL of hydrochloric acid (HCl, 1 + 1) and 1 mL of bromine water (which is bromine saturated water at 20 °C). Boil until the excess bromine is removed. Add 200 mL of HCl, cool, and dilute to 1 L in a volumetric flask. Dilute 100 mL of this solution to 1 L.

7.4.2 *Option 2*—Weigh 0.70 ± 0.01 g of iron (II) ammonium sulfate hexahydrate ($\text{FeSO}_4 \cdot (\text{NH}_4)_2\text{SO}_4 \cdot 6\text{H}_2\text{O}$, minimum purity, 99.5 %) in 500 mL of water containing 20 mL of sulfuric acid (H_2SO_4 , sp gr 1.84) and diluting to 1 L with water. Dilute 100 mL of this solution to 1 L.

7.4.3 *Option 3*—Purchased certified iron standard aqueous solution may also be utilized if its pH is less than 2.

7.5 *1,10-Phenanthroline Solution* (3 g/L)—Dissolve 0.9 g of 1,10-phenanthroline monohydrate in 30 mL of methanol and dilute to 300 mL with water.⁷

7.6 *Ammonium Acetate—Acetic Acid Solution*—Dissolve 100 g of ammonium acetate ($\text{CH}_3\text{COONH}_4$) in about 600 mL of water, filter, add 200 mL of glacial acetic acid to the filtrate, and dilute to 1 L with water.

⁵ ACS Reagent Chemicals, Specifications and Procedures for Reagents and Standard-Grade Reference Materials, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

⁶ This solution is used for calibration only.

⁷ Frederick, G., and Richter, F. P., *Phenanthrolines and Substituted Phenanthroline Indicators*, GFS Publication No. 205, 1944 (no charge).

7.7 *Sulfuric Acid*, ACS trace metals grade.

8. Sampling

8.1 Collect the sample in accordance with Practice E300.

8.2 Since this procedure is intended for trace levels of iron, care must be taken to ensure that the sample container and sampling point do not contribute rust (iron) to the sample.

9. Calibration

9.1 By means of suitable pipets or a buret, transfer 0 (reagent blank), 2, 4, 6, 8, and 10 mL, respectively, of the standard iron solution to each of six 100-mL, glass-stoppered volumetric flasks. These flasks contain 0, 20, 40, 60, 80, and 100 μg of iron, respectively. Dilute the contents of each flask to 80 mL with water. Develop the color and measure the absorbance of each calibration standard as described in 10.3 and 10.4.

9.2 Plot the results in an X-Y graph, with the micrograms of iron on the x-axis and the respective absorbances on the y-axis. Visually evaluate the calibration graph obtained for linearity and for the absence of obvious outlying values. If so, proceed to the next step. If not, investigate for an assignable cause.

9.2.1 Establish a linear regression function from the calibration data using the statistical method of least squares, for example, with the aid of a spreadsheet. The formula for a linear calibration function is:

$$y = a + bx \quad (1)$$

where:

b = slope of calibration line, and

a = intercept.

9.2.2 Evaluate the linearity of the calibration function by calculating the correlation coefficient r . A typical proper value is $r \geq \pm 0.9900$.

NOTE 2—If the photometer readings are percent transmittance, they may be converted to absorbance as follows:

$$A = \log\left(\frac{100}{T}\right) \quad (2)$$

where:

A = absorbance, and

T = percent transmittance.

10. Procedure

10.1 Weigh to three significant figures a sample containing 1 to 100 μg of iron into a 100-mL, glass-stoppered volumetric flask (Note 3). pH of sample should be less than 2; otherwise adjust with sulfuric acid, ACS trace metals grade. If the sample is water soluble, dissolve it in water and dilute to 80 mL with water. If the sample is not water soluble, methanol or another suitable solvent may be used (Note 4).

NOTE 3—The sample size should not exceed 80 mL. When using large samples, the miscibility of the samples and the reagents should be checked before the determination is made. In any case, preliminary tests must be made to determine if the sample or any impurities in the sample interfere in any way with the analysis. If a 1-cm cell is used, the sample must contain at least 5 μg of iron.

NOTE 4—Solvents that have been found suitable for use without