



Designation: E236 – 66 (Reapproved 2022)

Standard Specification for Apparatus for Microdetermination of Alkoxy Groups¹

This standard is issued under the fixed designation E236; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers apparatus for the microdetermination of methoxyl and ethoxyl groups in organic compounds.

NOTE 1—This specification was originally developed by the Committee on Microchemical Apparatus, Division of Analytical Chemistry, American Chemical Society.²

1.2 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Reaction Flask

2.1 The reaction flask shall be of borosilicate glass and shall conform to the requirements of Fig. 1.

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.01 on Laboratory Ware and Supplies.

Current edition approved Sept. 1, 2022. Published September 2022. Originally approved in 1964. Last previous edition approved in 2017 as E236 – 66(2017). DOI: 10.1520/E0236-66R22.

² Committee on Microchemical Analysis, Division of Analytical Chemistry, American Chemical Society. "Report on Recommended Specifications for Microchemical Apparatus, Alkoxy." *Analytical Chemistry*, Vol 28, 1956, p. 112.

3. Condenser with Scrubber

3.1 The West-type condenser with scrubber shall be of borosilicate glass and conform to the requirements of Fig. 2.

NOTE 2—Spent scrubbing solution may be removed from the scrubber by inserting through the side arm of the scrubber a drawn-out medicine dropper or small rubber tube connected to a vacuum source.

NOTE 3—The scrubber is constricted to prevent scrubbing solution from being sucked back into the reaction flask or carried over into the receiver.

4. Delivery Tube

4.1 The delivery tube shall be of borosilicate glass and conform to the requirements of Fig. 3.

5. Receivers

5.1 Two types of receivers are specified, as follows:


5.1.1 *Receiver for Volumetric Analysis*³ shall be of borosilicate glass and conform to the requirements of Fig. 4.

5.1.2 *Receiver for Gravimetric Analysis* shall be of borosilicate glass and conform to the requirements of Fig. 5.

6. Keywords

6.1 alkoxy; apparatus; groups; microdetermination

³ Used to establish the official A.O.A.C. microchemical method for alkoxy group determination, *Official Methods of Analysis*, 9th Ed., Association of Official Agricultural Chemists, Washington, DC, 1960, p. 646.




All dimensions are in millimetres.

FIG. 1 Reaction Flask

iTeh Standards
(<https://standards.itih.ai>)
Document Preview

[ASTM E236-66\(2022\)](https://standards.itih.ai/catalog/standards/sist/3b5dd17c-49dd-4d27-8965-decfd1975b3d/astm-e236-662022)

<https://standards.itih.ai/catalog/standards/sist/3b5dd17c-49dd-4d27-8965-decfd1975b3d/astm-e236-662022>



All dimensions are in millimetres.

FIG. 2 West-Type Condenser with Scrubber