



Designation: D801 – 02 (Reapproved 2022)

Standard Test Methods for Sampling and Testing Dipentene¹

This standard is issued under the fixed designation D801; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 These test methods cover procedures for sampling and testing dipentene and related terpene solvents, consisting chiefly of monocyclic terpene hydrocarbons distilling above the range for turpentine.

1.2 The procedures given in these test methods appear in the following order:

	Section
Sampling	4
Detection and Removal of Separated Water	5
Appearance	6
Color	7
Odor	8
Specific Gravity	9
Refractive Index	10
Composition	11
Flash Point	12
Moisture	13

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

¹ These test methods are under the jurisdiction of ASTM Committee D01 on Paint and Related Coatings, Materials, and Applications and are the direct responsibility of Subcommittee D01.34 on Pine Chemicals and Hydrocarbon Resins.

Current edition approved Dec. 1, 2022. Published December 2022. Originally approved in 1944. Last previous edition approved in 2017 as D801 – 02 (2017). DOI: 10.1520/D0801-02R22.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

- D268 Guide for Sampling and Testing Volatile Solvents and Chemical Intermediates for Use in Paint and Related Coatings and Material
- D890 Test Method for Water in Liquid Pine Chemicals
- D1209 Test Method for Color of Clear Liquids (Platinum-Cobalt Scale)
- D6166 Test Method for Color of Pine Chemicals and Related Products (Instrumental Determination of Gardner Color)
- D6387 Test Methods for Composition of Turpentine and Related Terpene Products by Capillary Gas Chromatography
- E300 Practice for Sampling Industrial Chemicals

3. Significance and Use

3.1 The testing procedures described in these test methods have been in use for many years and emphasize the physical properties rather than the chemical composition of dipentene. These procedures were sufficient when dipentene was used primarily as a solvent. Currently, however, dipentene finds application as a chemical raw material and a knowledge of its chemical composition is therefore important. Test Methods D6387 describe a capillary gas chromatography method which is suitable for determining both the major and minor components found in dipentene.

4. Sampling

4.1 Sample the material in accordance with the procedure described in Guide D268 and Practice E300. If a sample from close to the bottom of a tank shows a decided difference in color or appearance from samples taken at other depths, take an extra bottom sample and examine it separately from the composite sample. In such case the composite sample shall not include any portion of such bottom sample.

5. Detection and Removal of Separated Water

5.1 Draw a portion of the dipentene by means of a glass or metal container with a removable stopper or top,³ or with a thief, from the lowest part of the container, or by opening the bottom valve of the level tank car. If water is found to be

³ Detailed description of equipment suitable for such sampling is given in Practice E300.

present, draw it all out, record the quantity, and deduct it from the total volume of liquid delivered.

6. Appearance

6.1 Examine a portion of the sample after agitation to determine its clarity and freedom from foreign matter and separated water.

7. Color

7.1 Compare the color of the sample in any suitable or designated apparatus with the accepted or specific color standard. Accepted color standards are the Gardner color scale found in Test Method **D6166** and the platinum-cobalt scale found in Test Method **D1209**.

8. Odor

8.1 Compare the odor of the sample with an agreed upon water-free reference sample kept in the dark in a completely filled well-stoppered bottle. In the absence of such a reference sample, compare with samples of known purity similarly preserved.

8.2 The use of an odor panel made up of personnel trained to smell terpenic odors is recommended.

9. Specific Gravity

9.1 Determine the specific gravity at 15.6/15.6 °C by any convenient method, reporting the value to the nearest 0.0005. Correct determinations made at any other temperature, using apparatus standardized at 15.6 °C by adding to or subtracting from the observed reading 0.00082 for each degree Celsius that the temperature of the liquid is above or below 15.6 °C.

10. Refractive Index

10.1 Determine the refractive index with an instrument having a resolution of ± 0.0001 at 20 °C, if possible. If tested at any other temperature, correct the reading obtained to 20 °C by adding or subtracting 0.00045 for each degree Celsius that

the temperature at which the determination was made is respectively above or below 20 °C.

11. Composition

11.1 Determine the component concentration of dipentene using gas chromatography. The referenced method (Test Methods **D6387**) employs capillary column technology.

12. Flash Point

12.1 Determine the flash point by any acceptable means agreed upon between the producer and the customer.

13. Moisture

13.1 The recommended method for the determination of dissolved water uses the Karl Fischer reagent and is described in Test Method **D890**. This method is more rapid than the distillation method described in **13.2** and is applicable to oxidized products.

13.2 *Alternative Method—Distillation:*

13.2.1 The azeotropic distillation method described in the procedure section of Test Method **D890** can be used for moisture analysis of dipentene.

14. Precision and Bias

14.1 Some of the procedures described in these test methods were developed many years ago and were once used widely for the sampling and testing of dipentene. Currently these test methods are less widely used and it is not considered practical to determine the precision and bias of the individual test methods at this time.

14.2 Where reference is made to other ASTM test methods, these test methods should be examined to determine their precision and bias.

15. Keywords

15.1 dipentene; gas chromatography; sampling

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/