

TECHNICAL SPECIFICATION



**Medical electrical equipment –
Part 4-2: Guidance and interpretation – Electromagnetic immunity: performance
of medical electrical equipment and medical electrical systems**

Document Preview

[IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>





THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

International
Standards
Document Preview

[IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>

TECHNICAL SPECIFICATION



**Medical electrical equipment –
Part 4-2: Guidance and interpretation – Electromagnetic immunity: performance
of medical electrical equipment and medical electrical systems**

Document Preview

[IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.01; 33.100.20

ISBN 978-2-8322-8289-2

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	8
0.1 * General	8
0.2 Purpose of this document	8
0.3 How to use this document.....	8
0.4 IMMUNITY TEST LEVELS.....	9
1 Scope.....	10
2 Normative references	10
3 Terms and definitions	12
4 General recommendations.....	15
4.1 Concurrent and sequential testing.....	15
4.2 General test conditions	15
4.2.1 Configurations	15
4.2.2 Artificial hand	16
4.2.3 Power input voltages and frequencies.....	17
5 ME EQUIPMENT and ME SYSTEMS identification, marking and documents	17
5.1 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts for which the connector testing exemption specified in 8.13.2 d) is used.....	17
5.2 ACCOMPANYING DOCUMENTS	17
5.2.1 General	17
5.2.2 Instructions for use	17
5.2.3 Requirements applicable to ME EQUIPMENT and ME SYSTEMS for which the connector testing exemption specified in 8.13.2 d) is used.....	17
5.2.4 * Technical description	18
6 Documentation of the tests	18
6.1 Test plan.....	18
6.2 Test report	19
7 * EMISSIONS	19
8 IMMUNITY recommendations	19
8.1 General.....	19
8.2 PATIENT physiological simulation	20
8.3 Termination of PATIENT-COUPLED parts.....	21
8.4 HAND-HELD ME EQUIPMENT and parts intended to be HAND-HELD.....	21
8.5 Subsystems	21
8.6 PERMANENTLY INSTALLED LARGE ME EQUIPMENT and LARGE ME SYSTEMS.....	21
8.7 Operating modes	22
8.8 Non-ME EQUIPMENT	22
8.9 * Environments of INTENDED USE.....	22
8.10 * Performance criteria	23
8.11 * IMMUNITY TEST LEVELS	23
8.12 * IMMUNITY to proximity fields from RF wireless communications equipment	30
8.13 * ESD testing of connectors	31
8.13.1 Application of ESD to connectors.....	31
8.13.2 Exclusions	32
8.14 Test report	32
Annex A (informative) General guidance and rationale.....	34

Annex B (informative) Guide to labelling recommendations	40
B.1 ACCOMPANYING DOCUMENTS, instructions for use	40
B.2 ACCOMPANYING DOCUMENTS, technical description	40
Annex C (informative) Determination of IMMUNITY TEST LEVELS for SPECIAL ENVIRONMENTS	41
C.1 General	41
C.2 EM DISTURBANCE level determination	42
C.3 Assessment of EM DISTURBANCE sources	42
C.4 Test methods	42
C.5 Test plan	42
C.6 Examples of mitigations and special conditions	43
Annex D (informative) Identification of specific IMMUNITY performance criteria	44
D.1 General	44
D.2 IMMUNITY performance criteria principles	44
D.2.1 General	44
D.2.2 IMMUNITY performance criteria for non-ME EQUIPMENT used in an ME SYSTEM	44
D.2.3 IMMUNITY performance criteria determination	44
D.3 IMMUNITY performance criteria examples	44
D.3.1 General examples	44
D.3.2 Example of immunity performance criteria for a radiological table system	46
D.3.3 Example of immunity performance criteria for ultrasonic diagnostic equipment	46
Annex E (informative) Performance criteria specified by IEC 61000-6-x generic EMC standards	48
Annex F (informative) Mapping between this document and the elements of IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020	49
Bibliography	55
Index of defined terms used in this document	57
Figure 1 – RC element of the artificial hand	16
Figure 2 – * PORTS of ME EQUIPMENT and ME SYSTEMS	20
Figure 3 – Examples of locations within EM ENVIRONMENTS	25
Table 1 – Recommended minimum test plan	18
Table 2 – * ENCLOSURE PORT	26
Table 3 – * Input AC power PORT	27
Table 4 – Input DC power PORT	28
Table 5 – * PATIENT COUPLING PORT	29
Table 6 – SIP/SOP PORT	29
Table 7 – Test specifications for enclosure port immunity to RF wireless communications equipment	30
Table 8 – Parts of connectors to be tested for ESD, based on the connector shell and cover material	31
Table 9 – * Testing of connectors and pins while connected and disconnected	31
Table 10 – Test report minimum contents	32

Table A.1 – Assumptions used in determining IMMUNITY TEST LEVELS specified in Table 7 38

Table B.1 – ACCOMPANYING DOCUMENTS, instructions for use 40

Table B.2 – ACCOMPANYING DOCUMENTS, technical description..... 40

Table C.1 – Examples of adjusted IMMUNITY TEST LEVELS due to mitigations or special conditions 43

Table D.1 – Example of IMMUNITY performance criteria for a radiological table and gantry system 46

Table D.2 – Example of IMMUNITY performance criteria for ULTRASONIC DIAGNOSTIC EQUIPMENT 47

Table F.1 – Mapping between the elements of this document and IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020 49

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 4-2: Guidance and interpretation – Electromagnetic immunity:
performance of medical electrical equipment
and medical electrical systems**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC TS 60601-4-2 has been prepared by subcommittee 62A: Common aspects of medical equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is a Technical Specification.

This first edition cancels and replaces the first edition of IEC TR 60601-4-2 published in 2016. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to IEC TR 60601-4-2:2016:

- a) aligned with IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020,
- b) updated references,

- c) changed conducted disturbance cable length exclusion (Table 6) from 3 m to 1 m and added a citation of IEC 61000-4-6:2013, Annex B.

The text of this Technical Specification is based on the following documents:

Draft	Report on voting
62A/1532/DTS	62A/1560/RVDTS

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this Technical Specification is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- recommendations and definitions: roman type.
- *test instructions: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 1 includes 1.1, 1.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 1.1, 1.2 and 1.3.1 are all subclauses of Clause 1).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document; however, we chose to use it in this document only as described in 0.3;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>

INTRODUCTION

0.1 * General

MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS are used in the practice of medicine because they provide functions that are associated with the INTENDED USE. If MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM does not provide these functions because of a lack of IMMUNITY to EM DISTURBANCES, that are expected to occur in the environment(s) of INTENDED USE, this can interfere with the practice of medicine. This document provides guidance for IMMUNITY of functions of INTENDED USE that are not related to BASIC SAFETY or ESSENTIAL PERFORMANCE.

Examples of functions that might not be BASIC SAFETY or ESSENTIAL PERFORMANCE but that might be INTENDED USE include the following:

- the ability to print an ultrasound image remotely;
- the ability of a scale to accurately measure PATIENT weight;
- accuracy of X-RAY TUBE VOLTAGE in X-ray equipment for radiography and radioscopy, e.g. the error is less than 5 %.

See Annex A for more information.

In general, the IMMUNITY TEST LEVELS in IEC 60601-1-2 are based on reasonably foreseeable maximum levels of EM DISTURBANCES. In this document, IMMUNITY TEST LEVELS for performance are based on typical levels of EM DISTURBANCES. Rationale concerning test methodology can be found in Annex A of this document and in Annex A of IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020.

NOTE In general, typical IMMUNITY TEST LEVELS are equal to or lower than reasonably foreseeable maximum levels.

0.2 Purpose of this document

The purpose of this document is to provide a consistent method for evaluating the ability of MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM to perform without degradation of performance in the presence of typical EM DISTURBANCES.

0.3 How to use this document

This document can be used in conjunction with IEC 60601-1-2 and testing for conformity to both documents can be done at the same time. This allows IMMUNITY testing of BASIC SAFETY, ESSENTIAL PERFORMANCE and performance of the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM during one test, concurrently or sequentially. The main difference is the use of performance criteria instead of pass/fail criteria, and differences can also include modes and configurations. For BASIC SAFETY and ESSENTIAL PERFORMANCE, the pass/fail criteria are determined as specified by IEC 60601-1-2. For performance, the criteria are determined by the specifications, instructions and information provided by the MANUFACTURER.

This document uses "recommend" and "should" in place of "shall" in most cases. "Shall" is used where an action is required by other standards or something needs to be done in a prescribed way in order to be effective. Also, this document has "normative" references. They are "normative" because if you choose to follow the recommendations of this document, they are indispensable for that use. An example of this would be testing for radiated RF IMMUNITY. The test methods of IEC 61000-4-3 would be indispensable for this testing.

0.4 IMMUNITY TEST LEVELS

The IMMUNITY TEST LEVELS specified in this document are typical for the locations of INTENDED USE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS. However, Annex C provides a method for modifying the specified typical IMMUNITY TEST LEVELS for performance if necessary or for particular environments (e.g. SPECIAL ENVIRONMENTS) for which this document does not specify IMMUNITY TEST LEVELS.

iTeh Standards
(<https://standards.iteh.ai>)
Document Preview

[IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>

MEDICAL ELECTRICAL EQUIPMENT –

Part 4-2: Guidance and interpretation – Electromagnetic immunity: performance of medical electrical equipment and medical electrical systems

1 Scope

This document applies to the performance of MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM in the presence of EM DISTURBANCES. Hereafter, MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM are referred to as ME EQUIPMENT or an ME SYSTEM, respectively.

The object of this document is to provide guidance on the assessment of the performance of ME EQUIPMENT or an ME SYSTEM in the presence of EM DISTURBANCES.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

IEC 60417:2002, *Graphical symbols for use on equipment*, available from: <http://www.graphical-symbols.info/equipment>

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012

IEC 60601-1:2005/AMD2:2020

IEC 60601-1-2:2014, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: EM disturbances – Requirements and tests*

IEC 60601-1-2:2014/AMD1:2020

IEC 60601-1-8:2006, *Medical electrical equipment – Part 1-8: General requirements for basic safety and essential performance – Collateral standard: General requirements tests and guidance for alarm systems in medical electrical equipment and medical electrical systems*

IEC 60601-1-8:2006/AMD1:2012

IEC 60601-1-8:2006/AMD2:2020

IEC 60601-1-11:2015, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment*

IEC 60601-1-11:2015/AMD1:2020

IEC 60601-1-12:2014, *Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment*

IEC 60601-1-12:2014/AMD1:2020

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3:2006¹, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-3:2006/AMD1:2007

IEC 61000-4-3:2006/AMD2:2010

IEC 61000-4-4:2012, *Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*

IEC 61000-4-5:2014, *Electromagnetic compatibility (EMC) – Part 4-5: Testing and measurement techniques – Surge immunity test*

IEC 61000-4-5:2014/AMD1:2017

IEC 61000-4-6:2013², *Electromagnetic compatibility (EMC) – Part 4-6: Testing and measurement techniques – Immunity to conducted disturbances, induced by radio-frequency fields*

IEC 61000-4-8:2009, *Electromagnetic compatibility (EMC) – Part 4-8: Testing and measurement techniques – Power frequency magnetic field immunity test*

IEC 61000-4-11:2004³, *Electromagnetic compatibility (EMC) – Part 4-11: Testing and measurement techniques – Voltage dips, short interruptions and voltage variations immunity tests*

IEC 61000-4-11:2004/AMD1:2017 [IEC TS 60601-4-2:2024](https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024)

<https://standards.iteh.ai/catalog/standards/iec/0da29006-c829-4490-a7b5-9a926c40f79d/iec-ts-60601-4-2-2024>

CISPR 16-1-2:2014, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-2: Radio disturbance and immunity measuring apparatus – Coupling devices for conducted disturbance measurements*

CISPR 16-1-2:2014/AMD1:2017

ISO 7637-2:2011, *Road vehicles – Electrical disturbances from conduction and coupling – Part 2: Electrical transient conduction along supply lines only*

¹ This publication was withdrawn.

² This publication was withdrawn and replaced with IEC 61000-4-6:2023.

³ This publication was withdrawn.

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-8:2006, IEC 60601-1-8:2006/AMD1:2012, IEC 60601-1-8:2006/AMD2:2020, IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020, IEC 60601-1-12:2014, IEC 60601-1-12:2014/AMD1:2020 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

NOTE 1 Where the terms "voltage" and "current" are used in this document, they mean the RMS values of an alternating, direct or composite voltage or current unless stated otherwise.

NOTE 2 The term "electrical equipment" is used to mean ME EQUIPMENT or other electrical equipment. This document also uses the term "equipment" to mean ME EQUIPMENT or other electrical or non-electrical equipment in the context of an ME SYSTEM.

NOTE 3 The dictionary definition of "performance" applies.

NOTE 4 An index of defined terms is found at the end of the document.

3.1

ACCESSIBLE PART

part of electrical equipment other than an APPLIED PART that can be touched by means of the standard test finger

Note 1 to entry: See also 5.9.2.1 of IEC 60601-1:2005.

[SOURCE: IEC 60601-1:2005, 3,2, modified – the original NOTE has been modified to add a reference to IEC 60601-1:2005.]

3.2

APPLIED PART

part of ME EQUIPMENT that in NORMAL USE necessarily comes into physical contact with the PATIENT for ME EQUIPMENT or an ME SYSTEM to perform its function

Note 1 to entry: See Figure 3, Figure 4 and Figure A.1 to Figure A.7 (inclusive) of IEC 60601-1:2005.

[SOURCE: IEC 60601-1:2005, 3,8, modified – Note 1 has been modified to add a reference to IEC 60601-1:2005 and Note 2 and Note 3 have been deleted.]

3.3

ELECTROMAGNETIC DISTURBANCE

EM DISTURBANCE

any electromagnetic phenomenon that could degrade the performance of a device, equipment or system

Note 1 to entry: An EM DISTURBANCE can be ELECTROMAGNETIC NOISE, an unwanted signal or a change in the propagation medium itself.

[SOURCE: IEC 60601-1-2:2014, 3.3]

3.4

EMISSION

ELECTROMAGNETIC EMISSION

the phenomenon by which electromagnetic energy emanates from a source

[SOURCE: IEC 60601-1-2:2014, 3.4]