



**SLOVENSKI STANDARD**  
**SIST EN ISO 105-J01:1999**

**01-november-1999**

---

**Tekstilije - Preskušanje barvne obstojnosti - Del J01: Splošna načela merjenja barve na površini (ISO 105-J01:1997)**

Textiles - Tests for colour fastness - Part J01: General principles for measurement of surface colour (ISO 105-J01:1997)

Textilien - Farbechtheitsprüfungen - Teil J01: Grundlagen für die Messung von Körperfarben (ISO 105-J01:1997)

Textiles - Essais de solidité des teintures - Partie J01: Principes généraux du mesurage de la couleur de surface (ISO 105-J01:1997)

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>

**Ta slovenski standard je istoveten z: EN ISO 105-J01:1999**

---

**ICS:**

59.080.01      Tekstilije na splošno      Textiles in general

**SIST EN ISO 105-J01:1999**      en

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 105-J01:1999](#)

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>

EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

EN ISO 105-J01

May 1999

ICS 59.080.01

English version

Textiles - Tests for colour fastness - Part J01: General principles  
for measurement of surface colour (ISO 105-J01:1997)

Textiles - Essais de solidité des teintures - Partie J01:  
Principes généraux du mesurage de la couleur de surface  
(ISO 105-J01:1997)

Textilien - Farbechtheitsprüfungen - Teil J01: Grundlagen  
für die Messung von Körperfarben (ISO 105-J01:1997)

This European Standard was approved by CEN on 18 April 1999.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Page 2  
EN ISO 105-J01:1999

### Foreword

The text of the International Standard from Technical Committee ISO/TC 38 "Textiles" of the International Organization for Standardization (ISO) has been taken over as an European Standard by Technical Committee CEN/TC 248 "Textiles and textile products", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 1999, and conflicting national standards shall be withdrawn at the latest by November 1999.

According to the CEN/CENÉLEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

### Endorsement notice

The text of the International Standard ISO 105-J01:1997 has been approved by CEN as a European Standard without any modification.

NOTE: Normative references to International Standards are listed in annex ZA (normative).

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN ISO 105-J01:1999

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>

**Annex ZA (normative)**  
**Normative references to international publications**  
**with their relevant European publications**

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN</u>	<u>Year</u>
ISO 139	1973	Textiles - Standard atmospheres for conditioning and testing	EN 20139	1992

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

SIST EN ISO 105-J01:1999

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

[SIST EN ISO 105-J01:1999](#)

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>

INTERNATIONAL  
STANDARD

ISO  
105-J01

Fourth edition  
1997-12-15

---

---

**Textiles — Tests for colour fastness —**

**Part J01:**

General principles for measurement of surface  
colour

*Textiles — Essais de solidité des teintures —  
Partie J01: Principes généraux du mesurage de la couleur de surface*  
**(standards.iteh.ai)**

[SIST EN ISO 105-J01:1999](https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999)

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>



Reference number  
ISO 105-J01:1997(E)

**ISO 105-J01:1997(E)****Foreword**

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 105-J01 was prepared by Technical Committee ISO/TC 38, *Textiles*, Subcommittee SC 1, *Tests for coloured textiles and colorants*.

This fourth edition cancels and replaces the third edition (ISO 105-J01:1989), which has been technically revised.

ISO 105 was previously published in thirteen "parts", each designated by a letter (e.g. "Part A"), with publication dates between 1978 and 1985. Each part contained a series of "sections", each designated by the respective part letter and by a two-digit serial number (e.g. "Section A01"). These sections are now being republished as separate documents, themselves designated "parts" but retaining their earlier alphanumeric designations. A complete list of these parts is given in ISO 105-A01.

Annex A forms an integral part of this part of ISO 105. Annex B is for information only.

© ISO 1997

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization  
Case postale 56 • CH-1211 Genève 20 • Switzerland  
Internet central@iso.ch  
X.400 c=ch; a=400net; p=iso; o=isocs; s=central

Printed in Switzerland



# Textiles — Tests for colour fastness —

## Part J01:

### General principles for measurement of surface colour

#### 1 Scope

This part of ISO 105 is designed as a reference document to support the proper measurement of the colour of specimens by instrumental means as required in many parts of ISO 105. The document describes general concepts and problems associated with reflectance colour measurement.

Annex A specifies techniques and specimen handling procedures.

#### 2 Normative references

**iTeh STANDARD PREVIEW**  
**(standards.iteh.ai)**

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 105. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreement based on this part of ISO 105 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 139:1973, *Textiles — Standard atmospheres for conditioning and testing*.

CIE Publication No. 15.2, *Colorimetry*, 2nd ed. (1986)<sup>1)</sup>.

#### 3 Definitions

For the purposes of this part of ISO 105, the following definitions apply.

**3.1 colour measurement:** A numerical representation of the colour of a specimen obtained by use of a colour measuring instrument; a single measurement may represent an average of multiple readings of a specimen.

1) Available from the International Commission on Illumination Central Bureau, Kegelgasse 27, A-1030 Vienna, Austria.

**3.2 colour measuring instrument:** Any device (such as a colorimeter or spectrophotometer) used to measure the relative amount of radiation reflected from a specimen in the visible region of the spectrum (comprising the wavelengths from 360 nm to 780 nm, and including as a minimum the region from 400 nm to 700 nm).

**3.3 geometry (of a colour measuring instrument):** One of the following illumination/viewing conditions.

d/0	0/d	0/45	45/0
-----	-----	------	------

which describes the angle or manner in which a colour measuring instrument

1) illuminates the specimen:

d	0	0	45
---	---	---	----

2) views the resulting reflected light:

0(0°-10°)	d	45	0
-----------	---	----	---

d = diffuse; 0 = normal (0° to 10°); 45 (45° ± 2°) = tolerable range of the angle between the direction of illumination or viewing and the normal to the specimen.

NOTE — Instruments of different geometries may produce different colorimetric results on most textile materials.

SIST EN ISO 105-J01:1999

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c->

**3.4 area-of-view [optical aperture] (of a colour measuring instrument):** The dimensions (size and shape) of the surface area that a colour measuring instrument is capable of covering in a single colour measurement.

**3.5 fluorescence:** A phenomenon in which radiant flux of certain wavelengths is absorbed and re-emitted at other, usually longer, wavelengths.

**3.6 reflectance:** The ratio of the reflected radiant or luminous flux (light) to the incident flux in the given conditions.

**3.7 reflectance factor:** The ratio of the flux reflected from the specimen to the flux reflected from the perfect reflecting diffuser under the same geometric and spectral conditions of measurement.

**3.8 specular reflection:** The reflection without diffusion, in accordance with the laws of optical reflection, as in a mirror.

**3.9 standardization (of colour measuring instrument):** The act of measuring one or more calibrated materials with a colour measuring instrument for the purpose of calculating a set of correction factors to be applied to subsequent measurements.

NOTE — Calibration is typically performed by an instrument manufacturer to ensure that the instrument meets the criteria as established by national standardizing laboratories.

**3.10 verification standard:** In colour measurement, any stable material which is used for the purpose of confirming (or verifying) the validity of an instrument standardization. Colour measurements, which are made immediately following a standardization, are compared to original measurements of the standard to detect improper standardization.

## 4 Principle

Materials of an opaque or nearly opaque nature (but not translucent) are measured by reflectance methods in order to obtain a numerical representation of the colour of the specimen.

### NOTES

1 Proper equipment set-up, standardization of the colour measuring instrument and proper presentation of the test specimens to the instrument are required to achieve consistent, reliable and meaningful reflectance measurement results.

2 In general, instrumental colour measurement procedures are dictated by the type of specimen to be measured and the instrument with which it will be measured. Many types of colour measuring instrumentation are available, differing in such features as area-of-view, illumination method, and geometry. The user is cautioned that conflicting results may be obtained on comparisons of data acquired on instruments of different designs.

[SIST EN ISO 105-J01:1999](https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999)

<https://standards.iteh.ai/catalog/standards/sist/969669dc-8d1d-4b22-be3c-89a512c0bd7e/sist-en-iso-105-j01-1999>

## 5 Apparatus

**5.1 Reflectance colour measuring instrument,** for illuminating a specimen and measuring the amount of light which is reflected from the surface of the specimen. Illumination is usually polychromatic (white light); however monochromatic mode is acceptable for nonfluorescent specimens. Reflectance colour measuring instruments may be broadly divided into two groups:

- a) Spectrophotometers (typically diffuse/0, using polychromatic illumination) separate and measure the spectrum of light reflected from the specimen relative to a reference white at regular intervals (wavelength intervals of 5 nm, 10 nm and 20 nm are most common). These data may be used to calculate the desired tristimulus values ( $X, Y, Z$ ) for any given illuminant and observer. Some spectrophotometers (typically 0/diffuse) illuminate the sample with monochromatic light and measure the amount of light reflected from the surface as the sample is illuminated at regular wavelength intervals.
- b) Colorimeters measure the tristimulus values ( $X, Y, Z$ ) directly through broadband filters which are designed to produce colorimetric values for one illuminant and observer (typically  $C/2$ ). Measurement of reflectance factors at specific wavelengths is not possible with a colorimeter.

Within these two categories, the instruments are further defined by their geometry as defined in 3.3.