



Designation: 51607 – 22

## Standard Practice for Use of an Alanine-EPR Dosimetry System<sup>1</sup>

This standard is issued under the fixed designation 51607; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reappraisal.

### 1. Scope

1.1 This practice covers dosimeter materials, instrumentation, and procedures for using the alanine-EPR dosimetry system to measure the absorbed dose in the photon or electron radiation processing of materials. The alanine system is based on electron paramagnetic resonance (EPR) spectroscopy of free radicals derived from the amino acid alanine.<sup>2</sup>

1.2 The alanine dosimeter is classified as a type I dosimeter as it is affected by individual influence quantities in a well-defined way that can be expressed in terms of independent correction factors (see ISO/ASTM Practice 52628). The alanine dosimeter may be used in either a reference standard dosimetry system or in a routine dosimetry system.

1.3 This document is one of a set of standards that provides recommendations for properly implementing dosimetry in radiation processing, and describes a means of achieving compliance with the requirements of ISO/ASTM 52628 “Practice for Dosimetry in Radiation Processing” for alanine dosimetry system. It should be read in conjunction with ISO/ASTM 52628.

1.4 This practice covers the use of alanine-EPR dosimetry systems under the following conditions:

1.4.1 The absorbed dose range is between 0.001 kGy and 150 kGy.

1.4.2 The absorbed dose rate is up to  $1 \times 10^2$  Gy s<sup>-1</sup> for continuous radiation fields and up to  $3 \times 10^{10}$  Gy s<sup>-1</sup> for pulsed radiation fields (1-4).<sup>3</sup>

1.4.3 The radiation energy for photons and electrons is between 0.1 MeV and 30 MeV (1, 2, 5-8).

1.4.4 The irradiation temperature is between -78 °C and +70 °C (2, 9-12).

1.5 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.

1.6 This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

### 2. Referenced documents

#### 2.1 ASTM Standards:<sup>4</sup>

E3083 Terminology Relating to Radiation Processing: Dosimetry and Applications

#### 2.2 ISO/ASTM Standards:<sup>4</sup>

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51707 Guide for Estimation of Measurement Uncertainty in Dosimetry for Radiation Processing

52628 Practice for Dosimetry in Radiation Processing

52701 Guide for Performance Characterization of Dosimeters and Dosimetry Systems for Use in Radiation Processing

#### 2.3 International Commission on Radiation Units and Measurements (ICRU) Reports:<sup>5</sup>

ICRU Report 80 Dosimetry Systems for Use in Radiation Processing

ICRU Report 85a Fundamental Quantities and Units for Ionizing Radiation

#### 2.4 ISO Standard:<sup>6</sup>

12749-4 Nuclear energy – Vocabulary – Part 4: Dosimetry for radiation processing

<sup>1</sup> This practice is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.02 on Dosimetry Systems. Originally developed as a joint ASTM/ISO standard in conjunction with ISO/TC 85/WG 3.

Current edition approved Jan. 1, 2022. Published May 2024. Originally approved in 1994. Last previous edition approved in 2013 as ISO/ASTM 51607:2013(E). DOI: 10.1520/51607-22.

<sup>2</sup> The term “electron spin resonance” (ESR) is used interchangeably with electron paramagnetic resonance (EPR).

<sup>3</sup> The boldface numbers in parentheses refer to the bibliography at the end of this standard.

<sup>4</sup> For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard’s Document Summary page on the ASTM website.

<sup>5</sup> Available from International Commission on Radiation Units and Measurements (ICRU), 7910 Woodmont Ave., Suite 400, Bethesda, MD 20814-3095, U.S.A., <https://www.icru.org/>.

<sup>6</sup> Available from International Organization for Standardization (ISO), ISO Central Secretariat, Chemin de Blandonnet 8, CP 401, 1214 Vernier, Geneva, Switzerland, <https://www.iso.org>.

## 2.5 Joint Committee for Guides in Metrology (JCGM)

### Reports:

JCGM 100:2008, GUM 1995, with minor corrections, Evaluation of measurement data – Guide to the Expression of Uncertainty in Measurement<sup>7</sup>

JCGM 200:2012 (JCGM 200:2008 with minor revisions), VIM International Vocabulary of Metrology – Basic and General Concepts and Associated Terms<sup>8</sup>

## 3. Terminology

### 3.1 Definitions:

3.1.1 *alanine dosimeter*—specified quantity and physical form of the radiation-sensitive material alanine and any added inert substance such as a binder.

3.1.2 *alanine-EPR dosimetry system*—system used for determining absorbed dose, consisting of alanine dosimeters, an EPR spectrometer and its associated reference materials, and procedures for the system's use.

3.1.3 *alanine-EPR dosimeter response*—value resulting from applied adjustments to the EPR signal amplitude.

3.1.4 *check standard*—a standard prepared independently of the calibration standards that is measured to verify the performance of a dosimetry system.

3.1.5 *EPR intensity reference material*—a stable paramagnetic material whose measurement by EPR is applied to the dosimeter EPR signal amplitude as part of the dosimeter response determination.

3.1.6 *EPR signal amplitude*—peak-to-peak amplitude of the central signal of the EPR spectrum.

3.1.6.1 *Discussion*—This signal is proportional to the alanine-derived free radical concentration in the alanine dosimeter.

3.1.7 *EPR spectroscopy*—measurement of resonant absorption of electromagnetic energy resulting from the transition of unpaired electrons between different energy levels, upon application of radio frequencies to a paramagnetic substance in the presence of a magnetic field.

3.1.8 *EPR spectrum*—first derivative of the electron paramagnetic absorption spectrum measured as a function of the magnetic field.

3.1.9 *zero dose amplitude*—EPR signal amplitude of an unirradiated alanine dosimeter with the same EPR spectrometer parameters used for the lowest measurable absorbed dose value.

3.2 Definitions of other terms used in this standard that pertain to radiation measurement and dosimetry may be found in ASTM Terminology E3083. Definitions in Terminology E3083 are compatible with ICRU Report 85a; that document, therefore, may be used as an alternative reference.

<sup>7</sup> Document produced by Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1). Available free of charge at the BIPM website (<http://www.bipm.org>).

<sup>8</sup> Document produced by Working Group 2 of the Joint Committee for Guides in Metrology (JCGM/WG 2). Available free of charge at the BIPM website (<http://www.bipm.org>).

## 4. Significance and use

4.1 The alanine-EPR dosimetry system provides a means for measuring absorbed dose. It is based on the measurement of specific stable free radicals in crystalline alanine generated by ionizing radiation.

4.2 Alanine-EPR dosimetry systems are used in reference- or transfer-standard or routine dosimetry systems in radiation applications that include: sterilization of medical devices and pharmaceuticals, food irradiation, polymer modifications, medical therapy and radiation damage studies in materials (1, 13-15).

## 5. Overview

5.1 The dosimeter is prepared using  $\alpha$ -alanine,  $\text{CH}_3\text{-CH}(\text{NH}_2)\text{-COOH}$ , in the form of polycrystalline powder.

5.2 All stereoisomers of  $\alpha$ -alanine are suitable for dosimetry; L-alanine is used most commonly.

5.3 Usual physical shapes are films or pellets (cylinders), and pellets with a packaged form.

NOTE 1—Additives, capsules, or film support materials used in the preparation of dosimeters should not add any significant intrinsic or radiation-induced EPR signal. Examples of suitable binders are ethylene-propylene rubber, gelatin, paraffin, polyethylene, polyethylene vinyl acetate, polystyrene, polyvinylpyrrolidone, polyvinyl propylene, and stearin. Lubricants added in the dosimeter manufacturing process are optional. An example of a suitable lubricant is stearic acid (16-21).

5.4 The dosimeter contains crystalline alanine and registers the absorbed dose by the formation of alanine-derived free radicals (22). Identification and measurement of alanine-derived free radicals are performed by EPR spectroscopy. ICRU Report 80 provides information on the scientific basis and historical development of this dosimetry system.

5.5 The measurement of free radicals by EPR spectroscopy is nondestructive. This can be repeated and hence can be used for archival purposes (23-25).

## 6. Influence quantities

6.1 Factors other than absorbed dose which influence the dosimeter response are referred to as influence quantities, and are discussed in the following sections (see also ISO/ASTM Guide 52701). Examples of such influence quantities are temperature and dose rate.

### 6.2 Pre-Irradiation Conditions:

6.2.1 *Dosimeter Conditioning and Packaging*—Alanine dosimeter conditioning and packaging may be important under certain conditions (see 6.2.4).

NOTE 2—The sorting of alanine pellet dosimeters by mass into sub-lots will improve the measurement uncertainty.

6.2.2 *Time Since Manufacture*—There is no known influence of time since manufacture on alanine dosimeters when stored under recommended conditions.

6.2.3 *Temperature*—There is no known influence of pre-irradiation temperature. However, it is recommended that alanine dosimeters be stored at manufacturer recommended

temperatures. Exposure to temperatures outside the manufacturer's recommended range should be avoided to reduce the potential for adverse effects on dosimeter response.

**6.2.4 Relative Humidity**—The humidity during pre-irradiation storage may influence the EPR signal amplitude of alanine dosimeters (24, 25). The effect of humidity may be reduced by sealing dosimeters in a material impervious to water.

**6.2.5 Exposure to Light**—There is no known influence of ambient light.

### 6.3 Conditions During Irradiation:

**6.3.1 Irradiation Temperature**—The irradiation temperature influences the EPR signal amplitude of alanine dosimeters.

**NOTE 3**—The effect of irradiation temperature on the dosimeter EPR signal amplitude may be dependent on the dosimeter type. The temperature coefficient,  $R, (K^{-1})$  is described by the relationship,  $(\Delta m/m)/\Delta T$ , where  $m$  is the EPR signal amplitude (in arbitrary units) and  $T$  is the irradiation temperature (in K). For dosimeters with L-alanine, a positive temperature coefficient, expressed in percent, in the range of +0.1 to +0.2 % °C<sup>-1</sup> is typical for irradiation temperatures from -10 °C to +70 °C (10, 11, 26-28); refer to Ref (9, 12) for irradiation temperatures below -10 °C. The temperature coefficient for dosimeters prepared with the DL stereoisomer of alanine is more than 50 % higher than one prepared with L-alanine (29). A summary of published temperature coefficients is tabulated in Ref (26, 29).

**6.3.2 Absorbed-Dose Rate**—Under normal radiation processing conditions there is no measurable effect of absorbed dose rate; however, a dose dependent effect has been characterized for alanine dosimeters irradiated to high doses at low dose rates (30).

**NOTE 4**—The dose-rate effect is absorbed-dose dependent. Alanine dosimeters irradiated with gamma radiation to absorbed doses > 5 kGy at low dose rates (< 2 Gy/s) show a progressive decrease in EPR signal amplitudes relative to that found at dose rates greater than 2 Gy/s (30). This combined dose/dose rate effect may reach several percent and is irradiation temperature dependent; though relatively constant above 0 °C, no rate effect was measured at -10 °C and -40 °C (31).

**6.3.3 Dose Fractionation**—There is no known influence of dose fractionation.

**NOTE 5**—In some instances the fractionation of dose to alanine dosimeters may not be straightforward. Certain influence quantities that contribute to the dosimeter response may not be equivalent for the fractionated and non-fractionated irradiations. For example, the fractionation of dose imposes multiple temperature changes to the dosimeter that may not be equivalent to the irradiation temperature experienced by a dosimeter irradiated to a single dose (equal to the sum of the fractionated doses). An accurate comparison of fractionated and non-fractionated doses will depend greatly on an accurate knowledge of the irradiation temperature for the irradiations (see 6.3.1).

**6.3.4 Relative Humidity**—The humidity during irradiation may influence the EPR signal amplitude of alanine dosimeters. The effect of humidity may be reduced by sealing dosimeters in a material impervious to water.

**6.3.5 Exposure to Light**—There is no known influence of ambient light.

**6.3.6 Radiation Energy**—For most radiation processing applications there is no influence of radiation energy for photons and electrons.

**NOTE 6**—Differences have been reported between the absorbed dose to water response of alanine dosimeters irradiated by photons and electrons

over a range of energies (4, 6-8). The response in electron beams has been reported to be 1-2 % lower than in Co-60 beams (8) and the response in 150 kV X-ray beams has been reported to be ~15 % lower (7). The response to ~100 keV electrons was found to be equivalent to the response to high energy electrons (32, 33).

### 6.4 Post-Irradiation Conditions:

**6.4.1 Time**—The interval between irradiation and dosimeter reading shall be standardized and should conform to the manufacturer's recommendations (see 6.4.4). Alanine dosimeters are commonly regarded as stable over time periods as long as weeks or months. However, the degree of stability may be influenced by, but not limited to, absorbed dose, relative humidity, dosimeter composition and this should be characterized by the end user.

**6.4.2 Temperature**—There is no known influence of storage temperature on alanine dosimeters. However, it is recommended that alanine dosimeters be stored in accordance with manufacturer's recommendations.

**6.4.3 Conditioning Treatment**—Post-irradiation treatment is not applicable.

**6.4.4 Storage Relative Humidity**—The humidity during post-irradiation storage can influence the EPR signal amplitude of alanine dosimeters. Sufficient time should be allowed for dosimeters to equilibrate with ambient conditions before measurement (25).

**6.4.5 Exposure to Light**—There is no significant influence of ambient light.

### 6.5 Response Measurement Conditions:

**6.5.1 Exposure to Light**—There is no significant influence of ambient light.

**6.5.2 Temperature**—Controlled temperatures are recommended for measuring alanine dosimeters. Avoid exposure to temperatures outside the manufacturer's recommended range.

**6.5.3 Relative Humidity**—The humidity during measurement can influence the EPR signal amplitude of alanine dosimeters. During measurement, the effects of humidity can be compensated by measuring the ratio of the alanine signal to that of a humidity insensitive EPR intensity reference material (see 7.3.1). If a humidity sensitive EPR reference material is used (for example, an irradiation-calibrated alanine dosimeter) compensate for changes in humidity by standardizing the time of measurement (see 6.4.4) between the reference and dosimeter reading (see Note 8).

**NOTE 7**—Some commercial EPR instrumentation may automatically compensate for nominal changes in temperature and ambient humidity.

**NOTE 8**—The historical data for humidity effects on alanine dosimeters and quantitative EPR measurements have been compiled (25, 34-36).

## 7. Dosimetry system and its verification

**7.1** The following are components of the Alanine-EPR Dosimetry System:

**7.1.1** Alanine Dosimeters.

**7.1.2** EPR Spectrometer.

**7.1.2.1** An X-band EPR spectrometer is used to measure the EPR signal amplitude of an alanine dosimeter. To obtain the expanded uncertainty cited in 12.3, an EPR spectrometer should be capable of the following settings:

(1) microwave frequency 9 GHz to 10 GHz with automatic frequency locking (AFC);