



Designation: E1894 – 24

# Standard Guide for Selecting Dosimetry Systems for Application in Pulsed X-Ray Sources<sup>1</sup>

This standard is issued under the fixed designation E1894; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This guide provides assistance in selecting and using dosimetry systems in flash X-ray experiments. Both dose and dose rate techniques are described.

1.2 Operating characteristics of flash X-ray sources are given, with emphasis on the spectrum of the photon output.

1.3 Assistance is provided to relate the measured dose to the response of a device under test (DUT). The device is assumed to be a semiconductor electronic part or system.

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>2</sup>

**E170** Terminology Relating to Radiation Measurements and Dosimetry

**E666** Practice for Calculating Absorbed Dose From Gamma or X Radiation

**E668** Practice for Application of Thermoluminescence-Dosimetry (TLD) Systems for Determining Absorbed Dose in Radiation-Hardness Testing of Electronic Devices

**E1249** Practice for Minimizing Dosimetry Errors in Radiation Hardness Testing of Silicon Electronic Devices Using Co-60 Sources

### 2.2 ISO/ASTM Standards:<sup>3</sup>

**ISO/ASTM 51261** Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

**ISO/ASTM 51275** Practice for Use of a Radiochromic Film Dosimetry System

**ISO/ASTM 51310** Practice for Use of a Radiochromic Optical Waveguide Dosimetry System

### 2.3 International Commission on Radiation Units (ICRU) and Measurements Reports:<sup>4</sup>

**ICRU Report 14** Radiation Dosimetry: X rays and Gamma Rays with Maximum Photon Energies Between 0.6 and 50 MeV

**ICRU Report 17** Radiation Dosimetry: X rays Generated at Potentials of 5 to 150 kV

**ICRU Report 34** The Dosimetry of Pulsed Radiation

**ICRU Report 51** Quantities and Units in Radiation Protection Dosimetry

**ICRU Report 60** Fundamental Quantities and Units for Ionizing Radiation

**ICRU Report 76** Measurement Quality Assurance for Ionizing Radiation Dosimetry

**ICRU Report 77** Elastic Scattering of Electrons and Positrons

**ICRU Report 80** Dosimetry Systems for Use in Radiation Processing

**ICRU Report 85a** Fundamental Quantities and Units for Ionizing Radiation

## 3. Terminology

3.1 *absorbed dose enhancement*—increase (or decrease) in the absorbed dose (as compared to the equilibrium absorbed dose) at a location in a material of interest. This can be

<sup>1</sup> This guide is under the jurisdiction of ASTM Committee E10 on Nuclear Technology and Applications and is the direct responsibility of Subcommittee E10.07 on Radiation Dosimetry for Radiation Effects on Materials and Devices.

Current edition approved May 1, 2024. Published May 2024. Originally approved in 1997. Last previous edition approved in 2018 as E1894 – 18. DOI: 10.1520/E1894-24.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> For referenced ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>4</sup> Available from the International Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, U.S.A.

expected to occur near an interface with a material of higher or lower atomic number.

3.2 *converter*—a target for electron beams, generally a high atomic number material, in which bremsstrahlung X-rays are produced by radiative energy loss of the incident electrons.

3.3 *dosimetry system*—a system used for determining absorbed dose which consists of dosimeters, measurement instruments with their associated reference standards, and procedures for the system’s use.

3.4 *DUT*—device under test. This is the electronic component or system being tested to determine its performance during and/or after irradiation.

3.5 *endpoint energy*—endpoint energy refers to the peak energy of the electron beam, usually in MeV, generated in a flash X-ray source and is equal to the terminal voltage of the accelerator in megavolts (MV). The word “endpoint” refers to the highest photon energy of the bremsstrahlung spectra, and this endpoint is equal to the maximum or peak in the electron energy. For example, if the most energetic electron that strikes the converter is 10 MeV, this electron produces a range of bremsstrahlung photon energies but the maximum energy of any photon is equal to 10 MeV, the endpoint energy. Most photons have energies one tenth to one third of the maximum electron energy for typical flash X-ray sources in the 1 MV to 10 MV endpoint voltage region, respectively.

3.6 *endpoint voltage*—Endpoint voltage refers to the peak voltage across a bremsstrahlung diode in a flash X-ray source. For example, a 10-MV flash X-ray source is designed to reach a peak voltage of 10 MV across the anode-cathode gap which generates the electron beam for striking a converter to produce bremsstrahlung.

3.7 *equilibrium absorbed dose*—absorbed dose within some incremental volume in the target material in which the condition of electron equilibrium (the energies, number, and direction of charged particles induced by the radiation are constant throughout the volume) exists. For lower electron energies, where bremsstrahlung production is negligible, the equilibrium absorbed dose is equal to the kerma.

NOTE 1—For practical purposes, assuming the spatial gradient in the X-ray field is small over the range of the maximum energy secondary electrons generated by the incident photons, the equilibrium absorbed dose is the absorbed dose value that exists in a material at a distance from any interface with another material greater than this range.

#### 4. Significance and Use

4.1 Flash X-ray facilities provide intense bremsstrahlung radiation environments, usually in a single sub-microsecond pulse, which often fluctuates in amplitude, shape, and spectrum from shot to shot. Therefore, appropriate dosimetry must be fielded on every exposure to characterize the environment, see ICRU Report 34. These intense bremsstrahlung sources have a variety of applications which include the following:

- (1) Studies of the effects of X-rays and gamma rays on materials.
- (2) Studies of the effects of radiation on electronic devices such as transistors, diodes, and capacitors.
- (3) Computer code validation studies.

4.2 This guide is written to assist the experimenter in selecting the needed dosimetry systems for use at pulsed X-ray facilities. This guide also provides a brief summary on how to use each of the dosimetry systems. Other guides (see Section 2) provide more detailed information on selected dosimetry systems in radiation environments and should be consulted after an initial decision is made on the appropriate dosimetry system to use. There are many key parameters which describe a flash X-ray source, such as dose, dose rate, spectrum, pulse width, etc., such that typically no single dosimetry system can measure all the parameters simultaneously. However, it is frequently the case that not all key parameters must be measured in a given experiment.

#### 5. General Characteristics of Flash X-ray Sources

5.1 *Flash X-ray Overview*—Flash X-ray sources operate like a dental X-ray source but at much higher voltages and intensities and usually in a single, very short burst, see ICRU Report 17. A high voltage is developed across an anode-cathode gap (the diode) and field emission creates a pulsed electron beam traveling from the cathode to the anode. A high-atomic-number element such as tantalum is placed on the anode to maximize the production of bremsstrahlung created when the electrons strike the anode. Graphite or aluminum is usually placed downstream of the converter to stop the electron beam completely but let the X-radiation pass through. Finally, a debris shield made of Kevlar or low-density polyethylene is sometimes necessary to stop exploding converter material from leaving the source. All of these components taken together form what is commonly called a bremsstrahlung diode.

5.2 *Relationship Between Flash X-ray Diode Voltage and X-ray Energy of Bremsstrahlung*—Flash X-ray sources produce bremsstrahlung by generating an intense electron beam which then strikes a high atomic number ( $Z$ ) converter such as tantalum. The electron-solid interactions produce “braking” radiation or, in German, bremsstrahlung. Fig. 1 shows the typical range of photon energies produced by flash X-ray sources with different electron endpoint energies. The data in Fig. 1 is generated by tallying the photon spectrum using ITS with optimized tantalum/carbon bremsstrahlung converters (1).<sup>5</sup> If the average radiation produced is in the 20 to 100 keV region, the source is said to be a “medium X-ray simulator.” If the average photon energy is in the 100 to 300 keV region, the term used is “hard X-ray simulator.” At the high end of the flash X-ray range are sources which produce an average photon energy of around 2 MeV. Because this photon energy is in the typical gamma-ray spectral range, the source is called a “gamma-ray simulator.”

5.2.1 The average energy of the bremsstrahlung spectrum,  $\bar{E}_{\text{photon}}$ , through an optimized converter can be estimated using the following relationship (1):

$$\bar{E}_{\text{photon}} = k \cdot \sqrt{\epsilon} \quad \text{where } 5.1 < k < 18.9 \quad (1)$$

where  $\bar{E}_{\text{photon}}$  is the average energy of the bremsstrahlung photons in keV and  $\epsilon$  is the average energy of the electrons in

<sup>5</sup> The boldface numbers in parentheses refer to the list of references at the end of this standard.




FIG. 1 Range of Available Bremsstrahlung Spectra from Flash X-ray Sources with Optimized Bremsstrahlung Converters

the electron beam incident on the converter in keV. The value of  $k$  depends on the converter thickness; thin targets will have values at the lower end of the range, while thick targets optimized for higher incident energies will have values at the upper end. When an optimized bremsstrahlung converter is used, a rule-of-thumb may be used that the average photon energy is about  $\frac{1}{3}$  or  $\frac{1}{6}$  of the electron endpoint energy (1). For a fixed converter design, the photon energy away from the optimization point is roughly proportional to the square root of the electron endpoint energy with the proportionality factor varying between about 5 and 19 depending upon the design point (1). This equation and Fig. 1 indicate that most of the photons have energies much less than the endpoint electron energy, or in voltage units, the flash X-ray voltage. Additionally, the bremsstrahlung spectrum is very non-Gaussian so caution must be exercised in using the average energy of the distribution for dosimetry planning.

## 6. Measurement Principles

6.1 Typically in flash X-ray irradiations, one is interested in some physical change in a critical region of a device under test (DUT). The dosimetry associated with the study of such a physical change may be broken into three parts:

- (1) Determine the absorbed dose in a dosimeter.
- (2) Using the dosimeter measurement, estimate the absorbed dose in the region and material of interest in the DUT.
- (3) If required, relate the estimated absorbed dose in the DUT to the physical change of interest (holes trapped, interface states generated, photocurrent produced, etc.).

6.2 This section will be concerned with the first two of the above listed parts of dosimetry: (1) what is necessary to determine a meaningful absorbed dose for the dosimeter, and (2) what is necessary to extrapolate this measured dose to the estimated dose in the region of interest. The final step in dosimetry, associating the absorbed dose with a physical change of interest, is outside the scope of this guide.

### 6.3 Energy Deposition:

6.3.1 *Secondary Electrons*—Both in the case of absorbed dose in the DUT and absorbed dose in the dosimeter, the energy is deposited largely by secondary electrons. That is, the incident photons interact with the material of, or surrounding, the DUT or the dosimeter and lose energy to Compton electrons, photoelectrons, and Auger electrons. The energy which is finally deposited in the material is deposited by these secondary particles.

6.3.2 *Transport of Photons*—In some cases, it is necessary to consider the transport and loss of photons as they move to the region whose absorbed dose is being determined. A correction for the attenuation of an incident photon beam is an example of such a consideration.

6.3.3 *Transport of Electrons*—Electron transport may cause energy originally imparted to electrons in one region to be carried to a second region depending on the range of the electrons. As a result, it is necessary to consider the transport and loss of electrons as they move into and out of the regions whose absorbed dose is being determined. In particular, it is necessary to distinguish between equilibrium and non-equilibrium conditions for electron transport.

6.3.3.1 *Charged Particle Equilibrium*—Occurs when the numbers, energies, and angles of particles transported into a region of interest are approximately balanced by those transported out of that region. (See “Equilibrium Absorbed Dose” in 3.7.)

6.3.3.2 *Dose Enhancement*—Because photoelectron production per atom is roughly proportional to the atomic number raised to the fourth power for energies less than 100 keV (2), one expects more photoelectrons to be produced in high atomic number layers than in low atomic number layers for the same photon fluence and spectrum. Thus, there may be a net flow of energetic electrons from the high atomic number layers into the low atomic number layers. This non-equilibrium flow of electrons may result in an enhancement of the dose in the low atomic number layer. Dose enhancement problems are often



caused by high atomic number bonding layers (for example, gold) and metallization layers (for example, W-Si or Ta-Si).

#### 6.4 Absorbed Dose in Dosimeter:

##### 6.4.1 Equilibrium Absorbed Dose in Dosimeter:

6.4.1.1 It is frequently possible to use dosimeters under approximate equilibrium conditions. The interpretation of the output of the dosimeter is straightforward only when the energy deposition processes within the dosimeter are approximately in equilibrium. That is, when the absorbed dose within the dosimeter is an equilibrium absorbed dose.

6.4.1.2 It is possible to treat non-equilibrium energy deposition within a dosimeter, but such an analysis requires electron and photon transport calculations, often in the form of computer codes.

##### 6.4.2 Limiting Cases:

6.4.2.1 There are two limiting cases for which the dosimeter data can be analyzed in a straightforward manner.

##### 6.4.2.2 Limiting Case One: Short Electron Range:

(1) For this case, secondary electron ranges are small in comparison with the size of the dosimeter.

(2) Essentially all electrons which deposit energy within the dosimeter will be produced within the dosimeter.

(3) Non-equilibrium effects due to electron transport are negligible, but photon attenuation corrections may be necessary.

(4) An example of this limiting case is: 20 keV photons depositing energy in a typical (0.889 mm thick) thermoluminescence (TL) dosimeter (TLD). In this case, the secondary electrons have ranges which are small in comparison with the size of the TLD. As a result, it is not necessary to perform a correction for the effect of electron transport on absorbed dose. On the other hand, 20 keV photons may be significantly attenuated while traveling through a TLD depending on the material. Thus a correction due to this effect may be necessary.

##### 6.4.2.3 Limiting Case Two: Large Electron Range:

(1) When the maximum secondary electron range is large compared with the size of the dosimeter, the dosimeter must be surrounded by an equilibrating layer. This layer must be chosen to be of an appropriate thickness, density, and atomic number. Generally, the range of secondary electrons must be smaller than the thickness of the equilibrating layer. For further discussion of equilibrating layers, see Practice E668.

(2) Essentially all electrons which deposit energy within the dosimeter originate in the equilibrating layer.

(3) Bragg-Gray cavity theory applies. That is, the dose within the dosimeter is the equilibrium dose for the equilibrating layer (corrected by the differences in electron stopping power of the dosimeter and the equilibrating layer and any photon attenuation through the material).

(4) An example of this limiting case is 1 MeV photons incident on a typical TLD surrounded by an appropriate equilibrating layer. In this case, the range of the secondary electrons will be large in comparison to the size of the TLD. Thus the dose measured will be the equilibrium dose in the TLD (with a small correction for the differences in the stopping power for the electrons in the TLD material and the material of the equilibrating layer).

##### 6.4.2.4 Intermediate Cases:

(1) The intermediate cases, where secondary electron ranges are neither small nor large in comparison to the dosimeter size, are cases where non-equilibrium energy deposition is to be expected.

(2) An example of an intermediate case is: 100 keV photons incident on a typical TLD.

(3) The careful treatment of dosimetry for intermediate cases requires the use of combined photon/electron radiation transport calculations of the energy deposition in the dosimeters to get satisfactory measurements.

#### 6.5 Absorbed Dose in Device Under Test:

6.5.1 Absorbed Dose in Device Under Test—The conditions within a DUT during a flash X-ray irradiation are frequently far from equilibrium. In many cases, these experiments will observe dose enhancement effects. As a result, it is frequently necessary to perform a dose-enhancement correction to estimate the absorbed dose within the region of interest. Unless the DUT meets the two limiting cases discussed in the previous section, which is rarely true, the dose should be obtained using combined photon/electron radiation transport calculations of the energy deposition in order to get satisfactory dose estimates. Within a system being tested, a calculation of the radiation transport through the system to the individual devices of interest must be performed.

#### 6.6 Spectral Considerations:

6.6.1 Broad Energy Range—The set of available flash X-ray sources spans a very wide photon energy range. Useful intensities may be obtained for energies as low as 10 keV and as high as 10 MeV. Each individual flash X-ray source, of course, does not produce useful photons over such a wide range. Dosimetry for such flash X-ray sources can be simplified if the different flash spectra are categorized into three types: low energy, medium energy, and high energy. This categorization refers to the average energy of the photon spectrum and not the higher electron endpoint energy (5.2).

6.6.2 Dosimetry for Three Energy Bands—A summary of the dosimetry requirements for the three flash X-ray energy bands is provided in Table 1.

#### 6.7 Absorbed Dose Interpretation:

6.7.1 In 6.1, it is pointed out that the second of three steps in the dosimetry process is the use of the measured absorbed dose in the dosimeter to estimate the absorbed dose in a region of interest within the device under test.

6.7.2 In the previous sections the dose in both the dosimeter (6.4) and the DUT (6.5) is estimated. Although the dose can easily be calculated from tables or a simple photon transport code for the dosimeter, which is often in equilibrium, the dose in the DUT, which is rarely in equilibrium, often requires a sophisticated combined photon/electron radiation transport code. Assuming this has been accomplished, one can use the ratio of the calculated doses at a given fluence and spectrum as the correction factor for the measured dose in the dosimeter to estimate the actual dose in the DUT.

$$D_{\text{DUT}} = \text{Calculated} \left( \frac{D_{\text{DUT}}}{D_{\text{Dosimeter}}} \right) \times D_{\text{Dosimeter}} \quad (2)$$

where  $D_{\text{DUT}}$  is the absorbed dose in the device material, and  $D_{\text{Dosimeter}}$  is the equilibrium absorbed dose in the dosimeter.

TABLE 1 Flash X-ray Dosimetry Characteristics for Three Energy Bands

Flash X-Ray Type	Absorbed Dose in Dosimeter	Absorbed Dose in DUT
Low energy (average photon energy: 20–100 keV)	Can get electron equilibrium. <sup>A</sup> May need photon transport correction. <sup>B</sup>	Depends on DUT: May need an electron transport calculation. <sup>C</sup> May need a photon transport calculation. <sup>B</sup>
Medium energy (average photon energy: 100–300 keV)	Often cannot get electron equilibrium and therefore need electron transport calculation. <sup>C,D</sup>	Depends on DUT: May need an electron transport calculation, especially without use of beam filtration. <sup>C,E</sup> May need a photon transport calculation. <sup>B</sup>
High energy (average photon energy: 300–3000 keV)	Can get electron equilibrium with proper equilibration layer. <sup>F,D</sup>	Depends on DUT: May need an electron transport calculation. <sup>C</sup> Usually no photon transport calculation needed. <sup>D</sup>

<sup>A</sup> The dosimeter or region of interest is large compared to the maximum secondary electron range (6.4.2.2).  
<sup>B</sup> The dosimeter or region of interest is large compared to the photon range (6.4.2.2, Paragraph 4).  
<sup>C</sup> The dosimeter or region of interest is of comparable size to the maximum secondary electron range (6.4.2.4).  
<sup>D</sup> The dosimeter or region of interest is small compared to the photon range (6.4.2.3, Paragraph 4).  
<sup>E</sup> A filter may be used to essentially eliminate the lower energy portions of the flash X-ray spectrum. This makes the spectrum more nearly monochromatic and may simplify dosimetry.  
<sup>F</sup> The dosimeter or region of interest is small compared to the electron range (6.4.2.3).

## 7. Dosimetry Systems

7.1 *Introduction*—In this section a brief summary of eight different types of dosimetry systems is given. The intent of this guide is to provide enough information so that users can decide which system might be appropriate for their application. Users are expected to investigate in more detail the proper use and limitations of a system using references in this and other ASTM documents before undertaking any radiation measurements. The pros and cons of each system are quickly summarized in Tables 2 and 3. The values cited in these tables are indicative of what is typically obtained under normal testing conditions. Many of these parameters can be improved at the expenditure of considerable effort. Only the precision of each dosimeter is discussed here because the accuracy of the dosimeter will be determined by the quality of the calibration and specific usage and this is beyond the scope of this document.

### 7.2 Thermoluminescence Dosimeters:

7.2.1 *Introduction*—Thermoluminescence dosimeters (TLDs), see Terminology E170, are popular dosimeters because they are small, passive, inexpensive, and can retain accurate dose information for long periods of time between irradiation and readout. The dose range of TLDs is typically  $>10^{-4}$  to  $<5 \times 10^3$  Gy ( $>10^{-2}$  to  $<5 \times 10^5$  rad).

TABLE 2 Dose Measurements

Dosimeter Type	TLD	Calorimeter	Optichromic	Radiochromic Film
Dose Range (Gy)	$10^{-4}$ – $5 \times 10^3$	$10$ – $10^5$	$0.4$ – $2 \times 10^4$	$0.01$ – $5 \times 10^4$
Photon Energy Range (MeV)	$>0.01$	$>0.01$	$>0.01$	$>0.01$
Size	Small	Medium	Medium	Small to Large
Ease of calibration	Easy	Easy	Easy	Moderate
Precision	5–15 %	5 %	5 %	5 %
Active Material	CaF <sub>2</sub> , LiF	Au, Si, Al	Organic	Organic
Availability	Buy	Make or Buy	Buy	Buy
System Cost	Medium	Low	Low	Low to High

TABLE 3 Dose Rate Measurements

Dosimeter Type	PIN Diode	Compton Diode	Cerenkov/Scintillator-Photodetector	PCD
Dose Rate Range (Gy/s)	$2 \times 10^{-1}$ – $10^8$	$10^7$ – $10^{10}$	$10^4$ – $10^9$	$2 \times 10^5$ – $10^{10}$
Photon Energy Range (MeV)	$>0.01$	$>0.5$	$>0.02$	$>0.01$
Size	Small to Medium	Large	Very Large	Small to Medium
Ease of calibration	Moderate	Difficult	Moderate	Moderate
Precision	10 %	10 %	4 %	10 %
Active Material	Si, GaAs	Tungsten	Plastic	Diamond, GaAs
Availability	Buy	Buy or Make	Make	Buy or Make
System cost	Low	Medium	High	Medium

7.2.1.1 Many materials are available for thermoluminescence dosimetry; however, the favorites for radiation hardness testing are lithium fluoride (LiF), manganese activated calcium fluoride (CaF<sub>2</sub>:Mn), and dysprosium activated calcium fluoride (CaF<sub>2</sub>:Dy). Dosimeters are available as powder, chips made from polycrystalline material, and discs consisting of very fine powder uniformly dispersed throughout a polytetrafluoroethylene (PTFE) matrix. A commonly used size of chip is 3.2 by 3.2 by 0.9 mm (0.125 by 0.125 by 0.035 in.). A commonly used size of PTFE dosimeter is a disc, 6 mm in diameter and 0.4 mm thick. In addition, these dosimeter materials can be made into arrays for dose mapping.

7.2.2 *Principles*—Thermoluminescent materials consist of a crystalline insulator with added dopants which introduce stable electron traps into the forbidden band gap. Ionizing radiation creates electrons and holes which are trapped by stable traps in the band gap. The density of filled traps is proportional to the dose absorbed by the material. Subsequent heating of the material empties the electron traps, allowing electrons from F-centers to recombine with free holes at luminescence centers, emitting light. The integrated light output is proportional to the density of filled traps, and therefore to the absorbed dose in the TLD material.

7.2.2.1 To make an equilibrium dose measurement, the TLD must be enclosed in an equilibrium capsule of the appropriate material and thickness when it is exposed. Methods for determining equilibrium capsule thickness are given in Practice E666.

7.2.2.2 Readout of TLDs is accomplished by an instrument consisting of a heater, optical system, and photomultiplier detector to measure the light emitted by the TLD during a predetermined heating cycle, and an integrating picoammeter which can measure the current or the charge from the photomultiplier. The total integrated charge from the photomultiplier during part or all of the heating cycle is usually related to the absorbed dose in the dosimeter. TLD readout instruments are available from a number of manufacturers.

7.2.3 Applications—The choice of the TLD and capsule material depend on the application. For high-energy photon sources (photon energy  $\geq 200$  keV) such as high voltage flash X-ray generators and  $^{60}\text{Co}$  and  $^{137}\text{Cs}$  sources, and where the quantity of interest is Si or  $\text{SiO}_2$  equilibrium dose, a good combination of materials is a  $\text{CaF}_2:\text{Mn}$  TLD in an Al equilibrium shield (see Practice E1249). For these materials and photon energies, all mass energy absorption coefficients and mass stopping powers are so close to those of silicon that the silicon equilibrium dose can be determined using Bragg-Gray cavity theory. The response of  $\text{CaF}_2:\text{Mn}$  is nearly linear with absorbed dose and the dosimeters retain dose information for a long time with a small fading correction whose form is well known (7.2.5).

7.2.3.1 For radiation sources which have a significant portion of photons with energy below 200 keV, more detailed calculations are necessary regardless of which material is chosen for the TLD. In this case the choice of TLD and capsule material is flexible, and can be made on the basis of convenience or cost.

7.2.3.2 The choice of whether powder, chips, or PTFE discs are used is primarily one of convenience and cost. For specific instructions on the proper procedures for using TLD systems to determine absorbed dose in radiation hardness testing of electronic devices, see Practices E666 and E668.

7.2.4 Advantages—TLDs are small, inexpensive, and require no instrumentation during irradiation. They are thus ideally suited for measuring dose at many locations within a test object, or for measuring dose near the areas of interest in even very small ( $\sim 3$  mm) test objects. Their large sensitivity range makes it possible to measure mGy to kGy doses with one dosimetry system.

7.2.5 Limitations—Energy Range: TLDs are usable over the photon energy range from at least 10 keV to 10 MeV. (Indeed, their sensitivity extends down into the ultraviolet region; hence, TLDs must be protected from exposure to sunlight, etc., especially for low dose applications.) The dose response as a function of energy can be calculated for TLDs which consist of 100 % TL phosphor by using standard equilibrium dose and Bragg-Gray cavity theories (Practice E666). However, care must be exercised when using PTFE disc dosimeters at low energies. Since these dosimeters consist of fine grains of TL phosphor distributed throughout a PTFE matrix, these simple theories may not apply. At high energies (photon energy  $\geq 2$

MeV), the TLD can be treated using Bragg-Gray cavity theory, with all photon interactions assumed to take place in the equilibrium shield material, and the resultant secondary electrons depositing energy in the phosphor grains. There is a smooth transition to more moderate energies (photon energy  $\approx 200$  keV) where the photon interactions must be considered to occur primarily in the PTFE matrix. Finally, at the lowest photon energies (photon energy  $\leq 10$  to 20 keV), all photon interactions which eventually deposit energy in the phosphor grains must be assumed to originate in the individual phosphor grain itself. The dose response of PTFE TLDs as a function of photon energy is therefore a complex function of energy which is not easy to derive analytically. A modern electron/photon transport code such as ITS (3) can be helpful in calculating this relationship (4).

7.2.5.1 Most TL phosphors exhibit fading with time; that is, the observed TL response becomes progressively smaller as the time interval between irradiation and readout increases (5). For most of the common TL phosphors the fading behavior is well known. The fading of  $\text{CaF}_2:\text{Mn}$  TLDs has been reported to obey the relationship (6):

$$(TL)/TL_o = (t/t_o)^{-0.017} \quad (3)$$

where  $t$  is in hours and the subscript  $o$  refers to the time of irradiation, and unsubscripted quantities refer to the time of readout. The value of the coefficient 0.017 is only typical and can vary from batch to batch and must be determined experimentally. Use the procedures in 8.7 of Practice E668 to test for fading effects in the type of TLD chosen.

7.2.6 Sensitivity—TLDs can be used to measure dose from 100  $\mu\text{Gy}$  to 5 kGy (10 mrad to 500 krad). Most TL materials saturate in the range of 3 to 10 kGy (300 krad to 1 Mrad). The dose response of calcium fluoride TLDs is independent of dose rate for dose rates less than  $10^{10}$  Gy/s ( $10^{12}$  rad/s) (7). The dose response of lithium fluoride has been reported to be independent of dose rate only up to  $10^9$  Gy/s ( $10^{11}$  rad/s) (8).

7.2.7 Calibration—TLD systems (dosimeter plus reader) must be calibrated in a standard radiation field before use; see ICRU Report 14. The most convenient sources for this are calibrated  $^{60}\text{Co}$  or  $^{137}\text{Cs}$  sources. Care must be taken to expose the TLDs in an appropriate equilibrium shield. If the calibration source is calibrated in units of exposure rate, then the dose absorbed by the TLD can be found by:

$$D_{\text{TLD}} = \left( \frac{(\mu_{\text{en}}/\rho)_{\text{TLD}}}{(\mu_{\text{en}}/\rho)_{\text{air}}} \right) (D_{\text{air}}/X) X_0 \exp[-(\mu_{\text{en}}/\rho)_{\text{eqmat}} \rho_{\text{eqmat}} x] \quad (4)$$

where:

- $D_{\text{TLD}}$  = is the dose absorbed by the TLD,
- $X_o$  = is the free-in-air exposure of the source,
- $D_{\text{air}}/X$  = 33.68 Gy·kg/C (0.869 rad/R),
- $\mu_{\text{en}}/\rho$  = is the mass-energy absorption coefficient,
- $\rho_{\text{eqmat}}$  = is the density of the equilibrator material, and
- $x$  = is the thickness of the equilibrator material.

7.2.8 Reproducibility—The reproducibility of most TLD materials is about  $\pm 8\%$  at  $1\sigma$ . Better reproducibility can be achieved with TLD chips by irradiating a number of them to a low dose, reading them, and selecting those which fall into a tighter group. A reproducibility of  $\pm 2\%$  is often achievable, and  $\pm 1\%$  is possible by this means.