

**Medical electrical equipment –
Part 2-10: Particular requirements for the
safety of nerve and muscle stimulators**

Amendment 1

CORRIGENDUM 1

Page 7

42 Excessive temperatures

Replace the existing Clause 42 as follows:

42 Excessive temperatures

This clause of the General Standard applies except as follows:

Amend the compliance test for the requirements of subclauses 42.1 to 42.3 as follows:

5) *Test criteria*

Additional third paragraph:

Compliance with the requirements for maximum temperatures specified in the General Standard shall be checked under the conditions specified in Item aa) of subclause 7.1.