

INTERNATIONAL STANDARD



5943

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Cheese and processed cheese products – Determination of chloride content – Potentiometric titration method

Fromages et fromages fondus – Détermination de la teneur en chlorures – Méthode par titrage potentiométrique

iTeh STANDARD PREVIEW

First edition – 1978-10-15

(standards.iteh.ai)

ISO 5943:1978

<https://standards.iteh.ai/catalog/standards/sist/763216a2-39f8-4958-8807-476c9fad079/iso-5943-1978>

UDC 637.3 : 546.131 : 543.849

Ref. No. ISO 5943-1978 (E)

Descriptors : dairy products, cheeses, chemical analysis, determination of content, chlorides.

Price based on 2 pages

FOREWORD

ISO (the International Organization for Standardization) is a worldwide federation of national standards institutes (ISO member bodies). The work of developing International Standards is carried out through ISO technical committees. Every member body interested in a subject for which a technical committee has been set up has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council.

International Standard ISO 5943 was developed by Technical Committee ISO/TC 34, *Agricultural food products*, and was circulated to the member bodies in June 1977.

(standards.iteh.ai)

It has been approved by the member bodies of the following countries :

ISO 5943:1978		
Australia	Hungary	Portugal
Austria	India	Romania
Belgium	Iran	South Africa, Rep. of
Bulgaria	Israel	Spain
Canada	Italy	Turkey
Czechoslovakia	Korea, Rep. of	United Kingdom
Egypt, Arab Rep. of	Mexico	U.S.A.
France	Netherlands	U.S.S.R.
Germany, F.R.	New Zealand	Venezuela
Ghana	Peru	Yugoslavia

No member body expressed disapproval of the document.

NOTE — The method specified in this International Standard has been developed jointly with the IDF (International Dairy Federation) and the AOAC (Association of Official Analytical Chemists, U.S.A.). The text as approved by the above organizations will also be published by FAO/WHO (Code of Principles concerning Milk and Milk Products and Associated Standards), by the IDF and by the AOAC (Official Methods of Analysis).

Cheese and processed cheese products – Determination of chloride content – Potentiometric titration method

0 INTRODUCTION

The method described in this International Standard was originally intended for routine application since it is very fast and may be easily automated. However, there is some evidence that the precision of the method (repeatability and reproducibility) is comparable to, and may even be better than, that of the method described in ISO 2970, *Cheese – Determination of chloride content (Reference method)*. Studies are therefore being undertaken with a view to deciding whether the method described in the present document will replace that described in ISO 2970 as the reference method.

1 SCOPE AND FIELD OF APPLICATION

This International Standard specifies a potentiometric titration method for the determination of the chloride content of cheese and processed cheese products.

The method is applicable to all cheeses and processed cheese products containing more than 0,2% (*m/m*) of chloride ion.

2 REFERENCE

ISO/R 707, *Milk and milk products – Sampling*.

3 DEFINITION

chloride content of cheese and processed cheese products :
The substances determined by the procedure specified below, expressed as a percentage by mass of chloride ion or sodium chloride or any other chloride.

4 PRINCIPLE

Suspension of a test portion in water. Acidification with nitric acid and subsequent potentiometric titration of chloride ion with standard volumetric silver nitrate solution.

5 REAGENTS

All reagents shall be of analytical grade. Water used shall be distilled water or water of at least equivalent purity.

5.1 Silver nitrate, 0,08 to 0,12 N standard volumetric solution.

Dissolve 13,6 to 20,4 g of silver nitrate (AgNO_3) in water which is practically free from carbon dioxide and dilute to 1 000 ml. Standardize the solution against sodium chloride (NaCl), which has previously been dried at 300 °C, expressing the normality of the silver nitrate solution to four decimal places.

Store the solution away from direct sunlight.

5.2 Nitric acid, approximately 4 N solution.

ISO 5943:1976 APPARATUS

<https://standards.iteh.ai/catalog/standards/sist/763216a2-39f8-4958-8807->

6.1 Device for grinding or grating cheese, capable of being easily cleaned.

6.2 Analytical balance.

6.3 Blender.

6.4 Potentiometer provided with a measuring electrode suitable for the determination of chloride (for example a silver electrode) and a reference electrode [for example a mercury(II) sulphate electrode].

6.5 Vessel suitable for blending and titrating.

6.6 Graduated cylinders, capacities 50 and 10 ml.

6.7 Burette, graduated in 0,1 ml, capacity 50 ml, conforming to ISO/R 385, or an **automatic plunger burette**, readable to the nearest 0,01 ml. The burette, or automatic plunger burette, should preferably be made of brown glass.

6.8 Stirrer.

7 SAMPLING

See ISO/R 707.

Sample cheese in brine by taking fragments of at least 200 g. Before analysis, dry the sample with filter paper.

8 PROCEDURE

8.1 Preparation of the test sample

Prior to analysis, remove the rind or smear or mouldy surface layer of the cheese, in such a way as to provide a sample representative of the cheese as it is usually consumed. Grind or grate the sample by means of an appropriate device (6.1); mix the ground or grated mass quickly, and if possible grind or grate a second time and again mix thoroughly. If the sample cannot be ground or grated, mix it thoroughly by intensive stirring and kneading.

Transfer the test sample to an air-tight container to await analysis, which should be carried out as soon as possible after grinding. If delay is unavoidable, take all precautions to ensure proper preservation of the sample and to prevent condensation of moisture on the inside surface of the container. The storage temperature should be 10 to 12 °C.

Clean the device after grinding or grating each sample.

8.2 Test portion

Weigh, to the nearest 0,001 g, 2 to 5 g of the test sample (8.1) into the vessel (6.5).

8.3 Determination

8.3.1 Add 30 ml of water at about 55 °C. Suspend the test portion using the blender. Rinse the blender with approximately 10 ml of water, collecting the rinsings in the vessel.

8.3.2 Add 2 to 3 ml of the nitric acid solution (5.2). Put the measuring electrode and the reference electrode into the suspension.

Titrate the contents of the vessel with the standard volumetric silver nitrate solution (5.1) from the burette (6.7), stirring continuously, until the end point has nearly been reached. Then titrate cautiously until reaching the end point, which corresponds to the maximum potential difference observed between two successive equal additions (of about 0,05 ml) of the silver nitrate solution.

8.4 Blank test

Carry out a blank test using the reagents but omitting the test portion.

9 EXPRESSION OF RESULTS

9.1 Method of calculation and formula

Calculate the chloride content, as a percentage by mass, by means of the formula

$$\frac{(V_1 - V_0) \times T \times f}{m}$$

where

V_0 is the volume, in millilitres, of the silver nitrate solution used in the blank test;

V_1 is the volume, in millilitres, of the silver nitrate solution used in the determination;

T is the normality of the silver nitrate solution;

m is the mass, in grams, of the test portion;

f is the factor for expressing the result as a percentage of any chloride.

The numerical values are, for example :

$f = 3,55$ for expression as % Cl^-

$f = 5,84$ for expression as % NaCl

$f = 7,46$ for expression as % KCl.

Report the result to the second decimal place.

9.2 Repeatability

The difference between the results of two determinations carried out in rapid succession by the same analyst shall not exceed 0,04 g of Cl^- ion (or the equivalent quantity of any chloride) per 100 g of the cheese or of the processed cheese product.

10 TEST REPORT

The test report shall show the method used and the result obtained. It shall also mention all operating conditions not specified in this International Standard, or regarded as optional, as well as any circumstances that may have influenced the result.

The report shall include all details required for the complete identification of the sample.