INTERNATIONAL STANDARD

ISO 5961

Second edition 1994-01-15

Water quality — Determination of cadmium by atomic absorption spectrometry

iTeh STANDARD PREVIEW

Qualité de l'eau S Dosage du cadmium par spectrométrie d'absorption àtomique

ISO 5961:1994 https://standards.iteh.ai/catalog/standards/sist/ad1b5012-1cd5-4949-bd23-fcb9aa45233d/iso-5961-1994



ISO 5961:1994(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 5961 was prepared by Technical Committee ISO/TC 147, Water quality, Subcommittee SC 2, Physical, chemical, biochemical methods.

https://standards.iteh.ai/catalog/standards/sist/ad1b5012-1cd5-4949-bd23-

This second edition cancels and replaces 45the 1/is first 61-edition (ISO 5961:1985), of which it constitutes a technical revision.

Annex A of this International Standard is for information only.

© ISO 1994

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Water quality — Determination of cadmium by atomic absorption spectrometry

Section 1: General

iTeh STANDARD PREVIEW the concentration range is 0,3 μg/l to 3 μg/l. The range

1.1 Scope

This International Standard specifies two methods for the determination of cadmium: flame atomic absorp by the use of smaller dosing volumes. Cadmium can tion spectrometry (AAS) (Section 2) and be determined in sludges and sediments after an apelectrothermal atomization (AAS) (Section 3) a45233d/iso-59 propriate digestion procedure.

1.1.1 Determination of cadmium using AAS in an air-acetylene flame

The method is applicable to the analysis of water and waste water when the concentration of cadmium is between 0,05 mg/l and 1 mg/l. Higher concentrations can be determined after dilution of the sample. The range of application of the method can be extended to lower concentrations by carefully evaporating the water sample, previously acidified with nitric acid. Cadmium can be determined in sludges and sediments after an appropriate digestion procedure avoiding the formation of a precipitate.

1.1.2 Determination of cadmium by electrothermal atomization AAS

The method is suitable for the determination of cadmium in water when, with a dosing volume of 10 μ l,

1.2 Normative reference

The following standard contains provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the edition indicated was valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent edition of the standard indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 5667-3:—1), Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples.

¹⁾ To be published. (Revision of ISO 5667-3:1985)

ISO 5961:1994(E) © ISO

Section 2: Determination of cadmium using atomic absorption spectrometry in an air-acetylene flame

2.1 Interferences

The following ions will not interfere with the method as long as the mass concentrations specified below are not exceeded:

The total salt content of the measuring solution shall be less than 15 g/l and the electrical conductivity shall be lower than 20 000 mS/m. Samples of unpredictable matrix effects shall be examined appropriately. This influence shall be compensated for either by diluting the sample or by applying the method of standard additions (see 3.6.2.2).

3 000pm/g/lindards.iteh.ai/catalog/standar

fcb9aa45233d/

2.2 Principle

Titanium

Aspiration of the acidified sample into the airacetylene flame of an atomic absorption spectrometer. Measurement of the cadmium concentration at a wavelength of 228,8 nm.

2.3 Reagents

Use only reagents of recognized analytical grade and only distilled water or water of equivalent purity. The cadmium content of water used for blank determinations and for the preparation of standard solutions shall be negligibly low compared with the lowest mass concentration to be determined in the sample.

2.3.1 Nitric acid, $\rho = 1,40 \text{ g/ml}.$

2.3.2 Hydrogen peroxide, $w(H_2O_2) = 30 \% (m/m)$.

2.3.3 Cadmium stock solution I.

 $\rho(Cd) = 1 000 \text{ mg/l}.$

Dissolve 1,000 g \pm 0,002 g of cadmium in 10 ml of nitric acid (2.3.1) and 10 ml of water (see 2.3) in a 1 000 ml one-mark volumetric flask. Dilute to volume with water.

Store the solution in polyethylene or borosilicate glass containers. The solution is stable for 1 year.

Alternatively, use any commercially available stock solution containing 1,000 g/l \pm 0,002 g/l of cadmium.

TANDARD PREVIEW

standard 2.3.4 (Cadmium standard solution I,

 $\rho(Cd) = 10 \text{ mg/l}.$

O 596 1:1994 to 10 ml of the cadmium stock solution (2.3.3) tandards strain 1000 ml one-mark volumetric flask, add 10 ml 233d/solution (2.3.1) and dilute to volume with water.

Store the solution in polyethylene or borosilicate glass containers. The solution is stable for at least one month if stored at room temperature.

NOTE 1 The use of a microlitre pipette permits a standard solution of 100 ml to be prepared.

2.3.5 Cadmium calibration solutions

Prepare a minimum of five calibration solutions in accordance with the expected cadmium concentrations.

As an example, proceed as follows for the range from 0,05 mg/l to 1,0 mg/l:

Pipette 0,5 ml; 2,0 ml; 4,0 ml; 6,0 ml; 8,0 ml and 10,0 ml respectively of the cadmium standard solution (2.3.4) into 100 ml one-mark volumetric flasks.

Add 1 ml of nitric acid (2.3.1) to each of these solutions. Dilute to volume with water and mix.

The calibration solutions contain 0,05 mg/l; 0,2 mg/l; 0,4 mg/l; 0,6 mg/l; 0,8 mg/l and 1,0 mg/l of cadmium respectively.

2.3.6 Blank test solution

Pipette 1 ml of nitric acid (2.3.1) into a 100 ml one-mark volumetric flask, and dilute to volume with water (2.3.4).

If the sample requires a pretreatment by digestion, the blank shall be given the same pretreatment (see 2.5.2).

2.3.7 Solution for zero-setting the instrument

Use water (2.3.4) as a zero-setting solution. The blank solution (2.3.6) may also serve for zero-setting provided its cadmium concentration is negligibly low.

2.4 Apparatus

Immediately before use, clean the glassware with warm, dilute nitric acid, approximately 2 mol/l (e.g. by soaking for 24 h), followed by a thorough rinsing with water (2.3). Verify that each lot of pipette tips and single-use plastics vessels is free from potential cadmium contamination by carrying out blank measurements (see 2.6.1).

Usual laboratory apparatus and

2.4.1 Atomic absorption spectrometer, equipped with background correction and a radiation source for the determination of cadmium, operated according to the manufacturer's instructions.

2.4.2 Gas supply for air and acetylene. It is essential that the manufacturer's safety instructions be observed. The residual gas pressure for acetylene cylinders shall be at least 5×10^5 Pa.

2.4.3 Air-acetylene burner.

- **2.4.4 One-mark volumetric flasks** of capacity 10 ml, 100 ml and 1 000 ml.
- **2.4.5 One-mark pipettes**, of nominal capacity 1 ml, 2 ml, 3 ml, 4 ml, 5 ml, 6 ml, 8 ml, 10 ml, 20 ml, 30 ml and 40 ml.
- 2.4.6 Microlitre pipettes or diluters.
- 2.4.7 Beakers, of capacity 250 ml.

2.4.8 Heating device, for example a hotplate.

2.4.9 Membrane filtration device with filters, of pore size $0.45 \mu m$, washed thoroughly with dilute nitric acid and rinsed with water.

2.5 Sampling and sample pretreatment

See ISO 5667-3.

2.5.1 Sampling

PREVIEW

Collect the samples in polyethylene or borosilicate glass containers which have previously been cleaned with nitric acid and water.

2.5.2 Pretreatment and preparation of the sample solutions

2.5.2.1 Pretreatment for the determination of the content of dissolved cadmium

tus and ISO 5961:199 Filter the water sample as soon as possible after https://standards.iteh.ai/catalog/standards/sistsalimpling- (2d5.4) 4through a membrane filter of pore fcb9aa45233d/iso-59(size) 0,45 µm.

To stabilize the filtrate, add, for example, 10 ml of nitric acid (2.3.1) per litre of water sample to achieve a pH of less than 2; add more acid, if necessary, to ensure a pH of less than 2.

2.5.2.2 Pretreatment for the determination of cadmium after mineralization

Acidify the water sample as soon as possible after sampling by adding 1 ml of nitric acid (2.3.1) per litre of sample; add more acid, if necessary, to ensure a pH of less than 2.

Completely mix the sample, for example by thorough shaking.

Place 100 ml of the homogenized sample in a 250 ml beaker. Add 1 ml of nitric acid (2.3.1) and 1 ml of hydrogen peroxide (2.3.2).

Heat the beaker on a hotplate until about 0,5 ml remains.

It is essential that the sample is not reduced to dryness.

ISO 5961:1994(E) © ISO

If there is heavy organic contamination of the water sample, consider a repeated addition of hydrogen peroxide (Caution).

Dissolve the residue in 1 ml of nitric acid (2.3.1) and some water, transfer quantitatively to a 100 ml one-mark volumetric flask and dilute to volume with water.

If sample volumes other than those specified are used, the apparatus and reagents shall be adjusted to the respective ratio of volumes.

It is not necessary to carry out a digestion in cases where cadmium is quantitatively determinable without this treatment. In such circumstances, only acidify the water sample.

2.6 Procedure

Prior to measurement, adjust the instrumental parameters of the spectrometer (2.4.1) in accordance with the manufacturer's operating instructions.

Optimize flame conditions.

a mid-range calibration solution to check whether the calibration curve is still valid.

If the cadmium content of the measuring solution exceeds the range of validity of the calibration curve, dilute the measuring solution accordingly with blank solution.

2.7 Evaluation

2.7.1 Calculation

Establish the calibration function by linear regression using the data obtained from the measurement of the calibration solutions.

Calculate the mass concentration $\rho(Cd)$ of cadmium, expressed in milligrams per litre, in the water sample using the equation

ing instructions. $\rho(Cd) = \frac{(A_1 - A_0) \times V_1}{\text{TTEH STANDARD PREV}^{t_1}}$

Zero the spectrometer by aspirating the zero-setting ard where ai)

solution (2.3.7) into the flame.

ISO 5961:199A₀ is the absorbance of the blank solution;

https://standards.iteh.ai/catalog/standards/sist/ad1b5012-1cd5-4949-bd23-

2.6.1 Blank test

 $\label{eq:cb9aa45233d/iso-59A1-1984} \text{the absorbance of the measuring solution;}$

Carry out a blank test with the blank solution (2.3.6) at the same time as the determination, using the same procedure and the same quantities of all the reagents as in the sampling and determination, but replacing the test portion with water (2.3).

2.6.2 Calibration

To establish the calibration curve, spray in the calibration solutions (2.3.5) successively, in ascending order of concentration, and record the absorbance.

Check the zero with the zero-setting solution (2.3.7) after the measurement of each calibration solution

2.6.3 Test portion measurement

Continue by measuring the blank solution (2.6.1) and the measuring solutions (2.5.2) and record the absorbances.

After each measuring series, but at least after 10 to 20 measurements, re-analyse the blank solution and

- b is the slope of the calibration curve, in litres per milligram, according to 2.6.2;
- V₁ is the volume, in millilitres, of the measuring solution;
- V₂ is the volume, in millilitres, of the water sample applied to prepare the measuring solution.

Dilution steps which differ from those specified above shall be taken into account during the calculation.

2.7.2 Expression of results

The values shall be rounded to the nearest 0,01 mg/l.

EXAMPLES

Cadmium (Cd): 0,07 mg/l

Cadmium (Cd): 0,41 mg/l

© ISO ISO ISO ISO 5961:1994(E)

2.8 Test report

The test report shall contain the following information:

- a) a reference to this International Standard;
- b) complete identification of the water sample;
- c) expression of the results, according to 2.7.2;

- d) details of sample pretreatment, if applicable;
- e) any deviations from this procedure and all the circumstances which may have affected the results.

2.9 Precision

See table 1.

Table 1 — Precision data

Sample No.	l	n	n _a %	x mg/l	\overline{x} mg/l	σ _r mg/l	VC, %	σ _R mg/l	VC _R %	WFR %	Suitable for
A1) 2) B1) 2) C1) 3)	23 23 23	63 60 63	0 5 0	0,113 0,563 0,968	0,118 0,587 1,008	0,009 7 0,030 9 0,073 3	8,2 5,3 7,3	0,005 8 0,009 2 0,012 1	5,0 1,6 1,2	104,1 104,3 104,1	
			%	μg/l	μg/l	μg/l	%	μg/l	%	%	
D4) 2)	52 22 17 13	133 55 45 31	2 2 0 11	0,91 0,91 0,91 0,91	1,292 1,148 1,363 1,39	0,319 6 0,270 4 0,356 8 0,171 8	24,7 23,5 26,2 12,4	0,083 0 0,048 3 0,115 9 0,064 3	6,4 4,2 8,5 4,6	141,9 126,2 149,8 152,7	all methods ⁵⁾ method 3.6.2.1 addition: manual addition: automatic
E4) 2)	52 22 17 13	136 56 40 35	o Toeh o	2,70 2,70 2,70 2,70 2,70	2,96 2,78 2,99 3,22	0,639 9 0,631 8 0,344 5 0,559 2	21,6 22,8 11,5 17,4	0,266 3 0,199 4 0,195 1 0,158 3	9,0 7,2 6,5 4,9	109,6 102,9 110,7 119,2	all methods ⁵⁾ method 3.6.2.1 addition: manual addition: automatic
F4) 6)	53 23 17 13	135 56 43 38 _{tto}	7 11 7 •//st8nda	16,2 16,2 16,2 ds ¹ 6,2	18,34 17,14 18,3 <u>3</u> S(3,348 7 2,232 8 3,96917/9 3,107/7 _{jet}	18,3 13,0 1 21,7	1,420 8 0,756 7 1,497 6	7,7 4,4 8,2	113,2 105,8 113,1 124,5	all methods ⁵⁾ method 3.6.2.1 addition: manual addition: automatic

- fcb9aa45233d/iso-5961-1994
- is the number of participating laboratories
- n is the number of values
- $n_{\rm a}$ is the percentage of outliers
- x is the true value
- \bar{x} is the total mean

- is the repeatability standard deviation
- VC_r is the repeatability variation coefficient
- σ_{R} is the reproducibility standard deviation
- VC_R is the reproducibility variation coefficient
- WFR is the recovery rate

- 1) AAS method in an air-acetylene flame.
- 2) Industrial waste water with a low level of pollution.
- 3) C: spiked drinking water.
- 4) Electrothermal atomization AAS method.
- 5) "All methods" means that both the manual and the automated addition have been included in the calculation.
- 6) F: spiked distilled water.

ISO 5961:1994(E) © ISO

Section 3: Determination of cadmium by electrothermal atomization atomic absorption spectrometry

3.1 Interferences

The following ions will not, if present individually, interfere with the method as long as their mass concentrations do not exceed 100 mg/l: iron, copper, nickel, cobalt and lead.

No interferences could be observed up to mass concentrations of 1 000 mg/l for sodium, potassium, calcium, magnesium, sulfate and chloride.

Combinations of the above-mentioned ions, even if present in much lower concentrations, can lead to depression or enhancement of the measured signals. Some other ions, if present in low concentrations, can cause interferences. The method of standard additions should therefore be given preference to the normal calibration. Samples of unpredictable matrix behaviour need to be analysed using the method of standard additions. Interferences caused by background absorption can be eliminated to a large extent by the application of a background compensating system.

3.3.5 Cadmium standard solution II,

 $\rho(Cd) = 100 \, \mu g/l.$

Pipette 10,0 ml of the cadmium stock solution II (3.3.4) into a 1 000 ml one-mark volumetric flask, add 10 ml of nitric acid (3.3.1) and dilute to volume with water.

Store the solution in a polyethylene or borosilicate glass container. Prepare the solution immediately before use.

3.3.6 Cadmium standard solution III,

 $\rho(Cd) = 5 \mu g/I.$

Pipette 50 ml of the cadmium standard solution II (3.3.5) into a 1 000 ml one-mark volumetric flask, add 10 ml of nitric acid (3.3.1) and dilute to volume with

water (2.3). EVE

fcb9aa45233d/i3:3:771- Cadmium calibration solutions

Store the solution in a polyethylene or borosilicate container. Prepare the solution immediately before use.

ISO 5961:1994

https://standards.iteh.ai/catalog/standards/sist/ad1b5012-1cd5-4949-bd23-

3.2 Principle

Injection of the acidified sample into an electrically heated graphite tube of an electrothermal atomization atomic absorption spectrometer. Measurement of the absorbance at a wavelength of 228,8 nm. If necessary, application of the standard additions technique.

3.3 Reagents

The requirements for the purity of the reagents are given in 2.3.

- **3.3.1** Nitric acid, $\rho = 1,40$ g/ml.
- **3.3.2** Hydrogen peroxide, $w(H_2O_2) = 30 \% (m/m)$.

3.3.3 Cadmium stock solution I,

 $\rho(Cd) = 1 000 \text{ mg/l}.$

See 2.3.3.

3.3.4 Cadmium stock solution II, $\rho(Cd) = 10 \text{ mg/l}.$

See 2.3.4.

Prepare at least five calibration solutions using the cadmium standard solution II (3.3.5) and matching the expected cadmium concentrations.

As an example, proceed as follows for the range from $0.3 \mu g/l$ to $3 \mu g/l$.

Pipette 0,3 ml; 1,0 ml; 1,7 ml; 2,4 ml; and 3,0 ml respectively of the cadmium standard solution II (3.3.5) into five 100 ml graduated flasks.

Add 1 ml of nitric acid (3.3.1) to each of these solutions. Dilute to volume with water and mix.

The calibration solutions contain 0,3 μ g/l; 1,0 μ g/l; 1,7 μ g/l; 2,4 μ g/l; and 3,0 μ g/l of cadmium, respectively.

Prepare the calibration solutions immediately before use.

3.3.8 Blank test solution

Pipette 1 ml of nitric acid (3.3.1) into a 100 ml one-mark volumetric flask and dilute to volume with water.

© ISO ISO 5961:1994(E)

If the sample requires a pretreatment by digestion, the blank shall be given the same pretreatment (see 2.5.2).

3.3.9 Zero-setting solution

Use water (2.3) as a zero-setting solution. The blank solution (3.3.8) may also serve for zero-setting, provided its cadmium concentration is negligibly low.

3.3.10 Matrix-matching solutions (see [2] in annex A)

Dissolve 1,0 g of palladium powder in 3 ml of nitric acid (2.3.1) and 20 ml of hydrochloric acid $(\rho = 1,19 \text{ g/ml})$ whilst gently warming, and dilute to 100 ml with water.

Dissolve 10 g of ammonium nitrate (NH₄NO₃) in water and dilute to 100 ml with water.

3.3.10.1 Solution for waters of high salinity

Mix equal volumes of the palladium solution and the ammonium nitrate solution (see 3.3.10). AND ARI

10 μl of this solution contain 50 μg of Rd and 500 μg s.i.g.h. Blank test of NH₄NO₃.

pollution fcb9aa45233d/iso-

Mix 15 ml of the palladium solution and 15 ml of the ammonium nitrate solution in a 100 ml measuring flask, and dilute to volume with water.

 $10~\mu l$ of this solution contain $15~\mu g$ of Pd and $150~\mu g$ of NH₄NO₃.

3.4 Apparatus

For cleaning of glassware see 2.4.

Usual laboratory apparatus and

- 3.4.1 Atomic absorption spectrometer, equipped with background correction and a radiation source for the determination of cadmium.
- 3.4.2 Graphite furnace, equipped with a control unit.

3.4.3 Pyrolytically coated graphite tube with a platform.

NOTE 2 If no interference is expected, normal tubes may be used.

3.4.4 Gas supply for argon.

3.4.5 Other apparatus (see 2.4).

NOTE 3 For optimum precision an automatic sample injector is recommended.

3.5 Sampling and sample pretreatment

Proceed as specified in 2.5.

3.6 Procedure

Prior to measurement, adjust the instrumental parameters of the spectrometer (3.4.1) in accordance with the manufacturer's operating instructions.

Zero the spectrometer using the zero-setting solution (3.3.9).

3.3.10.2 Solution for water with a low level of 5961:199 carry out a blank test at the same time as the determination, using the same procedure and the same quantities of all the reagents as in the sampling and determination, but replacing the test portion with wa-

3.6.2 Calibration and determination

Before each batch of determinations, prepare from the cadmium standard solution II (3.3.5) at least five calibration solutions covering the range of concentrations to be determined.

3.6.2.1 Direct measurement

Direct measurement shall only be used if the matrix influences dealt with in 3.1 can be ignored. Otherwise, apply the procedure of standard additions as described in 3.6.2.2.

Measure the absorbances or the integrated absorbances (peak areas) of the calibration solutions (3.3.7), the blank solutions (3.3.8) and the sample solutions (see 2.5.2) in accordance with the instructions given by the manufacturer of the instrument.