International Standard



INTERNATIONAL ORGANIZATION FOR STANDARDIZATION®MEXCHAPOCHAR OPPAHUSALUUR TO CTAHDAPTUSALUU®ORGANISATION INTERNATIONALE DE NORMALISATION

Molecular absorption spectrometry — Vocabulary — General — Apparatus

Spectrométrie d'absorption moléculaire - Vocabulaire - Généralités - Appareillage

First edition – 1982-08-15 Teh STANDARD PREVIEW (standards.iteh.ai)

<u>ISO 6286:1982</u> https://standards.iteh.ai/catalog/standards/sist/0f606774-025f-4a79-b4b8-88a995194ac1/iso-6286-1982

Descriptors : spectrophotometry, molecular absorption spectrophotometry, vocabulary, apparatus, generalities.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards institutes (ISO member bodies). The work of developing International Standards is carried out through ISO technical committees. Every member body interested in a subject for which a technical committee has been set up has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council.

iTeh STANDARD PREVIEW

International Standard ISO 6286 was developed by Technical Committee ISO/TC 47, *Chemistry*, and was circulated to the member bodies in July 1980. OS. Iten. al)

It has been approved by the member bodies of the following countries: 982

https://standards.iteh.ai/catalog/standards/sist/0f606774-025f-4a79-b4b8-Germany, F.R. 88a9951Pelaridiso-6286-1982

Australia		
Austria		
Belgium		
Brazil		
China		
Egypt, Arab	Rep.	of
France		

Germany, F.R. Hungary India Italy Korea, Rep. of Mexico Netherlands

Portugal Romania South Africa, Rep. of Switzerland USSR

No member body expressed disapproval of the document.

Molecular absorption spectrometry — Vocabulary — General — Apparatus

Scope and field of application 1

This International Standard gives definitions of a certain number of terms, and some general information, relating to molecular absorption spectrometry of solutions, together with general data concerning the instruments used, and, in particular, specifies :

a) the terminology to be used to characterize, by description, these instruments;

homogeneous, isotropic, non-luminescent and non-diffusing contained in an optical cell (with two plane and parallel surfaces). In table 2

terms 14 and 15 relate to the theoretical aspect, and are a simple adaptation of terms 9 and 11 to the special case of molecular absorption spectrometry;

terms 16 to 20 relate to actual phenomena and measure-____

IEW b) the characteristics and qualities of an instrument, by terms 21 and 22 relate to the method of expression of giving a summary of the principles of certain methods of Cresults. verifying them.

ISO 6286:198A list of the French terms equivalent to those defined in tables 1 2 Terms, definitions symbols formulae and ards/sistend 8/sistend ards/sistend 8/sistend 8/sisten units

88a995194ac1/iso-6286-1982

Molecular absorption spectrometry is a technique applicable to both qualitative and quantitative analyses and it enables measurements to be made of the concentration of a compound dissolved in a solution; it is effective in the near ultraviolet, visible and near infra-red regions, which correspond to a wavelength interval from about 180 to 1 000 nm.

The terms given in tables 1 and 2 are classified so that they are defined before their use in later definitions.

Table 1 is given for the purposes of comprehensiveness; it collates terms from the Vocabulaire international de l'éclairage (International lighting vocabulary), account of which has been taken in the choice of terms and in the drawing up of the definitions forming the subject of table 2. In table 1

terms 1 to 8 relate to the interaction of any electromagnetic radiation of an optical nature (UV, visible, IR) with any medium observed from the outside;

terms 9 to 11 relate to the interaction of any electromagnetic radiation of an optical nature (UV, visible, IR) with a medium with plane and parallel surfaces, which is homogeneous, isotropic, non-luminescent and nonscattering, observed from the inside.

Table 2 is in line with the scope of this International Standard and therefore concerns the interaction of a beam of monochromatic luminous radiation striking, at normal incidence, a medium consisting of a solution which is

General 3

ments;

Molecular absorption spectrometry obeys the following laws.

3.1 Lambert-Bouguer's law

When a parallel beam of monochromatic radiation of flux ϕ_0 traverses, at normal incidence, an absorbing medium with plane, parallel surfaces and which is homogeneous, isotropic, non-luminescent and non-scattering, over an optical path length b, the transmitted flux Φ_{tr} is given by the equation

$$\Phi_{\rm tr} = \Phi_0 \, {\rm e}^{-kb}$$

where

е is the base of natural logarithms;

k is a linear absorption coefficient.

This equation is derived from integration of the differential equation

$$\mathrm{d}\Phi_{\mathrm{tr}} = -k \, \Phi_{\mathrm{tr}} \, \mathrm{d}x$$

where x varies from 0 to b, $d\Phi_{tr}$ is the reduction in radiant energy flux along an infinitely small optical path length dx, and $\Phi_{\rm tr}$ is the value of the flux at point *x*.

No.	Term	Definition	Symbol	Formula	International symbol for unit
1	incident flux ¹⁾	Radiant (lurninous) flux of the radiation striking an external surface of the medium	Φ_0		w
2	transmitted flux ¹⁾	Radiant (luminous) flux of the radiation emerging from the medium	$arPsi_{ m tr}$		w
3	transmission (CIE 45-05-65)	Passage of radiation through a medium without change of wavelength			
4	transmittance (CIE 45-20-085)	Ratio of the transmitted radiant (luminous) flux to the incident flux	τ	$\tau = \frac{\boldsymbol{\varphi}_{\rm tr}}{\boldsymbol{\varphi}_0}$	
5	absorbance; transmis- sion (optical) density (CIE 45-20-100)	Logarithm (to base ten) of the reciprocal of the transmittance	A	$A = \lg \frac{1}{\tau}$	
6	absorbed flux without phenomena other than absorption	Difference between the incident and transmitted flux	Φ_{a}	$arPsi_{ m o}-arPsi_{ m tr}$	w
7	absorption (CIE 45-05-070)	Transformation of radiant energy to a different form of energy by interaction with matter	D PREVI	EW	
8	absorptance (CIE 45-20-115)	Ratio of the absorbed radiant (luminous) flux to the incident flux	.iteh _a ai)	$\frac{\varPhi_{a}}{\varPhi_0}$	
9	internal transmittance (of a homogenous non- diffusing layer) (CIE 45-20-090)	Ratio of the radiant (luminous) flux- reaching the exit surface of the layer to the flux which leaves the entry surface	<u>1982</u> s/sist/0f606774-025f-+ -6286-1982	4a79-b4b8-	
10	internal absorbance; internal transmission density (CIE 45-20-105)	Logarithm (to base ten) of the reciprocal of the internal transmittance	A_{i}	$A_{i} = \lg \frac{1}{\tau_{i}}$	
11	internal absorptance (of a homogenous non- diffusing layer (CIE 45-20-120)	Ratio of the radiant (luminous) flux absorbed between the entry and the exit surfaces of the layer to the flux which leaves the entry surfaces	<i>a</i> _i		

Table 1 - Summary of definitions relating to radiation and to the optical properties of matter

1) The definition or this term, used, but not defined, in the Vocabulaire international de l'éclairage (International lighting vocabulary) has been deduced from the definitions which follow it.

NOTE - The references between parentheses are those of publication CIE No. 38, Vocabulaire internationale de l'éclairage (TC 23) 1977.

No.	Term	Definition	Symbol	Formula	International symbol for unit
12	thickness ¹⁾	Distance between the internal plane and parallel surfaces of an optical cell	I		mm or cm
13	optical path length ¹⁾	Distance traversed by a light ray between the entry and exit surfaces of a solution contained in an optical cell	b		mm or cm
14	internal transmittance	Ratio of the radiant (luminous) flux reaching the exit surface of a solution contained in an optical cell, to the flux which leaves the entry surface	$ au_{\mathrm{i}}$		
15	internal absorptance	Ratio of the radiant (luminous) flux absorbed between the entry and the exit surfaces of a solution contained in an optical cell, to the flux which leaves the entry surface	α_{i}		
16	reference flux	Radiant flux of the monochromatic radia- tion transmitted by the optical cell con- taining the solution used as reference and reaching the detector	${\pmb \Phi}_{ m r}$		W
17	sample flux	Radiant flux of the monochromatic radia- tion transmitted by the optical cell con- taining the solution on which the measurement is made and reaching the detector	PREVIEV h.ai),	V	w
18	percentage transmittance	The ratio, expressed as a percentage, of the sample flux to the reference flux 15/0 88a995104ac1/ac.6286	606774-025f-4a79-	b4b8- 100 $\frac{\Phi_{\rm s}}{\Phi_{\rm r}}$	
19	partial internal absorbance; partial internal transmission density ²⁾	Fraction of the internal absorbance of a solution due to certain of its consti- tuents. The absorbance of a solution for given experimental conditions is thus the difference between its internal absorb- ance and that of the solution used as reference	Ap	$\lg \frac{\phi_{\rm r}}{\phi_{\rm s}}$	
20	characteristic partial internal absorbance; characteristic partial internal transmission density ²⁾	The partial internal absorbance of a solution due to only one of its con- stituents (for example compound dissolved for analysis)	A _c	$\lg \frac{\Phi_{\rm r}}{\Phi_{\rm s}}$	
21	concentration	Ratio of the mass of compound dissolved to the volume of the solution According to the units used, one can dif- ferentiate 21.1 and 21.2			
21.1	mass concentration	Ratio of the mass of the compound dissolved to the volume of the solution	Q		kg/m ³
21.2	amount-of-substance concentration	Ratio of the amount-of-substance of compound dissolved to the volume of the solution	с		mol/l
1				I	

Table 2 – Terms and definitions relating to molecular absorption spectrometry

1) Terms 12 and 13 are equivalent when the incident light ray is normal. The product of the optical path length and the refractive index of the absorbing solution is called the "chemin optique" (literally "optical path").

2) The terms "partial internal transmission density" and "characteristic partial internal transmission density" are obsolete.

Table 2 (concluded)

No.	Term	Definition	Symbol	Formula	International symbol for unit
22	characteristic partial internal absorbance coefficient	Characteristic absorbance per unit thickness and per unit concentration of the dissolved compound under con- sideration			
		NOTE — By extension, and in general, the con- centration used is often that of the element or molecule being determined.			
		According to the units used, one can dif- ferentiate 22.1 and 22.2			
22.1	specific mass absorbance coefficient	Absorbance coefficient for which the thickness is usually expressed in cen- timetres and the concentration in grams of compound dissolved in one litre of solution	а	$\frac{A_{\rm c}}{l\varrho}$	cm ⁻¹ ·g ⁻¹ ·I
22.2	specific molar absorbance coefficient ¹⁾	Absorbance coefficient for which the thickness is usually expressed in cen- timetres and the concentration in moles of compound dissolved in one litre of solution	E	$\frac{A_{c}}{lc}$	cm ^{−1} ·mol ^{−1} ·l or m ² ·mol ^{−1}

1) Specific molar absorbance coefficient is often called "molar absorptivity"

NOTE — The units employed to express concentration and thickness should be specified, as also should the wavelength for the determination, the temperature of the solutions and, if possible, the width of the pass band. As the absorbance due to a compound depends upon the state of that compound in solution, its absorbance coefficient is only completely defined if the nature of the solvent and the physico-chemical properties of the solution are specified.

For terms 18, 19, 20 and 22, it is recommended that one of the two following typographical arrangements be used :

a)



https://standards.iteh.ai/catalog/standards/sist/0f606774-025f-4a79-b4b8-88a995194ac1/iso-6286-1982

which allows specification of :

- the temperature t, in degrees Celsius;

-the measurement wavelength λ , in nanometres;

-the nature of the solvent;

-the value x of the term in question with its subscript or unit, if necessary.

Example :

For specific mass absorbance coefficient



b)

x subscript or unit at t °C, λ nanometres, solvent

Example :

For specific mass absorbance coefficient

2 200 cm⁻¹·g⁻¹·l at 20 °C, 540 nm, in aqueous solution

3.2 Beer's law

The radiant flux of a beam of parallel monochromatic radiation decreases exponentially as the concentration of the absorbing compound increases, all other factors being constant :

$$\Phi_{\rm tr} = \Phi_0 \, {\rm e}^{-k_{\rm m} \, \varrho}$$

or

$$\Phi = \Phi_0 e^{-k_{\rm E}}$$

where

- is the incident flux; Φ_0
- is the transmitted flux; Φ_{tr}
- is the base of natural logarithms; е

 $k_{\rm m}$ and $k_{
m \epsilon}$ are absorption coefficients which are constant for given experimental conditions;

- is the mass concentration; 0

In practice, absorbances are generally measured so that the characteristic absorbance of the dissolved compound under consideration can be obtained by applying the additive law (see 3.3) in the form

$$A_2 - A_1 = A_c = ab\varrho = \varepsilon bc$$

where

is the absorbance of solution 1; A_1

is the absorbance of solution 2; A_2

A_c is the characteristic absorbance of the dissolved compound under consideration;

a, b, c, ε and ρ have the same meanings as in 3.2 and in the previous equation.

Apparatus

Sub-clause 4.1 is intended to facilitate understanding of the objectives mentioned in clause 1 and which are detailed in 4.2

c is the amount of substance concentration. and 4.3. **F**W

4.1 Components of molecular absorption spectrometers standards.i

3.3 Additive nature of the laws of Lambert-Bouguer and Beer

When a beam of parallel monochromatic radiation traverses, at normal incidence, an absorbing medium with plane, parallel iso-62 surfaces and which is homogeneous, isotropic, nonluminescent, and non-scattering, and consists of a solution of n dissolved compounds which do not react with one another, the total absorbance is equal to the sum of the n characteristic absorbances.

3.4 General law

The laws of Lambert-Bouguer and Beer can be expressed by a single equation :

$$\Phi_{tr} = \Phi_0 \, 10^{-ab\varrho}$$

or

$$\Phi_{tr} = \Phi_0 \, 10^{-\varepsilon bc}$$

where

a is the specific mass absorbance coefficient which is constant for given experimental conditions;

b is the optical path length;

 ε is the specific molar absorbance coefficient which is constant for given experimental conditions;

 $\Phi_0, \ \Phi_{\rm tr}, \ \varrho$ and c have the same meanings as in 3.2.

ISO 6286:198The components which make up molecular absorption spectrometers, i.e. instruments designed for the determination of absorbance or percentage transmittance, are intended to assure the three following functions :

- a) production of a beam of radiation of a selected band of wavelengths and control of the bandwidth;
- b) introduction of the solution to be examined into the beam of radiation:
- c) measurement.

To the three fundamental units assuring these three functions, one must add various associated devices (collimators, lenses, fixed or rotating mirrors, diaphragms, slits, etc.) which define the beam appropriately in space and direct it onto the various parts of these units.

A wavelength scale system, the graduations of which may correspond to wavelength (nm) or to wave number (cm⁻¹), completes the instrument.

4.1.1 Devices for the production of a beam of radiation

The beam of radiation is characterized by its spectral composition, its intensity, its configuration and its direction in space; consequently, its production involves a source of radiation, a selector of the wavelengths emitted, and various complementary devices.

The spectral characteristics of an instrument are not directly related to any one of these parts but result from their association.

4.1.1.1 Sources of radiation

Sources of radiation are differentiated principally by the emission spectra they produce. Such a spectrum may, in fact, cover a range of wavelengths, more or less expanded, and show a continuous or discontinuous profile with, as the case may be, bands of various width or line emissions.

For example, a tungsten filament lamp emits a continuous spectrum in the visible and near infra-red. Other examples are a hydrogen lamp emitting a continuous spectrum in the ultraviolet and metal vapour lamps (Hg, Na, Cd, etc.) which, under certain conditions, emit a line spectrum.

An instrument may be equipped with one or more sources.

Some types of sources of continuous radiation are associated with interferometers in order to obtain a temporal programming of the emission of radiation; this results in a particular exploitation of the signals received by the detector (Fourier transformation technique). This technique eliminates the need for wavelength selectors (4.1.1.2).

4.1.1.2 Wavelength selector

depending on the circumstances.

This is the part of the instrument comprising one or more ARD devices which allow the isolation of a range of wavelengths from the spectrum emitted by the radiation source. An ar 41.31 entrance slit and an exit slit are associated with these devices,

4.1.1.3 Complementary devices

These are the parts of the apparatus which give the radiation beam the appropriate spatial definition, i.e. cross-section, parallelism, focus, path-type (single or double beam) etc. Various devices, such as collimators, lenses, fixed or rotating mirrors, diaphragms and slits, are used for this purpose.

4.1.2 Devices for introducing the solution

In order to intercept the measuring beam, solutions are generally introduced in a transparent vessel called an "optical cell" or more simply "cell". There are several versions which differ in their geometry, and which allow for intermittent or continuous introduction of the solutions for measurement. The simplest are hollow rectangular prisms. The cells are usually supported and positioned by cell carriers which, in instruments for the visible and ultraviolet regions, are positioned after the wavelength selector.

The compartment in which the cell carrier is mounted is, according to the design of the instrument, more or less totally shaded from ambient light and can be equipped with ancillary devices such as thermal conditioning, cell-changer, etc.

ar 41.3 Measurement system

PREVIEW

The conversion of the information contained within the ISO 62 emiergent beam into signals which are intelligible to the user of

These devices (which may be used singly calternatively a log/standathe/sinstrument4 (excluding-telectronic stabilization devices) together in the same apparatus) may be classified according to 94ac1/isvolves 1982 the phenomena to which they relate :

- a) absorbing filters, the function of which is based on the selective absorption of certain radiations;
- b) interference filters, the function of which is based on the interference of radiation;

c) prisms and gratings, the function of which is based on the dispersion of radiation (the term "monochromator" is often used to designate a selection device to isolate a narrow wavelength region).

These devices may equally well be grouped into two categories according to their method of use :

 fixed pass band selectors, generally known as filters : the change from one wavelength to another is effected by placing a different filter in the path of the beam. Each filter corresponds to a definite wavelength band;

- selectors for continuous variation of wavelength, which are known as prisms, gratings and adjustable interference filters : a mechanical device works so that the different rays that they separate are dispersed in a continuous way (with reference to the exit slit when this is provided).

The exit slit of the selector is often adjustable and is generally adjusted so that it is of the same width as the entrance slit; its width is one of the parameters upon which the spectral purity of the measuring beam depends. a) reception of the beam by a detector;

b) amplification, if necessary, of the signal emitted by the detector;

c) presentation of the amplified signal by an indicator device.

4.1.3.1 Detector

In general, detectors transform the energy of the radiant energy flux carried by the emergent beam of radiation into electrical energy. Their functioning is based :

 $-\,$ either on photoemission phenomena : photoemissive cells;

- or on photoconductive phenomena : photoconductive cells and photovoltaic cells.

The different types of the corresponding detectors can be used either separately or together in the same apparatus :

a) photoemissive cells, also called "photoemitters", comprise vacuum-tubes, gas-filled tubes and photomultipliers (which are vacuum tubes with amplification by secondary emission); their useful wavelength range is limited by the metal or metals forming the cathode; they can be used between 200 and 1 000 nm, but use is generally between 200 and 500 nm; b) photoconductive cells, also called "photosensitive cells", comprise an element, the conductivity of which varies as a function of the intensity of the radiation; they are employed for wavelengths above 700 nm;

c) photovoltaic cells, also called "photopiles" or "barrierlayer cells"; they can be used from 230 to 1 000 nm, but generally use is between 400 and 800 nm.

As the intensity of the beam has to be adjusted to the characteristics of the measuring detector, a mechanical system capable of attenuating the intensity is generally fitted. Depending on the design, continuously or non-continuously variable devices such as diaphragms, gratings, absorbant glasses, etc., are used. These devices can be subjected to automatic control which, for example, maintains a constant luminous intensity whatever the wavelength used.

4.1.3.2 Amplifier

The electrical signals delivered by the detector are generally weak, and often very weak, and it is, therefore, necessary in most cases to amplify them.

Amplification is of either direct or alternating current. The latter indicates that the light beam received by the detector is modulated, the modulation being effected by means of mechanical, electronic, electromechanical or other devices.

4.1.3.3 Indicator device

The electrical signals emitted by/the detector are expressed ards/sist/0 after amplification, as percentage transmissionsory units of/iso-6286 absorbance. It is into one or other of these two forms, sometimes both, that the signals are generally converted by the indicator device which, in certain instruments fitted with a calculator, may, however, yield results directly expressed in concentration units.

Indicator devices may be of either analogue or digital type. Analogue indicator devices have a continuous scale with a pointer (spot or needle galvanometers, etc.) and digital indicator devices show a number (indicator tubes, light-emitting diodes, liquid crystal displays, etc.).

If the instrument produces a record of the results, the following can be distinguished

a) recorders giving values in analogue form (the recording may be "continuous" or "point-by-point");

b) printers giving values in numerical form.

In short, the indicating device is an electronic circuit, the nature of which varies with the instrument, which converts electrical signals into information which the operator can use.

4.2 Main types of instruments

Spectrometers designed for the determination of the absorbance or percentage transmittance of a solution may be classified according to various criteria, in particular according to the nature of the radiation selection device and/or the nature of the measuring device. NOTE – The term "photometer" is reserved to an instrument for measuring photometric quantities, fitted with detectors for photopic vision $V(\lambda)$.

The following are differentiated according to the nature of the wavelength selection device.

4.2.1 Instruments with selectors for continuous variation and discontinuous variation

4.2.1.1 Instruments with selectors for continuous variation

These are instruments in which a continuous variation of wavelength is possible. They comprise a continuous spectrum source of radiation, together with a radiation selector with continuous variation of wavelength.

4.2.1.2 Instruments with selectors for discontinuous variation

These are instruments in which continuous variation of wavelength is not possible. According to the origin of the discontinuity of the spectrum, the following may be differentiated :

a) line-spectrum instruments, in which the discontinuous spectrum is due to the radiation source;

of iso 6286 struments, in which the discontinuous spectrum is due to the wavelength selector.

4.2.2 Double-beam and single-beam instruments

The following types of instruments may be differentiated according to whether the sample flux is compared simultaneously or successively with the reference flux.

4.2.2.1 Double-beam instruments

The measurement result given by the instrument is obtained from a simultaneous comparison of the sample flux with the reference flux; the measurement process is therefore always carried out using two beams at once (with two optical cells).

In practice the manner in which this is achieved varies : usually, two beams of radiation issuing from the same source are directed either onto two detectors arranged in opposition, or onto a single detector which receives them alternately but at a frequency sufficiently high for the system to be regarded as comparing them simultaneously.

4.2.2.2 Single-beam instruments

The measurement result given by the instrument results from successive determinations of the reference flux and the sample flux; the beam traverses only one cell at a time.